


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030623\
 Data File : BP013950.D
 Acq On : 06 Mar 2023 10:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Mar 07 06:48:25 2023






AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.9	111643	PASS
68	69	0.00	2	1.1	1461	PASS
69	198	0.00	100	34.2	127738	PASS
70	69	0.00	2	0.5	646	PASS
127	198	10	80	41.6	155613	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	373791	PASS
199	198	5	9	7.2	26760	PASS
275	198	10	60	31.0	115915	PASS
365	198	1	100	4.3	16239	PASS
441	198	0.01	100	14.5	54091	PASS
442	442	100	100	100.0	338325	PASS
443	442	15	24	19.6	66475	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030623\
Data File : BP013950.D
Acq On : 06 Mar 2023 10:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Mar 07 06:18:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 07 03:35:43 2023
Response via : Initial Calibration



TIC: BP013950.D\data.ms

(70) Pentachlorophenol (C)

17.128min (0.000) 233515.23 ng




response 468007

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	64.16
264.00	61.50	63.16
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030623\
 Data File : BP013950.D
 Acq On : 06 Mar 2023 10:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Mar 07 06:18:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 07 03:35:43 2023
 Response via : Initial Calibration



TIC: BP013950.D\data.ms

(77) Benzidine

19.657min (+ 0.006) -20.00 ng

response 2320243

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.24
183.00	12.10	12.17
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/6/2023	BNA_P	<u>BP013950.D</u>
Compound Name	Response	Retention Time
DDT	1295686	20.851
DDD	16110	20.469
DDE	612	19.933
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
16722	1312408	1.27

Instrument :
BNA_P
ClientSampleId :
DFTPP