

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013962.D
 Acq On : 07 Mar 2023 15:20
 Operator : CG/JU
 Sample : SSTD00535
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Quant Time: Mar 08 00:42:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.093	152	113062	20.000	ng/ul	0.00
20) Naphthalene-d8	10.916	136	454326	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.728	164	329716	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.486	188	800011	20.000	ng/ul	0.00
79) Chrysene-d12	21.563	240	893390	20.000	ng/ul	0.00
88) Perylene-d12	24.115	264	1030144	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	5841	2.141	ng/uL	0.00
4) Pyridine-d5	3.881	84	35306	4.366	ng/ul	0.01
7) Phenol-d5	7.240	99	43317	4.300	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.410	67	26127	4.426	ng/ul	0.00
11) 2-Chlorophenol-d4	7.616	132	34277	4.622	ng/ul	0.00
15) 4-Methylphenol-d8	8.799	113	34075	4.138	ng/ul	0.00
21) Nitrobenzene-d5	9.263	128	15676	4.959	ng/ul	0.00
24) 2-Nitrophenol-d4	9.993	143	18305	5.015	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.534	165	35648	4.793	ng/ul	0.01
31) 4-Chloroaniline-d4	11.057	131	47101	4.468	ng/ul	0.01
46) Dimethylphthalate-d6	14.134	166	129505	4.987	ng/ul	0.00
49) Acenaphthylene-d8	14.428	160	135652	4.826	ng/ul	0.00
54) 4-Nitrophenol-d4	14.922	143	19006	4.108	ng/ul	0.00
60) Fluorene-d10	15.722	176	111747	5.052	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.839	200	23685	4.826	ng/ul	0.00
73) Anthracene-d10	17.586	188	184966	5.056	ng/ul	0.00
81) Pyrene-d10	19.804	212	237706	4.692	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.939	264	252802	4.898	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	6234	2.129	ng/uL	94
5) Pyridine	3.905	79	35024	4.382	ng/ul	96
6) Benzaldehyde	7.228	77	25920	5.706	ng/ul	90
8) Phenol	7.269	94	43983	4.253	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.505	93	36993	4.464	ng/ul	98
12) 2-Chlorophenol	7.652	128	34091	4.480	ng/ul	97
13) 2-Methylphenol	8.534	108	32155	4.074	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.616	45	41179	4.369	ng/ul	98
16) Acetophenone	8.922	105	59322	4.493	ng/ul	95
17) N-Nitroso-di-n-propyla...	8.899	70	31327	4.777	ng/ul	98
18) 4-Methylphenol	8.863	108	36380	4.175	ng/ul	98
19) Hexachloroethane	9.181	117	15191	4.876	ng/ul	95
22) Nitrobenzene	9.310	77	43441	5.223	ng/ul	97
23) Isophorone	9.834	82	82945	4.727	ng/ul	98
25) 2-Nitrophenol	10.022	139	19476	4.923	ng/ul	99
26) 2,4-Dimethylphenol	10.075	107	44340	5.034	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.316	93	50419	4.758	ng/ul	94
29) 2,4-Dichlorophenol	10.563	162	35354	4.828	ng/ul	99
30) Naphthalene	10.969	128	119367	4.935	ng/ul	98
32) 4-Chloroaniline	11.081	127	49301	4.639	ng/ul	96
33) Hexachlorobutadiene	11.246	225	29352	5.691	ng/ul	99
34) Caprolactam	11.863	113	9874	3.843	ng/ul	86
35) 4-Chloro-3-methylphenol	12.187	107	40281	4.765	ng/ul	96
36) 2-Methylnaphthalene	12.563	142	79707	4.688	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013962.D
 Acq On : 07 Mar 2023 15:20
 Operator : CG/JU
 Sample : SSTD00535
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Quant Time: Mar 08 00:42:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.781	142	82808	4.713	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.928	216	55335	5.326	ng/ul	97
40) Hexachlorocyclopentadiene	12.904	237	29401	4.407	ng/ul	98
41) 2,4,6-Trichlorophenol	13.169	196	31499	4.620	ng/ul	98
42) 2,4,5-Trichlorophenol	13.245	196	35503	4.811	ng/ul	91
43) 1,1'-Biphenyl	13.563	154	118720	4.799	ng/ul	99
44) 2-Chloronaphthalene	13.610	162	94157	4.838	ng/ul	99
45) 2-Nitroaniline	13.816	65	23757	5.068	ng/ul	96
47) Dimethylphthalate	14.181	163	130158	5.053	ng/ul	99
48) 2,6-Dinitrotoluene	14.304	165	22176	4.928	ng/ul	94
50) Acenaphthylene	14.451	152	144508	4.765	ng/ul	99
51) 3-Nitroaniline	14.634	138	21140	4.751	ng/ul	98
52) Acenaphthene	14.792	153	103469	4.875	ng/ul	98
53) 2,4-Dinitrophenol	14.845	184	13805	4.302	ng/ul	94
55) 4-Nitrophenol	14.939	109	19083	4.701	ng/ul	92
56) Dibenzofuran	15.128	168	151107	4.966	ng/ul	99
57) 2,4-Dinitrotoluene	15.086	165	34287	5.111	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.351	232	35464	5.098	ng/ul	96
59) Diethylphthalate	15.534	149	131524	5.085	ng/ul	100
61) Fluorene	15.775	166	126043	5.077	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.769	204	69224	5.363	ng/ul	95
63) 4-Nitroaniline	15.798	138	21199	5.032	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.851	198	22865	4.752	ng/ul#	94
67) N-Nitrosodiphenylamine	15.981	169	105141	4.835	ng/ul	99
68) 4-Bromophenyl-phenylether	16.669	248	44803	5.303	ng/ul	96
69) Hexachlorobenzene	16.786	284	53391	5.424	ng/ul	97
70) Atrazine	16.928	200	44141	5.076	ng/ul	97
71) Pentachlorophenol	17.133	266	29635	4.626	ng/ul	95
72) Phenanthrene	17.528	178	212168	5.012	ng/ul	100
74) Anthracene	17.616	178	214984	5.048	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.534	216	56257	5.268	ng/uL	98
76) Pentachlorobenzene	15.045	250	60745	5.512	ng/uL	96
77) Carbazole	17.886	167	185367	4.941	ng/ul	99
78) Di-n-butylphthalate	18.416	149	221372	4.865	ng/ul	99
80) Fluoranthene	19.480	202	266601	4.597	ng/ul	99
82) Pyrene	19.833	202	284395	4.599	ng/ul	99
83) Butylbenzylphthalate	20.686	149	106838	5.048	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.469	252	104194	5.182	ng/ul	98
85) Benzo(a)anthracene	21.545	228	300467	4.922	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.445	149	157022	4.845	ng/ul	100
87) Chrysene	21.604	228	289983	5.035	ng/ul	98
89) Di-n-octyl phthalate	22.416	149	272387	3.843	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	318540	4.637	ng/ul	99
91) Benzo(k)fluoranthene	23.374	252	312926	4.592	ng/ul	99
93) Benzo(a)pyrene	23.998	252	281429	4.834	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.792	276	383833	5.878	ng/ul	98
95) Dibenzo(a,h)anthracene	26.804	278	321140	5.887	ng/ul	99
96) Benzo(g,h,i)perylene	27.615	276	309438	5.952	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013962.D
 Acq On : 07 Mar 2023 15:20
 Operator : CG/JU
 Sample : SSTD00535
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Quant Time: Mar 08 00:42:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 2.129 ng/uL
 RT: 3.475 min Scan# 49
 Delta R.T. -0.000 min
 Lab File: BP013962.D
 Acq: 07 Mar 2023 15:20




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617


Tgt Ion: 88 Resp: 6234
 Ion Ratio Lower Upper
 88 100
 43 30.4 23.9 35.9
 58 55.6 50.2 75.2



#5
 Pyridine
 Concen: 4.382 ng/uL
 RT: 3.905 min Scan# 122
 Delta R.T. 0.012 min
 Lab File: BP013962.D
 Acq: 07 Mar 2023 15:20

Tgt Ion: 79 Resp: 35024
 Ion Ratio Lower Upper
 79 100
 52 71.5 55.1 82.7
 51 40.1 28.9 43.3









#6
Benzaldehyde
Concen: 5.706 ng/ul
RT: 7.228 min Scan# 61
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 77 Resp: 25920
Ion Ratio Lower Upper
77 100
105 98.6 74.2 111.2
106 99.9 69.8 104.8



#8
Phenol
Concen: 4.253 ng/ul
RT: 7.269 min Scan# 694
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion: 94 Resp: 43983
Ion Ratio Lower Upper
94 100
65 28.6 25.2 37.8
66 45.4 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 4.464 ng/ul
RT: 7.505 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 93 Resp: 36993
Ion Ratio Lower Upper
93 100
63 71.5 58.8 88.2
95 31.7 25.8 38.6



#12
2-Chlorophenol
Concen: 4.480 ng/ul
RT: 7.652 min Scan# 759
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion: 128 Resp: 34091
Ion Ratio Lower Upper
128 100
64 50.6 38.3 57.5
130 31.0 25.7 38.5





#13

2-Methylphenol

Concen: 4.074 ng/ul

RT: 8.534 min Scan# 908

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :





SSTD005617

Tgt Ion:108 Resp: 32155

Ion Ratio Lower Upper

108 100

107 88.2 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.369 ng/ul

RT: 8.616 min Scan# 923

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion: 45 Resp: 41179


Ion Ratio Lower Upper

45 100

77 16.8 14.2 21.2

79 13.7 10.6 15.8









#16
Acetophenone
Concen: 4.493 ng/ul
RT: 8.922 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:105 Resp: 59322
Ion Ratio Lower Upper
105 100
77 87.4 65.4 98.2
51 28.6 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 4.777 ng/ul
RT: 8.899 min Scan# 971
Delta R.T. -0.000 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion: 70 Resp: 31327
Ion Ratio Lower Upper
70 100
42 52.1 40.9 61.3
101 9.8 7.9 11.9
130 18.3 16.6 25.0









#18
4-Methylphenol
Concen: 4.175 ng/ul
RT: 8.863 min Scan# 90
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:108 Resp: 36380
Ion Ratio Lower Upper
108 100
107 109.9 89.8 134.6



#19
Hexachloroethane
Concen: 4.876 ng/ul
RT: 9.181 min Scan# 1019
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:117 Resp: 15191
Ion Ratio Lower Upper
117 100
201 122.9 91.7 137.5
199 72.7 59.1 88.7









#22
Nitrobenzene
Concen: 5.223 ng/ul
RT: 9.310 min Scan# 110
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 77 Resp: 43441
Ion Ratio Lower Upper
77 100
123 38.9 32.6 48.8
65 15.4 11.2 16.8



#23
Isophorone
Concen: 4.727 ng/ul
RT: 9.834 min Scan# 1130
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion: 82 Resp: 82945
Ion Ratio Lower Upper
82 100
95 8.6 7.1 10.7
138 15.3 13.3 19.9









#25
2-Nitrophenol
Concen: 4.923 ng/ul
RT: 10.022 min Scan# 1161
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion	Ratio	Lower	Upper
139	100		
65	59.5	46.6	70.0
109	43.9	34.5	51.7



#26
2,4-Dimethylphenol
Concen: 5.034 ng/ul
RT: 10.075 min Scan# 1171
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion	Ratio	Lower	Upper
107	100		
121	50.0	37.1	55.7
122	72.6	61.2	91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 4.758 ng/ul

RT: 10.316 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion: 93 Resp: 50419

Ion Ratio Lower Upper

93 100

95 36.1 25.6 38.4

123 11.1 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 4.828 ng/ul

RT: 10.563 min Scan# 1254

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:162 Resp: 35354


Ion Ratio Lower Upper

162 100

164 65.0 52.0 78.0

98 35.0 29.1 43.7









#30
Naphthalene
Concen: 4.935 ng/ul
RT: 10.969 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:128 Resp: 119367
Ion Ratio Lower Upper
128 100
129 11.8 8.5 12.7
127 14.2 11.0 16.4



#32
4-Chloroaniline
Concen: 4.639 ng/ul
RT: 11.081 min Scan# 1342
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:127 Resp: 49301
Ion Ratio Lower Upper
127 100
129 33.2 24.9 37.3





#33

Hexachlorobutadiene

Concen: 5.691 ng/ul

RT: 11.246 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:225 Resp: 29352

Ion Ratio Lower Upper

225 100

223 61.7 48.0 72.0

227 62.9 50.3 75.5



#34

Caprolactam

Concen: 3.843 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:113 Resp: 9874


Ion Ratio Lower Upper

113 100

55 134.9 124.0 186.0

56 113.3 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 4.765 ng/ul
RT: 12.187 min Scan# 1529
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:107 Resp: 40281
Ion Ratio Lower Upper
107 100
144 22.2 19.4 29.0
142 70.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 4.688 ng/ul
RT: 12.563 min Scan# 1594
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:142 Resp: 79707
Ion Ratio Lower Upper
142 100
141 87.6 71.7 107.5





#37

1-Methylnaphthalene

Concen: 4.713 ng/ul

RT: 12.781 min Scan# 1631

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:142 Resp: 82808

Ion Ratio Lower Upper

142 100

141 92.3 73.8 110.6

116 5.0 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 5.326 ng/ul

RT: 12.928 min Scan# 1656

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Tgt Ion:216 Resp: 55335




Ion Ratio Lower Upper


216 100

214 74.3 61.7 92.5

179 19.4 16.2 24.4

108 15.7 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 4.407 ng/ul

RT: 12.904 min Scan# 1652

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:237 Resp: 29401

Ion Ratio Lower Upper

237 100

235 63.2 49.0 73.6

272 13.0 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 4.620 ng/ul

RT: 13.169 min Scan# 1697

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:196 Resp: 31499


Ion Ratio Lower Upper

196 100

198 95.1 78.5 117.7

200 32.0 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 4.811 ng/ul

RT: 13.245 min Scan# 1708

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:196 Resp: 35503

Ion Ratio Lower Upper

196 100

198 87.5 77.6 116.4

200 28.3 25.3 37.9



#43

1,1'-Biphenyl

Concen: 4.799 ng/ul

RT: 13.563 min Scan# 1764

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:154 Resp: 118720


Ion Ratio Lower Upper

154 100

153 41.4 32.6 49.0

76 10.7 9.4 14.0





#44

2-Chloronaphthalene

Concen: 4.838 ng/ul

RT: 13.610 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:162 Resp: 94157

Ion Ratio Lower Upper

162 100

164 34.2 26.0 39.0

127 39.2 31.3 46.9



#45

2-Nitroaniline

Concen: 5.068 ng/ul

RT: 13.816 min Scan# 1807

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion: 65 Resp: 23757


Ion Ratio Lower Upper

65 100

92 73.6 56.3 84.5

138 105.5 81.4 122.2









#47
Dimethylphthalate
Concen: 5.053 ng/ul
RT: 14.181 min Scan# 1869
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:163 Resp: 130158
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 9.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 4.928 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:165 Resp: 22176
Ion Ratio Lower Upper
165 100
89 70.0 51.7 77.5
121 23.3 17.5 26.3





#50

Acenaphthylene

Concen: 4.765 ng/ul

RT: 14.451 min Scan# 1915

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:152 Resp: 144508

Ion Ratio Lower Upper

152 100

151 19.5 16.2 24.2

153 12.8 10.4 15.6



#51

3-Nitroaniline

Concen: 4.751 ng/ul

RT: 14.634 min Scan# 1946

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:138 Resp: 21140


Ion Ratio Lower Upper

138 100

108 12.9 10.5 15.7

92 121.0 98.5 147.7









#52
Acenaphthene
Concen: 4.875 ng/ul
RT: 14.792 min Scan# 1973
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:153 Resp: 103469
Ion Ratio Lower Upper
153 100
152 46.9 37.7 56.5
154 84.9 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 4.302 ng/ul
RT: 14.845 min Scan# 1982
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:184 Resp: 13805
Ion Ratio Lower Upper
184 100
63 57.1 51.9 77.9
154 71.8 58.8 88.2









#55
4-Nitrophenol
Concen: 4.701 ng/ul
RT: 14.939 min Scan# 1996
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:109 Resp: 19083
Ion Ratio Lower Upper
109 100
139 101.6 74.5 111.7
65 115.8 86.6 130.0



#56
Dibenzofuran
Concen: 4.966 ng/ul
RT: 15.128 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:168 Resp: 151107
Ion Ratio Lower Upper
168 100
139 39.2 31.8 47.6





#57

2,4-Dinitrotoluene

Concen: 5.111 ng/ul

RT: 15.086 min Scan# 2023

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:165 Resp: 34287

Ion Ratio Lower Upper

165 100

63 46.3 34.6 52.0

182 3.5 2.2 3.2#



#58

2,3,4,6-Tetrachlorophenol

Concen: 5.098 ng/ul

RT: 15.351 min Scan# 2068

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Tgt Ion:232 Resp: 35464




Ion Ratio Lower Upper


232 100

131 45.2 34.1 51.1

130 2.1 2.1 3.1

166 30.6 23.0 34.6









#59
Diethylphthalate
Concen: 5.085 ng/ul
RT: 15.534 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:149 Resp: 131524
Ion Ratio Lower Upper
149 100
177 22.9 18.2 27.2
150 12.5 10.0 15.0



#61
Fluorene
Concen: 5.077 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:166 Resp: 126043
Ion Ratio Lower Upper
166 100
165 99.5 78.9 118.3
167 13.1 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 5.363 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:204 Resp: 69224
Ion Ratio Lower Upper
204 100
206 34.0 26.6 39.8
141 55.8 49.2 73.8



#63
4-Nitroaniline
Concen: 5.032 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:138 Resp: 21199
Ion Ratio Lower Upper
138 100
92 56.0 44.5 66.7
108 97.9 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 4.752 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. -0.000 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:198 Resp: 22865
Ion Ratio Lower Upper
198 100
182 7.1 3.0 4.6#
77 28.2 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 4.835 ng/ul
RT: 15.981 min Scan# 2175
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:169 Resp: 105141
Ion Ratio Lower Upper
169 100
168 66.4 53.7 80.5
167 35.1 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 5.303 ng/ul
RT: 16.669 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:248 Resp: 44803
Ion Ratio Lower Upper
248 100
250 96.1 78.8 118.2
141 70.0 60.8 91.2



#69
Hexachlorobenzene
Concen: 5.424 ng/ul
RT: 16.786 min Scan# 2312
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:284 Resp: 53391
Ion Ratio Lower Upper
284 100
142 31.1 23.2 34.8
249 31.1 25.7 38.5





#70

Atrazine

Concen: 5.076 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:200 Resp: 44141

Ion Ratio Lower Upper

200 100

173 26.2 19.0 28.4

215 46.8 38.5 57.7



#71

Pentachlorophenol

Concen: 4.626 ng/ul

RT: 17.133 min Scan# 2371

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:266 Resp: 29635


Ion Ratio Lower Upper

266 100

264 66.3 48.4 72.6

268 61.1 50.4 75.6









#72
Phenanthrene
Concen: 5.012 ng/ul
RT: 17.528 min Scan# 2437
Delta R.T. 0.006 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:178 Resp: 212168
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 19.2 15.4 23.0



#74
Anthracene
Concen: 5.048 ng/ul
RT: 17.616 min Scan# 2453
Delta R.T. -0.000 min
Lab File: BP013962.D
Acq: 07 Mar 2023 15:20

Tgt Ion:178 Resp: 214984
Ion Ratio Lower Upper
178 100
179 15.0 12.5 18.7
176 19.0 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 5.268 ng/uL

RT: 13.534 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:216 Resp: 56257

Ion Ratio Lower Upper

216 100

214 77.6 61.4 92.2

218 51.4 39.1 58.7



#76

Pentachlorobenzene

Concen: 5.512 ng/uL

RT: 15.045 min Scan# 2016

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:250 Resp: 60745


Ion Ratio Lower Upper

250 100

248 62.0 51.0 76.4

252 59.1 50.8 76.2





#77

Carbazole

Concen: 4.941 ng/ul

RT: 17.886 min Scan# 2498

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:167 Resp: 185367

Ion Ratio Lower Upper

167 100

166 22.5 17.5 26.3

139 13.2 10.5 15.7



#78

Di-n-butylphthalate

Concen: 4.865 ng/ul

RT: 18.416 min Scan# 2589

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:149 Resp: 221372


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2

104 5.9 4.6 6.8





#80

Fluoranthene

Concen: 4.597 ng/ul

RT: 19.480 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:202 Resp: 266601

Ion Ratio Lower Upper

202 100

101 7.7 6.6 10.0

100 6.1 5.0 7.4



#82

Pyrene

Concen: 4.599 ng/ul

RT: 19.833 min Scan# 2830

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:202 Resp: 284395


Ion Ratio Lower Upper

202 100

101 9.3 7.8 11.8

100 8.3 6.5 9.7





#83

Butylbenzylphthalate

Concen: 5.048 ng/ul

RT: 20.686 min Scan# 2974

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:149 Resp: 106838

Ion Ratio Lower Upper

149 100

91 69.8 57.9 86.9

206 23.1 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 5.182 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:252 Resp: 104194


Ion Ratio Lower Upper

252 100

254 62.9 51.3 76.9

126 9.5 6.9 10.3





#85

Benzo(a)anthracene

Concen: 4.922 ng/ul

RT: 21.545 min Scan# 3121

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:228 Resp: 300467

Ion Ratio Lower Upper

228 100

229 19.8 15.5 23.3

226 26.0 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.845 ng/ul

RT: 21.445 min Scan# 3104

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:149 Resp: 157022


Ion Ratio Lower Upper

149 100

167 27.9 22.3 33.5

279 5.5 4.3 6.5





#87

Chrysene

Concen: 5.035 ng/ul

RT: 21.604 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:228 Resp: 289983

Ion Ratio Lower Upper

228 100

226 29.1 24.2 36.2

229 19.1 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 3.843 ng/ul




RT: 22.416 min Scan# 3269


Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Tgt Ion:149 Resp: 272387





#90

Benzo(b)fluoranthene

Concen: 4.637 ng/ul

RT: 23.327 min Scan# 3423

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:252 Resp: 318540

Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 7.2 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 4.592 ng/ul

RT: 23.374 min Scan# 3432

Delta R.T. -0.000 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:252 Resp: 312926


Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 7.4 5.5 8.3





#93

Benzo(a)pyrene

Concen: 4.834 ng/ul

RT: 23.998 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:252 Resp: 281429

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.2

125 7.5 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 5.878 ng/ul

RT: 26.792 min Scan# 4013

Delta R.T. 0.012 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20




Tgt Ion:276 Resp: 383833


Ion Ratio Lower Upper

276 100

138 17.5 13.1 19.7

277 24.5 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 5.887 ng/ul

RT: 26.804 min Scan# 4014

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:278 Resp: 321140

Ion Ratio Lower Upper

278 100

139 11.0 8.7 13.1

279 24.5 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 5.952 ng/ul

RT: 27.615 min Scan# 4153

Delta R.T. 0.006 min

Lab File: BP013962.D

Acq: 07 Mar 2023 15:20

Tgt Ion:276 Resp: 309438

Ion Ratio Lower Upper

276 100

138 15.8 11.9 17.9

277 24.0 18.7 28.1

