

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013963.D
 Acq On : 07 Mar 2023 15:54
 Operator : CG/JU
 Sample : SST01036
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010618

Quant Time: Mar 08 00:43:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	110606	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	472698	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.722	164	343292	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	879988	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	979475	20.000	ng/u1	0.00
88) Perylene-d12	24.104	264	1114197	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	12271	4.598	ng/uL	0.00
4) Pyridine-d5	3.875	84	80040	10.117	ng/u1	0.00
7) Phenol-d5	7.234	99	87497	8.879	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	54986	9.522	ng/u1	0.00
11) 2-Chlorophenol-d4	7.611	132	70478	9.715	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	72613	9.015	ng/u1	0.00
21) Nitrobenzene-d5	9.258	128	33894	10.306	ng/u1	0.00
24) 2-Nitrophenol-d4	9.987	143	43476	11.449	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.528	165	82326	10.638	ng/u1	0.00
31) 4-Chloroaniline-d4	11.046	131	107047	9.759	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	291898	10.796	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	299421	10.230	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	46227	9.597	ng/u1	0.00
60) Fluorene-d10	15.716	176	251068	10.902	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.834	200	57997	10.743	ng/u1	0.00
73) Anthracene-d10	17.575	188	419087	10.415	ng/u1	0.00
81) Pyrene-d10	19.798	212	524605	9.445	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.933	264	570673	10.222	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	13289	4.639	ng/uL	94
5) Pyridine	3.893	79	80075	10.242	ng/u1	92
6) Benzaldehyde	7.216	77	53879	12.124	ng/u1	98
8) Phenol	7.263	94	93809	9.273	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.499	93	73458	9.062	ng/u1	99
12) 2-Chlorophenol	7.646	128	71362	9.587	ng/u1	99
13) 2-Methylphenol	8.528	108	69534	9.005	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.605	45	42042	4.560	ng/u1	100
16) Acetophenone	8.916	105	124206	9.615	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.893	70	62841	9.796	ng/u1	99
18) 4-Methylphenol	8.858	108	75380	8.842	ng/u1	97
19) Hexachloroethane	9.175	117	32059	10.518	ng/u1	99
22) Nitrobenzene	9.299	77	92856	10.731	ng/u1	99
23) Isophorone	9.828	82	194251	10.641	ng/u1	99
25) 2-Nitrophenol	10.016	139	44663	10.851	ng/u1	98
26) 2,4-Dimethylphenol	10.069	107	99463	10.854	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.305	93	112898	10.241	ng/u1	98
29) 2,4-Dichlorophenol	10.552	162	82228	10.793	ng/u1	98
30) Naphthalene	10.957	128	254522	10.115	ng/u1	99
32) 4-Chloroaniline	11.069	127	108206	9.785	ng/u1	96
33) Hexachlorobutadiene	11.240	225	64865	12.087	ng/u1	92
34) Caprolactam	11.851	113	24050	8.996	ng/u1	96
35) 4-Chloro-3-methylphenol	12.181	107	93900	10.676	ng/u1	99
36) 2-Methylnaphthalene	12.557	142	182442	10.312	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013963.D
 Acq On : 07 Mar 2023 15:54
 Operator : CG/JU
 Sample : SST01036
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010618

Quant Time: Mar 08 00:43:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.775	142	181904	9.951	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.922	216	121154	11.199	ng/ul	99
40) Hexachlorocyclopentadiene	12.899	237	72990	10.508	ng/ul	98
41) 2,4,6-Trichlorophenol	13.157	196	75715	10.666	ng/ul	97
42) 2,4,5-Trichlorophenol	13.234	196	81593	10.618	ng/ul	99
43) 1,1'-Biphenyl	13.557	154	264063	10.251	ng/ul	98
44) 2-Chloronaphthalene	13.604	162	207514	10.240	ng/ul	98
45) 2-Nitroaniline	13.810	65	57386	11.757	ng/ul	99
47) Dimethylphthalate	14.169	163	288377	10.753	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	54374	11.605	ng/ul	97
50) Acenaphthylene	14.445	152	316930	10.038	ng/ul	99
51) 3-Nitroaniline	14.628	138	48176	10.400	ng/ul	98
52) Acenaphthene	14.787	153	225250	10.192	ng/ul	100
53) 2,4-Dinitrophenol	14.834	184	32092	9.605	ng/ul	95
55) 4-Nitrophenol	14.928	109	51521	12.190	ng/ul	98
56) Dibenzofuran	15.122	168	332122	10.482	ng/ul	99
57) 2,4-Dinitrotoluene	15.081	165	82362	11.791	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.345	232	80745	11.148	ng/ul	97
59) Diethylphthalate	15.528	149	295771	10.983	ng/ul	99
61) Fluorene	15.769	166	277216	10.725	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.763	204	153779	11.442	ng/ul	97
63) 4-Nitroaniline	15.792	138	50428	11.496	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.845	198	56997	10.769	ng/ul#	94
67) N-Nitrosodiphenylamine	15.975	169	244218	10.210	ng/ul	99
68) 4-Bromophenyl-phenylether	16.657	248	102087	10.985	ng/ul	97
69) Hexachlorobenzene	16.775	284	120234	11.104	ng/ul	95
70) Atrazine	16.928	200	102097	10.674	ng/ul	99
71) Pentachlorophenol	17.128	266	71678	10.172	ng/ul	99
72) Phenanthrene	17.522	178	470648	10.107	ng/ul	99
74) Anthracene	17.610	178	475249	10.144	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.528	216	126375	10.758	ng/uL	99
76) Pentachlorobenzene	15.040	250	132226	10.908	ng/uL	98
77) Carbazole	17.881	167	421785	10.220	ng/ul	98
78) Di-n-butylphthalate	18.410	149	518577	10.361	ng/ul	99
80) Fluoranthene	19.475	202	602657	9.478	ng/ul	99
82) Pyrene	19.828	202	637222	9.399	ng/ul	99
83) Butylbenzylphthalate	20.680	149	243996	10.516	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.463	252	247080	11.208	ng/ul	99
85) Benzo(a)anthracene	21.539	228	677835	10.128	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.439	149	372694	10.490	ng/ul	100
87) Chrysene	21.598	228	649817	10.291	ng/ul	99
89) Di-n-octyl phthalate	22.410	149	608490	7.937	ng/ul	100
90) Benzo(b)fluoranthene	23.321	252	708026	9.530	ng/ul	100
91) Benzo(k)fluoranthene	23.369	252	673087	9.132	ng/ul	99
93) Benzo(a)pyrene	23.986	252	622950	9.893	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.780	276	834668	11.817	ng/ul	99
95) Dibenzo(a,h)anthracene	26.792	278	690873	11.710	ng/ul	100
96) Benzo(g,h,i)perylene	27.604	276	673702	11.980	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
Data File : BP013963.D
Acq On : 07 Mar 2023 15:54
Operator : CG/JU
Sample : SST010618
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD010618

Quant Time: Mar 08 00:43:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 00:39:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.639 ng/uL
RT: 3.475 min Scan# 49
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:	88	Resp:	13289
Ion	Ratio	Lower	Upper
88	100		
43	34.7	23.9	35.9
58	65.9	50.2	75.2



#5
Pyridine
Concen: 10.242 ng/uL
RT: 3.893 min Scan# 120
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:	79	Resp:	80075
Ion	Ratio	Lower	Upper
79	100		
52	78.0	55.1	82.7
51	37.3	28.9	43.3









#6
Benzaldehyde
Concen: 12.124 ng/ul
RT: 7.216 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion: 77 Resp: 53879
Ion Ratio Lower Upper
77 100
105 96.1 74.2 111.2
106 87.9 69.8 104.8



#8
Phenol
Concen: 9.273 ng/ul
RT: 7.263 min Scan# 693
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion: 94 Resp: 93809
Ion Ratio Lower Upper
94 100
65 31.2 25.2 37.8
66 45.2 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 9.062 ng/ul
RT: 7.499 min Scan# 733
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:	93	Resp:	73458
Ion	Ratio	Lower	Upper
93	100		
63	72.7	58.8	88.2
95	31.8	25.8	38.6



#12
2-Chlorophenol
Concen: 9.587 ng/ul
RT: 7.646 min Scan# 758
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:	128	Resp:	71362
Ion	Ratio	Lower	Upper
128	100		
64	48.9	38.3	57.5
130	32.3	25.7	38.5





#13

2-Methylphenol

Concen: 9.005 ng/ul

RT: 8.528 min Scan# 908

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :





SSTD010618

Tgt Ion:108 Resp: 69534

Ion Ratio Lower Upper

108 100

107 87.9 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.560 ng/ul

RT: 8.605 min Scan# 921

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion: 45 Resp: 42042


Ion Ratio Lower Upper

45 100

77 17.7 14.2 21.2

79 13.3 10.6 15.8









#16
Acetophenone
Concen: 9.615 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:105 Resp: 124206
Ion Ratio Lower Upper
105 100
77 80.7 65.4 98.2
51 28.3 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 9.796 ng/ul
RT: 8.893 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion: 70 Resp: 62841
Ion Ratio Lower Upper
70 100
42 51.3 40.9 61.3
101 10.8 7.9 11.9
130 19.9 16.6 25.0









#18
4-Methylphenol
Concen: 8.842 ng/ul
RT: 8.858 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:108 Resp: 75380
Ion Ratio Lower Upper
108 100
107 109.1 89.8 134.6



#19
Hexachloroethane
Concen: 10.518 ng/ul
RT: 9.175 min Scan# 1018
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:117 Resp: 32059
Ion Ratio Lower Upper
117 100
201 114.8 91.7 137.5
199 76.4 59.1 88.7









#22
Nitrobenzene
Concen: 10.731 ng/ul
RT: 9.299 min Scan# 1039
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion: 77	Resp: 92856
Ion Ratio	Lower Upper
77 100	
123 39.9	32.6 48.8
65 13.6	11.2 16.8



#23
Isophorone
Concen: 10.641 ng/ul
RT: 9.828 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion: 82	Resp: 194251
Ion Ratio	Lower Upper
82 100	
95 8.3	7.1 10.7
138 17.0	13.3 19.9









#25
2-Nitrophenol
Concen: 10.851 ng/ul
RT: 10.016 min Scan# 1161
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:139 Resp: 44663
Ion Ratio Lower Upper
139 100
65 58.5 46.6 70.0
109 45.2 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 10.854 ng/ul
RT: 10.069 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:107 Resp: 99463
Ion Ratio Lower Upper
107 100
121 47.6 37.1 55.7
122 77.4 61.2 91.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.241 ng/ul
 RT: 10.305 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BP013963.D
 Acq: 07 Mar 2023 15:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010618


Tgt Ion: 93 Resp: 112898
 Ion Ratio Lower Upper
 93 100
 95 33.4 25.6 38.4
 123 10.7 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 10.793 ng/ul
 RT: 10.552 min Scan# 1252
 Delta R.T. -0.000 min
 Lab File: BP013963.D
 Acq: 07 Mar 2023 15:54

Tgt Ion: 162 Resp: 82228
 Ion Ratio Lower Upper
 162 100
 164 62.9 52.0 78.0
 98 35.7 29.1 43.7









#30
Naphthalene
Concen: 10.115 ng/ul
RT: 10.957 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:128 Resp: 254522
Ion Ratio Lower Upper
128 100
129 10.5 8.5 12.7
127 13.4 11.0 16.4



#32
4-Chloroaniline
Concen: 9.785 ng/ul
RT: 11.069 min Scan# 1340
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:127 Resp: 108206
Ion Ratio Lower Upper
127 100
129 33.5 24.9 37.3









#33
Hexachlorobutadiene
Concen: 12.087 ng/ul
RT: 11.240 min Scan# 1369
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:225 Resp: 64865
Ion Ratio Lower Upper
225 100
223 67.4 48.0 72.0
227 68.1 50.3 75.5



#34
Caprolactam
Concen: 8.996 ng/ul
RT: 11.851 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:113 Resp: 24050
Ion Ratio Lower Upper
113 100
55 147.9 124.0 186.0
56 124.9 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 10.676 ng/ul
RT: 12.181 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:107 Resp: 93900
Ion Ratio Lower Upper
107 100
144 24.4 19.4 29.0
142 73.4 59.6 89.4



#36
2-Methylnaphthalene
Concen: 10.312 ng/ul
RT: 12.557 min Scan# 1593
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:142 Resp: 182442
Ion Ratio Lower Upper
142 100
141 89.1 71.7 107.5





#37

1-Methylnaphthalene

Concen: 9.951 ng/ul

RT: 12.775 min Scan# 1630

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:142 Resp: 181904

Ion Ratio Lower Upper

142 100

141 93.8 73.8 110.6

116 4.4 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 11.199 ng/ul

RT: 12.922 min Scan# 1655

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Tgt Ion:216 Resp: 121154




Ion Ratio Lower Upper


216 100

214 76.6 61.7 92.5

179 20.9 16.2 24.4

108 16.6 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 10.508 ng/ul

RT: 12.899 min Scan# 1651

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:237 Resp: 72990

Ion Ratio Lower Upper

237 100

235 62.2 49.0 73.6

272 13.7 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 10.666 ng/ul

RT: 13.157 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:196 Resp: 75715


Ion Ratio Lower Upper

196 100

198 101.0 78.5 117.7

200 29.2 25.4 38.0









#42
2,4,5-Trichlorophenol
Concen: 10.618 ng/ul
RT: 13.234 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:196 Resp: 81593
Ion Ratio Lower Upper
196 100
198 97.8 77.6 116.4
200 32.4 25.3 37.9



#43
1,1'-Biphenyl
Concen: 10.251 ng/ul
RT: 13.557 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:154 Resp: 264063
Ion Ratio Lower Upper
154 100
153 39.9 32.6 49.0
76 10.8 9.4 14.0









#44
2-Chloronaphthalene
Concen: 10.240 ng/ul
RT: 13.604 min Scan# 1771
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:162 Resp: 207514
Ion Ratio Lower Upper
162 100
164 34.0 26.0 39.0
127 40.2 31.3 46.9



#45
2-Nitroaniline
Concen: 11.757 ng/ul
RT: 13.810 min Scan# 1806
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion: 65 Resp: 57386
Ion Ratio Lower Upper
65 100
92 70.2 56.3 84.5
138 100.7 81.4 122.2









#47
Dimethylphthalate
Concen: 10.753 ng/ul
RT: 14.169 min Scan# 1867
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:163 Resp: 288377
Ion Ratio Lower Upper
163 100
194 4.6 3.6 5.4
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 11.605 ng/ul
RT: 14.298 min Scan# 1889
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:165 Resp: 54374
Ion Ratio Lower Upper
165 100
89 62.3 51.7 77.5
121 20.6 17.5 26.3





#50

Acenaphthylene

Concen: 10.038 ng/ul

RT: 14.445 min Scan# 1915

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:152 Resp: 316930

Ion Ratio Lower Upper

152 100

151 19.9 16.2 24.2

153 12.5 10.4 15.6



#51

3-Nitroaniline

Concen: 10.400 ng/ul

RT: 14.628 min Scan# 1945

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:138 Resp: 48176


Ion Ratio Lower Upper

138 100

108 12.4 10.5 15.7

92 124.8 98.5 147.7









#52
Acenaphthene
Concen: 10.192 ng/ul
RT: 14.787 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:153 Resp: 225250
Ion Ratio Lower Upper
153 100
152 47.2 37.7 56.5
154 87.6 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 9.605 ng/ul
RT: 14.834 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:184 Resp: 32092
Ion Ratio Lower Upper
184 100
63 67.8 51.9 77.9
154 69.1 58.8 88.2









#55
4-Nitrophenol
Concen: 12.190 ng/ul
RT: 14.928 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:109 Resp: 51521
Ion Ratio Lower Upper
109 100
139 90.8 74.5 111.7
65 107.4 86.6 130.0



#56
Dibenzofuran
Concen: 10.482 ng/ul
RT: 15.122 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:168 Resp: 332122
Ion Ratio Lower Upper
168 100
139 40.1 31.8 47.6





#57

2,4-Dinitrotoluene

Concen: 11.791 ng/ul

RT: 15.081 min Scan# 2023

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:165 Resp: 82362

Ion Ratio Lower Upper

165 100

63 48.8 34.6 52.0

182 2.8 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 11.148 ng/ul

RT: 15.345 min Scan# 2067

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Tgt Ion:232 Resp: 80745




Ion Ratio Lower Upper


232 100

131 44.3 34.1 51.1

130 2.5 2.1 3.1

166 30.8 23.0 34.6









#59
Diethylphthalate
Concen: 10.983 ng/ul
RT: 15.528 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:149 Resp: 295771
Ion Ratio Lower Upper
149 100
177 23.2 18.2 27.2
150 12.0 10.0 15.0



#61
Fluorene
Concen: 10.725 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:166 Resp: 277216
Ion Ratio Lower Upper
166 100
165 101.0 78.9 118.3
167 14.0 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 11.442 ng/ul
RT: 15.763 min Scan# 2137
Delta R.T. 0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:204 Resp: 153779
Ion Ratio Lower Upper
204 100
206 32.7 26.6 39.8
141 58.4 49.2 73.8



#63
4-Nitroaniline
Concen: 11.496 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:138 Resp: 50428
Ion Ratio Lower Upper
138 100
92 60.2 44.5 66.7
108 104.8 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 10.769 ng/ul
RT: 15.845 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:198 Resp: 56997
Ion Ratio Lower Upper
198 100
182 5.2 3.0 4.6#
77 28.5 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 10.210 ng/ul
RT: 15.975 min Scan# 2174
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:169 Resp: 244218
Ion Ratio Lower Upper
169 100
168 67.7 53.7 80.5
167 36.6 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 10.985 ng/ul
RT: 16.657 min Scan# 210
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:248 Resp: 102087
Ion Ratio Lower Upper
248 100
250 95.8 78.8 118.2
141 73.1 60.8 91.2



#69
Hexachlorobenzene
Concen: 11.104 ng/ul
RT: 16.775 min Scan# 2310
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:284 Resp: 120234
Ion Ratio Lower Upper
284 100
142 33.3 23.2 34.8
249 31.1 25.7 38.5









#70
Atrazine
Concen: 10.674 ng/ul
RT: 16.928 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:200 Resp: 102097
Ion Ratio Lower Upper
200 100
173 24.2 19.0 28.4
215 47.7 38.5 57.7



#71
Pentachlorophenol
Concen: 10.172 ng/ul
RT: 17.128 min Scan# 2370
Delta R.T. 0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:266 Resp: 71678
Ion Ratio Lower Upper
266 100
264 59.7 48.4 72.6
268 63.0 50.4 75.6









#72
Phenanthrene
Concen: 10.107 ng/ul
RT: 17.522 min Scan# 2437
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:178 Resp: 470648
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.4
176 19.7 15.4 23.0



#74
Anthracene
Concen: 10.144 ng/ul
RT: 17.610 min Scan# 2452
Delta R.T. -0.006 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:178 Resp: 475249
Ion Ratio Lower Upper
178 100
179 15.7 12.5 18.7
176 18.3 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.758 ng/uL
RT: 13.528 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:216 Resp: 126375
Ion Ratio Lower Upper
216 100
214 75.3 61.4 92.2
218 48.4 39.1 58.7



#76
Pentachlorobenzene
Concen: 10.908 ng/uL
RT: 15.040 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:250 Resp: 132226
Ion Ratio Lower Upper
250 100
248 64.8 51.0 76.4
252 65.3 50.8 76.2









#77
 Carbazole
 Concen: 10.220 ng/ul
 RT: 17.881 min Scan# 2498
 Delta R.T. -0.000 min
 Lab File: BP013963.D
 Acq: 07 Mar 2023 15:54




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010618


Tgt Ion:167 Resp: 421785
 Ion Ratio Lower Upper
 167 100
 166 22.8 17.5 26.3
 139 12.7 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 10.361 ng/ul
 RT: 18.410 min Scan# 2588
 Delta R.T. -0.000 min
 Lab File: BP013963.D
 Acq: 07 Mar 2023 15:54

Tgt Ion:149 Resp: 518577
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.4 11.2
 104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 9.478 ng/ul
RT: 19.475 min Scan# 2769
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54




Instrument :
BNA_P
ClientSampleId :
SSTD010618


Tgt Ion:202 Resp: 602657
Ion Ratio Lower Upper
202 100
101 8.1 6.6 10.0
100 6.6 5.0 7.4



#82
Pyrene
Concen: 9.399 ng/ul
RT: 19.828 min Scan# 2829
Delta R.T. -0.000 min
Lab File: BP013963.D
Acq: 07 Mar 2023 15:54

Tgt Ion:202 Resp: 637222
Ion Ratio Lower Upper
202 100
101 9.5 7.8 11.8
100 7.7 6.5 9.7





#83

Butylbenzylphthalate

Concen: 10.516 ng/ul

RT: 20.680 min Scan# 2974

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:149 Resp: 243996

Ion Ratio Lower Upper

149 100

91 71.5 57.9 86.9

206 22.5 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 11.208 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:252 Resp: 247080


Ion Ratio Lower Upper

252 100

254 64.7 51.3 76.9

126 8.8 6.9 10.3





#85

Benzo(a)anthracene

Concen: 10.128 ng/ul

RT: 21.539 min Scan# 3121

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:228 Resp: 677835

Ion Ratio Lower Upper

228 100

229 19.5 15.5 23.3

226 26.2 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 10.490 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:149 Resp: 372694


Ion Ratio Lower Upper

149 100

167 27.8 22.3 33.5

279 5.7 4.3 6.5





#87

Chrysene

Concen: 10.291 ng/ul

RT: 21.598 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:228 Resp: 649817

Ion Ratio Lower Upper

228 100

226 29.5 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 7.937 ng/ul




RT: 22.410 min Scan# 3268


Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Tgt Ion:149 Resp: 608490





#90

Benzo(b)fluoranthene

Concen: 9.530 ng/ul

RT: 23.321 min Scan# 3423

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:252 Resp: 708026

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 6.9 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 9.132 ng/ul

RT: 23.369 min Scan# 3431

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:252 Resp: 673087


Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0

125 7.1 5.5 8.3





#93

Benzo(a)pyrene

Concen: 9.893 ng/ul

RT: 23.986 min Scan# 3536

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:252 Resp: 622950

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 7.4 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 11.817 ng/ul

RT: 26.780 min Scan# 4011

Delta R.T. -0.000 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54




Tgt Ion:276 Resp: 834668


Ion Ratio Lower Upper

276 100

138 16.1 13.1 19.7

277 25.7 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 11.710 ng/ul

RT: 26.792 min Scan# 4014

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Instrument :

BNA_P

ClientSampleId :

SSTD010618





Tgt Ion:278 Resp: 690873

Ion Ratio Lower Upper

278 100

139 11.0 8.7 13.1

279 24.0 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 11.980 ng/ul

RT: 27.604 min Scan# 4151

Delta R.T. -0.006 min

Lab File: BP013963.D

Acq: 07 Mar 2023 15:54

Tgt Ion:276 Resp: 673702

Ion Ratio Lower Upper

276 100

138 14.5 11.9 17.9

277 22.9 18.7 28.1

