

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013964.D
 Acq On : 07 Mar 2023 16:29
 Operator : CG/JU
 Sample : SSTD02037
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020619

Quant Time: Mar 08 00:43:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	122329	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	517604	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.722	164	361554	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	873571	20.000	ng/u1	0.00
79) Chrysene-d12	21.563	240	999709	20.000	ng/u1	0.00
88) Perylene-d12	24.104	264	1136410	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	26397	8.944	ng/uL	0.00
4) Pyridine-d5	3.870	84	182005	20.801	ng/u1	0.00
7) Phenol-d5	7.234	99	219532	20.142	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	133701	20.934	ng/u1	0.00
11) 2-Chlorophenol-d4	7.611	132	174367	21.732	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	185227	20.792	ng/u1	0.00
21) Nitrobenzene-d5	9.258	128	89149	24.754	ng/u1	0.00
24) 2-Nitrophenol-d4	9.987	143	103729	24.946	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	196091	23.141	ng/u1	0.00
31) 4-Chloroaniline-d4	11.046	131	255551	21.276	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	660676	23.200	ng/u1	0.00
49) Acenaphthylene-d8	14.422	160	701251	22.749	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	113625	22.397	ng/u1	0.00
60) Fluorene-d10	15.716	176	554575	22.864	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.834	200	138004	25.750	ng/u1	0.00
73) Anthracene-d10	17.581	188	894780	22.400	ng/u1	0.00
81) Pyrene-d10	19.798	212	1147901	20.248	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.939	264	1282377	22.521	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	28798	9.089	ng/uL	100
5) Pyridine	3.893	79	186103	21.522	ng/u1	100
6) Benzaldehyde	7.216	77	134351	27.334	ng/u1	100
8) Phenol	7.264	94	227197	20.307	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.499	93	180729	20.159	ng/u1	100
12) 2-Chlorophenol	7.646	128	175912	21.368	ng/u1	100
13) 2-Methylphenol	8.528	108	175603	20.563	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.611	45	203615	19.966	ng/u1	100
16) Acetophenone	8.916	105	312847	21.898	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.899	70	159901	22.538	ng/u1	100
18) 4-Methylphenol	8.858	108	194264	20.604	ng/u1	100
19) Hexachloroethane	9.175	117	76593	22.720	ng/u1	100
22) Nitrobenzene	9.299	77	236006	24.909	ng/u1	100
23) Isophorone	9.828	82	452032	22.613	ng/u1	100
25) 2-Nitrophenol	10.016	139	107474	23.847	ng/u1	100
26) 2,4-Dimethylphenol	10.069	107	237360	23.654	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.310	93	256457	21.245	ng/u1	100
29) 2,4-Dichlorophenol	10.552	162	195424	23.426	ng/u1	100
30) Naphthalene	10.963	128	592902	21.518	ng/u1	100
32) 4-Chloroaniline	11.069	127	255201	21.076	ng/u1	100
33) Hexachlorobutadiene	11.240	225	151998	25.866	ng/u1	100
34) Caprolactam	11.852	113	60229	20.575	ng/u1	100
35) 4-Chloro-3-methylphenol	12.181	107	223562	23.213	ng/u1	100
36) 2-Methylnaphthalene	12.557	142	419272	21.643	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013964.D
 Acq On : 07 Mar 2023 16:29
 Operator : CG/JU
 Sample : SST02037
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020619

Quant Time: Mar 08 00:43:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 00:39:16 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.775	142	419947	20.980	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.922	216	280310	24.603	ng/ul	100
40) Hexachlorocyclopentadiene	12.899	237	187367	25.612	ng/ul	100
41) 2,4,6-Trichlorophenol	13.157	196	176425	23.597	ng/ul	100
42) 2,4,5-Trichlorophenol	13.234	196	196813	24.319	ng/ul	100
43) 1,1'-Biphenyl	13.557	154	597778	22.035	ng/ul	100
44) 2-Chloronaphthalene	13.604	162	471150	22.076	ng/ul	100
45) 2-Nitroaniline	13.810	65	144514	28.113	ng/ul	100
47) Dimethylphthalate	14.175	163	651937	23.082	ng/ul	100
48) 2,6-Dinitrotoluene	14.298	165	130719	26.490	ng/ul	100
50) Acenaphthylene	14.451	152	735177	22.108	ng/ul	100
51) 3-Nitroaniline	14.628	138	114579	23.485	ng/ul	100
52) Acenaphthene	14.787	153	504395	21.671	ng/ul	100
53) 2,4-Dinitrophenol	14.834	184	88392	25.119	ng/ul	100
55) 4-Nitrophenol	14.928	109	122316	27.477	ng/ul	100
56) Dibenzofuran	15.122	168	739039	22.147	ng/ul	100
57) 2,4-Dinitrotoluene	15.087	165	194139	26.389	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.345	232	180204	23.624	ng/ul	100
59) Diethylphthalate	15.534	149	657448	23.181	ng/ul	100
61) Fluorene	15.769	166	613659	22.542	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.757	204	338431	23.909	ng/ul	100
63) 4-Nitroaniline	15.793	138	124649	26.980	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.851	198	130819	24.898	ng/ul	100
67) N-Nitrosodiphenylamine	15.975	169	532033	22.407	ng/ul	100
68) 4-Bromophenyl-phenylether	16.657	248	224210	24.304	ng/ul	100
69) Hexachlorobenzene	16.781	284	266303	24.774	ng/ul	100
70) Atrazine	16.928	200	229324	24.151	ng/ul	100
71) Pentachlorophenol	17.122	266	166150	23.752	ng/ul	100
72) Phenanthrene	17.522	178	1003340	21.705	ng/ul	100
74) Anthracene	17.616	178	1027012	22.082	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.528	216	289065	24.788	ng/ul	100
76) Pentachlorobenzene	15.040	250	296864	24.669	ng/ul	100
77) Carbazole	17.881	167	911345	22.245	ng/ul	100
78) Di-n-butylphthalate	18.410	149	1145006	23.044	ng/ul	100
80) Fluoranthene	19.475	202	1318717	20.319	ng/ul	100
82) Pyrene	19.828	202	1374324	19.861	ng/ul	100
83) Butylbenzylphthalate	20.680	149	560073	23.650	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.469	252	552096	24.537	ng/ul	100
85) Benzo(a)anthracene	21.545	228	1471976	21.549	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.439	149	848586	23.400	ng/ul	100
87) Chrysene	21.598	228	1405150	21.803	ng/ul	100
89) Di-n-octyl phthalate	22.410	149	1460127	18.672	ng/ul	100
90) Benzo(b)fluoranthene	23.322	252	1574276	20.775	ng/ul	100
91) Benzo(k)fluoranthene	23.374	252	1557287	20.715	ng/ul	100
93) Benzo(a)pyrene	23.992	252	1392040	21.674	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.780	276	1852845	25.720	ng/ul	100
95) Dibenzo(a,h)anthracene	26.798	278	1535491	25.516	ng/ul	100
96) Benzo(g,h,i)perylene	27.610	276	1488484	25.952	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
Data File : BP013964.D
Acq On : 07 Mar 2023 16:29
Operator : CG/JU
Sample : SST02037
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020619

Quant Time: Mar 08 00:43:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 00:39:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.089 ng/uL
RT: 3.475 min Scan# 49
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion: 88 Resp: 28798
Ion Ratio Lower Upper
88 100
43 29.9 23.9 35.9
58 62.7 50.2 75.2



#5
Pyridine
Concen: 21.522 ng/uL
RT: 3.893 min Scan# 120
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion: 79 Resp: 186103
Ion Ratio Lower Upper
79 100
52 68.9 55.1 82.7
51 36.1 28.9 43.3









#6
Benzaldehyde
Concen: 27.334 ng/ul
RT: 7.216 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion: 77 Resp: 134351
Ion Ratio Lower Upper
77 100
105 92.7 74.2 111.2
106 87.3 69.8 104.8



#8
Phenol
Concen: 20.307 ng/ul
RT: 7.264 min Scan# 693
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion: 94 Resp: 227197
Ion Ratio Lower Upper
94 100
65 31.5 25.2 37.8
66 46.5 37.2 55.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.159 ng/ul
 RT: 7.499 min Scan# 733
 Delta R.T. 0.000 min
 Lab File: BP013964.D
 Acq: 07 Mar 2023 16:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020619


Tgt Ion: 93 Resp: 180729
 Ion Ratio Lower Upper
 93 100
 63 73.5 58.8 88.2
 95 32.2 25.8 38.6



#12
 2-Chlorophenol
 Concen: 21.368 ng/ul
 RT: 7.646 min Scan# 758
 Delta R.T. 0.000 min
 Lab File: BP013964.D
 Acq: 07 Mar 2023 16:29

Tgt Ion: 128 Resp: 175912
 Ion Ratio Lower Upper
 128 100
 64 47.9 38.3 57.5
 130 32.1 25.7 38.5





#13

2-Methylphenol

Concen: 20.563 ng/ul

RT: 8.528 min Scan# 908

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :





SSTD020619

Tgt Ion:108 Resp: 175603

Ion Ratio Lower Upper

108 100

107 88.1 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.966 ng/ul

RT: 8.611 min Scan# 922

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion: 45 Resp: 203615


Ion Ratio Lower Upper

45 100

77 17.7 14.2 21.2

79 13.2 10.6 15.8









#16
Acetophenone
Concen: 21.898 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:105 Resp: 312847
Ion Ratio Lower Upper
105 100
77 81.8 65.4 98.2
51 27.3 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.538 ng/ul
RT: 8.899 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion: 70 Resp: 159901
Ion Ratio Lower Upper
70 100
42 51.1 40.9 61.3
101 9.9 7.9 11.9
130 20.8 16.6 25.0









#18
4-Methylphenol
Concen: 20.604 ng/ul
RT: 8.858 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:108 Resp: 194264
Ion Ratio Lower Upper
108 100
107 112.2 89.8 134.6



#19
Hexachloroethane
Concen: 22.720 ng/ul
RT: 9.175 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:117 Resp: 76593
Ion Ratio Lower Upper
117 100
201 114.6 91.7 137.5
199 73.9 59.1 88.7









#22
Nitrobenzene
Concen: 24.909 ng/ul
RT: 9.299 min Scan# 1039
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion: 77 Resp: 236006
Ion Ratio Lower Upper
77 100
123 40.7 32.6 48.8
65 14.0 11.2 16.8



#23
Isophorone
Concen: 22.613 ng/ul
RT: 9.828 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion: 82 Resp: 452032
Ion Ratio Lower Upper
82 100
95 8.9 7.1 10.7
138 16.6 13.3 19.9









#25
2-Nitrophenol
Concen: 23.847 ng/ul
RT: 10.016 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:139 Resp: 107474
Ion Ratio Lower Upper
139 100
65 58.3 46.6 70.0
109 43.1 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 23.654 ng/ul
RT: 10.069 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:107 Resp: 237360
Ion Ratio Lower Upper
107 100
121 46.4 37.1 55.7
122 76.5 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.245 ng/ul

RT: 10.310 min Scan# 1211

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion: 93 Resp: 256457

Ion Ratio Lower Upper

93 100

95 32.0 25.6 38.4

123 10.6 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 23.426 ng/ul

RT: 10.552 min Scan# 1252

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:162 Resp: 195424


Ion Ratio Lower Upper

162 100

164 65.0 52.0 78.0

98 36.4 29.1 43.7









#30
Naphthalene
Concen: 21.518 ng/ul
RT: 10.963 min Scan# 1322
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:128 Resp: 592902
Ion Ratio Lower Upper
128 100
129 10.6 8.5 12.7
127 13.7 11.0 16.4



#32
4-Chloroaniline
Concen: 21.076 ng/ul
RT: 11.069 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:127 Resp: 255201
Ion Ratio Lower Upper
127 100
129 31.1 24.9 37.3





#33

Hexachlorobutadiene

Concen: 25.866 ng/ul

RT: 11.240 min Scan# 1369

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:225 Resp: 151998

Ion Ratio Lower Upper

225 100

223 60.0 48.0 72.0

227 62.9 50.3 75.5



#34

Caprolactam

Concen: 20.575 ng/ul

RT: 11.852 min Scan# 1473

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:113 Resp: 60229


Ion Ratio Lower Upper

113 100

55 155.0 124.0 186.0

56 127.5 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 23.213 ng/ul
RT: 12.181 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:107 Resp: 223562
Ion Ratio Lower Upper
107 100
144 24.2 19.4 29.0
142 74.5 59.6 89.4



#36
2-Methylnaphthalene
Concen: 21.643 ng/ul
RT: 12.557 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:142 Resp: 419272
Ion Ratio Lower Upper
142 100
141 89.6 71.7 107.5





#37

1-Methylnaphthalene

Concen: 20.980 ng/ul

RT: 12.775 min Scan# 1630

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:142 Resp: 419947

Ion Ratio Lower Upper

142 100

141 92.2 73.8 110.6

116 4.6 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.603 ng/ul

RT: 12.922 min Scan# 1655

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Tgt Ion:216 Resp: 280310




Ion Ratio Lower Upper


216 100

214 77.1 61.7 92.5

179 20.3 16.2 24.4

108 16.5 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 25.612 ng/ul

RT: 12.899 min Scan# 1651

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:237 Resp: 187367

Ion Ratio Lower Upper

237 100

235 61.3 49.0 73.6

272 12.0 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 23.597 ng/ul

RT: 13.157 min Scan# 1695

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:196 Resp: 176425


Ion Ratio Lower Upper

196 100

198 98.1 78.5 117.7

200 31.7 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 24.319 ng/ul

RT: 13.234 min Scan# 1708

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:196 Resp: 196813

Ion Ratio Lower Upper

196 100

198 97.0 77.6 116.4

200 31.6 25.3 37.9



#43

1,1'-Biphenyl

Concen: 22.035 ng/ul

RT: 13.557 min Scan# 1763

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:154 Resp: 597778


Ion Ratio Lower Upper

154 100

153 40.8 32.6 49.0

76 11.7 9.4 14.0









#44
2-Chloronaphthalene
Concen: 22.076 ng/ul
RT: 13.604 min Scan# 1771
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:162 Resp: 471150
Ion Ratio Lower Upper
162 100
164 32.5 26.0 39.0
127 39.1 31.3 46.9



#45
2-Nitroaniline
Concen: 28.113 ng/ul
RT: 13.810 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion: 65 Resp: 144514
Ion Ratio Lower Upper
65 100
92 70.4 56.3 84.5
138 101.8 81.4 122.2









#47
Dimethylphthalate
Concen: 23.082 ng/ul
RT: 14.175 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:163 Resp: 651937
Ion Ratio Lower Upper
163 100
194 4.5 3.6 5.4
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 26.490 ng/ul
RT: 14.298 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:165 Resp: 130719
Ion Ratio Lower Upper
165 100
89 64.6 51.7 77.5
121 21.9 17.5 26.3









#50
Acenaphthylene
Concen: 22.108 ng/ul
RT: 14.451 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:152 Resp: 735177
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 23.485 ng/ul
RT: 14.628 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:138 Resp: 114579
Ion Ratio Lower Upper
138 100
108 13.1 10.5 15.7
92 123.1 98.5 147.7





#52

Acenaphthene

Concen: 21.671 ng/ul

RT: 14.787 min Scan# 1972

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:153 Resp: 504395

Ion Ratio Lower Upper

153 100

152 47.1 37.7 56.5

154 88.0 70.4 105.6



#53

2,4-Dinitrophenol

Concen: 25.119 ng/ul

RT: 14.834 min Scan# 1980

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:184 Resp: 88392


Ion Ratio Lower Upper

184 100

63 64.9 51.9 77.9

154 73.5 58.8 88.2









#55
4-Nitrophenol
Concen: 27.477 ng/ul
RT: 14.928 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:109 Resp: 122316
Ion Ratio Lower Upper
109 100
139 93.1 74.5 111.7
65 108.3 86.6 130.0



#56
Dibenzofuran
Concen: 22.147 ng/ul
RT: 15.122 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:168 Resp: 739039
Ion Ratio Lower Upper
168 100
139 39.7 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 26.389 ng/ul
RT: 15.087 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:165 Resp: 194139
Ion Ratio Lower Upper
165 100
63 43.3 34.6 52.0
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.624 ng/ul
RT: 15.345 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:232 Resp: 180204
Ion Ratio Lower Upper
232 100
131 42.6 34.1 51.1
130 2.6 2.1 3.1
166 28.8 23.0 34.6









#59
Diethylphthalate
Concen: 23.181 ng/ul
RT: 15.534 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:149 Resp: 657448
Ion Ratio Lower Upper
149 100
177 22.7 18.2 27.2
150 12.5 10.0 15.0



#61
Fluorene
Concen: 22.542 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:166 Resp: 613659
Ion Ratio Lower Upper
166 100
165 98.6 78.9 118.3
167 13.8 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 23.909 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:204 Resp: 338431
Ion Ratio Lower Upper
204 100
206 33.2 26.6 39.8
141 61.5 49.2 73.8



#63
4-Nitroaniline
Concen: 26.980 ng/ul
RT: 15.793 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:138 Resp: 124649
Ion Ratio Lower Upper
138 100
92 55.6 44.5 66.7
108 109.0 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 24.898 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:198 Resp: 130819
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.6
77 25.4 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 22.407 ng/ul
RT: 15.975 min Scan# 2174
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:169 Resp: 532033
Ion Ratio Lower Upper
169 100
168 67.1 53.7 80.5
167 35.8 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 24.304 ng/ul
RT: 16.657 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020619


Tgt Ion:248 Resp: 224210
Ion Ratio Lower Upper
248 100
250 98.5 78.8 118.2
141 76.0 60.8 91.2



#69
Hexachlorobenzene
Concen: 24.774 ng/ul
RT: 16.781 min Scan# 2311
Delta R.T. 0.000 min
Lab File: BP013964.D
Acq: 07 Mar 2023 16:29

Tgt Ion:284 Resp: 266303
Ion Ratio Lower Upper
284 100
142 29.0 23.2 34.8
249 32.1 25.7 38.5





#70

Atrazine

Concen: 24.151 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:200 Resp: 229324

Ion Ratio Lower Upper

200 100

173 23.7 19.0 28.4

215 48.1 38.5 57.7



#71

Pentachlorophenol

Concen: 23.752 ng/ul

RT: 17.122 min Scan# 2369

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:266 Resp: 166150


Ion Ratio Lower Upper

266 100

264 60.5 48.4 72.6

268 63.0 50.4 75.6





#72

Phenanthrene

Concen: 21.705 ng/ul

RT: 17.522 min Scan# 2437

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:178 Resp: 1003340

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.2 15.4 23.0



#74

Anthracene

Concen: 22.082 ng/ul

RT: 17.616 min Scan# 2453

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:178 Resp: 1027012


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.9 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.788 ng/uL

RT: 13.528 min Scan# 1758

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:216 Resp: 289065

Ion Ratio Lower Upper

216 100

214 76.8 61.4 92.2

218 48.9 39.1 58.7



#76

Pentachlorobenzene

Concen: 24.669 ng/uL

RT: 15.040 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:250 Resp: 296864


Ion Ratio Lower Upper

250 100

248 63.7 51.0 76.4

252 63.5 50.8 76.2









#77
 Carbazole
 Concen: 22.245 ng/ul
 RT: 17.881 min Scan# 2498
 Delta R.T. 0.000 min
 Lab File: BP013964.D
 Acq: 07 Mar 2023 16:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020619


Tgt Ion:167 Resp: 911345
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 13.1 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 23.044 ng/ul
 RT: 18.410 min Scan# 2588
 Delta R.T. 0.000 min
 Lab File: BP013964.D
 Acq: 07 Mar 2023 16:29

Tgt Ion:149 Resp: 1145006
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.7 4.6 6.8





#80

Fluoranthene

Concen: 20.319 ng/ul

RT: 19.475 min Scan# 2769

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:202 Resp: 1318717

Ion Ratio Lower Upper

202 100

101 8.3 6.6 10.0

100 6.2 5.0 7.4



#82

Pyrene

Concen: 19.861 ng/ul

RT: 19.828 min Scan# 2829

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:202 Resp: 1374324


Ion Ratio Lower Upper

202 100

101 9.8 7.8 11.8

100 8.1 6.5 9.7





#83

Butylbenzylphthalate

Concen: 23.650 ng/ul

RT: 20.680 min Scan# 2974

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:149 Resp: 560073

Ion Ratio Lower Upper

149 100

91 72.4 57.9 86.9

206 22.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 24.537 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:252 Resp: 552096


Ion Ratio Lower Upper

252 100

254 64.1 51.3 76.9

126 8.6 6.9 10.3





#85

Benzo(a)anthracene

Concen: 21.549 ng/ul

RT: 21.545 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:228 Resp: 1471976

Ion Ratio Lower Upper

228 100

229 19.4 15.5 23.3

226 27.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.400 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:149 Resp: 848586


Ion Ratio Lower Upper

149 100

167 27.9 22.3 33.5

279 5.4 4.3 6.5





#87

Chrysene

Concen: 21.803 ng/ul

RT: 21.598 min Scan# 3130

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:228 Resp: 1405150

Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 18.672 ng/ul




RT: 22.410 min Scan# 3268


Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Tgt Ion:149 Resp: 1460127





#90

Benzo(b)fluoranthene

Concen: 20.775 ng/ul

RT: 23.322 min Scan# 3423

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:252 Resp: 1574276

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 6.8 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 20.715 ng/ul

RT: 23.374 min Scan# 3432

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:252 Resp: 1557287


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 6.9 5.5 8.3





#93

Benzo(a)pyrene

Concen: 21.674 ng/ul

RT: 23.992 min Scan# 3537

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:252 Resp: 1392040

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 7.7 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 25.720 ng/ul

RT: 26.780 min Scan# 4011

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29




Tgt Ion:276 Resp: 1852845


Ion Ratio Lower Upper

276 100

138 16.4 13.1 19.7

277 25.2 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 25.516 ng/ul

RT: 26.798 min Scan# 4014

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020619





Tgt Ion:278 Resp: 1535491

Ion Ratio Lower Upper

278 100

139 10.9 8.7 13.1

279 23.7 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 25.952 ng/ul

RT: 27.610 min Scan# 4152

Delta R.T. 0.000 min

Lab File: BP013964.D

Acq: 07 Mar 2023 16:29

Tgt Ion:276 Resp: 1488484

Ion Ratio Lower Upper

276 100

138 14.9 11.9 17.9

277 23.4 18.7 28.1

