

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013968.D
 Acq On : 07 Mar 2023 18:46
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV623

Quant Time: Mar 08 01:25:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 01:24:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.087	152	125251	20.000	ng/u1	0.00
20) Naphthalene-d8	10.910	136	503234	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.722	164	364993	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.481	188	956627	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	1164878	20.000	ng/u1	0.00
88) Perylene-d12	24.098	264	1315867	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	23182	6.993	ng/uL	0.00
4) Pyridine-d5	3.870	84	157083	17.114	ng/u1	0.00
7) Phenol-d5	7.234	99	194184	17.568	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	130060	20.176	ng/u1	0.00
11) 2-Chlorophenol-d4	7.611	132	152436	18.102	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	163735	18.135	ng/u1	0.00
21) Nitrobenzene-d5	9.258	128	78770	19.618	ng/u1	0.00
24) 2-Nitrophenol-d4	9.987	143	92060	19.153	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.528	165	172329	19.055	ng/u1	0.00
31) 4-Chloroaniline-d4	11.046	131	212308	17.739	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	575638	18.583	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	618758	18.857	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	104464	19.137	ng/u1	0.00
60) Fluorene-d10	15.716	176	485057	18.477	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.834	200	125090	17.301	ng/u1	0.00
73) Anthracene-d10	17.581	188	822803	17.718	ng/u1	0.00
81) Pyrene-d10	19.798	212	1057760	16.464	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.933	264	1187628	16.968	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.476	88	23683	6.618	ng/uL	92
5) Pyridine	3.893	79	153056	16.443	ng/u1	97
6) Benzaldehyde	7.216	77	144490	26.272	ng/u1	97
8) Phenol	7.264	94	209155	18.349	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.499	93	160663	18.398	ng/u1	98
12) 2-Chlorophenol	7.646	128	162578	19.162	ng/u1	99
13) 2-Methylphenol	8.528	108	157432	18.574	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.616	45	181636	19.260	ng/u1	96
16) Acetophenone	8.916	105	282145	19.344	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.899	70	144444	19.358	ng/u1	99
18) 4-Methylphenol	8.858	108	173719	18.634	ng/u1	97
19) Hexachloroethane	9.175	117	67990	18.297	ng/u1	93
22) Nitrobenzene	9.299	77	209386	19.494	ng/u1	100
23) Isophorone	9.828	82	406327	19.644	ng/u1	99
25) 2-Nitrophenol	10.016	139	100609	20.310	ng/u1	98
26) 2,4-Dimethylphenol	10.069	107	212571	19.605	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.310	93	228553	19.103	ng/u1	98
29) 2,4-Dichlorophenol	10.552	162	173424	19.440	ng/u1	98
30) Naphthalene	10.957	128	514132	18.605	ng/u1	99
32) 4-Chloroaniline	11.069	127	219315	18.210	ng/u1	97
33) Hexachlorobutadiene	11.240	225	128391	18.086	ng/u1	97
34) Caprolactam	11.852	113	40072	13.834	ng/u1	92
35) 4-Chloro-3-methylphenol	12.181	107	202789	19.658	ng/u1	96
36) 2-Methylnaphthalene	12.557	142	376643	19.432	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030723\
 Data File : BP013968.D
 Acq On : 07 Mar 2023 18:46
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV623


Quant Time: Mar 08 01:25:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 08 01:24:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.775	142	359527	18.455	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.922	216	254507	19.283	ng/ul	97
40) Hexachlorocyclopentadiene	12.899	237	166632	17.630	ng/ul	100
41) 2,4,6-Trichlorophenol	13.157	196	161819	19.517	ng/ul	97
42) 2,4,5-Trichlorophenol	13.234	196	172219	18.927	ng/ul	99
43) 1,1'-Biphenyl	13.557	154	518863	18.465	ng/ul	99
44) 2-Chloronaphthalene	13.604	162	403291	18.070	ng/ul	99
45) 2-Nitroaniline	13.810	65	129158	19.567	ng/ul	96
47) Dimethylphthalate	14.175	163	579010	18.884	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	124205	20.490	ng/ul	98
50) Acenaphthylene	14.446	152	658646	19.126	ng/ul	99
51) 3-Nitroaniline	14.628	138	122924	22.903	ng/ul	97
52) Acenaphthene	14.787	153	454363	18.873	ng/ul	99
53) 2,4-Dinitrophenol	14.834	184	85841	18.963	ng/ul	97
55) 4-Nitrophenol	14.928	109	115500	19.588	ng/ul	99
56) Dibenzofuran	15.122	168	660902	18.871	ng/ul	100
57) 2,4-Dinitrotoluene	15.087	165	181309	20.019	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.345	232	182386	20.951	ng/ul	95
59) Diethylphthalate	15.534	149	605979	19.486	ng/ul	100
61) Fluorene	15.769	166	551123	18.984	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.757	204	304905	19.029	ng/ul	98
63) 4-Nitroaniline	15.793	138	132809	24.073	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.851	198	126767	18.378	ng/ul	96
67) N-Nitrosodiphenylamine	15.975	169	493048	18.249	ng/ul	98
68) 4-Bromophenyl-phenylether	16.657	248	202282	17.491	ng/ul	100
69) Hexachlorobenzene	16.781	284	244111	17.913	ng/ul	99
70) Atrazine	16.928	200	216493	18.578	ng/ul	99
71) Pentachlorophenol	17.128	266	156378	17.555	ng/ul	98
72) Phenanthrene	17.522	178	917549	17.585	ng/ul	99
74) Anthracene	17.616	178	950107	17.959	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.528	216	241678	16.462	ng/uL	99
76) Pentachlorobenzene	15.040	250	260390	17.074	ng/uL	98
77) Carbazole	17.881	167	871314	18.672	ng/ul	99
78) Di-n-butylphthalate	18.410	149	1082905	18.568	ng/ul	100
80) Fluoranthene	19.475	202	1207019	16.483	ng/ul	100
82) Pyrene	19.828	202	1276959	16.716	ng/ul	100
83) Butylbenzylphthalate	20.680	149	534559	17.355	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.463	252	577498	19.366	ng/ul	99
85) Benzo(a)anthracene	21.539	228	1418666	17.315	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.433	149	813802	17.672	ng/ul	98
87) Chrysene	21.598	228	1340280	17.164	ng/ul	99
89) Di-n-octyl phthalate	22.404	149	1402470	16.891	ng/ul	100
90) Benzo(b)fluoranthene	23.316	252	1522691	17.366	ng/ul	100
91) Benzo(k)fluoranthene	23.369	252	1425788	17.053	ng/ul	100
93) Benzo(a)pyrene	23.986	252	1304193	17.101	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.780	276	1666585	16.507	ng/ul	99
95) Dibenzo(a,h)anthracene	26.792	278	1397488	16.654	ng/ul	100
96) Benzo(g,h,i)perylene	27.604	276	1322311	16.386	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 08 01:25:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 08 01:24:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.618 ng/uL
RT: 3.476 min Scan# 49
Delta R.T. 0.001 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion: 88	Resp: 23683
Ion Ratio	Lower Upper
88 100	
43 30.2	23.9 35.9
58 71.6	50.2 75.2



#5
Pyridine
Concen: 16.443 ng/uL
RT: 3.893 min Scan# 120
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 79	Resp: 153056
Ion Ratio	Lower Upper
79 100	
52 70.6	55.1 82.7
51 38.1	28.9 43.3









#6
Benzaldehyde
Concen: 26.272 ng/ul
RT: 7.216 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion: 77 Resp: 144490
Ion Ratio Lower Upper
77 100
105 94.8 74.2 111.2
106 90.9 69.8 104.8



#8
Phenol
Concen: 18.349 ng/ul
RT: 7.264 min Scan# 693
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 94 Resp: 209155
Ion Ratio Lower Upper
94 100
65 31.6 25.2 37.8
66 45.6 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.398 ng/ul
RT: 7.499 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion: 93 Resp: 160663
Ion Ratio Lower Upper
93 100
63 75.5 58.8 88.2
95 32.3 25.8 38.6



#12
2-Chlorophenol
Concen: 19.162 ng/ul
RT: 7.646 min Scan# 758
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 128 Resp: 162578
Ion Ratio Lower Upper
128 100
64 48.6 38.3 57.5
130 31.0 25.7 38.5









#13
2-Methylphenol
Concen: 18.574 ng/ul
RT: 8.528 min Scan# 908
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:108 Resp: 157432
Ion Ratio Lower Upper
108 100
107 93.2 70.5 105.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.260 ng/ul
RT: 8.616 min Scan# 923
Delta R.T. 0.006 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 45 Resp: 181636
Ion Ratio Lower Upper
45 100
77 19.5 14.2 21.2
79 14.4 10.6 15.8









#16
Acetophenone
Concen: 19.344 ng/ul
RT: 8.916 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:105 Resp: 282145
Ion Ratio Lower Upper
105 100
77 80.4 65.4 98.2
51 26.8 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.358 ng/ul
RT: 8.899 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 70 Resp: 144444
Ion Ratio Lower Upper
70 100
42 52.4 40.9 61.3
101 10.0 7.9 11.9
130 20.3 16.6 25.0









#18
4-Methylphenol
Concen: 18.634 ng/ul
RT: 8.858 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:108 Resp: 173719
Ion Ratio Lower Upper
108 100
107 109.4 89.8 134.6



#19
Hexachloroethane
Concen: 18.297 ng/ul
RT: 9.175 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:117 Resp: 67990
Ion Ratio Lower Upper
117 100
201 124.7 91.7 137.5
199 75.7 59.1 88.7









#22
Nitrobenzene
Concen: 19.494 ng/ul
RT: 9.299 min Scan# 1039
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion: 77 Resp: 209386
Ion Ratio Lower Upper
77 100
123 40.8 32.6 48.8
65 14.3 11.2 16.8



#23
Isophorone
Concen: 19.644 ng/ul
RT: 9.828 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion: 82 Resp: 406327
Ion Ratio Lower Upper
82 100
95 8.3 7.1 10.7
138 16.5 13.3 19.9









#25
2-Nitrophenol
Concen: 20.310 ng/ul
RT: 10.016 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:139 Resp: 100609
Ion Ratio Lower Upper
139 100
65 57.2 46.6 70.0
109 40.7 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 19.605 ng/ul
RT: 10.069 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:107 Resp: 212571
Ion Ratio Lower Upper
107 100
121 47.9 37.1 55.7
122 75.5 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 19.103 ng/ul

RT: 10.310 min Scan# 1211

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion: 93 Resp: 228553

Ion Ratio Lower Upper

93 100

95 33.3 25.6 38.4

123 10.9 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 19.440 ng/ul

RT: 10.552 min Scan# 1252

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:162 Resp: 173424


Ion Ratio Lower Upper

162 100

164 65.7 52.0 78.0

98 38.6 29.1 43.7









#30
Naphthalene
Concen: 18.605 ng/ul
RT: 10.957 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:128 Resp: 514132
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 12.8 11.0 16.4



#32
4-Chloroaniline
Concen: 18.210 ng/ul
RT: 11.069 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:127 Resp: 219315
Ion Ratio Lower Upper
127 100
129 32.7 24.9 37.3









#33
Hexachlorobutadiene
Concen: 18.086 ng/ul
RT: 11.240 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:225 Resp: 128391
Ion Ratio Lower Upper
225 100
223 63.9 48.0 72.0
227 63.7 50.3 75.5



#34
Caprolactam
Concen: 13.834 ng/ul
RT: 11.852 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:113 Resp: 40072
Ion Ratio Lower Upper
113 100
55 141.0 124.0 186.0
56 122.8 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 19.658 ng/ul
RT: 12.181 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:107 Resp: 202789
Ion Ratio Lower Upper
107 100
144 23.6 19.4 29.0
142 70.7 59.6 89.4



#36
2-Methylnaphthalene
Concen: 19.432 ng/ul
RT: 12.557 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:142 Resp: 376643
Ion Ratio Lower Upper
142 100
141 89.9 71.7 107.5





#37

1-Methylnaphthalene

Concen: 18.455 ng/ul

RT: 12.775 min Scan# 1630

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:142 Resp: 359527

Ion Ratio Lower Upper

142 100

141 94.0 73.8 110.6

116 4.8 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.283 ng/ul

RT: 12.922 min Scan# 1655

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Tgt Ion:216 Resp: 254507




Ion Ratio Lower Upper


216 100

214 80.5 61.7 92.5

179 20.6 16.2 24.4

108 15.7 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 17.630 ng/ul
RT: 12.899 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:237 Resp: 166632
Ion Ratio Lower Upper
237 100
235 61.4 49.0 73.6
272 12.5 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 19.517 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:196 Resp: 161819
Ion Ratio Lower Upper
196 100
198 95.0 78.5 117.7
200 31.5 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 18.927 ng/ul

RT: 13.234 min Scan# 1708

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:196 Resp: 172219

Ion Ratio Lower Upper

196 100

198 98.3 77.6 116.4

200 31.3 25.3 37.9



#43

1,1'-Biphenyl

Concen: 18.465 ng/ul

RT: 13.557 min Scan# 1763

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:154 Resp: 518863


Ion Ratio Lower Upper

154 100

153 40.2 32.6 49.0

76 11.3 9.4 14.0





#44

2-Chloronaphthalene

Concen: 18.070 ng/ul

RT: 13.604 min Scan# 1771

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:162 Resp: 403291

Ion Ratio Lower Upper

162 100

164 33.4 26.0 39.0

127 39.9 31.3 46.9



#45

2-Nitroaniline

Concen: 19.567 ng/ul

RT: 13.810 min Scan# 1806

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion: 65 Resp: 129158


Ion Ratio Lower Upper

65 100

92 69.4 56.3 84.5

138 96.2 81.4 122.2









#47
Dimethylphthalate
Concen: 18.884 ng/ul
RT: 14.175 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:163 Resp: 579010
Ion Ratio Lower Upper
163 100
194 4.3 3.6 5.4
164 9.8 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.490 ng/ul
RT: 14.298 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:165 Resp: 124205
Ion Ratio Lower Upper
165 100
89 62.5 51.7 77.5
121 21.2 17.5 26.3









#50
Acenaphthylene
Concen: 19.126 ng/ul
RT: 14.446 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:152 Resp: 658646
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 22.903 ng/ul
RT: 14.628 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:138 Resp: 122924
Ion Ratio Lower Upper
138 100
108 11.7 10.5 15.7
92 126.3 98.5 147.7









#52
Acenaphthene
Concen: 18.873 ng/ul
RT: 14.787 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:153 Resp: 454363
Ion Ratio Lower Upper
153 100
152 47.9 37.7 56.5
154 87.0 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 18.963 ng/ul
RT: 14.834 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:184 Resp: 85841
Ion Ratio Lower Upper
184 100
63 67.9 51.9 77.9
154 72.3 58.8 88.2









#55
4-Nitrophenol
Concen: 19.588 ng/ul
RT: 14.928 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:109 Resp: 115500
Ion Ratio Lower Upper
109 100
139 93.0 74.5 111.7
65 106.3 86.6 130.0



#56
Dibenzofuran
Concen: 18.871 ng/ul
RT: 15.122 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:168 Resp: 660902
Ion Ratio Lower Upper
168 100
139 40.0 31.8 47.6





#57

2,4-Dinitrotoluene

Concen: 20.019 ng/ul

RT: 15.087 min Scan# 2023

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:165 Resp: 181309

Ion Ratio Lower Upper

165 100

63 43.5 34.6 52.0

182 3.0 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.951 ng/ul

RT: 15.345 min Scan# 2067

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Tgt Ion:232 Resp: 182386




Ion Ratio Lower Upper


232 100

131 38.9 34.1 51.1

130 2.2 2.1 3.1

166 27.3 23.0 34.6









#59
Diethylphthalate
Concen: 19.486 ng/ul
RT: 15.534 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:149 Resp: 605979
Ion Ratio Lower Upper
149 100
177 22.7 18.2 27.2
150 12.7 10.0 15.0



#61
Fluorene
Concen: 18.984 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:166 Resp: 551123
Ion Ratio Lower Upper
166 100
165 96.5 78.9 118.3
167 13.4 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 19.029 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:204 Resp: 304905
Ion Ratio Lower Upper
204 100
206 32.6 26.6 39.8
141 59.9 49.2 73.8



#63
4-Nitroaniline
Concen: 24.073 ng/ul
RT: 15.793 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:138 Resp: 132809
Ion Ratio Lower Upper
138 100
92 60.4 44.5 66.7
108 113.8 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 18.378 ng/ul
RT: 15.851 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:198 Resp: 126767
Ion Ratio Lower Upper
198 100
182 3.5 3.0 4.6
77 23.3 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 18.249 ng/ul
RT: 15.975 min Scan# 2174
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:169 Resp: 493048
Ion Ratio Lower Upper
169 100
168 65.6 53.7 80.5
167 36.4 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 17.491 ng/ul
RT: 16.657 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:248 Resp: 202282
Ion Ratio Lower Upper
248 100
250 98.8 78.8 118.2
141 76.2 60.8 91.2



#69
Hexachlorobenzene
Concen: 17.913 ng/ul
RT: 16.781 min Scan# 2311
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:284 Resp: 244111
Ion Ratio Lower Upper
284 100
142 29.1 23.2 34.8
249 31.0 25.7 38.5





#70

Atrazine

Concen: 18.578 ng/ul

RT: 16.928 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:200 Resp: 216493

Ion Ratio Lower Upper

200 100

173 24.0 19.0 28.4

215 47.5 38.5 57.7



#71

Pentachlorophenol

Concen: 17.555 ng/ul

RT: 17.128 min Scan# 2370

Delta R.T. 0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:266 Resp: 156378


Ion Ratio Lower Upper

266 100

264 63.0 48.4 72.6

268 63.9 50.4 75.6





#72

Phenanthrene

Concen: 17.585 ng/ul

RT: 17.522 min Scan# 2437

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:178 Resp: 917549

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.6 15.4 23.0



#74

Anthracene

Concen: 17.959 ng/ul

RT: 17.616 min Scan# 2453

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:178 Resp: 950107


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.9 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.462 ng/uL
RT: 13.528 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:216 Resp: 241678
Ion Ratio Lower Upper
216 100
214 77.5 61.4 92.2
218 49.3 39.1 58.7



#76
Pentachlorobenzene
Concen: 17.074 ng/uL
RT: 15.040 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:250 Resp: 260390
Ion Ratio Lower Upper
250 100
248 61.7 51.0 76.4
252 64.8 50.8 76.2





#77

Carbazole

Concen: 18.672 ng/ul

RT: 17.881 min Scan# 2498

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:167 Resp: 871314

Ion Ratio Lower Upper

167 100

166 21.4 17.5 26.3

139 13.0 10.5 15.7



#78

Di-n-butylphthalate

Concen: 18.568 ng/ul

RT: 18.410 min Scan# 2588

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:149 Resp: 1082905


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 16.483 ng/ul
RT: 19.475 min Scan# 2769
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46




Instrument :
BNA_P
ClientSampleId :
SICV623


Tgt Ion:202 Resp: 1207019
Ion Ratio Lower Upper
202 100
101 8.0 6.6 10.0
100 6.3 5.0 7.4



#82
Pyrene
Concen: 16.716 ng/ul
RT: 19.828 min Scan# 2829
Delta R.T. 0.000 min
Lab File: BP013968.D
Acq: 07 Mar 2023 18:46

Tgt Ion:202 Resp: 1276959
Ion Ratio Lower Upper
202 100
101 9.8 7.8 11.8
100 7.8 6.5 9.7





#83

Butylbenzylphthalate

Concen: 17.355 ng/ul

RT: 20.680 min Scan# 2974

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:149 Resp: 534559

Ion Ratio Lower Upper

149 100

91 71.9 57.9 86.9

206 22.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 19.366 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:252 Resp: 577498


Ion Ratio Lower Upper

252 100

254 65.1 51.3 76.9

126 9.0 6.9 10.3





#85

Benzo(a)anthracene

Concen: 17.315 ng/ul

RT: 21.539 min Scan# 3121

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:228 Resp: 1418666

Ion Ratio Lower Upper

228 100

229 19.5 15.5 23.3

226 27.2 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.672 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:149 Resp: 813802


Ion Ratio Lower Upper

149 100

167 27.1 22.3 33.5

279 4.8 4.3 6.5





#87

Chrysene

Concen: 17.164 ng/ul

RT: 21.598 min Scan# 3130

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:228 Resp: 1340280

Ion Ratio Lower Upper

228 100

226 29.7 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 16.891 ng/ul




RT: 22.404 min Scan# 3267


Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Tgt Ion:149 Resp: 1402470





#90

Benzo(b)fluoranthene

Concen: 17.366 ng/ul

RT: 23.316 min Scan# 3422

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:252 Resp: 1522691

Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 7.0 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 17.053 ng/ul

RT: 23.369 min Scan# 3431

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:252 Resp: 1425788


Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 6.9 5.5 8.3





#93

Benzo(a)pyrene

Concen: 17.101 ng/ul

RT: 23.986 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:252 Resp: 1304193

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.2

125 7.7 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.507 ng/ul

RT: 26.780 min Scan# 4011

Delta R.T. 0.000 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46




Tgt Ion:276 Resp: 1666585


Ion Ratio Lower Upper

276 100

138 16.6 13.1 19.7

277 26.0 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 16.654 ng/ul

RT: 26.792 min Scan# 4014

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Instrument :

BNA_P

ClientSampleId :

SICV623





Tgt Ion:278 Resp: 1397488

Ion Ratio Lower Upper

278 100

139 11.3 8.7 13.1

279 23.8 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 16.386 ng/ul

RT: 27.604 min Scan# 4151

Delta R.T. -0.006 min

Lab File: BP013968.D

Acq: 07 Mar 2023 18:46

Tgt Ion:276 Resp: 1322311

Ion Ratio Lower Upper

276 100

138 16.0 11.9 17.9

277 24.7 18.7 28.1

