

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030922\
 Data File : BP009318.D
 Acq On : 09 Mar 2022 11:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 10 04:08:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	33.456	16.4	103	0.00
3	Pyridine	40.000	35.763	10.6	106	0.00
4	n-Nitrosodimethylamine	40.000	37.513	6.2	109	0.00
5 S	2-Fluorophenol	80.000	71.813	10.2	109	0.00
6	Aniline	40.000	37.037	7.4	111	0.00
7 S	Phenol-d6	80.000	73.717	7.9	111	0.00
8	2-Chlorophenol	40.000	37.491	6.3	111	0.00
9	Benzaldehyde	40.000	36.301	9.2	108	0.00
10 C	Phenol	40.000	36.817	8.0	111	0.00
11	bis(2-Chloroethyl)ether	40.000	37.473	6.3	111	0.00
12	1,3-Dichlorobenzene	40.000	35.643	10.9	110	0.00
13 C	1,4-Dichlorobenzene	40.000	35.765	10.6	110	0.00
14	1,2-Dichlorobenzene	40.000	35.916	10.2	110	0.00
15	Benzyl Alcohol	40.000	38.684	3.3	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.279	11.8	108	0.00
17	2-Methylphenol	40.000	38.334	4.2	115	0.00
18	Hexachloroethane	40.000	36.964	7.6	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.573	-1.4	120	0.00
20	3+4-Methylphenols	40.000	38.779	3.1	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	37.046	7.4	116	0.00
23 S	Nitrobenzene-d5	80.000	77.995	2.5	118	0.00
24	Nitrobenzene	40.000	37.660	5.9	116	0.00
25	Isophorone	40.000	39.875	0.3	122	0.00
26 C	2-Nitrophenol	40.000	45.906	-14.8	134	0.00
27	2,4-Dimethylphenol	40.000	36.913	7.7	115	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.069	4.8	119	0.00
29 C	2,4-Dichlorophenol	40.000	39.345	1.6	119	0.00
30	1,2,4-Trichlorobenzene	40.000	37.028	7.4	118	0.00
31	Naphthalene	40.000	36.499	8.8	115	0.00
32	Benzoic acid	40.000	51.900	-29.8#	154	0.02
33	4-Chloroaniline	40.000	38.353	4.1	119	0.00
34 C	Hexachlorobutadiene	40.000	38.718	3.2	123	0.00
35	Caprolactam	40.000	45.830	-14.6	134	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.549	-1.4	125	0.00
37	2-Methylnaphthalene	40.000	37.998	5.0	119	0.00
38	1-Methylnaphthalene	40.000	38.091	4.8	120	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	126	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.375	6.6	126	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.479	6.3	125	0.00
42 S	2,4,6-Tribromophenol	80.000	94.718	-18.4	154	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.527	-1.3	131	0.00
44	2,4,5-Trichlorophenol	40.000	40.161	-0.4	131	0.00
45 S	2-Fluorobiphenyl	80.000	72.996	8.8	125	0.00
46	1,1'-Biphenyl	40.000	36.526	8.7	124	0.00
47	2-Chloronaphthalene	40.000	36.764	8.1	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030922\
 Data File : BP009318.D
 Acq On : 09 Mar 2022 11:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 10 04:08:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.337	-13.3	141	0.00
49	Acenaphthylene	40.000	37.822	5.4	126	0.00
50	Dimethylphthalate	40.000	39.205	2.0	132	0.00
51	2,6-Dinitrotoluene	40.000	42.850	-7.1	135	0.00
52 C	Acenaphthene	40.000	38.015	5.0	128	0.00
53	3-Nitroaniline	40.000	42.990	-7.5	136	0.00
54 P	2,4-Dinitrophenol	40.000	51.253	-28.1#	158	0.00
55	Dibenzofuran	40.000	37.421	6.4	128	0.00
56 P	4-Nitrophenol	40.000	43.024	-7.6	134	0.00
57	2,4-Dinitrotoluene	40.000	46.048	-15.1	142	0.00
58	Fluorene	40.000	38.154	4.6	131	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.698	-6.7	141	0.00
60	Diethylphthalate	40.000	41.313	-3.3	138	0.00
61	4-Chlorophenyl-phenylether	40.000	39.163	2.1	134	0.00
62	4-Nitroaniline	40.000	44.924	-12.3	141	0.00
63	Azobenzene	40.000	39.000	2.5	130	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	133	0.00
65	4,6-Dinitro-2-methylphenol	40.000	48.447	-21.1	157	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.902	7.7	131	0.00
67	4-Bromophenyl-phenylether	40.000	40.826	-2.1	145	0.00
68	Hexachlorobenzene	40.000	39.880	0.3	143	0.00
69	Atrazine	40.000	41.895	-4.7	143	0.00
70 C	Pentachlorophenol	40.000	42.860	-7.1	143	0.00
71	Phenanthrene	40.000	36.402	9.0	131	0.00
72	Anthracene	40.000	37.592	6.0	133	0.00
73	Carbazole	40.000	37.111	7.2	130	0.00
74	Di-n-butylphthalate	40.000	43.151	-7.9	146	0.00
75 C	Fluoranthene	40.000	37.883	5.3	132	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	130	0.00
77	Benzydine	40.000	41.521	-3.8	127	0.00
78	Pyrene	40.000	37.990	5.0	131	0.00
79 S	Terphenyl-d14	80.000	78.086	2.4	136	0.00
80	Butylbenzylphthalate	40.000	46.395	-16.0	149	0.00
81	Benzo(a)anthracene	40.000	37.398	6.5	130	0.00
82	3,3'-Dichlorobenzidine	40.000	40.542	-1.4	133	0.00
83	Chrysene	40.000	36.996	7.5	130	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.086	-10.2	144	0.00
85 c	Di-n-octyl phthalate	40.000	46.385	-16.0	148	0.00
86 I	Perylene-d12	20.000	20.000	0.0	126	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.687	5.8	126	0.00
88	Benzo(b)fluoranthene	40.000	37.461	6.3	124	0.00
89	Benzo(k)fluoranthene	40.000	38.877	2.8	132	0.00
90 C	Benzo(a)pyrene	40.000	38.324	4.2	127	0.00
91	Dibenzo(a,h)anthracene	40.000	37.847	5.4	126	0.00
92	Benzo(g,h,i)perylene	40.000	37.435	6.4	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030922\
Data File : BP009318.D
Acq On : 09 Mar 2022 11:25
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 10 04:08:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 01:55:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0