

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
 Data File : BP014035.D
 Acq On : 10 Mar 2023 17:48
 Operator : CG/JU
 Sample : 01804-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA4

Quant Time: Mar 10 21:31:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.075	152	135122	20.000	ng/ul	0.00
20) Naphthalene-d8	10.899	136	559068	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.716	164	400945	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.469	188	927723	20.000	ng/ul	0.00
79) Chrysene-d12	21.545	240	996339	20.000	ng/ul	-0.01
88) Perylene-d12	24.086	264	1008554	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	13367	3.738	ng/uL	0.00
4) Pyridine-d5	3.876	84	58959	5.954	ng/ul	0.00
7) Phenol-d5	7.228	99	57196	4.797	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.399	67	182186	26.198	ng/ul	0.00
11) 2-Chlorophenol-d4	7.605	132	168372	18.533	ng/ul	0.00
15) 4-Methylphenol-d8	8.781	113	111915	11.490	ng/ul	-0.01
21) Nitrobenzene-d5	9.252	128	111857	25.076	ng/ul	0.00
24) 2-Nitrophenol-d4	9.975	143	121938	22.836	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.516	165	204868	20.390	ng/ul	0.00
31) 4-Chloroaniline-d4	11.040	131	234773	17.657	ng/ul	0.00
46) Dimethylphthalate-d6	14.116	166	868800	25.533	ng/ul	0.00
49) Acenaphthylene-d8	14.410	160	890052	24.692	ng/ul	0.00
54) 4-Nitrophenol-d4	14.922	143	19835	3.308	ng/ul	0.00
60) Fluorene-d10	15.704	176	753657	26.134	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.822	200	162714	23.206	ng/ul	-0.01
73) Anthracene-d10	17.569	188	1296459	28.787	ng/ul	0.00
81) Pyrene-d10	19.792	212	1682714	30.622	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.922	264	1618273	30.165	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
Data File : BP014035.D
Acq On : 10 Mar 2023 17:48
Operator : CG/JU
Sample : 01804-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AA4

Quant Time: Mar 10 21:31:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 09 01:10:12 2023
Response via : Initial Calibration

