

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
 Data File : BP014061.D
 Acq On : 11 Mar 2023 10:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776

Quant Time: Mar 13 00:51:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 10 22:47:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.075	152	177224	20.000	ng/ul	0.00
20) Naphthalene-d8	10.898	136	738415	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.710	164	532123	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.469	188	1263049	20.000	ng/ul	0.00
79) Chrysene-d12	21.545	240	1288657	20.000	ng/ul	0.00
88) Perylene-d12	24.080	264	1246539	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	38538	8.216	ng/uL	0.00
4) Pyridine-d5	3.863	84	273315	21.045	ng/ul	0.00
7) Phenol-d5	7.222	99	334888	21.413	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.393	67	205800	22.563	ng/ul	0.00
11) 2-Chlorophenol-d4	7.599	132	252708	21.208	ng/ul	0.00
15) 4-Methylphenol-d8	8.781	113	277063	21.688	ng/ul	0.00
21) Nitrobenzene-d5	9.246	128	125196	21.249	ng/ul	0.00
24) 2-Nitrophenol-d4	9.969	143	143205	20.305	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	262830	19.806	ng/ul	0.00
31) 4-Chloroaniline-d4	11.034	131	358703	20.425	ng/ul	0.00
46) Dimethylphthalate-d6	14.116	166	895134	19.821	ng/ul	0.00
49) Acenaphthylene-d8	14.404	160	1004912	21.006	ng/ul	0.00
54) 4-Nitrophenol-d4	14.904	143	157817	19.831	ng/ul	0.00
60) Fluorene-d10	15.704	176	754727	19.720	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	178365	18.684	ng/ul	0.00
73) Anthracene-d10	17.563	188	1248500	20.362	ng/ul	0.00
81) Pyrene-d10	19.786	212	1511713	21.270	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.909	264	1373221	20.710	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.469	88	41079	8.112	ng/uL#	86
5) Pyridine	3.881	79	287936	21.862	ng/ul	94
6) Benzaldehyde	7.204	77	204598	26.292	ng/ul	98
8) Phenol	7.252	94	350110	21.708	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.487	93	274987	22.255	ng/ul	97
12) 2-Chlorophenol	7.634	128	260958	21.737	ng/ul	95
13) 2-Methylphenol	8.516	108	262509	21.889	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.604	45	325898	24.423	ng/ul	98
16) Acetophenone	8.904	105	456615	22.125	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.887	70	237528	22.498	ng/ul	97
18) 4-Methylphenol	8.846	108	289527	21.948	ng/ul	100
19) Hexachloroethane	9.157	117	111600	21.225	ng/ul	94
22) Nitrobenzene	9.287	77	345805	21.941	ng/ul	99
23) Isophorone	9.816	82	607298	20.009	ng/ul	98
25) 2-Nitrophenol	10.004	139	152948	21.042	ng/ul	96
26) 2,4-Dimethylphenol	10.057	107	326301	20.509	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.298	93	349942	19.933	ng/ul	100
29) 2,4-Dichlorophenol	10.540	162	258595	19.755	ng/ul	99
30) Naphthalene	10.945	128	825414	20.356	ng/ul	99
32) 4-Chloroaniline	11.057	127	361629	20.464	ng/ul	98
33) Hexachlorobutadiene	11.222	225	201871	19.380	ng/ul	99
34) Caprolactam	11.840	113	50004	11.765	ng/ul	98
35) 4-Chloro-3-methylphenol	12.169	107	315907	20.870	ng/ul	96
36) 2-Methylnaphthalene	12.545	142	597962	21.025	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
 Data File : BP014061.D
 Acq On : 11 Mar 2023 10:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776

Quant Time: Mar 13 00:51:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 10 22:47:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.763	142	596849	20.879	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.910	216	380307	19.764	ng/ul	100
40) Hexachlorocyclopentadiene	12.887	237	233906	16.975	ng/ul	100
41) 2,4,6-Trichlorophenol	13.145	196	243347	20.131	ng/ul	100
42) 2,4,5-Trichlorophenol	13.222	196	274595	20.699	ng/ul	98
43) 1,1'-Biphenyl	13.545	154	843079	20.579	ng/ul	98
44) 2-Chloronaphthalene	13.592	162	669505	20.576	ng/ul	99
45) 2-Nitroaniline	13.798	65	220735	22.938	ng/ul	91
47) Dimethylphthalate	14.163	163	890344	19.918	ng/ul	99
48) 2,6-Dinitrotoluene	14.286	165	184870	20.919	ng/ul	99
50) Acenaphthylene	14.434	152	1052185	20.958	ng/ul	99
51) 3-Nitroaniline	14.616	138	170472	21.786	ng/ul	93
52) Acenaphthene	14.775	153	729036	20.772	ng/ul	99
53) 2,4-Dinitrophenol	14.822	184	110907	16.805	ng/ul	94
55) 4-Nitrophenol	14.922	109	162086	18.855	ng/ul	91
56) Dibenzofuran	15.110	168	1030675	20.186	ng/ul	97
57) 2,4-Dinitrotoluene	15.075	165	268216	20.313	ng/ul#	92
58) 2,3,4,6-Tetrachlorophenol	15.333	232	253110	19.943	ng/ul	97
59) Diethylphthalate	15.522	149	901602	19.886	ng/ul	99
61) Fluorene	15.757	166	854477	20.189	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.745	204	459181	19.656	ng/ul	99
63) 4-Nitroaniline	15.786	138	175262	21.790	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.833	198	171227	18.801	ng/ul	92
67) N-Nitrosodiphenylamine	15.963	169	732528	20.535	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	298026	19.518	ng/ul	98
69) Hexachlorobenzene	16.763	284	351686	19.546	ng/ul	96
70) Atrazine	16.916	200	302137	19.638	ng/ul	98
71) Pentachlorophenol	17.110	266	218504	18.578	ng/ul	98
72) Phenanthrene	17.510	178	1419615	20.607	ng/ul	100
74) Anthracene	17.598	178	1440652	20.625	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.510	216	398499	20.558	ng/ul	99
76) Pentachlorobenzene	15.028	250	401240	19.927	ng/ul	98
77) Carbazole	17.869	167	1279653	20.769	ng/ul	99
78) Di-n-butylphthalate	18.398	149	1558700	20.243	ng/ul	100
80) Fluoranthene	19.463	202	1773365	21.891	ng/ul	98
82) Pyrene	19.816	202	1836385	21.730	ng/ul	98
83) Butylbenzylphthalate	20.668	149	742733	21.797	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.451	252	652880	19.791	ng/ul	100
85) Benzo(a)anthracene	21.527	228	1868919	20.619	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.421	149	1077735	21.155	ng/ul	99
87) Chrysene	21.580	228	1771466	20.507	ng/ul	100
89) Di-n-octyl phthalate	22.392	149	1868275	23.753	ng/ul	100
90) Benzo(b)fluoranthene	23.298	252	1796561	21.629	ng/ul	99
91) Benzo(k)fluoranthene	23.351	252	1783286	22.515	ng/ul	99
93) Benzo(a)pyrene	23.962	252	1493682	20.675	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.745	276	1700227	17.777	ng/ul	97
95) Dibenzo(a,h)anthracene	26.756	278	1420197	17.866	ng/ul	98
96) Benzo(g,h,i)perylene	27.562	276	1342312	17.559	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 8.112 ng/uL
 RT: 3.469 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BP014061.D
 Acq: 11 Mar 2023 10:21




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776


Tgt Ion: 88 Resp: 41079
 Ion Ratio Lower Upper
 88 100
 43 31.0 23.9 35.9
 58 77.4 50.2 75.2#



#5
 Pyridine
 Concen: 21.862 ng/uL
 RT: 3.881 min Scan# 118
 Delta R.T. -0.006 min
 Lab File: BP014061.D
 Acq: 11 Mar 2023 10:21

Tgt Ion: 79 Resp: 287936
 Ion Ratio Lower Upper
 79 100
 52 76.1 55.1 82.7
 51 35.9 28.9 43.3









#6
Benzaldehyde
Concen: 26.292 ng/ul
RT: 7.204 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 77 Resp: 204598
Ion Ratio Lower Upper
77 100
105 95.0 74.2 111.2
106 87.9 69.8 104.8



#8
Phenol
Concen: 21.708 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion: 94 Resp: 350110
Ion Ratio Lower Upper
94 100
65 31.6 25.2 37.8
66 45.9 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 22.255 ng/ul
RT: 7.487 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 93 Resp: 274987
Ion Ratio Lower Upper
93 100
63 77.1 58.8 88.2
95 32.0 25.8 38.6



#12
2-Chlorophenol
Concen: 21.737 ng/ul
RT: 7.634 min Scan# 756
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:128 Resp: 260958
Ion Ratio Lower Upper
128 100
64 52.2 38.3 57.5
130 33.2 25.7 38.5









#13
2-Methylphenol
Concen: 21.889 ng/ul
RT: 8.516 min Scan# 907
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:108 Resp: 262509
Ion Ratio Lower Upper
108 100
107 89.8 70.5 105.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.423 ng/ul
RT: 8.604 min Scan# 921
Delta R.T. -0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion: 45 Resp: 325898
Ion Ratio Lower Upper
45 100
77 16.1 14.2 21.2
79 13.1 10.6 15.8









#16
Acetophenone
Concen: 22.125 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:105 Resp: 456615
Ion Ratio Lower Upper
105 100
77 83.2 65.4 98.2
51 27.1 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.498 ng/ul
RT: 8.887 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion: 70 Resp: 237528
Ion Ratio Lower Upper
70 100
42 53.2 40.9 61.3
101 9.8 7.9 11.9
130 19.7 16.6 25.0









#18
4-Methylphenol
Concen: 21.948 ng/ul
RT: 8.846 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:108 Resp: 289527
Ion Ratio Lower Upper
108 100
107 112.7 89.8 134.6



#19
Hexachloroethane
Concen: 21.225 ng/ul
RT: 9.157 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:117 Resp: 111600
Ion Ratio Lower Upper
117 100
201 109.1 91.7 137.5
199 68.2 59.1 88.7









#22
Nitrobenzene
Concen: 21.941 ng/ul
RT: 9.287 min Scan# 1037
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion: 77 Resp: 345805
Ion Ratio Lower Upper
77 100
123 40.8 32.6 48.8
65 14.9 11.2 16.8



#23
Isophorone
Concen: 20.009 ng/ul
RT: 9.816 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion: 82 Resp: 607298
Ion Ratio Lower Upper
82 100
95 8.5 7.1 10.7
138 15.5 13.3 19.9









#25
2-Nitrophenol
Concen: 21.042 ng/ul
RT: 10.004 min Scan# 1159
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:139 Resp: 152948
Ion Ratio Lower Upper
139 100
65 61.6 46.6 70.0
109 44.7 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 20.509 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:107 Resp: 326301
Ion Ratio Lower Upper
107 100
121 45.9 37.1 55.7
122 74.9 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 19.933 ng/ul

RT: 10.298 min Scan# 1209

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion: 93 Resp: 349942

Ion Ratio Lower Upper

93 100

95 31.9 25.6 38.4

123 10.6 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 19.755 ng/ul

RT: 10.540 min Scan# 1250

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:162 Resp: 258595


Ion Ratio Lower Upper

162 100

164 64.2 52.0 78.0

98 37.6 29.1 43.7









#30
Naphthalene
Concen: 20.356 ng/ul
RT: 10.945 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:128 Resp: 825414
Ion Ratio Lower Upper
128 100
129 10.5 8.5 12.7
127 13.1 11.0 16.4



#32
4-Chloroaniline
Concen: 20.464 ng/ul
RT: 11.057 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:127 Resp: 361629
Ion Ratio Lower Upper
127 100
129 32.3 24.9 37.3









#33
Hexachlorobutadiene
Concen: 19.380 ng/ul
RT: 11.222 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:225 Resp: 201871
Ion Ratio Lower Upper
225 100
223 61.1 48.0 72.0
227 62.6 50.3 75.5



#34
Caprolactam
Concen: 11.765 ng/ul
RT: 11.840 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:113 Resp: 50004
Ion Ratio Lower Upper
113 100
55 154.9 124.0 186.0
56 123.2 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 20.870 ng/ul
RT: 12.169 min Scan# 1527
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:107 Resp: 315907
Ion Ratio Lower Upper
107 100
144 23.0 19.4 29.0
142 70.2 59.6 89.4



#36
2-Methylnaphthalene
Concen: 21.025 ng/ul
RT: 12.545 min Scan# 1591
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:142 Resp: 597962
Ion Ratio Lower Upper
142 100
141 89.6 71.7 107.5





#37

1-Methylnaphthalene

Concen: 20.879 ng/ul

RT: 12.763 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:142 Resp: 596849

Ion Ratio Lower Upper

142 100

141 93.4 73.8 110.6

116 4.6 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.764 ng/ul

RT: 12.910 min Scan# 1653

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Tgt Ion:216 Resp: 380307




Ion Ratio Lower Upper


216 100

214 77.4 61.7 92.5

179 20.6 16.2 24.4

108 16.8 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 16.975 ng/ul

RT: 12.887 min Scan# 1649

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:237 Resp: 233906

Ion Ratio Lower Upper

237 100

235 61.1 49.0 73.6

272 12.4 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 20.131 ng/ul

RT: 13.145 min Scan# 1693

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:196 Resp: 243347


Ion Ratio Lower Upper

196 100

198 98.4 78.5 117.7

200 31.7 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 20.699 ng/ul

RT: 13.222 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:196 Resp: 274595

Ion Ratio Lower Upper

196 100

198 95.7 77.6 116.4

200 29.9 25.3 37.9



#43

1,1'-Biphenyl

Concen: 20.579 ng/ul

RT: 13.545 min Scan# 1761

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:154 Resp: 843079


Ion Ratio Lower Upper

154 100

153 41.9 32.6 49.0

76 12.7 9.4 14.0





#44

2-Chloronaphthalene

Concen: 20.576 ng/ul

RT: 13.592 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:162 Resp: 669505

Ion Ratio Lower Upper

162 100

164 32.5 26.0 39.0

127 40.2 31.3 46.9



#45

2-Nitroaniline

Concen: 22.938 ng/ul

RT: 13.798 min Scan# 1804

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion: 65 Resp: 220735


Ion Ratio Lower Upper

65 100

92 64.9 56.3 84.5

138 91.0 81.4 122.2









#47
Dimethylphthalate
Concen: 19.918 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:163 Resp: 890344
Ion Ratio Lower Upper
163 100
194 4.3 3.6 5.4
164 9.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.919 ng/ul
RT: 14.286 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:165 Resp: 184870
Ion Ratio Lower Upper
165 100
89 65.8 51.7 77.5
121 21.8 17.5 26.3









#50
Acenaphthylene
Concen: 20.958 ng/ul
RT: 14.434 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:152 Resp: 1052185
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.5 10.4 15.6



#51
3-Nitroaniline
Concen: 21.786 ng/ul
RT: 14.616 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:138 Resp: 170472
Ion Ratio Lower Upper
138 100
108 12.7 10.5 15.7
92 131.3 98.5 147.7









#52
Acenaphthene
Concen: 20.772 ng/ul
RT: 14.775 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:153 Resp: 729036
Ion Ratio Lower Upper
153 100
152 48.5 37.7 56.5
154 87.5 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 16.805 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:184 Resp: 110907
Ion Ratio Lower Upper
184 100
63 74.0 51.9 77.9
154 74.2 58.8 88.2









#55
4-Nitrophenol
Concen: 18.855 ng/ul
RT: 14.922 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:109 Resp: 162086
Ion Ratio Lower Upper
109 100
139 102.5 74.5 111.7
65 117.9 86.6 130.0



#56
Dibenzofuran
Concen: 20.186 ng/ul
RT: 15.110 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:168 Resp: 1030675
Ion Ratio Lower Upper
168 100
139 41.3 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 20.313 ng/ul
RT: 15.075 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:165 Resp: 268216
Ion Ratio Lower Upper
165 100
63 48.3 34.6 52.0
182 3.3 2.2 3.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.943 ng/ul
RT: 15.333 min Scan# 2065
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:232 Resp: 253110
Ion Ratio Lower Upper
232 100
131 44.5 34.1 51.1
130 2.4 2.1 3.1
166 30.9 23.0 34.6









#59
Diethylphthalate
Concen: 19.886 ng/ul
RT: 15.522 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:149 Resp: 901602
Ion Ratio Lower Upper
149 100
177 22.0 18.2 27.2
150 12.8 10.0 15.0



#61
Fluorene
Concen: 20.189 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:166 Resp: 854477
Ion Ratio Lower Upper
166 100
165 97.9 78.9 118.3
167 13.5 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 19.656 ng/ul
RT: 15.745 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:204 Resp: 459181
Ion Ratio Lower Upper
204 100
206 32.8 26.6 39.8
141 60.9 49.2 73.8



#63
4-Nitroaniline
Concen: 21.790 ng/ul
RT: 15.786 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:138 Resp: 175262
Ion Ratio Lower Upper
138 100
92 56.0 44.5 66.7
108 109.3 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 18.801 ng/ul
RT: 15.833 min Scan# 2150
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:198 Resp: 171227
Ion Ratio Lower Upper
198 100
182 4.2 3.0 4.6
77 29.7 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.535 ng/ul
RT: 15.963 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:169 Resp: 732528
Ion Ratio Lower Upper
169 100
168 66.1 53.7 80.5
167 35.4 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 19.518 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:248 Resp: 298026
Ion Ratio Lower Upper
248 100
250 98.8 78.8 118.2
141 79.0 60.8 91.2



#69
Hexachlorobenzene
Concen: 19.546 ng/ul
RT: 16.763 min Scan# 2308
Delta R.T. -0.006 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:284 Resp: 351686
Ion Ratio Lower Upper
284 100
142 32.4 23.2 34.8
249 31.6 25.7 38.5





#70

Atrazine

Concen: 19.638 ng/ul

RT: 16.916 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:200 Resp: 302137

Ion Ratio Lower Upper

200 100

173 26.0 19.0 28.4

215 49.0 38.5 57.7



#71

Pentachlorophenol

Concen: 18.578 ng/ul

RT: 17.110 min Scan# 2367

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:266 Resp: 218504


Ion Ratio Lower Upper

266 100

264 62.4 48.4 72.6

268 63.4 50.4 75.6





#72

Phenanthrene

Concen: 20.607 ng/ul

RT: 17.510 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:178 Resp: 1419615

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.3 15.4 23.0



#74

Anthracene

Concen: 20.625 ng/ul

RT: 17.598 min Scan# 2450

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:178 Resp: 1440652


Ion Ratio Lower Upper

178 100

179 15.8 12.5 18.7

176 18.6 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.558 ng/uL

RT: 13.510 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:216 Resp: 398499

Ion Ratio Lower Upper

216 100

214 77.2 61.4 92.2

218 47.7 39.1 58.7



#76

Pentachlorobenzene

Concen: 19.927 ng/uL

RT: 15.028 min Scan# 2013

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:250 Resp: 401240


Ion Ratio Lower Upper

250 100

248 61.3 51.0 76.4

252 62.8 50.8 76.2









#77
 Carbazole
 Concen: 20.769 ng/ul
 RT: 17.869 min Scan# 2497
 Delta R.T. -0.000 min
 Lab File: BP014061.D
 Acq: 11 Mar 2023 10:21




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020776


Tgt Ion:167 Resp: 1279653
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.5 26.3
 139 13.2 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 20.243 ng/ul
 RT: 18.398 min Scan# 2586
 Delta R.T. -0.000 min
 Lab File: BP014061.D
 Acq: 11 Mar 2023 10:21

Tgt Ion:149 Resp: 1558700
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 5.9 4.6 6.8









#80
Fluoranthene
Concen: 21.891 ng/ul
RT: 19.463 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21




Instrument :
BNA_P
ClientSampleId :
SSTD020776


Tgt Ion:202 Resp: 1773365
Ion Ratio Lower Upper
202 100
101 8.9 6.6 10.0
100 6.8 5.0 7.4



#82
Pyrene
Concen: 21.730 ng/ul
RT: 19.816 min Scan# 2827
Delta R.T. 0.000 min
Lab File: BP014061.D
Acq: 11 Mar 2023 10:21

Tgt Ion:202 Resp: 1836385
Ion Ratio Lower Upper
202 100
101 10.5 7.8 11.8
100 8.7 6.5 9.7





#83

Butylbenzylphthalate

Concen: 21.797 ng/ul

RT: 20.668 min Scan# 2972

Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:149 Resp: 742733

Ion Ratio Lower Upper

149 100

91 73.8 57.9 86.9

206 22.7 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 19.791 ng/ul

RT: 21.451 min Scan# 3105

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:252 Resp: 652880


Ion Ratio Lower Upper

252 100

254 64.2 51.3 76.9

126 9.7 6.9 10.3





#85

Benzo(a)anthracene

Concen: 20.619 ng/ul

RT: 21.527 min Scan# 3118

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:228 Resp: 1868919

Ion Ratio Lower Upper

228 100

229 19.7 15.5 23.3

226 27.2 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.155 ng/ul

RT: 21.421 min Scan# 3100

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:149 Resp: 1077735


Ion Ratio Lower Upper

149 100

167 27.6 22.3 33.5

279 4.9 4.3 6.5





#87

Chrysene

Concen: 20.507 ng/ul

RT: 21.580 min Scan# 3129

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:228 Resp: 1771466

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 23.753 ng/ul




RT: 22.392 min Scan# 3265


Delta R.T. 0.000 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Tgt Ion:149 Resp: 1868275





#90

Benzo(b)fluoranthene

Concen: 21.629 ng/ul

RT: 23.298 min Scan# 3419

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:252 Resp: 1796561

Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 7.8 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 22.515 ng/ul

RT: 23.351 min Scan# 3428

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:252 Resp: 1783286


Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.0

125 7.2 5.5 8.3





#93

Benzo(a)pyrene

Concen: 20.675 ng/ul

RT: 23.962 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:252 Resp: 1493682

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.2

125 8.6 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.777 ng/ul

RT: 26.745 min Scan# 4005

Delta R.T. -0.006 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21




Tgt Ion:276 Resp: 1700227


Ion Ratio Lower Upper

276 100

138 18.6 13.1 19.7

277 25.7 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 17.866 ng/ul

RT: 26.756 min Scan# 4012

Delta R.T. -0.012 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Instrument :

BNA_P

ClientSampleId :

SSTD020776





Tgt Ion:278 Resp: 1420197

Ion Ratio Lower Upper

278 100

139 12.6 8.7 13.1

279 23.5 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.559 ng/ul

RT: 27.562 min Scan# 4144

Delta R.T. -0.012 min

Lab File: BP014061.D

Acq: 11 Mar 2023 10:21

Tgt Ion:276 Resp: 1342312

Ion Ratio Lower Upper

276 100

138 16.4 11.9 17.9

277 23.7 18.7 28.1

