

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
 Data File : BP014053.D
 Acq On : 11 Mar 2023 04:39
 Operator : CG/JU
 Sample : 01804-11
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AB0

Quant Time: Mar 11 05:11:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 10 22:47:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.075	152	133840	20.000	ng/u1	0.00
20) Naphthalene-d8	10.898	136	576578	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	409885	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	984572	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	1060243	20.000	ng/u1	0.00
88) Perylene-d12	24.080	264	1054506	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	16223	4.580	ng/uL	0.00
4) Pyridine-d5	3.875	84	65743	6.703	ng/u1	0.01
7) Phenol-d5	7.228	99	78931	6.683	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.393	67	209952	30.480	ng/u1	0.00
11) 2-Chlorophenol-d4	7.599	132	206773	22.978	ng/u1	0.00
15) 4-Methylphenol-d8	8.781	113	151168	15.669	ng/u1	0.00
21) Nitrobenzene-d5	9.246	128	133730	29.069	ng/u1	0.00
24) 2-Nitrophenol-d4	9.975	143	149851	27.211	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	251856	24.306	ng/u1	0.00
31) 4-Chloroaniline-d4	11.034	131	314264	22.918	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	1006375	28.931	ng/u1	0.00
49) Acenaphthylene-d8	14.410	160	1064812	28.896	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	29049	4.739	ng/u1	0.01
60) Fluorene-d10	15.704	176	896691	30.416	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	186523	25.065	ng/u1	0.00
73) Anthracene-d10	17.569	188	1637882	34.268	ng/u1	0.00
81) Pyrene-d10	19.786	212	2053317	35.114	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.915	264	1991547	35.505	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031023\
Data File : BP014053.D
Acq On : 11 Mar 2023 04:39
Operator : CG/JU
Sample : 01804-11
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AB0

Quant Time: Mar 11 05:11:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 10 22:47:08 2023
Response via : Initial Calibration

