

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031323\
 Data File : BP014074.D
 Acq On : 13 Mar 2023 21:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 04:25:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 04:15:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	118	0.00
2	1,4-Dioxane	40.000	40.040	-0.1	110	0.00
3	Pyridine	40.000	40.525	-1.3	108	0.00
4	n-Nitrosodimethylamine	40.000	40.461	-1.2	107	0.00
5 S	2-Fluorophenol	80.000	79.762	0.3	109	0.00
6	Aniline	40.000	40.238	-0.6	111	0.00
7 S	Phenol-d6	80.000	80.150	-0.2	110	0.00
8	2-Chlorophenol	40.000	38.983	2.5	108	0.00
9	Benzaldehyde	40.000	37.862	5.3	106	0.00
10 C	Phenol	40.000	40.116	-0.3	109	0.00
11	bis(2-Chloroethyl)ether	40.000	40.805	-2.0	114	0.00
12	1,3-Dichlorobenzene	40.000	39.250	1.9	110	0.00
13 C	1,4-Dichlorobenzene	40.000	39.060	2.3	109	0.00
14	1,2-Dichlorobenzene	40.000	39.323	1.7	111	0.00
15	Benzyl Alcohol	40.000	40.568	-1.4	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	46.611	-16.5	135	0.00
17	2-Methylphenol	40.000	40.341	-0.9	113	0.00
18	Hexachloroethane	40.000	39.571	1.1	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.970	-7.4	121	0.00
20	3+4-Methylphenols	40.000	40.586	-1.5	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	119	0.00
22	Acetophenone	40.000	41.140	-2.9	115	0.00
23 S	Nitrobenzene-d5	80.000	84.136	-5.2	113	0.00
24	Nitrobenzene	40.000	42.139	-5.3	117	0.00
25	Isophorone	40.000	40.926	-2.3	114	0.00
26 C	2-Nitrophenol	40.000	40.797	-2.0	111	0.00
27	2,4-Dimethylphenol	40.000	41.152	-2.9	114	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.954	-4.9	120	0.00
29 C	2,4-Dichlorophenol	40.000	39.388	1.5	111	0.00
30	1,2,4-Trichlorobenzene	40.000	38.673	3.3	109	0.00
31	Naphthalene	40.000	40.564	-1.4	115	0.00
32	Benzoic acid	40.000	34.515	13.7	100	0.00
33	4-Chloroaniline	40.000	40.630	-1.6	114	0.00
34 C	Hexachlorobutadiene	40.000	37.470	6.3	104	0.00
35	Caprolactam	40.000	40.246	-0.6	117	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.274	-3.2	117	0.00
37	2-Methylnaphthalene	40.000	39.469	1.3	114	0.00
38	1-Methylnaphthalene	40.000	39.675	0.8	115	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.893	2.8	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.762	5.6	102	0.00
42 S	2,4,6-Tribromophenol	80.000	74.272	7.2	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.549	1.1	109	0.00
44	2,4,5-Trichlorophenol	40.000	39.789	0.5	111	0.00
45 S	2-Fluorobiphenyl	80.000	80.890	-1.1	115	0.00
46	1,1'-Biphenyl	40.000	40.893	-2.2	113	0.00
47	2-Chloronaphthalene	40.000	40.408	-1.0	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031323\
 Data File : BP014074.D
 Acq On : 13 Mar 2023 21:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 04:25:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 04:15:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.058	-12.6	121	0.00
49	Acenaphthylene	40.000	40.698	-1.7	114	0.00
50	Dimethylphthalate	40.000	39.509	1.2	116	0.00
51	2,6-Dinitrotoluene	40.000	40.569	-1.4	115	0.00
52 C	Acenaphthene	40.000	40.549	-1.4	114	0.00
53	3-Nitroaniline	40.000	41.587	-4.0	116	0.00
54 P	2,4-Dinitrophenol	40.000	35.473	11.3	106	0.00
55	Dibenzofuran	40.000	40.740	-1.9	114	0.00
56 P	4-Nitrophenol	40.000	39.081	2.3	113	0.00
57	2,4-Dinitrotoluene	40.000	40.362	-0.9	112	0.00
58	Fluorene	40.000	39.966	0.1	114	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.849	5.4	107	0.00
60	Diethylphthalate	40.000	39.541	1.1	115	0.00
61	4-Chlorophenyl-phenylether	40.000	38.924	2.7	110	0.00
62	4-Nitroaniline	40.000	41.004	-2.5	115	0.00
63	Azobenzene	40.000	43.813	-9.5	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	119	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.874	2.8	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.348	-3.4	113	0.00
67	4-Bromophenyl-phenylether	40.000	38.463	3.8	106	0.00
68	Hexachlorobenzene	40.000	38.384	4.0	104	0.00
69	Atrazine	40.000	38.974	2.6	115	0.00
70 C	Pentachlorophenol	40.000	39.112	2.2	105	0.00
71	Phenanthrene	40.000	40.295	-0.7	113	0.00
72	Anthracene	40.000	40.343	-0.9	112	0.00
73	Carbazole	40.000	40.085	-0.2	113	0.00
74	Di-n-butylphthalate	40.000	39.523	1.2	113	0.00
75 C	Fluoranthene	40.000	37.406	6.5	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	37.526	6.2	84	0.00
78	Pyrene	40.000	43.593	-9.0	107	0.00
79 S	Terphenyl-d14	80.000	84.379	-5.5	104	0.00
80	Butylbenzylphthalate	40.000	40.709	-1.8	95	0.00
81	Benzo(a)anthracene	40.000	40.245	-0.6	93	0.00
82	3,3'-Dichlorobenzidine	40.000	38.120	4.7	84	0.00
83	Chrysene	40.000	40.453	-1.1	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.001	5.0	86	0.00
85 c	Di-n-octyl phthalate	40.000	35.988	10.0	76	0.00
86 I	Perylene-d12	20.000	20.000	0.0	80	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.875	-7.2	75	0.00
88	Benzo(b)fluoranthene	40.000	40.884	-2.2	78	0.00
89	Benzo(k)fluoranthene	40.000	41.698	-4.2	80	0.00
90 C	Benzo(a)pyrene	40.000	41.178	-2.9	77	0.00
91	Dibenzo(a,h)anthracene	40.000	42.882	-7.2	75	0.00
92	Benzo(g,h,i)perylene	40.000	43.891	-9.7	76	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031323\
Data File : BP014074.D
Acq On : 13 Mar 2023 21:25
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 14 04:25:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 14 04:15:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0