

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031422\
 Data File : BP009379.D
 Acq On : 14 Mar 2022 13:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 00:58:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	63	0.00
2	1,4-Dioxane	40.000	37.243	6.9	65	0.00
3	Pyridine	40.000	40.986	-2.5	69	0.00
4	n-Nitrosodimethylamine	40.000	40.532	-1.3	67	0.00
5 S	2-Fluorophenol	80.000	79.605	0.5	68	0.01
6	Aniline	40.000	40.206	-0.5	69	0.00
7 S	Phenol-d6	80.000	80.603	-0.8	69	0.03
8	2-Chlorophenol	40.000	41.729	-4.3	70	0.00
9	Benzaldehyde	40.000	38.260	4.4	64	0.00
10 C	Phenol	40.000	39.766	0.6	68	0.02
11	bis(2-Chloroethyl)ether	40.000	39.253	1.9	66	0.00
12	1,3-Dichlorobenzene	40.000	40.083	-0.2	70	0.00
13 C	1,4-Dichlorobenzene	40.000	40.090	-0.2	70	0.00
14	1,2-Dichlorobenzene	40.000	40.468	-1.2	70	0.00
15	Benzyl Alcohol	40.000	42.409	-6.0	72	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.386	16.5	58	0.00
17	2-Methylphenol	40.000	41.682	-4.2	71	0.02
18	Hexachloroethane	40.000	40.013	-0.0	69	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.044	-2.6	69	0.00
20	3+4-Methylphenols	40.000	41.748	-4.4	70	0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	62	0.00
22	Acetophenone	40.000	41.352	-3.4	70	0.00
23 S	Nitrobenzene-d5	80.000	87.463	-9.3	71	0.00
24	Nitrobenzene	40.000	41.715	-4.3	69	0.00
25	Isophorone	40.000	42.209	-5.5	70	0.00
26 C	2-Nitrophenol	40.000	53.663	-34.2#	85	0.00
27	2,4-Dimethylphenol	40.000	41.673	-4.2	70	0.01
28	bis(2-Chloroethoxy)methane	40.000	40.043	-0.1	67	0.00
29 C	2,4-Dichlorophenol	40.000	44.837	-12.1	73	0.01
30	1,2,4-Trichlorobenzene	40.000	42.626	-6.6	73	0.00
31	Naphthalene	40.000	41.076	-2.7	70	0.00
32	Benzoic acid	40.000	52.282	-30.7#	84	0.06
33	4-Chloroaniline	40.000	43.052	-7.6	72	0.00
34 C	Hexachlorobutadiene	40.000	45.269	-13.2	78	-0.01
35	Caprolactam	40.000	52.394	-31.0#	83	0.02
36 C	4-Chloro-3-methylphenol	40.000	45.424	-13.6	75	0.03
37	2-Methylnaphthalene	40.000	41.833	-4.6	71	0.00
38	1-Methylnaphthalene	40.000	42.029	-5.1	72	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	67	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.919	-7.3	78	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.854	2.9	70	0.00
42 S	2,4,6-Tribromophenol	80.000	116.376	-45.5#	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.182	-15.5	80	0.00
44	2,4,5-Trichlorophenol	40.000	45.651	-14.1	80	0.02
45 S	2-Fluorobiphenyl	80.000	81.830	-2.3	75	0.00
46	1,1'-Biphenyl	40.000	40.007	-0.0	73	0.00
47	2-Chloronaphthalene	40.000	41.241	-3.1	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031422\
 Data File : BP009379.D
 Acq On : 14 Mar 2022 13:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 00:58:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	50.088	-25.2#	83	0.00
49	Acenaphthylene	40.000	42.205	-5.5	75	0.00
50	Dimethylphthalate	40.000	42.881	-7.2	78	0.00
51	2,6-Dinitrotoluene	40.000	47.911	-19.8	81	0.00
52 C	Acenaphthene	40.000	41.744	-4.4	75	0.00
53	3-Nitroaniline	40.000	48.762	-21.9	83	0.01
54 P	2,4-Dinitrophenol	40.000	49.686	-24.2	82	0.01
55	Dibenzofuran	40.000	41.736	-4.3	76	0.00
56 P	4-Nitrophenol	40.000	48.530	-21.3	81	0.05
57	2,4-Dinitrotoluene	40.000	52.289	-30.7#	86	0.00
58	Fluorene	40.000	42.650	-6.6	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	49.006	-22.5	87	0.01
60	Diethylphthalate	40.000	44.046	-10.1	79	0.00
61	4-Chlorophenyl-phenylether	40.000	44.000	-10.0	81	0.00
62	4-Nitroaniline	40.000	53.201	-33.0#	89	0.00
63	Azobenzene	40.000	40.062	-0.2	71	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	73	0.00
65	4,6-Dinitro-2-methylphenol	40.000	55.604	-39.0#	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.626	-1.6	79	0.00
67	4-Bromophenyl-phenylether	40.000	46.247	-15.6	90	0.00
68	Hexachlorobenzene	40.000	46.495	-16.2	90	0.00
69	Atrazine	40.000	45.759	-14.4	85	0.00
70 C	Pentachlorophenol	40.000	45.763	-14.4	83	0.01
71	Phenanthrene	40.000	40.743	-1.9	79	0.00
72	Anthracene	40.000	41.692	-4.2	80	0.00
73	Carbazole	40.000	41.592	-4.0	79	0.01
74	Di-n-butylphthalate	40.000	43.712	-9.3	80	0.00
75 C	Fluoranthene	40.000	43.614	-9.0	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	76	0.00
77	Benzydine	40.000	41.587	-4.0	75	0.00
78	Pyrene	40.000	40.586	-1.5	82	0.00
79 S	Terphenyl-d14	80.000	85.094	-6.4	87	0.00
80	Butylbenzylphthalate	40.000	45.412	-13.5	86	0.00
81	Benzo(a)anthracene	40.000	40.833	-2.1	83	0.00
82	3,3'-Dichlorobenzidine	40.000	47.454	-18.6	91	0.00
83	Chrysene	40.000	41.045	-2.6	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.787	-2.0	78	0.00
85 c	Di-n-octyl phthalate	40.000	44.151	-10.4	83	0.00
86 I	Perylene-d12	20.000	20.000	0.0	79	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.843	-4.6	87	0.01
88	Benzo(b)fluoranthene	40.000	41.109	-2.8	84	0.01
89	Benzo(k)fluoranthene	40.000	39.227	1.9	83	0.00
90 C	Benzo(a)pyrene	40.000	41.221	-3.1	85	0.00
91	Dibenzo(a,h)anthracene	40.000	41.783	-4.5	87	0.02
92	Benzo(g,h,i)perylene	40.000	42.174	-5.4	87	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031422\
Data File : BP009379.D
Acq On : 14 Mar 2022 13:22
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 15 00:58:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 01:55:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1