


Benzaldehyde


Response Ratio



$R = -4.758e-002 A^2 + 7.938e-001 A + 8.574e-002$
Coef of Det (r^2) = 0.991439 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP031722.M
Calibration Table Last Updated: Thu Mar 17 18:26:53 2022

Hexachlorocyclopentadiene

Response Ratio




$$\text{Response} = 3.178\text{e-}001 * \text{Amt} - 1.158\text{e-}001$$

Coef of Det (r^2) = 0.996944 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP031722.M

Calibration Table Last Updated: Thu Mar 17 18:26:53 2022

2,4-Dinitrophenol



Response = 2.092e-001 * Amt - 4.740e-002
 Coef of Det (r^2) = 0.990711 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP031722.M
 Calibration Table Last Updated: Thu Mar 17 18:26:53 2022