

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
 Data File : BP009452.D
 Acq On : 18 Mar 2022 11:53
 Operator : CG/JU
 Sample : PB143161BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB143161BL

Quant Time: Mar 18 14:16:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 18:26:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	299876	20.000	ng	0.00
21) Naphthalene-d8	10.593	136	1161990	20.000	ng	0.00
39) Acenaphthene-d10	14.445	164	747760	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1559287	20.000	ng	0.00
76) Chrysene-d12	21.316	240	1363912	20.000	ng	0.00
86) Perylene-d12	23.721	264	1328330	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.405	112	2258340	131.485	ng	0.00
7) Phenol-d6	6.993	99	2745950	118.219	ng	0.00
23) Nitrobenzene-d5	8.958	82	1762520	80.604	ng	0.00
42) 2,4,6-Tribromophenol	15.940	330	1413698	114.568	ng	0.00
45) 2-Fluorobiphenyl	13.063	172	4373469	81.081	ng	0.00
79) Terphenyl-d14	19.780	244	6785050	89.290	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
Data File : BP009452.D
Acq On : 18 Mar 2022 11:53
Operator : CG/JU
Sample : PB143161BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB143161BL

Quant Time: Mar 18 14:16:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 18:26:53 2022
Response via : Initial Calibration

