

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
 Data File : BP009451.D
 Acq On : 18 Mar 2022 11:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 18:26:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.454	0.457	-0.7	128	0.00
3	Pyridine	1.222	1.217	0.4	124	0.00
4	n-Nitrosodimethylamine	0.450	0.449	0.2	126	0.00
5 S	2-Fluorophenol	1.146	1.131	1.3	124	0.00
6	Aniline	1.766	1.703	3.6	121	0.00
7 S	Phenol-d6	1.549	1.477	4.6	120	0.00
8	2-Chlorophenol	1.306	1.242	4.9	120	0.00
9	Benzaldehyde	0.821	0.725	11.7	126	0.00
10 C	Phenol	1.554	1.472	5.3	119	0.00
11	bis(2-Chloroethyl)ether	1.230	1.176	4.4	121	0.00
12	1,3-Dichlorobenzene	1.474	1.412	4.2	121	0.00
13 C	1,4-Dichlorobenzene	1.507	1.439	4.5	121	0.00
14	1,2-Dichlorobenzene	1.457	1.388	4.7	121	0.00
15	Benzyl Alcohol	1.235	1.153	6.6	117	0.00
16	2,2'-oxybis(1-Chloropropane	1.333	1.286	3.5	122	0.00
17	2-Methylphenol	1.071	1.003	6.3	118	0.00
18	Hexachloroethane	0.528	0.502	4.9	120	0.00
19 P	n-Nitroso-di-n-propylamine	0.980	0.910	7.1	118	0.00
20	3+4-Methylphenols	1.473	1.377	6.5	118	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
22	Acetophenone	0.495	0.477	3.6	119	0.00
23 S	Nitrobenzene-d5	0.376	0.367	2.4	119	0.00
24	Nitrobenzene	0.356	0.346	2.8	119	0.00
25	Isophorone	0.642	0.616	4.0	119	0.00
26 C	2-Nitrophenol	0.185	0.180	2.7	117	0.00
27	2,4-Dimethylphenol	0.261	0.253	3.1	118	0.00
28	bis(2-Chloroethoxy)methane	0.418	0.398	4.8	117	0.00
29 C	2,4-Dichlorophenol	0.319	0.300	6.0	114	0.00
30	1,2,4-Trichlorobenzene	0.370	0.355	4.1	117	0.00
31	Naphthalene	1.030	0.987	4.2	118	0.00
32	Benzoic acid	0.204	0.207	-1.5	121	0.00
33	4-Chloroaniline	0.420	0.400	4.8	117	0.00
34 C	Hexachlorobutadiene	0.250	0.238	4.8	117	0.00
35	Caprolactam	0.108	0.102	5.6	120	0.01
36 C	4-Chloro-3-methylphenol	0.343	0.326	5.0	119	0.00
37	2-Methylnaphthalene	0.723	0.683	5.5	118	0.00
38	1-Methylnaphthalene	0.704	0.662	6.0	117	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	0.664	0.638	3.9	116	0.00
41 P	Hexachlorocyclopentadiene	0.234	0.245	-4.7	120	0.00
42 S	2,4,6-Tribromophenol	0.330	0.307	7.0	113	0.00
43 C	2,4,6-Trichlorophenol	0.440	0.430	2.3	117	0.00
44	2,4,5-Trichlorophenol	0.478	0.466	2.5	117	0.00
45 S	2-Fluorobiphenyl	1.443	1.390	3.7	116	0.00
46	1,1'-Biphenyl	1.492	1.436	3.8	117	0.00
47	2-Chloronaphthalene	1.175	1.133	3.6	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
 Data File : BP009451.D
 Acq On : 18 Mar 2022 11:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 18:26:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.320	0.318	0.6	119	0.00
49	Acenaphthylene	1.813	1.755	3.2	118	0.00
50	Dimethylphthalate	1.515	1.416	6.5	114	0.00
51	2,6-Dinitrotoluene	0.318	0.304	4.4	115	0.00
52 C	Acenaphthene	1.083	1.043	3.7	118	0.00
53	3-Nitroaniline	0.335	0.327	2.4	118	0.00
54 P	2,4-Dinitrophenol	0.174	0.174	0.0	118	0.00
55	Dibenzofuran	1.820	1.743	4.2	117	0.00
56 P	4-Nitrophenol	0.250	0.242	3.2	113	0.00
57	2,4-Dinitrotoluene	0.437	0.419	4.1	115	0.00
58	Fluorene	1.431	1.350	5.7	116	0.00
59	2,3,4,6-Tetrachlorophenol	0.433	0.421	2.8	118	0.00
60	Diethylphthalate	1.517	1.399	7.8	113	0.00
61	4-Chlorophenyl-phenylether	0.788	0.740	6.1	115	0.00
62	4-Nitroaniline	0.352	0.330	6.2	112	0.00
63	Azobenzene	1.319	1.272	3.6	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	0.129	0.133	-3.1	114	0.00
66 c	n-Nitrosodiphenylamine	0.592	0.577	2.5	115	0.00
67	4-Bromophenyl-phenylether	0.245	0.236	3.7	115	0.00
68	Hexachlorobenzene	0.282	0.270	4.3	114	0.00
69	Atrazine	0.208	0.204	1.9	112	0.00
70 C	Pentachlorophenol	0.151	0.158	-4.6	117	0.00
71	Phenanthrene	1.071	1.042	2.7	116	0.00
72	Anthracene	1.057	1.031	2.5	116	0.00
73	Carbazole	1.034	0.991	4.2	114	0.00
74	Di-n-butylphthalate	1.211	1.157	4.5	112	0.00
75 C	Fluoranthene	1.296	1.261	2.7	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
77	Benzidine	0.490	0.457	6.7	122	0.00
78	Pyrene	1.318	1.225	7.1	116	0.00
79 S	Terphenyl-d14	1.114	1.023	8.2	114	0.00
80	Butylbenzylphthalate	0.560	0.522	6.8	117	0.00
81	Benzo(a)anthracene	1.314	1.274	3.0	124	0.00
82	3,3'-Dichlorobenzidine	0.508	0.483	4.9	122	0.00
83	Chrysene	1.244	1.201	3.5	123	0.00
84	Bis(2-ethylhexyl)phthalate	0.781	0.729	6.7	118	0.00
85 c	Di-n-octyl phthalate	1.346	1.295	3.8	123	0.00
86 I	Perylene-d12	1.000	1.000	0.0	128	0.00
87	Indeno(1,2,3-cd)pyrene	1.517	1.474	2.8	128	0.01
88	Benzo(b)fluoranthene	1.194	1.154	3.4	129	0.00
89	Benzo(k)fluoranthene	1.169	1.140	2.5	128	0.00
90 C	Benzo(a)pyrene	1.022	1.002	2.0	130	0.00
91	Dibenzo(a,h)anthracene	1.265	1.226	3.1	127	0.01
92	Benzo(g,h,i)perylene	1.244	1.213	2.5	128	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
Data File : BP009451.D
Acq On : 18 Mar 2022 11:19
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 18:26:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0