

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
 Data File : BP009451.D
 Acq On : 18 Mar 2022 11:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 18:26:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	121	0.00
2	1,4-Dioxane	40.000	40.255	-0.6	128	0.00
3	Pyridine	40.000	39.836	0.4	124	0.00
4	n-Nitrosodimethylamine	40.000	39.908	0.2	126	0.00
5 S	2-Fluorophenol	80.000	79.010	1.2	124	0.00
6	Aniline	40.000	38.584	3.5	121	0.00
7 S	Phenol-d6	80.000	76.251	4.7	120	0.00
8	2-Chlorophenol	40.000	38.040	4.9	120	0.00
9	Benzaldehyde	40.000	38.933	2.7	126	0.00
10 C	Phenol	40.000	37.889	5.3	119	0.00
11	bis(2-Chloroethyl)ether	40.000	38.231	4.4	121	0.00
12	1,3-Dichlorobenzene	40.000	38.299	4.3	121	0.00
13 C	1,4-Dichlorobenzene	40.000	38.179	4.6	121	0.00
14	1,2-Dichlorobenzene	40.000	38.115	4.7	121	0.00
15	Benzyl Alcohol	40.000	37.336	6.7	117	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.573	3.6	122	0.00
17	2-Methylphenol	40.000	37.449	6.4	118	0.00
18	Hexachloroethane	40.000	37.981	5.0	120	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.139	7.2	118	0.00
20	3+4-Methylphenols	40.000	37.415	6.5	118	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	119	0.00
22	Acetophenone	40.000	38.557	3.6	119	0.00
23 S	Nitrobenzene-d5	80.000	78.073	2.4	119	0.00
24	Nitrobenzene	40.000	38.936	2.7	119	0.00
25	Isophorone	40.000	38.406	4.0	119	0.00
26 C	2-Nitrophenol	40.000	38.794	3.0	117	0.00
27	2,4-Dimethylphenol	40.000	38.692	3.3	118	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.084	4.8	117	0.00
29 C	2,4-Dichlorophenol	40.000	37.653	5.9	114	0.00
30	1,2,4-Trichlorobenzene	40.000	38.438	3.9	117	0.00
31	Naphthalene	40.000	38.322	4.2	118	0.00
32	Benzoic acid	40.000	40.508	-1.3	121	0.00
33	4-Chloroaniline	40.000	38.082	4.8	117	0.00
34 C	Hexachlorobutadiene	40.000	38.054	4.9	117	0.00
35	Caprolactam	40.000	37.815	5.5	120	0.01
36 C	4-Chloro-3-methylphenol	40.000	38.047	4.9	119	0.00
37	2-Methylnaphthalene	40.000	37.773	5.6	118	0.00
38	1-Methylnaphthalene	40.000	37.612	6.0	117	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.458	3.9	116	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.123	4.7	120	0.00
42 S	2,4,6-Tribromophenol	80.000	74.320	7.1	113	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.090	2.3	117	0.00
44	2,4,5-Trichlorophenol	40.000	38.988	2.5	117	0.00
45 S	2-Fluorobiphenyl	80.000	77.077	3.7	116	0.00
46	1,1'-Biphenyl	40.000	38.487	3.8	117	0.00
47	2-Chloronaphthalene	40.000	38.585	3.5	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
 Data File : BP009451.D
 Acq On : 18 Mar 2022 11:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 18:26:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.732	0.7	119	0.00
49	Acenaphthylene	40.000	38.703	3.2	118	0.00
50	Dimethylphthalate	40.000	37.364	6.6	114	0.00
51	2,6-Dinitrotoluene	40.000	38.303	4.2	115	0.00
52 C	Acenaphthene	40.000	38.532	3.7	118	0.00
53	3-Nitroaniline	40.000	39.039	2.4	118	0.00
54 P	2,4-Dinitrophenol	40.000	37.856	5.4	118	0.00
55	Dibenzofuran	40.000	38.313	4.2	117	0.00
56 P	4-Nitrophenol	40.000	38.669	3.3	113	0.00
57	2,4-Dinitrotoluene	40.000	38.402	4.0	115	0.00
58	Fluorene	40.000	37.731	5.7	116	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.900	2.8	118	0.00
60	Diethylphthalate	40.000	36.898	7.8	113	0.00
61	4-Chlorophenyl-phenylether	40.000	37.538	6.2	115	0.00
62	4-Nitroaniline	40.000	37.497	6.3	112	0.00
63	Azobenzene	40.000	38.577	3.6	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.306	-3.3	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.014	2.5	115	0.00
67	4-Bromophenyl-phenylether	40.000	38.546	3.6	115	0.00
68	Hexachlorobenzene	40.000	38.349	4.1	114	0.00
69	Atrazine	40.000	39.278	1.8	112	0.00
70 C	Pentachlorophenol	40.000	42.104	-5.3	117	0.00
71	Phenanthrene	40.000	38.919	2.7	116	0.00
72	Anthracene	40.000	39.010	2.5	116	0.00
73	Carbazole	40.000	38.345	4.1	114	0.00
74	Di-n-butylphthalate	40.000	38.204	4.5	112	0.00
75 C	Fluoranthene	40.000	38.927	2.7	116	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	124	0.00
77	Benzydine	40.000	37.279	6.8	122	0.00
78	Pyrene	40.000	37.183	7.0	116	0.00
79 S	Terphenyl-d14	80.000	73.449	8.2	114	0.00
80	Butylbenzylphthalate	40.000	37.297	6.8	117	0.00
81	Benzo(a)anthracene	40.000	38.793	3.0	124	0.00
82	3,3'-Dichlorobenzidine	40.000	38.058	4.9	122	0.00
83	Chrysene	40.000	38.608	3.5	123	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.321	6.7	118	0.00
85 c	Di-n-octyl phthalate	40.000	38.490	3.8	123	0.00
86 I	Perylene-d12	20.000	20.000	0.0	128	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.865	2.8	128	0.01
88	Benzo(b)fluoranthene	40.000	38.673	3.3	129	0.00
89	Benzo(k)fluoranthene	40.000	39.009	2.5	128	0.00
90 C	Benzo(a)pyrene	40.000	39.241	1.9	130	0.00
91	Dibenzo(a,h)anthracene	40.000	38.776	3.1	127	0.01
92	Benzo(g,h,i)perylene	40.000	38.990	2.5	128	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031822\
Data File : BP009451.D
Acq On : 18 Mar 2022 11:19
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 18 14:15:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 18:26:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0