

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031825\
 Data File : BP024185.D
 Acq On : 18 Mar 2025 13:57
 Operator : RC/JU
 Sample : Q1583-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EFF-WASTE WATER

Quant Time: Mar 18 14:18:07 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 13 05:55:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.758	152	457444	20.000	ng	-0.01
21) Naphthalene-d8	10.546	136	1832298	20.000	ng	0.00
39) Acenaphthene-d10	14.393	164	1103670	20.000	ng	-0.01
64) Phenanthrene-d10	17.192	188	1992010	20.000	ng	-0.01
76) Chrysene-d12	21.645	240	1940679	20.000	ng	-0.01
86) Perylene-d12	25.039	264	2196024	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1406191m	49.070	ng	0.00
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	8.952	82	2842794	87.537	ng	0.03
42) 2,4,6-Tribromophenol	15.910	330	1282041	92.219	ng	0.00
45) 2-Fluorobiphenyl	13.004	172	5726243	76.504	ng	-0.01
79) Terphenyl-d14	19.916	244	9002320	95.754	ng	0.00
Target Compounds						
32) Benzoic acid	10.228	122	9911000m	516.442	ng	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031825\
Data File : BP024185.D
Acq On : 18 Mar 2025 13:57
Operator : RC/JU
Sample : Q1583-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EFF-WASTE WATER

Quant Time: Mar 18 14:18:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 13 05:55:58 2025
Response via : Initial Calibration

