

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032025\
 Data File : BP024192.D
 Acq On : 20 Mar 2025 14:52
 Operator : RC/JU
 Sample : PB167183BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK183

Quant Time: Mar 20 15:21:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030425.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 04 14:51:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.740	152	325699	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.516	136	1246057	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.381	164	695981	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.192	188	1333124	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	1406883	20.000	ng/u1	-0.01
88) Perylene-d12	25.027	264	1590659	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	58547	7.160	ng/uL	0.00
4) Pyridine-d5	3.675	84	765469	32.545	ng/u1	0.00
7) Phenol-d5	6.916	99	864344	29.911	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.075	67	520691	32.719	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.281	132	723465	32.613	ng/u1	-0.01
15) 4-Methylphenol-d8	8.440	113	667627	29.260	ng/u1	-0.02
21) Nitrobenzene-d5	8.887	128	330164	32.815	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.604	143	344819	32.826	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.146	165	585651	31.007	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.657	131	875150	30.316	ng/u1	-0.01
46) Dimethylphthalate-d6	13.787	166	1719392	34.092	ng/u1	-0.01
49) Acenaphthylene-d8	14.069	160	2011864	34.393	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.598	143	259181	25.102	ng/u1	0.00
60) Fluorene-d10	15.392	176	1487981	34.726	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.522	200	218671	25.980	ng/u1	0.00
73) Anthracene-d10	17.298	188	2297345	35.312	ng/u1	0.00
81) Pyrene-d10	19.669	212	2650720	37.607	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.798	264	3010808	36.510	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032025\
Data File : BP024192.D
Acq On : 20 Mar 2025 14:52
Operator : RC/JU
Sample : PB167183BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK183

Quant Time: Mar 20 15:21:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030425.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 04 14:51:41 2025
Response via : Initial Calibration

