

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032221\
 Data File : BP005018.D
 Acq On : 22 Mar 2021 17:06
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP032221

Quant Time: Mar 22 17:42:20 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 17:08:54 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	44.436	-11.1	101	0.00
3	Pyridine	40.000	41.136	-2.8	100	0.00
4	n-Nitrosodimethylamine	40.000	42.864	-7.2	104	0.00
5 S	2-Fluorophenol	80.000	81.283	-1.6	101	0.00
6	Aniline	40.000	41.665	-4.2	100	0.00
7 S	Phenol-d6	80.000	82.821	-3.5	101	0.00
8	2-Chlorophenol	40.000	40.934	-2.3	99	0.00
9	Benzaldehyde	40.000	45.852	-14.6	102	0.00
10 C	Phenol	40.000	41.249	-3.1	100	0.00
11	bis(2-Chloroethyl)ether	40.000	41.896	-4.7	102	0.00
12	1,3-Dichlorobenzene	40.000	40.129	-0.3	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.709	0.7	97	0.00
14	1,2-Dichlorobenzene	40.000	39.931	0.2	98	0.00
15	Benzyl Alcohol	40.000	43.102	-7.8	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.911	-2.3	101	0.00
17	2-Methylphenol	40.000	40.875	-2.2	97	0.00
18	Hexachloroethane	40.000	40.683	-1.7	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.027	-5.1	98	0.00
20	3+4-Methylphenols	40.000	41.386	-3.5	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	40.393	-1.0	96	0.00
23 S	Nitrobenzene-d5	80.000	83.474	-4.3	100	0.00
24	Nitrobenzene	40.000	41.042	-2.6	100	0.01
25	Isophorone	40.000	41.957	-4.9	98	0.00
26 C	2-Nitrophenol	40.000	41.831	-4.6	97	0.00
27	2,4-Dimethylphenol	40.000	41.378	-3.4	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.035	-2.6	97	0.00
29 C	2,4-Dichlorophenol	40.000	41.178	-2.9	96	0.00
30	1,2,4-Trichlorobenzene	40.000	39.832	0.4	94	0.01
31	Naphthalene	40.000	40.499	-1.2	98	0.00
32	Benzoic acid	40.000	42.661	-6.7	103	0.00
33	4-Chloroaniline	40.000	41.530	-3.8	96	0.00
34 C	Hexachlorobutadiene	40.000	40.317	-0.8	94	0.00
35	Caprolactam	40.000	44.567	-11.4	98	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.478	-6.2	98	0.00
37	2-Methylnaphthalene	40.000	40.649	-1.6	95	0.00
38	1-Methylnaphthalene	40.000	41.110	-2.8	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	37.629	5.9	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.038	2.4	88	0.01
42 S	2,4,6-Tribromophenol	80.000	74.494	6.9	82	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.722	3.2	91	0.00
44	2,4,5-Trichlorophenol	40.000	38.310	4.2	88	0.00
45 S	2-Fluorobiphenyl	80.000	76.887	3.9	93	0.00
46	1,1'-Biphenyl	40.000	38.397	4.0	94	0.00
47	2-Chloronaphthalene	40.000	38.252	4.4	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032221\
 Data File : BP005018.D
 Acq On : 22 Mar 2021 17:06
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP032221

Quant Time: Mar 22 17:42:20 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 17:08:54 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.319	-3.3	98	0.00
49	Acenaphthylene	40.000	39.001	2.5	94	0.00
50	Dimethylphthalate	40.000	38.969	2.6	93	0.00
51	2,6-Dinitrotoluene	40.000	39.534	1.2	92	0.00
52 C	Acenaphthene	40.000	38.538	3.7	94	0.00
53	3-Nitroaniline	40.000	40.136	-0.3	92	0.01
54 P	2,4-Dinitrophenol	40.000	40.307	-0.8	92	0.00
55	Dibenzofuran	40.000	38.711	3.2	93	0.01
56 P	4-Nitrophenol	40.000	40.686	-1.7	98	0.00
57	2,4-Dinitrotoluene	40.000	40.881	-2.2	95	0.01
58	Fluorene	40.000	38.286	4.3	92	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	38.429	3.9	87	0.00
60	Diethylphthalate	40.000	39.687	0.8	93	0.00
61	4-Chlorophenyl-phenylether	40.000	38.514	3.7	92	0.00
62	4-Nitroaniline	40.000	41.392	-3.5	92	0.00
63	Azobenzene	40.000	39.879	0.3	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.01
65	4,6-Dinitro-2-methylphenol	40.000	42.378	-5.9	86	0.01
66 c	n-Nitrosodiphenylamine	40.000	40.920	-2.3	92	0.00
67	4-Bromophenyl-phenylether	40.000	40.708	-1.8	88	0.00
68	Hexachlorobenzene	40.000	39.536	1.2	86	0.01
69	Atrazine	40.000	42.860	-7.1	89	0.00
70 C	Pentachlorophenol	40.000	40.975	-2.4	84	0.01
71	Phenanthrene	40.000	40.438	-1.1	90	0.00
72	Anthracene	40.000	40.905	-2.3	91	0.00
73	Carbazole	40.000	41.053	-2.6	92	0.00
74	Di-n-butylphthalate	40.000	42.567	-6.4	93	0.00
75 C	Fluoranthene	40.000	40.357	-0.9	89	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzydine	40.000	47.720	-19.3	86	0.00
78	Pyrene	40.000	41.882	-4.7	89	0.00
79 S	Terphenyl-d14	80.000	82.497	-3.1	86	0.01
80	Butylbenzylphthalate	40.000	43.849	-9.6	92	0.00
81	Benzo(a)anthracene	40.000	40.941	-2.4	88	0.01
82	3,3'-Dichlorobenzidine	40.000	41.820	-4.6	85	0.00
83	Chrysene	40.000	40.495	-1.2	85	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.689	-9.2	92	0.00
85 c	Di-n-octyl phthalate	40.000	43.677	-9.2	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	83	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.891	-2.2	84	0.01
88	Benzo(b)fluoranthene	40.000	42.185	-5.5	87	0.01
89	Benzo(k)fluoranthene	40.000	40.086	-0.2	82	0.01
90 C	Benzo(a)pyrene	40.000	41.255	-3.1	85	0.01
91	Dibenzo(a,h)anthracene	40.000	40.975	-2.4	84	0.01
92	Benzo(g,h,i)perylene	40.000	40.460	-1.2	83	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032221\
Data File : BP005018.D
Acq On : 22 Mar 2021 17:06
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP032221

Quant Time: Mar 22 17:42:20 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 22 17:08:54 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0