

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014226.D
 Acq On : 22 Mar 2023 20:29
 Operator : CG/JU
 Sample : 01977-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EYG89

Quant Time: Mar 23 02:35:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 02:31:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	164918	20.000	ng/ul	0.00
20) Naphthalene-d8	10.863	136	744744	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.686	164	549583	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.445	188	1301558	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.527	240	1335032	20.000	ng/ul	# 0.00
88) Perylene-d12	24.051	264	1385181	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	28857	6.611	ng/uL	0.00
4) Pyridine-d5	3.840	84	389880	32.261	ng/ul	0.00
7) Phenol-d5	7.199	99	552925	37.992	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.363	67	349724	41.204	ng/ul	0.00
11) 2-Chlorophenol-d4	7.569	132	406653	36.675	ng/ul	0.00
15) 4-Methylphenol-d8	8.757	113	465339	39.144	ng/ul	0.00
21) Nitrobenzene-d5	9.216	128	225277	37.911	ng/ul	0.00
24) 2-Nitrophenol-d4	9.946	143	266739	37.499	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	448729	33.527	ng/ul	0.00
31) 4-Chloroaniline-d4	11.004	131	620986	35.060	ng/ul	0.00
46) Dimethylphthalate-d6	14.092	166	1725572	36.996	ng/ul	0.00
49) Acenaphthylene-d8	14.381	160	1839644	37.233	ng/ul	0.00
54) 4-Nitrophenol-d4	14.892	143	276189	33.602	ng/ul	0.00
60) Fluorene-d10	15.681	176	1458158	36.888	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	308802	31.391	ng/ul	0.00
73) Anthracene-d10	17.545	188	2392506	37.865	ng/ul	0.00
81) Pyrene-d10	19.774	212	2909342	39.512	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.892	264	2818452	38.252	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
Data File : BP014226.D
Acq On : 22 Mar 2023 20:29
Operator : CG/JU
Sample : 01977-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EYG89

Quant Time: Mar 23 02:35:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 23 02:31:48 2023
Response via : Initial Calibration

