

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014263.D
 Acq On : 23 Mar 2023 18:42
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020789

Quant Time: Mar 24 00:52:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	161022	20.000	ng/u1	0.00
20) Naphthalene-d8	10.857	136	764216	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.681	164	553108	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1298983	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.522	240	1344143	20.000	ng/u1	# 0.00
88) Perylene-d12	24.039	264	1407552	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	30323	7.115	ng/uL	0.00
4) Pyridine-d5	3.834	84	216345	18.335	ng/u1	0.00
7) Phenol-d5	7.193	99	293273	20.639	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.358	67	186232	22.472	ng/u1	0.00
11) 2-Chlorophenol-d4	7.564	132	218535	20.186	ng/u1	0.00
15) 4-Methylphenol-d8	8.752	113	251256	21.647	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	117726	19.307	ng/u1	0.00
24) 2-Nitrophenol-d4	9.934	143	138219	18.936	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	251748	18.330	ng/u1	0.00
31) 4-Chloroaniline-d4	10.999	131	350130	19.264	ng/u1	0.00
46) Dimethylphthalate-d6	14.087	166	876171	18.665	ng/u1	0.00
49) Acenaphthylene-d8	14.375	160	960783	19.321	ng/u1	0.00
54) 4-Nitrophenol-d4	14.887	143	146705	17.735	ng/u1	0.00
60) Fluorene-d10	15.675	176	741927	18.650	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	170470	17.363	ng/u1	0.00
73) Anthracene-d10	17.539	188	1207593	19.150	ng/u1	0.00
81) Pyrene-d10	19.763	212	1469269	19.819	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	1420739	18.976	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	32114	6.980	ng/uL	92
5) Pyridine	3.858	79	222996	18.635	ng/u1	96
6) Benzaldehyde	7.175	77	175365	24.803	ng/u1	99
8) Phenol	7.222	94	307704	20.998	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.452	93	244336	21.764	ng/u1	95
12) 2-Chlorophenol	7.599	128	225319	20.657	ng/u1	95
13) 2-Methylphenol	8.481	108	235809	21.641	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.569	45	318061	26.234	ng/u1	99
16) Acetophenone	8.869	105	412326	21.990	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.846	70	229033	23.876	ng/u1	96
18) 4-Methylphenol	8.816	108	263203	21.960	ng/u1	99
19) Hexachloroethane	9.122	117	96572	20.215	ng/u1	97
22) Nitrobenzene	9.252	77	325903	19.980	ng/u1	98
23) Isophorone	9.781	82	643014	20.471	ng/u1	98
25) 2-Nitrophenol	9.969	139	144351	19.188	ng/u1	97
26) 2,4-Dimethylphenol	10.022	107	326317	19.817	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.258	93	374287	20.600	ng/u1	99
29) 2,4-Dichlorophenol	10.505	162	251069	18.532	ng/u1	95
30) Naphthalene	10.910	128	813674	19.389	ng/u1	100
32) 4-Chloroaniline	11.022	127	358650	19.610	ng/u1	98
33) Hexachlorobutadiene	11.181	225	179381	16.639	ng/u1	98
34) Caprolactam	11.810	113	46547	10.581	ng/u1	98
35) 4-Chloro-3-methylphenol	12.140	107	320006	20.427	ng/u1	96
36) 2-Methylnaphthalene	12.510	142	570365	19.377	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014263.D
 Acq On : 23 Mar 2023 18:42
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020789

Quant Time: Mar 24 00:52:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.728	142	573317	19.379	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.875	216	351910	17.594	ng/ul	99
40) Hexachlorocyclopentadiene	12.851	237	207815	14.510	ng/ul	100
41) 2,4,6-Trichlorophenol	13.116	196	233174	18.558	ng/ul	99
42) 2,4,5-Trichlorophenol	13.193	196	253353	18.373	ng/ul	99
43) 1,1'-Biphenyl	13.516	154	825979	19.397	ng/ul	98
44) 2-Chloronaphthalene	13.557	162	648711	19.181	ng/ul	97
45) 2-Nitroaniline	13.769	65	214365	21.431	ng/ul	92
47) Dimethylphthalate	14.134	163	880735	18.956	ng/ul	99
48) 2,6-Dinitrotoluene	14.263	165	178720	19.456	ng/ul	94
50) Acenaphthylene	14.404	152	1020655	19.558	ng/ul	99
51) 3-Nitroaniline	14.593	138	169208	20.804	ng/ul	94
52) Acenaphthene	14.746	153	712395	19.527	ng/ul	99
53) 2,4-Dinitrophenol	14.804	184	115546	16.844	ng/ul	94
55) 4-Nitrophenol	14.904	109	154542	17.296	ng/ul	85
56) Dibenzofuran	15.081	168	1015272	19.129	ng/ul	100
57) 2,4-Dinitrotoluene	15.051	165	263191	19.177	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.310	232	237080	17.971	ng/ul	99
59) Diethylphthalate	15.493	149	896447	19.022	ng/ul	99
61) Fluorene	15.728	166	844444	19.194	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.722	204	446189	18.375	ng/ul	97
63) 4-Nitroaniline	15.763	138	177233	21.199	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.810	198	165180	17.635	ng/ul	93
67) N-Nitrosodiphenylamine	15.934	169	720037	19.627	ng/ul	99
68) 4-Bromophenyl-phenylether	16.616	248	287742	18.323	ng/ul	98
69) Hexachlorobenzene	16.740	284	331700	17.925	ng/ul	98
70) Atrazine	16.892	200	290755	18.375	ng/ul	99
71) Pentachlorophenol	17.087	266	218134	18.034	ng/ul	99
72) Phenanthrene	17.481	178	1378630	19.459	ng/ul	100
74) Anthracene	17.575	178	1415532	19.705	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.481	216	367924	18.456	ng/ul	100
76) Pentachlorobenzene	14.998	250	377240	18.217	ng/ul	99
77) Carbazole	17.845	167	1256670	19.832	ng/ul	100
78) Di-n-butylphthalate	18.375	149	1576186	19.903	ng/ul	99
80) Fluoranthene	19.439	202	1701456	20.136	ng/ul#	95
82) Pyrene	19.792	202	1785283	20.253	ng/ul#	94
83) Butylbenzylphthalate	20.645	149	744384	20.944	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.427	252	629084	18.282	ng/ul#	99
85) Benzo(a)anthracene	21.504	228	1830957	19.367	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.398	149	1098548	20.673	ng/ul	98
87) Chrysene	21.557	228	1726019	19.156	ng/ul	99
89) Di-n-octyl phthalate	22.363	149	1884531	21.219	ng/ul	100
90) Benzo(b)fluoranthene	23.316	252	1749529	18.653	ng/ul	98
91) Benzo(k)fluoranthene	23.316	252	1753133	19.602	ng/ul	98
93) Benzo(a)pyrene	23.927	252	1555273	19.065	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.692	276	1815714	16.813	ng/ul#	96
95) Dibenzo(a,h)anthracene	26.704	278	1524122	16.980	ng/ul	98
96) Benzo(g,h,i)perylene	27.504	276	1416863	16.414	ng/ul#	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 6.980 ng/uL
RT: 3.440 min Scan# 43
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion: 88 Resp: 32114
Ion Ratio Lower Upper
88 100
43 27.8 23.9 35.9
58 70.2 50.2 75.2



#5
Pyridine
Concen: 18.635 ng/uL
RT: 3.858 min Scan# 114
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion: 79 Resp: 222996
Ion Ratio Lower Upper
79 100
52 73.2 55.1 82.7
51 36.4 28.9 43.3









#6
Benzaldehyde
Concen: 24.803 ng/ul
RT: 7.175 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion: 77 Resp: 175365
Ion Ratio Lower Upper
77 100
105 92.9 74.2 111.2
106 89.0 69.8 104.8



#8
Phenol
Concen: 20.998 ng/ul
RT: 7.222 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion: 94 Resp: 307704
Ion Ratio Lower Upper
94 100
65 30.6 25.2 37.8
66 44.3 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 21.764 ng/ul
RT: 7.452 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion: 93 Resp: 244336
Ion Ratio Lower Upper
93 100
63 78.6 58.8 88.2
95 31.4 25.8 38.6



#12
2-Chlorophenol
Concen: 20.657 ng/ul
RT: 7.599 min Scan# 750
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:128 Resp: 225319
Ion Ratio Lower Upper
128 100
64 53.6 38.3 57.5
130 31.8 25.7 38.5





#13

2-Methylphenol

Concen: 21.641 ng/ul

RT: 8.481 min Scan# 901

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :





SSTD020789

Tgt Ion:108 Resp: 235809

Ion Ratio Lower Upper

108 100

107 88.9 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 26.234 ng/ul

RT: 8.569 min Scan# 915

Delta R.T. 0.000 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion: 45 Resp: 318061


Ion Ratio Lower Upper

45 100

77 17.4 14.2 21.2

79 14.1 10.6 15.8









#16
Acetophenone
Concen: 21.990 ng/ul
RT: 8.869 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:105 Resp: 412326
Ion Ratio Lower Upper
105 100
77 82.5 65.4 98.2
51 27.6 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 23.876 ng/ul
RT: 8.846 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion: 70 Resp: 229033
Ion Ratio Lower Upper
70 100
42 53.3 40.9 61.3
101 9.6 7.9 11.9
130 17.4 16.6 25.0









#18
4-Methylphenol
Concen: 21.960 ng/ul
RT: 8.816 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:108 Resp: 263203
Ion Ratio Lower Upper
108 100
107 111.5 89.8 134.6



#19
Hexachloroethane
Concen: 20.215 ng/ul
RT: 9.122 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:117 Resp: 96572
Ion Ratio Lower Upper
117 100
201 116.0 91.7 137.5
199 69.7 59.1 88.7









#22
Nitrobenzene
Concen: 19.980 ng/ul
RT: 9.252 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion: 77 Resp: 325903
Ion Ratio Lower Upper
77 100
123 39.5 32.6 48.8
65 14.2 11.2 16.8



#23
Isophorone
Concen: 20.471 ng/ul
RT: 9.781 min Scan# 1121
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion: 82 Resp: 643014
Ion Ratio Lower Upper
82 100
95 7.8 7.1 10.7
138 15.8 13.3 19.9









#25
2-Nitrophenol
Concen: 19.188 ng/ul
RT: 9.969 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:139 Resp: 144351
Ion Ratio Lower Upper
139 100
65 58.9 46.6 70.0
109 38.7 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 19.817 ng/ul
RT: 10.022 min Scan# 1162
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:107 Resp: 326317
Ion Ratio Lower Upper
107 100
121 46.4 37.1 55.7
122 75.4 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.600 ng/ul

RT: 10.258 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion: 93 Resp: 374287

Ion Ratio Lower Upper

93 100

95 31.3 25.6 38.4

123 10.0 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 18.532 ng/ul

RT: 10.505 min Scan# 1244

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:162 Resp: 251069


Ion Ratio Lower Upper

162 100

164 63.3 52.0 78.0

98 42.2 29.1 43.7









#30
Naphthalene
Concen: 19.389 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:128 Resp: 813674
Ion Ratio Lower Upper
128 100
129 10.5 8.5 12.7
127 13.7 11.0 16.4



#32
4-Chloroaniline
Concen: 19.610 ng/ul
RT: 11.022 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:127 Resp: 358650
Ion Ratio Lower Upper
127 100
129 32.0 24.9 37.3





#33

Hexachlorobutadiene

Concen: 16.639 ng/ul

RT: 11.181 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:225 Resp: 179381

Ion Ratio Lower Upper

225 100

223 63.1 48.0 72.0

227 63.4 50.3 75.5



#34

Caprolactam

Concen: 10.581 ng/ul

RT: 11.810 min Scan# 1466

Delta R.T. -0.012 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:113 Resp: 46547


Ion Ratio Lower Upper

113 100

55 156.9 124.0 186.0

56 125.6 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 20.427 ng/ul
RT: 12.140 min Scan# 1522
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:107 Resp: 320006
Ion Ratio Lower Upper
107 100
144 21.7 19.4 29.0
142 71.1 59.6 89.4



#36
2-Methylnaphthalene
Concen: 19.377 ng/ul
RT: 12.510 min Scan# 1585
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:142 Resp: 570365
Ion Ratio Lower Upper
142 100
141 89.3 71.7 107.5





#37

1-Methylnaphthalene

Concen: 19.379 ng/ul

RT: 12.728 min Scan# 1622

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:142 Resp: 573317

Ion Ratio Lower Upper

142 100

141 93.9 73.8 110.6

116 4.4 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.594 ng/ul

RT: 12.875 min Scan# 1647

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Tgt Ion:216 Resp: 351910




Ion Ratio Lower Upper


216 100

214 77.7 61.7 92.5

179 21.2 16.2 24.4

108 17.5 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 14.510 ng/ul

RT: 12.851 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:237 Resp: 207815

Ion Ratio Lower Upper

237 100

235 61.1 49.0 73.6

272 12.0 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 18.558 ng/ul

RT: 13.116 min Scan# 1688

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:196 Resp: 233174


Ion Ratio Lower Upper

196 100

198 96.9 78.5 117.7

200 31.1 25.4 38.0









#42
2,4,5-Trichlorophenol
Concen: 18.373 ng/ul
RT: 13.193 min Scan# 1701
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:196 Resp: 253353
Ion Ratio Lower Upper
196 100
198 95.7 77.6 116.4
200 30.8 25.3 37.9



#43
1,1'-Biphenyl
Concen: 19.397 ng/ul
RT: 13.516 min Scan# 1756
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:154 Resp: 825979
Ion Ratio Lower Upper
154 100
153 39.8 32.6 49.0
76 12.5 9.4 14.0









#44
2-Chloronaphthalene
Concen: 19.181 ng/ul
RT: 13.557 min Scan# 1799
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:162 Resp: 648711
Ion Ratio Lower Upper
162 100
164 33.2 26.0 39.0
127 41.5 31.3 46.9



#45
2-Nitroaniline
Concen: 21.431 ng/ul
RT: 13.769 min Scan# 1799
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion: 65 Resp: 214365
Ion Ratio Lower Upper
65 100
92 69.0 56.3 84.5
138 89.9 81.4 122.2









#47
Dimethylphthalate
Concen: 18.956 ng/ul
RT: 14.134 min Scan# 1861
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:163 Resp: 880735
Ion Ratio Lower Upper
163 100
194 4.2 3.6 5.4
164 9.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 19.456 ng/ul
RT: 14.263 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:165 Resp: 178720
Ion Ratio Lower Upper
165 100
89 58.5 51.7 77.5
121 21.1 17.5 26.3









#50
Acenaphthylene
Concen: 19.558 ng/ul
RT: 14.404 min Scan# 1907
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:152 Resp: 1020655
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 20.804 ng/ul
RT: 14.593 min Scan# 1939
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:138 Resp: 169208
Ion Ratio Lower Upper
138 100
108 13.2 10.5 15.7
92 131.0 98.5 147.7









#52
Acenaphthene
Concen: 19.527 ng/ul
RT: 14.746 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:153 Resp: 712395
Ion Ratio Lower Upper
153 100
152 47.5 37.7 56.5
154 88.6 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 16.844 ng/ul
RT: 14.804 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:184 Resp: 115546
Ion Ratio Lower Upper
184 100
63 72.3 51.9 77.9
154 70.2 58.8 88.2









#55
4-Nitrophenol
Concen: 17.296 ng/ul
RT: 14.904 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:109 Resp: 154542
Ion Ratio Lower Upper
109 100
139 107.2 74.5 111.7
65 125.0 86.6 130.0



#56
Dibenzofuran
Concen: 19.129 ng/ul
RT: 15.081 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:168 Resp: 1015272
Ion Ratio Lower Upper
168 100
139 39.6 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 19.177 ng/ul
RT: 15.051 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:165 Resp: 263191
Ion Ratio Lower Upper
165 100
63 46.0 34.6 52.0
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.971 ng/ul
RT: 15.310 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:232 Resp: 237080
Ion Ratio Lower Upper
232 100
131 41.9 34.1 51.1
130 2.4 2.1 3.1
166 29.1 23.0 34.6









#59
Diethylphthalate
Concen: 19.022 ng/ul
RT: 15.493 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:149 Resp: 896447
Ion Ratio Lower Upper
149 100
177 22.0 18.2 27.2
150 12.2 10.0 15.0



#61
Fluorene
Concen: 19.194 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:166 Resp: 844444
Ion Ratio Lower Upper
166 100
165 97.1 78.9 118.3
167 13.2 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 18.375 ng/ul
RT: 15.722 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:204 Resp: 446189
Ion Ratio Lower Upper
204 100
206 32.6 26.6 39.8
141 58.7 49.2 73.8



#63
4-Nitroaniline
Concen: 21.199 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:138 Resp: 177233
Ion Ratio Lower Upper
138 100
92 58.0 44.5 66.7
108 95.6 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 17.635 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:198 Resp: 165180
Ion Ratio Lower Upper
198 100
182 4.1 3.0 4.6
77 29.2 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 19.627 ng/ul
RT: 15.934 min Scan# 2167
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:169 Resp: 720037
Ion Ratio Lower Upper
169 100
168 67.3 53.7 80.5
167 37.3 28.6 43.0





#68

4-Bromophenyl-phenylether

Concen: 18.323 ng/ul

RT: 16.616 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:248 Resp: 287742

Ion Ratio Lower Upper

248 100

250 95.7 78.8 118.2

141 76.7 60.8 91.2



#69

Hexachlorobenzene

Concen: 17.925 ng/ul

RT: 16.740 min Scan# 2304

Delta R.T. 0.000 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:284 Resp: 331700


Ion Ratio Lower Upper

284 100

142 30.5 23.2 34.8

249 31.1 25.7 38.5









#70
Atrazine
Concen: 18.375 ng/ul
RT: 16.892 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:200 Resp: 290755
Ion Ratio Lower Upper
200 100
173 24.5 19.0 28.4
215 47.2 38.5 57.7



#71
Pentachlorophenol
Concen: 18.034 ng/ul
RT: 17.087 min Scan# 2363
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:266 Resp: 218134
Ion Ratio Lower Upper
266 100
264 61.4 48.4 72.6
268 63.6 50.4 75.6









#72
Phenanthrene
Concen: 19.459 ng/ul
RT: 17.481 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:178 Resp: 1378630
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 19.2 15.4 23.0



#74
Anthracene
Concen: 19.705 ng/ul
RT: 17.575 min Scan# 2446
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:178 Resp: 1415532
Ion Ratio Lower Upper
178 100
179 15.1 12.5 18.7
176 18.5 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.456 ng/uL

RT: 13.481 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:216 Resp: 367924

Ion Ratio Lower Upper

216 100

214 77.1 61.4 92.2

218 48.6 39.1 58.7



#76

Pentachlorobenzene

Concen: 18.217 ng/uL

RT: 14.998 min Scan# 2008

Delta R.T. 0.000 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:250 Resp: 377240


Ion Ratio Lower Upper

250 100

248 63.5 51.0 76.4

252 65.3 50.8 76.2









#77
 Carbazole
 Concen: 19.832 ng/ul
 RT: 17.845 min Scan# 2492
 Delta R.T. 0.000 min
 Lab File: BP014263.D
 Acq: 23 Mar 2023 18:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020789


Tgt Ion:167 Resp: 1256670
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.5 26.3
 139 13.2 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 19.903 ng/ul
 RT: 18.375 min Scan# 2582
 Delta R.T. 0.000 min
 Lab File: BP014263.D
 Acq: 23 Mar 2023 18:42

Tgt Ion:149 Resp: 1576186
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.2
 104 5.5 4.6 6.8









#80
Fluoranthene
Concen: 20.136 ng/ul
RT: 19.439 min Scan# 2763
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:202 Resp: 1701456
Ion Ratio Lower Upper
202 100
101 10.0 6.6 10.0#
100 7.6 5.0 7.4#



#82
Pyrene
Concen: 20.253 ng/ul
RT: 19.792 min Scan# 2823
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:202 Resp: 1785283
Ion Ratio Lower Upper
202 100
101 12.3 7.8 11.8#
100 9.9 6.5 9.7#





#83

Butylbenzylphthalate

Concen: 20.944 ng/ul

RT: 20.645 min Scan# 2968

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:149 Resp: 744384

Ion Ratio Lower Upper

149 100

91 77.7 57.9 86.9

206 22.1 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.282 ng/ul

RT: 21.427 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:252 Resp: 629084


Ion Ratio Lower Upper

252 100

254 63.7 51.3 76.9

126 10.7 6.9 10.3#









#85
Benzo(a)anthracene
Concen: 19.367 ng/ul
RT: 21.504 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:228 Resp: 1830957
Ion Ratio Lower Upper
228 100
229 19.5 15.5 23.3
226 27.0 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.673 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:149 Resp: 1098548
Ion Ratio Lower Upper
149 100
167 27.2 22.3 33.5
279 4.6 4.3 6.5





#87

Chrysene

Concen: 19.156 ng/ul

RT: 21.557 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:228 Resp: 1726019

Ion Ratio Lower Upper

228 100

226 29.4 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 21.219 ng/ul




RT: 22.363 min Scan# 3260


Delta R.T. 0.000 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Tgt Ion:149 Resp: 1884531









#90
Benzo(b)fluoranthene
Concen: 18.653 ng/ul
RT: 23.316 min Scan# 3414
Delta R.T. 0.047 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42




Instrument :
BNA_P
ClientSampleId :
SSTD020789


Tgt Ion:252 Resp: 1749529
Ion Ratio Lower Upper
252 100
253 22.3 17.3 25.9
125 7.8 5.4 8.2



#91
Benzo(k)fluoranthene
Concen: 19.602 ng/ul
RT: 23.316 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP014263.D
Acq: 23 Mar 2023 18:42

Tgt Ion:252 Resp: 1753133
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.0
125 7.8 5.5 8.3





#93

Benzo(a)pyrene

Concen: 19.065 ng/ul

RT: 23.927 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:252 Resp: 1555273

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 9.6 6.2 9.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.813 ng/ul

RT: 26.692 min Scan# 3996

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42




Tgt Ion:276 Resp: 1815714


Ion Ratio Lower Upper

276 100

138 19.8 13.1 19.7#

277 25.8 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 16.980 ng/ul

RT: 26.704 min Scan# 3999

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Instrument :

BNA_P

ClientSampleId :

SSTD020789





Tgt Ion:278 Resp: 1524122

Ion Ratio Lower Upper

278 100

139 12.9 8.7 13.1

279 24.1 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 16.414 ng/ul

RT: 27.504 min Scan# 4134

Delta R.T. -0.006 min

Lab File: BP014263.D

Acq: 23 Mar 2023 18:42

Tgt Ion:276 Resp: 1416863

Ion Ratio Lower Upper

276 100

138 17.9 11.9 17.9#

277 24.7 18.7 28.1

