

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014266.D
 Acq On : 23 Mar 2023 20:25
 Operator : CG/JU
 Sample : AR1242-50PPM
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 AR1242-50PPM

Quant Time: Mar 24 00:53:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	164189	20.000	ng/ul	0.00
20) Naphthalene-d8	10.858	136	648677	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.681	164	458039	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.440	188	1151688	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.516	240	1004112	20.000	ng/ul	# 0.00
88) Perylene-d12	24.039	264	933877	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.399	96	60	0.014	ng/uL	0.00
4) Pyridine-d5	3.828	84	216	0.018	ng/ul	-0.01
7) Phenol-d5	7.199	99	13	0.001	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	79	0.009	ng/ul	-0.02
11) 2-Chlorophenol-d4	7.569	132	71	0.006	ng/ul	0.00
15) 4-Methylphenol-d8	8.746	113	12	0.001	ng/ul	0.00
21) Nitrobenzene-d5	9.199	128	18	0.003	ng/ul	-0.02
24) 2-Nitrophenol-d4	9.940	143	25	0.004	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.463	165	8	0.001	ng/ul	-0.02
31) 4-Chloroaniline-d4	10.993	131	8	0.001	ng/ul	-0.01
46) Dimethylphthalate-d6	14.093	166	8	0.000	ng/ul	0.00
49) Acenaphthylene-d8	14.393	160	82	0.002	ng/ul	0.01
54) 4-Nitrophenol-d4	14.893	143	15	0.002	ng/ul	0.00
60) Fluorene-d10	15.687	176	595	0.018	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.810	200	16	0.002	ng/ul	0.00
73) Anthracene-d10	17.534	188	5604	0.100	ng/ul	0.00
81) Pyrene-d10	19.775	212	458	0.008	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.039	264	799598	16.097	ng/ul	0.16

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
Data File : BP014266.D
Acq On : 23 Mar 2023 20:25
Operator : CG/JU
Sample : AR1242-50PPM
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
AR1242-50PPM

Quant Time: Mar 24 00:53:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 23 04:56:06 2023
Response via : Initial Calibration

