

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014268.D
 Acq On : 23 Mar 2023 21:34
 Operator : CG/JU
 Sample : PB151602BL
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK602

Quant Time: Mar 24 00:54:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	136840	20.000	ng/ul	0.00
20) Naphthalene-d8	10.857	136	516886	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.681	164	378682	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.439	188	983781	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.516	240	990423	20.000	ng/ul	# 0.00
88) Perylene-d12	24.039	264	1033073	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	148	0.041	ng/uL	0.01
4) Pyridine-d5	3.840	84	45	0.004	ng/ul	0.00
7) Phenol-d5	7.187	99	17	0.001	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.352	67	43	0.006	ng/ul	0.00
11) 2-Chlorophenol-d4	7.563	132	6	0.001	ng/ul	0.00
15) 4-Methylphenol-d8	8.752	113	22	0.002	ng/ul	0.00
21) Nitrobenzene-d5	9.210	128	14	0.003	ng/ul	0.00
24) 2-Nitrophenol-d4	9.940	143	4	0.001	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.469	165	18	0.002	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.993	131	34	0.003	ng/ul	-0.01
46) Dimethylphthalate-d6	14.087	166	6	0.000	ng/ul	0.00
49) Acenaphthylene-d8	14.381	160	4	0.000	ng/ul	0.00
54) 4-Nitrophenol-d4	14.875	143	23	0.004	ng/ul	-0.02
60) Fluorene-d10	15.675	176	12	0.000	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	10	0.001	ng/ul	-0.01
73) Anthracene-d10	17.439	188	983781	20.599	ng/ul	-0.10
81) Pyrene-d10	19.751	212	20	0.000	ng/ul	-0.02
92) Benzo(a)pyrene-d12	24.039	264	898622	16.353	ng/ul	0.16

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014268.D
 Acq On : 23 Mar 2023 21:34
 Operator : CG/JU
 Sample : PB151602BL
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK602

Quant Time: Mar 24 00:54:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

