

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014269.D
 Acq On : 23 Mar 2023 22:08
 Operator : CG/JU
 Sample : 01960-15
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EYFG8

Quant Time: Mar 24 00:54:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 04:56:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.034	152	155937	20.000	ng/ul	0.00
20) Naphthalene-d8	10.857	136	609502	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.681	164	440704	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.439	188	1113042	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.516	240	1003687	20.000	ng/ul	# 0.00
88) Perylene-d12	24.039	264	976953	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	114	0.028	ng/uL	0.01
4) Pyridine-d5	3.852	84	115	0.010	ng/ul	0.01
7) Phenol-d5	7.193	99	10	0.001	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.352	67	31	0.004	ng/ul	0.00
11) 2-Chlorophenol-d4	7.563	132	18	0.002	ng/ul	0.00
15) 4-Methylphenol-d8	8.734	113	12	0.001	ng/ul	-0.02
21) Nitrobenzene-d5	9.199	128	25	0.005	ng/ul	-0.02
24) 2-Nitrophenol-d4	9.946	143	14	0.002	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	16	0.001	ng/ul	-0.19
31) 4-Chloroaniline-d4	10.993	131	16	0.001	ng/ul	-0.01
46) Dimethylphthalate-d6	14.098	166	31	0.001	ng/ul	0.01
49) Acenaphthylene-d8	14.369	160	11	0.000	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.892	143	4	0.001	ng/ul	0.00
60) Fluorene-d10	15.651	176	9	0.000	ng/ul	-0.02
65) 4,6-Dinitro-2-methylph...	15.804	200	4	0.000	ng/ul	0.00
73) Anthracene-d10	17.604	188	13747	0.254	ng/ul	0.06
81) Pyrene-d10	19.763	212	146	0.003	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.039	264	850375	16.364	ng/ul	0.16

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
Data File : BP014269.D
Acq On : 23 Mar 2023 22:08
Operator : CG/JU
Sample : 01960-15
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EYFG8

Quant Time: Mar 24 00:54:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 23 04:56:06 2023
Response via : Initial Calibration

