

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
 Data File : BP014234.D
 Acq On : 23 Mar 2023 01:02
 Operator : CG/JU
 Sample : 01977-11
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EYG97

Quant Time: Mar 23 03:00:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 23 02:31:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	136126	20.000	ng/u1	0.00
20) Naphthalene-d8	10.863	136	613553	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.687	164	452971	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1090254	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.522	240	1124134	20.000	ng/u1	#-0.01
88) Perylene-d12	24.045	264	1189745	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	18928	5.254	ng/uL	0.00
4) Pyridine-d5	3.840	84	262822	26.347	ng/u1	0.00
7) Phenol-d5	7.199	99	380352	31.662	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.364	67	247456	35.321	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	279673	30.558	ng/u1	0.00
15) 4-Methylphenol-d8	8.752	113	318540	32.463	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	150169	30.675	ng/u1	0.00
24) 2-Nitrophenol-d4	9.940	143	182859	31.204	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.481	165	306802	27.824	ng/u1	0.00
31) 4-Chloroaniline-d4	11.005	131	420466	28.815	ng/u1	0.00
46) Dimethylphthalate-d6	14.087	166	1209796	31.470	ng/u1	0.00
49) Acenaphthylene-d8	14.381	160	1280369	31.441	ng/u1	0.00
54) 4-Nitrophenol-d4	14.893	143	159206	23.501	ng/u1	0.00
60) Fluorene-d10	15.681	176	1031018	31.646	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	190035	23.062	ng/u1	0.00
73) Anthracene-d10	17.539	188	1690586	31.942	ng/u1	0.00
81) Pyrene-d10	19.769	212	2073488	33.444	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	2048484	32.369	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032223\
Data File : BP014234.D
Acq On : 23 Mar 2023 01:02
Operator : CG/JU
Sample : 01977-11
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EYG97

Quant Time: Mar 23 03:00:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 23 02:31:48 2023
Response via : Initial Calibration

