

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032422\
 Data File : BP009531.D
 Acq On : 24 Mar 2022 16:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 06:34:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 06:34:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	137	0.00
2	1,4-Dioxane	40.000	41.599	-4.0	150	0.00
3	Pyridine	40.000	42.091	-5.2	148	0.00
4	n-Nitrosodimethylamine	40.000	41.559	-3.9	148	0.00
5 S	2-Fluorophenol	80.000	80.792	-1.0	143	0.00
6	Aniline	40.000	40.579	-1.4	144	0.00
7 S	Phenol-d6	80.000	80.215	-0.3	143	0.00
8	2-Chlorophenol	40.000	39.508	1.2	141	0.00
9	Benzaldehyde	40.000	41.002	-2.5	149	0.00
10 C	Phenol	40.000	40.273	-0.7	143	0.00
11	bis(2-Chloroethyl)ether	40.000	41.154	-2.9	147	0.00
12	1,3-Dichlorobenzene	40.000	39.306	1.7	140	0.00
13 C	1,4-Dichlorobenzene	40.000	39.192	2.0	140	0.00
14	1,2-Dichlorobenzene	40.000	39.183	2.0	140	0.00
15	Benzyl Alcohol	40.000	32.557	18.6	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.809	-12.0	160	0.00
17	2-Methylphenol	40.000	39.582	1.0	141	0.00
18	Hexachloroethane	40.000	39.683	0.8	142	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.529	-1.3	145	0.00
20	3+4-Methylphenols	40.000	39.904	0.2	142	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	138	0.00
22	Acetophenone	40.000	39.896	0.3	142	0.00
23 S	Nitrobenzene-d5	80.000	79.851	0.2	141	0.00
24	Nitrobenzene	40.000	39.942	0.1	142	0.00
25	Isophorone	40.000	41.010	-2.5	147	0.00
26 C	2-Nitrophenol	40.000	39.995	0.0	140	0.00
27	2,4-Dimethylphenol	40.000	40.814	-2.0	144	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.227	-5.6	150	0.00
29 C	2,4-Dichlorophenol	40.000	40.592	-1.5	142	0.00
30	1,2,4-Trichlorobenzene	40.000	39.446	1.4	139	0.00
31	Naphthalene	40.000	39.805	0.5	142	0.00
32	Benzoic acid	40.000	41.351	-3.4	142	0.00
33	4-Chloroaniline	40.000	39.938	0.2	141	0.00
34 C	Hexachlorobutadiene	40.000	38.616	3.5	137	0.00
35	Caprolactam	40.000	36.625	8.4	135	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.965	5.1	137	0.00
37	2-Methylnaphthalene	40.000	39.459	1.4	142	0.00
38	1-Methylnaphthalene	40.000	39.540	1.2	142	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	137	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.693	0.8	138	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.321	1.7	144	0.00
42 S	2,4,6-Tribromophenol	80.000	66.905	16.4	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.058	-0.1	138	0.00
44	2,4,5-Trichlorophenol	40.000	39.161	2.1	136	0.00
45 S	2-Fluorobiphenyl	80.000	78.738	1.6	137	0.00
46	1,1'-Biphenyl	40.000	40.049	-0.1	140	0.00
47	2-Chloronaphthalene	40.000	39.664	0.8	139	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032422\
 Data File : BP009531.D
 Acq On : 24 Mar 2022 16:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 06:34:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 06:34:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.871	2.8	134	0.00
49	Acenaphthylene	40.000	39.261	1.8	138	0.00
50	Dimethylphthalate	40.000	37.990	5.0	134	0.00
51	2,6-Dinitrotoluene	40.000	37.591	6.0	130	0.00
52 C	Acenaphthene	40.000	38.872	2.8	137	0.00
53	3-Nitroaniline	40.000	37.540	6.2	131	0.00
54 P	2,4-Dinitrophenol	40.000	35.151	12.1	126	0.00
55	Dibenzofuran	40.000	38.315	4.2	136	0.00
56 P	4-Nitrophenol	40.000	36.578	8.6	124	0.00
57	2,4-Dinitrotoluene	40.000	36.614	8.5	127	0.00
58	Fluorene	40.000	37.314	6.7	133	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.722	8.2	129	0.00
60	Diethylphthalate	40.000	37.337	6.7	132	0.00
61	4-Chlorophenyl-phenylether	40.000	37.529	6.2	133	0.00
62	4-Nitroaniline	40.000	35.257	11.9	122	0.00
63	Azobenzene	40.000	39.760	0.6	140	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	125	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.676	-1.7	122	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.271	-3.2	132	0.00
67	4-Bromophenyl-phenylether	40.000	40.050	-0.1	130	0.00
68	Hexachlorobenzene	40.000	38.302	4.2	123	0.00
69	Atrazine	40.000	40.514	-1.3	125	0.00
70 C	Pentachlorophenol	40.000	40.384	-1.0	122	0.00
71	Phenanthrene	40.000	39.458	1.4	128	0.00
72	Anthracene	40.000	39.749	0.6	128	0.00
73	Carbazole	40.000	38.412	4.0	124	0.00
74	Di-n-butylphthalate	40.000	40.781	-2.0	130	0.00
75 C	Fluoranthene	40.000	37.213	7.0	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	119	0.00
77	Benzydine	40.000	26.075	34.8#	82	0.00
78	Pyrene	40.000	40.173	-0.4	120	0.00
79 S	Terphenyl-d14	80.000	78.927	1.3	117	0.00
80	Butylbenzylphthalate	40.000	42.001	-5.0	126	0.00
81	Benzo(a)anthracene	40.000	39.715	0.7	121	0.00
82	3,3'-Dichlorobenzidine	40.000	38.354	4.1	117	0.00
83	Chrysene	40.000	39.504	1.2	120	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.150	-10.4	133	0.00
85 c	Di-n-octyl phthalate	40.000	44.435	-11.1	136	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.627	3.4	115	0.00
88	Benzo(b)fluoranthene	40.000	39.202	2.0	118	0.00
89	Benzo(k)fluoranthene	40.000	41.433	-3.6	123	0.00
90 C	Benzo(a)pyrene	40.000	40.022	-0.1	120	0.00
91	Dibenzo(a,h)anthracene	40.000	38.656	3.4	115	0.00
92	Benzo(g,h,i)perylene	40.000	38.461	3.8	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032422\
Data File : BP009531.D
Acq On : 24 Mar 2022 16:08
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 25 06:34:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 06:34:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0