

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032521\
 Data File : BP005041.D
 Acq On : 25 Mar 2021 10:02
 Operator : CG/JU
 Sample : PB135064BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Mar 25 10:36:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 10:21:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	27163	20.00	ng	0.00
21) Naphthalene-d8	10.510	136	102631	20.00	ng	# 0.00
39) Acenaphthene-d10	14.375	164	66678	20.00	ng	0.00
64) Phenanthrene-d10	17.134	188	145249	20.00	ng	0.00
76) Chrysene-d12	21.222	240	151781	20.00	ng	0.00
86) Perylene-d12	23.510	264	153240	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.317	112	241828	137.01	ng	0.00
7) Phenol-d6	6.899	99	307513	121.63	ng	0.00
23) Nitrobenzene-d5	8.875	82	201835	84.38	ng	0.00
42) 2,4,6-Tribromophenol	15.869	330	110151	126.79	ng	0.00
45) 2-Fluorobiphenyl	12.987	172	410739	83.33	ng	0.00
79) Terphenyl-d14	19.704	244	712780	85.49	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	175	0.23	ng	# 64
3) Pyridine	3.664	79	6	0.00	ng	# 1
4) n-Nitrosodimethylamine	3.552	42	156	0.23	ng	# 1
6) Aniline	7.064	93	58	0.02	ng	# 84
8) 2-Chlorophenol	7.193	128	22	0.01	ng	# 81
9) Benzaldehyde	6.899	77	656	0.51	ng	# 1
10) Phenol	6.917	94	95	0.04	ng	# 1
11) bis(2-Chloroethyl)ether	7.193	93	21	0.01	ng	# 83
12) 1,3-Dichlorobenzene	7.616	146	16	0.01	ng	# 56
13) 1,4-Dichlorobenzene	7.811	146	12	0.01	ng	# 91
14) 1,2-Dichlorobenzene	7.811	146	12	0.01	ng	# 93
15) Benzyl Alcohol	7.946	79	33	0.02	ng	# 1
16) 2,2'-oxybis(1-Chloropr...	8.299	45	25	0.02	ng	# 79
17) 2-Methylphenol	8.163	107	13	0.01	ng	# 19
18) Hexachloroethane	8.793	117	12	0.01	ng	# 38
19) n-Nitroso-di-n-propyla...	8.540	70	23	0.01	ng	# 52
20) 3+4-Methylphenols	8.528	107	43	0.02	ng	# 2
22) Acetophenone	8.522	105	38	0.01	ng	# 14
24) Nitrobenzene	8.881	77	969	0.38	ng	# 21
25) Isophorone	9.305	82	56	0.01	ng	# 92
26) 2-Nitrophenol	9.628	139	8	0.01	ng	# 92
27) 2,4-Dimethylphenol	9.693	122	5	0.00	ng	# 60
28) bis(2-Chloroethoxy)met...	9.916	93	11	0.00	ng	# 76
29) 2,4-Dichlorophenol	10.152	162	5	0.00	ng	# 26
30) 1,2,4-Trichlorobenzene	10.369	180	9	0.00	ng	# 69
31) Naphthalene	10.557	128	42	0.01	ng	# 40
32) Benzoic acid	9.852	122	6	0.01	ng	# 1
33) 4-Chloroaniline	10.669	127	5	0.00	ng	# 1
34) Hexachlorobutadiene	10.834	225	13	0.01	ng	# 44
35) Caprolactam	11.510	113	11	0.02	ng	# 95
36) 4-Chloro-3-methylphenol	11.822	107	9	0.00	ng	# 58
37) 2-Methylnaphthalene	12.181	142	35	0.01	ng	# 85
38) 1-Methylnaphthalene	12.399	142	29	0.01	ng	# 44
40) 1,2,4,5-Tetrachloroben...	12.546	216	6	0.00	ng	# 1
41) Hexachlorocyclopentadiene	12.516	237	5	0.00	ng	# 79

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032521\
 Data File : BP005041.D
 Acq On : 25 Mar 2021 10:02
 Operator : CG/JU
 Sample : PB135064BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Mar 25 10:36:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 10:21:09 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.793	196	6	0.00	ng	# 9
44) 2,4,5-Trichlorophenol	12.863	196	6	0.00	ng	# 60
46) 1,1'-Biphenyl	13.199	154	182	0.04	ng	# 51
47) 2-Chloronaphthalene	13.334	162	20	0.00	ng	95
48) 2-Nitroaniline	13.457	65	13	0.01	ng	# 40
49) Acenaphthylene	14.098	152	27	0.00	ng	# 68
50) Dimethylphthalate	13.851	163	46	0.01	ng	# 55
51) 2,6-Dinitrotoluene	13.704	165	26	0.02	ng	93
52) Acenaphthene	14.428	154	18	0.00	ng	# 52
53) 3-Nitroaniline	14.275	138	16	0.01	ng	# 77
54) 2,4-Dinitrophenol	14.493	184	6	0.01	ng	# 61
55) Dibenzofuran	14.781	168	32	0.00	ng	# 38
56) 4-Nitrophenol	14.728	139	8	0.01	ng	81
57) 2,4-Dinitrotoluene	14.740	165	15	0.01	ng	# 50
58) Fluorene	15.445	166	25	0.00	ng	# 34
59) 2,3,4,6-Tetrachlorophenol	15.010	232	14	0.01	ng	# 36
60) Diethylphthalate	15.210	149	89	0.02	ng	# 73
61) 4-Chlorophenyl-phenyle...	15.410	204	13	0.00	ng	# 76
62) 4-Nitroaniline	15.475	138	5	0.00	ng	88
63) Azobenzene	15.604	77	74	0.01	ng	85
65) 4,6-Dinitro-2-methylph...	15.522	198	5	0.00	ng	# 1
66) n-Nitrosodiphenylamine	15.869	169	3705	0.79	ng	# 54
67) 4-Bromophenyl-phenylether	16.287	248	7	0.00	ng	88
68) Hexachlorobenzene	16.428	284	7	0.00	ng	# 45
69) Atrazine	16.657	200	14	0.01	ng	98
70) Pentachlorophenol	16.775	266	4	0.00	ng	# 1
71) Phenanthrene	17.175	178	23	0.00	ng	# 4
72) Anthracene	17.287	178	41	0.00	ng	# 1
73) Carbazole	17.557	167	20	0.00	ng	# 1
74) Di-n-butylphthalate	18.104	149	129	0.02	ng	# 88
75) Fluoranthene	19.157	202	103	0.01	ng	80
77) Benzidine	19.310	184	27	0.01	ng	# 39
78) Pyrene	19.516	202	207	0.02	ng	# 86
80) Butylbenzylphthalate	20.363	149	244	0.06	ng	# 43
81) Benzo(a)anthracene	21.216	228	655	0.06	ng	# 71
82) 3,3'-Dichlorobenzidine	21.139	252	125	0.04	ng	# 38
83) Chrysene	21.263	228	495	0.05	ng	# 58
84) Bis(2-ethylhexyl)phtha...	21.145	149	206	0.04	ng	# 56
85) Di-n-octyl phthalate	22.016	149	253	0.03	ng	# 11
87) Indeno(1,2,3-cd)pyrene	25.827	276	19	0.00	ng	# 1
88) Benzo(b)fluoranthene	22.816	252	387	0.04	ng	# 40
89) Benzo(k)fluoranthene	22.851	252	149	0.01	ng	# 39
90) Benzo(a)pyrene	23.404	252	171	0.02	ng	# 1
91) Dibenzo(a,h)anthracene	25.904	278	283	0.03	ng	# 81
92) Benzo(g,h,i)perylene	26.586	276	264	0.03	ng	# 58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_P
ClientSampleId :

Quant Time: Mar 25 10:36:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 25 10:21:09 2021
Response via : Initial Calibration









Scan 788 (7.722 min): BP005024.D\data.ms (-779) #1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.722 min Scan# 788
 Delta R.T. 0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	152	Resp:	27163
Ion	Ratio	Lower	Upper
152	100		
150	149.1	123.8	185.6
115	64.5	43.8	65.6



Scan 28 (3.252 min): BP005024.D\data.ms (-23) #2
 1,4-Dioxane
 Concen: 0.23 ng
 RT: 3.281 min Scan# 33
 Delta R.T. 0.030 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	88	Resp:	175
Ion	Ratio	Lower	Upper
88	100		
58	65.7	44.2	66.4
43	62.9	15.2	22.8#









Pyridine
 Concen: 0.00 ng
 RT: 3.664 min Scan# 98
 Delta R.T. 0.012 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	79	Resp:	6
Ion	Ratio	Lower	Upper
79	100		
52	61.5	41.3	61.9
51	646.2	21.2	31.8#



n-Nitrosodimethylamine
 Concen: 0.23 ng
 RT: 3.552 min Scan# 79
 Delta R.T. -0.012 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	42	Resp:	156
Ion	Ratio	Lower	Upper
42	100		
74	32.3	148.6	223.0#
44	212.5	5.3	7.9#









2-Fluorophenol
 Concen: 137.01 ng
 RT: 5.317 min Scan# 379
 Delta R.T. 0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	112	Resp:	241828
Ion	Ratio	Lower	Upper
112	100		
64	56.0	32.8	49.2#
63	29.5	17.0	25.4#



Aniline
 Concen: 0.02 ng
 RT: 7.064 min Scan# 676
 Delta R.T. 0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	93	Resp:	58
Ion	Ratio	Lower	Upper
93	100		
66	24.6	25.4	38.0#
65	24.6	12.5	18.7#









Phenol-d6
 Concen: 121.63 ng
 RT: 6.899 min Scan# 648
 Delta R.T. 0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	99	Resp:	307513
Ion	Ratio	Lower	Upper
99	100		
42	14.4	8.6	13.0#
71	38.8	23.4	35.2#



2-Chlorophenol
 Concen: 0.01 ng
 RT: 7.193 min Scan# 698
 Delta R.T. -0.094 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	128	Resp:	22
Ion	Ratio	Lower	Upper
128	100		
130	35.0	13.0	53.0
64	50.0	10.4	50.4









Benzaldehyde
 Concen: 0.51 ng
 RT: 6.899 min Scan# 648
 Delta R.T. 0.030 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	77	Resp:	656
Ion Ratio	Lower	Upper	
77	100		
105	2.6	97.1	137.1#
106	2.6	93.7	133.7#



Phenol
 Concen: 0.04 ng
 RT: 6.917 min Scan# 651
 Delta R.T. -0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	94	Resp:	95
Ion Ratio	Lower	Upper	
94	100		
65	413.2	2.3	42.3#
66	1431.6	10.9	50.9#









bis(2-Chloroethyl)ether
Concen: 0.01 ng
RT: 7.193 min Scan# 698
Delta R.T. 0.041 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Instrument :
BNA_P
ClientSampled :


Tgt Ion:	93	Resp:	21
Ion	Ratio	Lower	Upper
93	100		
63	42.1	39.0	79.0
95	36.8	12.8	52.8



1,3-Dichlorobenzene
Concen: 0.01 ng
RT: 7.616 min Scan# 770
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	146	Resp:	16
Ion	Ratio	Lower	Upper
146	100		
148	81.5	52.5	78.7#
75	77.8	18.1	27.1#









1,4-Dichlorobenzene
Concen: 0.01 ng
RT: 7.811 min Scan# 803
Delta R.T. 0.059 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Instrument :
BNA_P
ClientSampled :

Tgt Ion:	146	Resp:	12
Ion	Ratio	Lower	Upper
146	100		
148	55.6	51.3	76.9
111	38.9	28.9	43.3




1,2-Dichlorobenzene
Concen: 0.01 ng
RT: 7.811 min Scan# 803
Delta R.T. -0.259 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	146	Resp:	12
Ion	Ratio	Lower	Upper
146	100		
148	55.6	51.6	77.4
111	38.9	31.0	46.6





Abundance Scan 830 (7.969 min): BP005024.D\data.ms (-821) #15




Benzyl Alcohol
Concen: 0.02 ng
RT: 7.946 min Scan# 826
Delta R.T. -0.018 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	79	Resp:	33
Ion	Ratio	Lower	Upper
79	100		
108	0.0	73.5	110.3#
77	160.0	50.3	75.5#



Abundance Scan 826 (7.946 min): BP005041.D\data.ms (-778) (-)





Abundance Scan 878 (8.252 min): BP005024.D\data.ms (-869) #16





2,2'-oxybis(1-Chloropropane)
Concen: 0.02 ng
RT: 8.299 min Scan# 886
Delta R.T. 0.053 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	45	Resp:	25
Ion	Ratio	Lower	Upper
45	100		
77	27.3	2.3	42.3
79	31.8	0.0	36.6



Abundance Scan 886 (8.299 min): BP005041.D\data.ms (-826) (-)









2-Methylphenol
 Concen: 0.01 ng
 RT: 8.163 min Scan# 863
 Delta R.T. 0.000 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	107	Resp:	13
Ion	Ratio	Lower	Upper
107	100		
108	53.8	89.6	134.4#
77	84.6	31.8	47.6#
79	138.5	31.1	46.7#



Hexachloroethane
 Concen: 0.01 ng
 RT: 8.793 min Scan# 970
 Delta R.T. 0.000 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02


Tgt Ion:	117	Resp:	12
Ion	Ratio	Lower	Upper
117	100		
119	50.0	78.8	118.2#
201	35.0	92.3	138.5#








n-Nitroso-di-n-propylamine
Concen: 0.01 ng
RT: 8.540 min Scan# 927
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :





Tgt Ion	Ratio	Lower	Upper
70	100		
42	81.0	32.8	49.2#
101	42.9	9.6	14.4#
130	38.1	24.3	36.5#




3+4-Methylphenols
Concen: 0.02 ng
RT: 8.528 min Scan# 925
Delta R.T. 0.035 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion	Ratio	Lower	Upper
107	100		
108	0.0	69.0	109.0#
77	91.7	8.5	48.5#
79	66.7	4.8	44.8#









Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.510 min Scan# 1262
 Delta R.T. 0.000 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :


Tgt Ion:	136	Resp:	102631
Ion	Ratio	Lower	Upper
136	100		
137	12.4	8.6	13.0
54	7.0	3.6	5.4#
68	6.3	3.3	4.9#



Acetophenone
 Concen: 0.01 ng
 RT: 8.522 min Scan# 924
 Delta R.T. -0.023 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02

Tgt Ion:	105	Resp:	38
Ion	Ratio	Lower	Upper
105	100		
71	40.0	2.6	3.8#
51	75.0	13.5	20.3#
120	0.0	19.6	29.4#









Nitrobenzene-d5
Concen: 84.38 ng
RT: 8.875 min Scan# 984
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Instrument :
BNA_P
ClientSampled :

Tgt Ion:	82	Resp:	201835
Ion	Ratio	Lower	Upper
82	100		
128	36.9	45.6	68.4#
54	41.8	30.1	45.1




Nitrobenzene
Concen: 0.38 ng
RT: 8.881 min Scan# 985
Delta R.T. -0.041 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	77	Resp:	969
Ion	Ratio	Lower	Upper
77	100		
123	1.9	48.6	72.8#
65	46.2	10.0	15.0#





Abundance Scan 1082 (9.452 min): BP005024.D\data.ms (-1075-)




Isophorone
Concen: 0.01 ng
RT: 9.305 min Scan# 1057
Delta R.T. -0.141 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	82	Resp:	56
Ion	Ratio	Lower	Upper
82	100		
95	6.9	6.4	9.6
138	18.8	18.7	28.1



Abundance Scan 1057 (9.305 min): BP005041.D\data.ms (-1030-)





Abundance Scan 1112 (9.628 min): BP005024.D\data.ms (-1109-)




2-Nitrophenol
Concen: 0.01 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	139	Resp:	8
Ion	Ratio	Lower	Upper
139	100		
109	34.8	22.1	33.1#
65	34.8	26.5	39.7



Abundance Scan 1112 (9.628 min): BP005041.D\data.ms (-1061-)





Abundance Scan 1123 (9.693 min): BP005024.D\data.ms (-1127-)




2,4-Dimethylphenol
Concen: 0.00 ng
RT: 9.693 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	122	Resp:	5
Ion	Ratio	Lower	Upper
122	100		
107	128.6	87.3	130.9
121	0.0	47.3	70.9#



Abundance Scan 1123 (9.693 min): BP005041.D\data.ms (-1072-)





Abundance Scan 1164 (9.934 min): BP005024.D\data.ms (-11528-)





bis(2-Chloroethoxy)methane
Concen: 0.00 ng
RT: 9.916 min Scan# 1161
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	93	Resp:	11
Ion	Ratio	Lower	Upper
93	100		
95	40.0	26.7	40.1
123	35.0	10.5	15.7#



Abundance Scan 1161 (9.916 min): BP005041.D\data.ms (-1112-)









2,4-Dichlorophenol
Concen: 0.00 ng
RT: 10.152 min Scan# 1201
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Instrument :
BNA_P
ClientSampled :


Tgt Ion:	162	Resp:	5
Ion	Ratio	Lower	Upper
162	100		
164	0.0	45.8	85.8#
98	0.0	10.2	50.2#



1,2,4-Trichlorobenzene
Concen: 0.00 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	180	Resp:	9
Ion	Ratio	Lower	Upper
180	100		
182	58.8	76.2	114.2#
145	35.3	23.8	35.8









#81 (-)
Naphthalene
Concen: 0.01 ng
RT: 10.557 min Scan# 1270
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Instrument :
BNA_P
ClientSampled :

Tgt Ion:	128	Resp:	42
Ion	Ratio	Lower	Upper
128	100		
129	50.0	8.8	13.2#
127	23.1	10.6	15.8#



#82 (-)
Benzoic acid
Concen: 0.01 ng
RT: 9.852 min Scan# 1150
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	122	Resp:	6
Ion	Ratio	Lower	Upper
122	100		
105	0.0	94.1	134.1#
77	387.5	50.1	90.1#





Abundance Scan 1291 (10.681 min): BP005024.D\data.ms (-1289) (-)

4-Chloroaniline
Concen: 0.00 ng
RT: 10.669 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	127	Resp:	5
Ion	Ratio	Lower	Upper
127	100		
129	112.5	25.7	38.5#
65	0.0	17.3	25.9#
92	137.5	12.9	19.3#




Abundance Scan 1289 (10.669 min): BP005041.D\data.ms (-1239) (-)





Abundance Scan 1319 (10.846 min): BP005024.D\data.ms (-1304) (-)


Hexachlorobutadiene
Concen: 0.01 ng
RT: 10.834 min Scan# 1317
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	225	Resp:	13
Ion	Ratio	Lower	Upper
225	100		
223	142.9	50.1	75.1#
227	57.1	51.4	77.2




Abundance Scan 1317 (10.834 min): BP005041.D\data.ms (-1268) (-)





Abundance Scan 1426 (11.475 min): BP005024.D\data.ms (-1435) (-)

Caprolactam
Concen: 0.02 ng
RT: 11.510 min Scan# 1432
Delta R.T. 0.041 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	113	Resp:	11
Ion	Ratio	Lower	Upper
113	100		
55	90.9	77.3	117.3
56	72.7	55.9	95.9





Abundance Scan 1432 (11.510 min): BP005041.D\data.ms (-1374) (-)




Abundance Scan 1483 (11.810 min): BP005024.D\data.ms (-1435) (-)

4-Chloro-3-methylphenol
Concen: 0.00 ng
RT: 11.822 min Scan# 1485
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	107	Resp:	9
Ion	Ratio	Lower	Upper
107	100		
144	50.0	22.7	34.1#
142	127.8	70.7	106.1#




Abundance Scan 1485 (11.822 min): BP005041.D\data.ms (-1432) (-)





Abundance Scan 1546 (12.181 min): BP005024.D\data.ms (-1537) (-)

2-Methylnaphthalene
Concen: 0.01 ng
RT: 12.181 min Scan# 1546
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	142	Resp:	35
Ion	Ratio	Lower	Upper
142	100		
141	93.8	71.4	107.2
115	56.3	25.8	38.6#




Abundance Scan 1546 (12.181 min): BP005041.D\data.ms (-1495) (-)





Abundance Scan 1583 (12.399 min): BP005024.D\data.ms (-1528) (-)


1-Methylnaphthalene
Concen: 0.01 ng
RT: 12.399 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Tgt Ion:	142	Resp:	29
Ion	Ratio	Lower	Upper
142	100		
141	41.7	74.4	111.6#
116	41.7	2.9	4.3#



Abundance Scan 1583 (12.399 min): BP005041.D\data.ms (-1532) (-)





Abundance Scan 1919 (14.375 min): BP005024.D\data.ms (-1868) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.375 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.5	79.4	119.2
160	44.4	36.1	54.1



Abundance Scan 1919 (14.375 min): BP005041.D\data.ms (-1868) (-)





Abundance Scan 1609 (12.551 min): BP005024.D\data.ms (-1600) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 0.00 ng
RT: 12.546 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion	Ratio	Lower	Upper
216	100		
214	233.3	63.0	94.6#
179	150.0	16.9	25.3#
108	116.7	12.6	19.0#




Abundance Scan 1608 (12.546 min): BP005041.D\data.ms (-1558) (-)





Abundance Scan 1605 (12.528 min): BP005024.D\data.ms (-1554) (-)

Hexachlorocyclopentadiene
Concen: 0.00 ng
RT: 12.516 min Scan# 1603
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	237	Resp:	5
Ion	Ratio	Lower	Upper
237	100		
235	77.8	43.2	83.2
272	0.0	0.0	33.3




Abundance Scan 1603 (12.516 min): BP005041.D\data.ms (-1554) (-)





Abundance Scan 2174 (15.875 min): BP005024.D\data.ms (-2162) (-)


2,4,6-Tribromophenol
Concen: 126.79 ng
RT: 15.869 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	330	Resp:	110151
Ion	Ratio	Lower	Upper
330	100		
332	94.5	77.1	115.7
141	35.5	23.4	35.2#




Abundance Scan 2173 (15.869 min): BP005041.D\data.ms (-2122) (-)





Abundance Scan 1651 (12.798 min): BP005024.D\data.ms (-1648) (-)

2,4,6-Trichlorophenol
Concen: 0.00 ng
RT: 12.793 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	196	Resp:	6
Ion	Ratio	Lower	Upper
196	100		
198	0.0	77.0	115.6#
200	70.0	25.1	37.7#






Abundance Scan 1650 (12.793 min): BP005041.D\data.ms (-1600) (-)




Abundance Scan 1663 (12.869 min): BP005024.D\data.ms (-1657) (-)

2,4,5-Trichlorophenol
Concen: 0.00 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	196	Resp:	6
Ion	Ratio	Lower	Upper
196	100		
198	88.9	77.3	115.9
97	77.8	31.0	46.4#
132	88.9	20.9	31.3#




Abundance Scan 1662 (12.863 min): BP005041.D\data.ms (-1611) (-)






Abundance Scan 1684 (12.993 min): BP005024.D\data.ms (-16745) (-)

2-Fluorobiphenyl
Concen: 83.33 ng
RT: 12.987 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :




Tgt Ion:	172	Resp:	410739
Ion Ratio	Lower	Upper	
172	100		
171	34.4	28.5	42.7
170	22.3	19.2	28.8






Abundance Scan 1720 (13.204 min): BP005024.D\data.ms (-17006) (-)

1,1'-Biphenyl
Concen: 0.04 ng
RT: 13.199 min Scan# 1719
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02




Tgt Ion:	154	Resp:	182
Ion Ratio	Lower	Upper	
154	100		
153	5.5	20.4	60.4#
76	3.5	0.0	29.3





Abundance Scan 1727 (13.245 min): BP005024.D\data.ms (-1717) (-)

2-Chloronaphthalene
Concen: 0.00 ng
RT: 13.334 min Scan# 1742
Delta R.T. 0.094 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion	Ratio	Lower	Upper
162	100		
127	38.1	29.9	44.9
164	38.1	26.1	39.1






Abundance Scan 1742 (13.334 min): BP005041.D\data.ms (-1675) (-)




Abundance Scan 1763 (13.457 min): BP005024.D\data.ms (-1711) (-)

2-Nitroaniline
Concen: 0.01 ng
RT: 13.457 min Scan# 1763
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion	Ratio	Lower	Upper
65	100		
92	175.0	70.6	106.0#
138	108.3	133.6	200.4#



Abundance Scan 1763 (13.457 min): BP005041.D\data.ms (-1711) (-)





Abundance Scan 1871 (14.092 min): BP005024.D\data.ms (-1871) (-)




Acenaphthylene
Concen: 0.00 ng
RT: 14.098 min Scan# 1872
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:152	Resp:	27
Ion Ratio	Lower	Upper
152 100		
151 28.6	15.9	23.9#
153 33.3	10.5	15.7#



Abundance Scan 1872 (14.098 min): BP005041.D\data.ms (-1820) (-)





Abundance Scan 1828 (13.840 min): BP005024.D\data.ms (-1858) (-)




Dimethylphthalate
Concen: 0.01 ng
RT: 13.851 min Scan# 1830
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:163	Resp:	46
Ion Ratio	Lower	Upper
163 100		
194 37.5	3.3	4.9#
164 18.8	8.2	12.2#




Abundance Scan 1830 (13.851 min): BP005041.D\data.ms (-1777) (-)





Abundance Scan 1848 (13.957 min): BP005024.D\data.ms (-1850) (-)

2,6-Dinitrotoluene
Concen: 0.02 ng
RT: 13.704 min Scan# 1805
Delta R.T. -0.253 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	165	Resp:	26
Ion	Ratio	Lower	Upper
165	100		
63	38.1	27.8	41.6
89	38.1	34.6	52.0




Abundance Scan 1805 (13.704 min): BP005041.D\data.ms (-1797) (-)





Abundance Scan 1930 (14.439 min): BP005024.D\data.ms (-1928) (-)


Acenaphthene
Concen: 0.00 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	154	Resp:	18
Ion	Ratio	Lower	Upper
154	100		
153	66.7	91.0	136.6#
152	95.2	43.0	64.6#




Abundance Scan 1928 (14.428 min): BP005041.D\data.ms (-1879) (-)





Abundance Scan 1904 (14.287 min): BP005024.D\data.ms (-1853) (-)

3-Nitroaniline
Concen: 0.01 ng
RT: 14.275 min Scan# 1902
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	138	Resp:	16
Ion	Ratio	Lower	Upper
138	100		
108	0.0	7.5	11.3#
92	71.4	75.0	112.6#




Abundance Scan 1902 (14.275 min): BP005041.D\data.ms (-1853) (-)





Abundance Scan 1939 (14.492 min): BP005024.D\data.ms (-1851) (-)


2,4-Dinitrophenol
Concen: 0.01 ng
RT: 14.493 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	184	Resp:	6
Ion	Ratio	Lower	Upper
184	100		
63	0.0	32.3	48.5#
154	77.8	49.1	73.7#




Abundance Scan 1939 (14.493 min): BP005041.D\data.ms (-1888) (-)





Abundance Scan 1987 (14.775 min): BP005024.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 0.00 ng
RT: 14.781 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	168	Resp:	32
Ion	Ratio	Lower	Upper
168	100		
139	0.0	29.1	43.7#
169	38.1	10.4	15.6#





Abundance Scan 1988 (14.781 min): BP005041.D\data.ms (-1936) (-)




Abundance Scan 1957 (14.598 min): BP005024.D\data.ms (-1956) (-)

4-Nitrophenol
Concen: 0.01 ng
RT: 14.728 min Scan# 1979
Delta R.T. 0.130 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	139	Resp:	8
Ion	Ratio	Lower	Upper
139	100		
109	77.8	44.6	84.6
65	77.8	41.6	81.6




Abundance Scan 1979 (14.728 min): BP005041.D\data.ms (-1906) (-)





Abundance Scan 1982 (14.745 min): BP005024.D\data.ms (-1952) (-)

2,4-Dinitrotoluene
Concen: 0.01 ng
RT: 14.740 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	165	Resp:	15
Ion	Ratio	Lower	Upper
165	100		
63	52.0	20.6	31.0#
89	20.0	42.6	63.8#
182	36.0	2.4	3.6#




Abundance Scan 1981 (14.740 min): BP005041.D\data.ms (-1931) (-)





Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2058) (-)


Fluorene
Concen: 0.00 ng
RT: 15.445 min Scan# 2101
Delta R.T. 0.018 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	166	Resp:	25
Ion	Ratio	Lower	Upper
166	100		
165	38.9	77.8	116.8#
167	61.1	11.0	16.4#




Abundance Scan 2101 (15.445 min): BP005041.D\data.ms (-2047) (-)





Abundance Scan 2026 (15.004 min): BP005024.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 0.01 ng
RT: 15.010 min Scan# 2027
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	232	Resp:	14
Ion	Ratio	Lower	Upper
232	100		
131	0.0	34.1	51.1#
130	42.9	1.8	2.8#
166	0.0	24.6	36.8#





Abundance Scan 2027 (15.010 min): BP005041.D\data.ms (-1975) (-)




Abundance Scan 2061 (15.210 min): BP005024.D\data.ms (-2059) (-)

Diethylphthalate
Concen: 0.02 ng
RT: 15.210 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	149	Resp:	89
Ion	Ratio	Lower	Upper
149	100		
177	5.6	18.0	27.0#
150	16.9	10.0	15.0#




Abundance Scan 2061 (15.210 min): BP005041.D\data.ms (-2010) (-)





Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2069) (-)

4-Chlorophenyl-phenylether
Concen: 0.00 ng
RT: 15.410 min Scan# 2095
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Instrument :
BNA_P
ClientSampled :



Tgt Ion:	204	Resp:	13
Ion	Ratio	Lower	Upper
204	100		
206	0.0	25.9	38.9#
141	53.8	46.5	69.7






Abundance Scan 2095 (15.410 min): BP005041.D\data.ms (-2046) (-)




Abundance Scan 2103 (15.457 min): BP005024.D\data.ms (-2062) (-)

4-Nitroaniline
Concen: 0.00 ng
RT: 15.475 min Scan# 2106
Delta R.T. 0.018 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	138	Resp:	5
Ion	Ratio	Lower	Upper
138	100		
92	54.5	20.7	60.7
108	63.6	47.8	87.8



Abundance Scan 2106 (15.475 min): BP005041.D\data.ms (-2052) (-)





Abundance Scan 2148 (15.722 min): BP005024.D\data.ms (-2148) (-)




Azobenzene
Concen: 0.01 ng
RT: 15.604 min Scan# 2128
Delta R.T. -0.112 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion: 77	Resp: 74
Ion Ratio	Lower Upper
77 100	
182 18.4	16.9 56.9
105 26.5	5.6 45.6
51 24.5	2.9 42.9



Abundance Scan 2128 (15.604 min): BP005041.D\data.ms (-2096) (-)





Abundance Scan 2388 (17.133 min): BP005024.D\data.ms (-2388) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.134 min Scan# 2388
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion: 188	Resp: 145249
Ion Ratio	Lower Upper
188 100	
94 8.7	6.2 9.2
80 9.0	6.2 9.2



Abundance Scan 2388 (17.134 min): BP005041.D\data.ms (-2337) (-)





Abundance Scan 2112 (15.510 min): BP005024.D\data.ms (-2105) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.522 min Scan# 2114
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:198 Resp: 5
Ion Ratio Lower Upper
198 100
51 150.0 6.4 46.4#
105 225.0 18.9 58.9#



Abundance Scan 2114 (15.522 min): BP005041.D\data.ms (-2061) (-)





Abundance Scan 2135 (15.645 min): BP005024.D\data.ms (-2105) (-)




n-Nitrosodiphenylamine
Concen: 0.79 ng
RT: 15.869 min Scan# 2173
Delta R.T. 0.230 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:169 Resp: 3705
Ion Ratio Lower Upper
169 100
168 10.9 53.8 80.6#
167 34.7 28.9 43.3



Abundance Scan 2173 (15.869 min): BP005041.D\data.ms (-2083) (-)





Abundance Scan 2250 (16.322 min): BP005024.D\data.ms (-2247) (-)




4-Bromophenyl-phenylether
Concen: 0.00 ng
RT: 16.287 min Scan# 2244
Delta R.T. -0.035 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	248	Resp:	7
Ion	Ratio	Lower	Upper
248	100		
250	81.8	76.7	115.1
141	72.7	53.0	79.6



Abundance Scan 2244 (16.287 min): BP005041.D\data.ms (-2199) (-)





Abundance Scan 2269 (16.433 min): BP005024.D\data.ms (-2268) (-)




Hexachlorobenzene
Concen: 0.00 ng
RT: 16.428 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	284	Resp:	7
Ion	Ratio	Lower	Upper
284	100		
142	0.0	24.2	36.4#
249	0.0	23.8	35.8#



Abundance Scan 2268 (16.428 min): BP005041.D\data.ms (-2218) (-)





Abundance Scan 2298 (16.604 min): BP005024.D\data.ms (-2298) (-)




Atrazine
Concen: 0.01 ng
RT: 16.657 min Scan# 2307
Delta R.T. 0.053 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	200	Resp:	14
Ion Ratio	Lower	Upper	
200	100		
173	28.0	5.9	45.9
215	48.0	28.4	68.4



Abundance Scan 2307 (16.657 min): BP005041.D\data.ms (-2247) (-)





Abundance Scan 2328 (16.781 min): BP005024.D\data.ms (-2319) (-)




Pentachlorophenol
Concen: 0.00 ng
RT: 16.775 min Scan# 2327
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	266	Resp:	4
Ion Ratio	Lower	Upper	
266	100		
268	112.5	51.1	76.7#
264	200.0	50.6	76.0#



Abundance Scan 2327 (16.775 min): BP005041.D\data.ms (-2277) (-)





Abundance Scan 2395 (17.175 min): BP005024.D\data.ms (-2343) (-)




Phenanthrene
Concen: 0.00 ng
RT: 17.175 min Scan# 2395
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	178	Resp:	23
Ion	Ratio	Lower	Upper
178	100		
176	26.9	15.3	22.9#
179	96.2	12.5	18.7#



Abundance Scan 2395 (17.175 min): BP005041.D\data.ms (-2344) (-)





Abundance Scan 2411 (17.269 min): BP005024.D\data.ms (-2474) (-)




Anthracene
Concen: 0.00 ng
RT: 17.287 min Scan# 2414
Delta R.T. 0.018 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	178	Resp:	41
Ion	Ratio	Lower	Upper
178	100		
176	81.5	15.0	22.4#
179	81.5	12.5	18.7#



Abundance Scan 2414 (17.287 min): BP005041.D\data.ms (-2360) (-)





Abundance Scan 2458 (17.545 min): BP005024.D\data.ms (-2478) (-)




Carbazole
Concen: 0.00 ng
RT: 17.557 min Scan# 2460
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	167	Resp:	20
Ion	Ratio	Lower	Upper
167	100		
166	111.1	17.6	26.4#
139	38.9	10.2	15.4#



Abundance Scan 2460 (17.557 min): BP005041.D\data.ms (-2407) (-)





Abundance Scan 2552 (18.098 min): BP005024.D\data.ms (-2573) (-)




Di-n-butylphthalate
Concen: 0.02 ng
RT: 18.104 min Scan# 2553
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	149	Resp:	129
Ion	Ratio	Lower	Upper
149	100		
150	6.4	7.3	10.9#
104	11.0	3.4	5.0#



Abundance Scan 2553 (18.104 min): BP005041.D\data.ms (-2501) (-)





Abundance Scan 2731 (19.151 min): BP005024.D\data.ms (-2731) (-)




Fluoranthene
Concen: 0.01 ng
RT: 19.157 min Scan# 2732
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	202	Resp:	103
Ion Ratio	Lower	Upper	
202	100		
101	0.0	0.0	29.4
203	10.0	0.0	37.8



Abundance Scan 2732 (19.157 min): BP005041.D\data.ms (-2680) (-)





Abundance Scan 3084 (21.227 min): BP005024.D\data.ms (-3073) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.222 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	240	Resp:	151781
Ion Ratio	Lower	Upper	
240	100		
120	9.1	7.8	11.6
236	25.3	20.8	31.2



Abundance Scan 3083 (21.222 min): BP005041.D\data.ms (-3033) (-)





Abundance Scan 2763 (19.339 min): BP005024.D\data.ms (-2757) (-)




Benzidine
Concen: 0.01 ng
RT: 19.310 min Scan# 2758
Delta R.T. -0.029 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	184	Resp:	27
Ion	Ratio	Lower	Upper
184	100		
185	0.0	11.5	17.3#
183	47.4	9.6	14.4#



Abundance Scan 2758 (19.310 min): BP005041.D\data.ms (-2712) (-)





Abundance Scan 2791 (19.504 min): BP005024.D\data.ms (-2783) (-)




Pyrene
Concen: 0.02 ng
RT: 19.516 min Scan# 2793
Delta R.T. 0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	202	Resp:	207
Ion	Ratio	Lower	Upper
202	100		
200	17.2	17.1	25.7
203	8.6	14.2	21.2#



Abundance Scan 2793 (19.516 min): BP005041.D\data.ms (-2740) (-)





Abundance Scan 2825 (19.704 min): BP005024.D\data.ms (-2819) (-)




Terphenyl-d14
Concen: 85.49 ng
RT: 19.704 min Scan# 2825
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	244	Resp:	712780
Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.8	8.6
122	9.5	7.8	11.8



Abundance Scan 2825 (19.704 min): BP005041.D\data.ms (-2774) (-)





Abundance Scan 2941 (20.386 min): BP005024.D\data.ms (-2929) (-)




Butylbenzylphthalate
Concen: 0.06 ng
RT: 20.363 min Scan# 2937
Delta R.T. -0.023 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	149	Resp:	244
Ion	Ratio	Lower	Upper
149	100		
91	10.2	45.4	68.0#
206	3.1	17.1	25.7#



Abundance Scan 2937 (20.363 min): BP005041.D\data.ms (-2890) (-)





Abundance Scan 3081 (21.210 min): BP005024.D\data.ms (-30720) (-)




Benzo(a)anthracene
Concen: 0.06 ng
RT: 21.216 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	228	Resp:	655
Ion	Ratio	Lower	Upper
228	100		
226	3.7	22.1	33.1#
229	17.1	15.9	23.9



Abundance Scan 3082 (21.216 min): BP005041.D\data.ms (-3031) (-)





Abundance Scan 3071 (21.151 min): BP005024.D\data.ms (-30620) (-)





3,3'-Dichlorobenzidine
Concen: 0.04 ng
RT: 21.139 min Scan# 3069
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	252	Resp:	125
Ion	Ratio	Lower	Upper
252	100		
254	15.4	51.0	76.4#
126	31.7	7.9	11.9#



Abundance Scan 3069 (21.139 min): BP005041.D\data.ms (-3020) (-)









Chrysene
 Concen: 0.05 ng
 RT: 21.263 min Scan# 3090
 Delta R.T. 0.000 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02




Instrument :
 BNA_P
 ClientSampled :

Tgt Ion	Ratio	Lower	Upper
228	100		
226	4.6	24.4	36.6#
229	4.9	15.8	23.6#



Bis(2-ethylhexyl)phthalate
 Concen: 0.04 ng
 RT: 21.145 min Scan# 3070
 Delta R.T. 0.006 min
 Lab File: BP005041.D
 Acq: 25 Mar 2021 10:02


Tgt Ion	Ratio	Lower	Upper
149	100		
167	3.5	22.6	34.0#
279	14.2	4.2	6.2#






Abundance Scan 3220 (22.027 min): BP005024.D\data.ms (-3218) (-)

Di-n-octyl phthalate
Concen: 0.03 ng
RT: 22.016 min Scan# 3218
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :




Tgt Ion:	149	Resp:	253
Ion Ratio	Lower	Upper	
149	100		
167	5.1	1.3	1.9#
43	43.1	3.7	5.5#






Abundance Scan 3473 (23.515 min): BP005024.D\data.ms (-3462) (-)


Perylene-d12
Concen: 20.00 ng
RT: 23.510 min Scan# 3472
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02



Tgt Ion:	264	Resp:	153240
Ion Ratio	Lower	Upper	
264	100		
260	23.6	18.8	28.2
265	21.7	17.4	26.2





Abundance Scan 3872 (25.862 min): BP005024.D\data.ms (-3872) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 25.827 min Scan# 3866
Delta R.T. -0.035 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :


Tgt Ion:	276	Resp:	19
Ion Ratio	Lower	Upper	
276	100		
138	478.9	18.7	28.1#
277	31.6	20.4	30.6#



Abundance Scan 3866 (25.827 min): BP005041.D\data.ms (-3821) (-)





Abundance Scan 3355 (22.821 min): BP005024.D\data.ms (-3355) (-)




Benzo(b)fluoranthene
Concen: 0.04 ng
RT: 22.816 min Scan# 3354
Delta R.T. -0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02


Tgt Ion:	252	Resp:	387
Ion Ratio	Lower	Upper	
252	100		
253	50.4	17.6	26.4#
125	30.7	7.3	10.9#



Abundance Scan 3354 (22.816 min): BP005041.D\data.ms (-3304) (-)






Abundance Scan 3363 (22.868 min): BP005024.D\data.ms (-3359) (-)




Benzo(k)fluoranthene
Concen: 0.01 ng
RT: 22.851 min Scan# 3360
Delta R.T. -0.017 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :

Tgt Ion:	252	Resp:	149
Ion Ratio	Lower	Upper	
252	100		
253	47.8	17.8	26.8#
125	38.8	7.4	11.2#






Abundance Scan 3456 (23.415 min): BP005024.D\data.ms (-3449) (-)




Benzo(a)pyrene
Concen: 0.02 ng
RT: 23.404 min Scan# 3454
Delta R.T. -0.012 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	252	Resp:	171
Ion Ratio	Lower	Upper	
252	100		
253	129.9	17.4	26.0#
125	67.1	8.2	12.4#






Abundance Scan 3875 (25.880 min): BP005024.D\data.ms (-3875) (-)




Dibenzo(a,h)anthracene
Concen: 0.03 ng
RT: 25.904 min Scan# 3879
Delta R.T. 0.024 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Instrument :
BNA_P
ClientSampled :

Tgt Ion:	278	Resp:	283
Ion Ratio	Lower	Upper	
278	100		
139	3.0	12.4	18.6#
279	17.5	19.0	28.6#



Abundance Scan 3995 (26.586 min): BP005024.D\data.ms (-3995) (-)



Benzo(g,h,i)perylene
Concen: 0.03 ng
RT: 26.586 min Scan# 3995
Delta R.T. 0.006 min
Lab File: BP005041.D
Acq: 25 Mar 2021 10:02

Tgt Ion:	276	Resp:	264
Ion Ratio	Lower	Upper	
276	100		
277	0.0	19.0	28.6#
138	4.8	16.6	24.8#

