

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032521\
 Data File : BP005040.D
 Acq On : 25 Mar 2021 09:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 10:19:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 10:18:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.566	0.554	2.1	96	0.00
3	Pyridine	1.407	1.550	-10.2	95	0.00
4	n-Nitrosodimethylamine	0.503	0.530	-5.4	95	0.00
5 S	2-Fluorophenol	1.300	1.326	-2.0	94	0.00
6	Aniline	2.110	2.197	-4.1	94	0.00
7 S	Phenol-d6	1.862	1.942	-4.3	95	0.00
8	2-Chlorophenol	1.371	1.395	-1.8	92	0.00
9	Benzaldehyde	0.948	0.925	2.4	94	0.00
10 C	Phenol	1.908	1.953	-2.4	93	0.00
11	bis(2-Chloroethyl)ether	1.397	1.455	-4.2	95	0.00
12	1,3-Dichlorobenzene	1.523	1.571	-3.2	94	0.00
13 C	1,4-Dichlorobenzene	1.546	1.579	-2.1	94	0.00
14	1,2-Dichlorobenzene	1.483	1.548	-4.4	97	0.00
15	Benzyl Alcohol	1.436	1.507	-4.9	93	0.00
16	2,2'-oxybis(1-Chloropropane	1.025	1.064	-3.8	94	0.00
17	2-Methylphenol	1.218	1.279	-5.0	95	0.00
18	Hexachloroethane	0.589	0.606	-2.9	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.208	1.281	-6.0	94	0.00
20	3+4-Methylphenols	1.663	1.753	-5.4	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
22	Acetophenone	0.575	0.583	-1.4	93	0.00
23 S	Nitrobenzene-d5	0.466	0.475	-1.9	93	0.00
24	Nitrobenzene	0.491	0.497	-1.2	94	0.00
25	Isophorone	0.860	0.901	-4.8	95	0.00
26 C	2-Nitrophenol	0.191	0.207	-8.4	98	0.00
27	2,4-Dimethylphenol	0.306	0.325	-6.2	98	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.460	-5.7	95	0.00
29 C	2,4-Dichlorophenol	0.343	0.359	-4.7	95	0.00
30	1,2,4-Trichlorobenzene	0.385	0.397	-3.1	95	0.00
31	Naphthalene	1.130	1.159	-2.6	95	0.00
32	Benzoic acid	0.227	0.224	1.3	86	0.00
33	4-Chloroaniline	0.457	0.478	-4.6	96	0.00
34 C	Hexachlorobutadiene	0.248	0.254	-2.4	93	0.00
35	Caprolactam	0.112	0.117	-4.5	94	0.00
36 C	4-Chloro-3-methylphenol	0.414	0.438	-5.8	96	0.00
37	2-Methylnaphthalene	0.816	0.838	-2.7	94	0.00
38	1-Methylnaphthalene	0.765	0.804	-5.1	96	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	0.651	0.662	-1.7	94	0.00
41 P	Hexachlorocyclopentadiene	0.354	0.374	-5.6	94	0.00
42 S	2,4,6-Tribromophenol	0.261	0.271	-3.8	96	0.00
43 C	2,4,6-Trichlorophenol	0.450	0.463	-2.9	93	0.00
44	2,4,5-Trichlorophenol	0.488	0.507	-3.9	96	0.00
45 S	2-Fluorobiphenyl	1.478	1.488	-0.7	94	0.00
46	1,1'-Biphenyl	1.525	1.543	-1.2	94	0.00
47	2-Chloronaphthalene	1.242	1.264	-1.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032521\
 Data File : BP005040.D
 Acq On : 25 Mar 2021 09:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 25 10:19:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 25 10:18:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.396	0.416	-5.1	95	0.00
49	Acenaphthylene	1.931	1.990	-3.1	95	0.00
50	Dimethylphthalate	1.540	1.566	-1.7	96	0.00
51	2,6-Dinitrotoluene	0.363	0.372	-2.5	94	0.00
52 C	Acenaphthene	1.188	1.213	-2.1	96	0.00
53	3-Nitroaniline	0.334	0.349	-4.5	92	0.00
54 P	2,4-Dinitrophenol	0.220	0.228	-3.6	93	0.00
55	Dibenzofuran	1.948	2.009	-3.1	97	0.00
56 P	4-Nitrophenol	0.274	0.297	-8.4	95	0.00
57	2,4-Dinitrotoluene	0.479	0.506	-5.6	95	0.00
58	Fluorene	1.573	1.597	-1.5	95	0.00
59	2,3,4,6-Tetrachlorophenol	0.450	0.454	-0.9	92	0.00
60	Diethylphthalate	1.507	1.548	-2.7	94	0.00
61	4-Chlorophenyl-phenylether	0.835	0.837	-0.2	96	0.00
62	4-Nitroaniline	0.343	0.372	-8.5	96	0.00
63	Azobenzene	1.722	1.731	-0.5	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	0.146	0.164	-12.3	96	0.00
66 c	n-Nitrosodiphenylamine	0.643	0.670	-4.2	95	0.00
67	4-Bromophenyl-phenylether	0.232	0.246	-6.0	95	0.00
68	Hexachlorobenzene	0.261	0.271	-3.8	95	0.00
69	Atrazine	0.214	0.232	-8.4	94	0.00
70 C	Pentachlorophenol	0.172	0.185	-7.6	97	0.00
71	Phenanthrene	1.159	1.215	-4.8	95	0.00
72	Anthracene	1.133	1.192	-5.2	96	0.00
73	Carbazole	1.064	1.126	-5.8	96	0.00
74	Di-n-butylphthalate	1.171	1.254	-7.1	95	0.00
75 C	Fluoranthene	1.364	1.425	-4.5	96	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
77	Benzydine	0.439	0.487	-10.9	89	0.00
78	Pyrene	1.374	1.452	-5.7	96	0.00
79 S	Terphenyl-d14	1.099	1.158	-5.4	96	0.00
80	Butylbenzylphthalate	0.510	0.562	-10.2	95	0.00
81	Benzo(a)anthracene	1.359	1.439	-5.9	98	0.00
82	3,3'-Dichlorobenzidine	0.457	0.491	-7.4	95	0.00
83	Chrysene	1.334	1.404	-5.2	97	0.00
84	Bis(2-ethylhexyl)phthalate	0.748	0.815	-9.0	94	0.00
85 c	Di-n-octyl phthalate	1.261	1.388	-10.1	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	1.500	1.584	-5.6	98	0.00
88	Benzo(b)fluoranthene	1.420	1.500	-5.6	98	0.00
89	Benzo(k)fluoranthene	1.361	1.428	-4.9	96	0.00
90 C	Benzo(a)pyrene	1.273	1.346	-5.7	97	0.00
91	Dibenzo(a,h)anthracene	1.297	1.379	-6.3	98	0.00
92	Benzo(g,h,i)perylene	1.251	1.311	-4.8	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032521\
Data File : BP005040.D
Acq On : 25 Mar 2021 09:28
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 25 10:19:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 25 10:18:57 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0