

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032621\
 Data File : BP005059.D
 Acq On : 26 Mar 2021 13:05
 Operator : CG/JU
 Sample : M1458-09
 Misc : MDL-WATER--8PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Quant Time: Mar 26 14:19:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:21:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	28929	20.00	ng	# 0.00
21) Naphthalene-d8	10.504	136	118038	20.00	ng	# 0.00
39) Acenaphthene-d10	14.369	164	77804	20.00	ng	0.00
64) Phenanthrene-d10	17.134	188	170386	20.00	ng	0.00
76) Chrysene-d12	21.227	240	175611	20.00	ng	0.00
86) Perylene-d12	23.516	264	176279	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.311	112	142470	75.79	ng	0.00
7) Phenol-d6	6.893	99	199145	73.96	ng	0.00
23) Nitrobenzene-d5	8.875	82	164110	59.65	ng	0.00
42) 2,4,6-Tribromophenol	15.869	330	77863	76.81	ng	0.00
45) 2-Fluorobiphenyl	12.987	172	352615	61.31	ng	0.00
79) Terphenyl-d14	19.704	244	600828	62.28	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.258	88	6521	7.97	ng	# 96
3) Pyridine	3.675	79	13847	6.80	ng	92
4) n-Nitrosodimethylamine	3.570	42	5345	7.34	ng	80
6) Aniline	7.052	93	25588	8.39	ng	# 88
8) 2-Chlorophenol	7.281	128	15035	7.58	ng	# 79
9) Benzaldehyde	6.869	77	15876	7.59	ng	# 79
10) Phenol	6.916	94	21455	7.77	ng	78
11) bis(2-Chloroethyl)ether	7.152	93	15896	7.87	ng	88
12) 1,3-Dichlorobenzene	7.605	146	17773	8.07	ng	# 93
13) 1,4-Dichlorobenzene	7.752	146	17732	7.93	ng	# 85
14) 1,2-Dichlorobenzene	8.063	146	17536	8.17	ng	# 90
15) Benzyl Alcohol	7.963	79	15157	7.30	ng	# 82
16) 2,2'-oxybis(1-Chloropr...	8.240	45	12520	8.45	ng	92
17) 2-Methylphenol	8.163	107	13630	7.74	ng	# 90
18) Hexachloroethane	8.787	117	6908	8.11	ng	88
19) n-Nitroso-di-n-propyla...	8.522	70	13936	7.98	ng	# 89
20) 3+4-Methylphenols	8.487	107	17971	7.47	ng	88
22) Acetophenone	8.540	105	25546	7.53	ng	# 94
24) Nitrobenzene	8.916	77	21822	7.53	ng	# 79
25) Isophorone	9.446	82	36660	7.22	ng	# 87
26) 2-Nitrophenol	9.622	139	7948	7.07	ng	# 62
27) 2,4-Dimethylphenol	9.693	122	14063	7.78	ng	94
28) bis(2-Chloroethoxy)met...	9.928	93	21639	8.42	ng	# 97
29) 2,4-Dichlorophenol	10.157	162	14187	7.01	ng	90
30) 1,2,4-Trichlorobenzene	10.369	180	16852	7.42	ng	97
31) Naphthalene	10.557	128	48292	7.24	ng	96
32) Benzoic acid	9.763	122	4455	3.32	ng	# 45
33) 4-Chloroaniline	10.675	127	20143	7.47	ng	# 89
34) Hexachlorobutadiene	10.840	225	10649	7.28	ng	97
35) Caprolactam	11.451	113	4764	7.18	ng	97
36) 4-Chloro-3-methylphenol	11.804	107	17161	7.02	ng	# 86
37) 2-Methylnaphthalene	12.175	142	35532	7.38	ng	# 95
38) 1-Methylnaphthalene	12.398	142	33203	7.36	ng	# 93
40) 1,2,4,5-Tetrachloroben...	12.546	216	20039	7.91	ng	# 97
41) Hexachlorocyclopentadiene	12.528	237	11886	8.62	ng	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032621\
 Data File : BP005059.D
 Acq On : 26 Mar 2021 13:05
 Operator : CG/JU
 Sample : M1458-09
 Misc : MDL-WATER--8PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Quant Time: Mar 26 14:19:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:21:24 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.793	196	12248	6.99	ng	94
44) 2,4,5-Trichlorophenol	12.863	196	13397	7.06	ng	# 86
46) 1,1'-Biphenyl	13.198	154	47424	7.99	ng	96
47) 2-Chloronaphthalene	13.240	162	36749	7.61	ng	97
48) 2-Nitroaniline	13.451	65	10378	6.74	ng	# 53
49) Acenaphthylene	14.092	152	54616	7.27	ng	99
50) Dimethylphthalate	13.834	163	46054	7.69	ng	# 98
51) 2,6-Dinitrotoluene	13.945	165	9759	6.91	ng	# 58
52) Acenaphthene	14.434	154	33587	7.27	ng	96
53) 3-Nitroaniline	14.287	138	8964	6.90	ng	# 69
54) 2,4-Dinitrophenol	14.492	184	4356	5.09	ng	# 83
55) Dibenzofuran	14.775	168	56079	7.40	ng	97
56) 4-Nitrophenol	14.592	139	6256	5.87	ng	# 54
57) 2,4-Dinitrotoluene	14.740	165	13183	7.07	ng	# 60
58) Fluorene	15.422	166	45644	7.46	ng	93
59) 2,3,4,6-Tetrachlorophenol	14.916	232	12157	6.95	ng	93
60) Diethylphthalate	15.204	149	44740	7.63	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	24940	7.68	ng	98
62) 4-Nitroaniline	15.451	138	9031	6.76	ng	# 61
63) Azobenzene	15.716	77	47406	7.08	ng	85
65) 4,6-Dinitro-2-methylph...	15.510	198	7526	6.05	ng	91
66) n-Nitrosodiphenylamine	15.639	169	38900	7.10	ng	98
67) 4-Bromophenyl-phenylether	16.322	248	15017	7.61	ng	94
68) Hexachlorobenzene	16.428	284	16406	7.39	ng	98
69) Atrazine	16.598	200	12044	6.62	ng	85
70) Pentachlorophenol	16.781	266	9337	6.37	ng	94
71) Phenanthrene	17.175	178	74467	7.54	ng	98
72) Anthracene	17.263	178	70290	7.28	ng	99
73) Carbazole	17.539	167	63041	6.95	ng	99
74) Di-n-butylphthalate	18.098	149	70813	7.10	ng	# 96
75) Fluoranthene	19.151	202	83720	7.20	ng	97
77) Benzidine	19.339	184	29390	7.63	ng	96
78) Pyrene	19.504	202	88436	7.33	ng	98
80) Butylbenzylphthalate	20.386	149	30056	6.71	ng	# 80
81) Benzo(a)anthracene	21.210	228	89021	7.46	ng	96
82) 3,3'-Dichlorobenzidine	21.151	252	31085	7.75	ng	99
83) Chrysene	21.263	228	85573	7.31	ng	98
84) Bis(2-ethylhexyl)phtha...	21.145	149	45390	6.91	ng	94
85) Di-n-octyl phthalate	22.027	149	76095	6.87	ng	# 93
87) Indeno(1,2,3-cd)pyrene	25.862	276	97601	7.38	ng	98
88) Benzo(b)fluoranthene	22.816	252	92109	7.36	ng	96
89) Benzo(k)fluoranthene	22.863	252	86623	7.22	ng	100
90) Benzo(a)pyrene	23.410	252	79385	7.07	ng	97
91) Dibenzo(a,h)anthracene	25.868	278	82131	7.19	ng	# 95
92) Benzo(g,h,i)perylene	26.574	276	79309	7.19	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032621\
 Data File : BP005059.D
 Acq On : 26 Mar 2021 13:05
 Operator : CG/JU
 Sample : M1458-09
 Misc : MDL-WATER--8PPM
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Quant Time: Mar 26 14:19:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:21:24 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.716 min Scan# 787
 Delta R.T. 0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021


Tgt Ion:	152	Resp:	28929
Ion	Ratio	Lower	Upper
152	100		
150	174.1	123.8	185.6
115	73.8	43.8	65.6#



1,4-Dioxane
 Concen: 7.97 ng
 RT: 3.258 min Scan# 29
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	88	Resp:	6521
Ion	Ratio	Lower	Upper
88	100		
58	56.8	44.2	66.4
43	23.2	15.2	22.8#









Pyridine
Concen: 6.80 ng
RT: 3.675 min Scan# 100
Delta R.T. 0.023 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Tgt Ion:	79	Resp:	13847
Ion	Ratio	Lower	Upper
79	100		
52	45.9	41.3	61.9
51	30.5	21.2	31.8



n-Nitrosodimethylamine
Concen: 7.34 ng
RT: 3.570 min Scan# 82
Delta R.T. 0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	42	Resp:	5345
Ion	Ratio	Lower	Upper
42	100		
74	156.5	148.6	223.0
44	7.6	5.3	7.9









2-Fluorophenol
 Concen: 75.79 ng
 RT: 5.311 min Scan# 378
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021


Tgt Ion:	112	Resp:	142470
Ion	Ratio	Lower	Upper
112	100		
64	56.5	32.8	49.2#
63	31.7	17.0	25.4#



Aniline
 Concen: 8.39 ng
 RT: 7.052 min Scan# 674
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	93	Resp:	25588
Ion	Ratio	Lower	Upper
93	100		
66	37.4	25.4	38.0
65	21.7	12.5	18.7#









Phenol-d6
Concen: 73.96 ng
RT: 6.893 min Scan# 647
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Tgt Ion:	99	Resp:	199145
Ion	Ratio	Lower	Upper
99	100		
42	14.5	8.6	13.0#
71	40.3	23.4	35.2#



2-Chlorophenol
Concen: 7.58 ng
RT: 7.281 min Scan# 713
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	128	Resp:	15035
Ion	Ratio	Lower	Upper
128	100		
130	29.9	13.0	53.0
64	50.6	10.4	50.4#









Benzaldehyde
 Concen: 7.59 ng
 RT: 6.869 min Scan# 643
 Delta R.T. 0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021


Tgt Ion:	77	Resp:	15876
Ion	Ratio	Lower	Upper
77	100		
105	98.9	97.1	137.1
106	85.5	93.7	133.7#



Phenol
 Concen: 7.77 ng
 RT: 6.916 min Scan# 651
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	94	Resp:	21455
Ion	Ratio	Lower	Upper
94	100		
65	31.4	2.3	42.3
66	44.0	10.9	50.9









#1
 bis(2-Chloroethyl)ether
 Concen: 7.87 ng
 RT: 7.152 min Scan# 691
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021


Tgt Ion:	93	Resp:	15896
Ion	Ratio	Lower	Upper
93	100		
63	70.9	39.0	79.0
95	35.9	12.8	52.8



#12
 1,3-Dichlorobenzene
 Concen: 8.07 ng
 RT: 7.605 min Scan# 768
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	146	Resp:	17773
Ion	Ratio	Lower	Upper
146	100		
148	65.0	52.5	78.7
75	35.1	18.1	27.1#









1,4-Dichlorobenzene
Concen: 7.93 ng
RT: 7.752 min Scan# 793
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Tgt Ion:	146	Resp:	17732
Ion	Ratio	Lower	Upper
146	100		
148	69.1	51.3	76.9
111	53.2	28.9	43.3#



1,2-Dichlorobenzene
Concen: 8.17 ng
RT: 8.063 min Scan# 846
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	146	Resp:	17536
Ion	Ratio	Lower	Upper
146	100		
148	68.3	51.6	77.4
111	50.5	31.0	46.6#









Benzyl Alcohol
 Concen: 7.30 ng
 RT: 7.963 min Scan# 829
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021


Tgt Ion:	79	Resp:	15157
Ion	Ratio	Lower	Upper
79	100		
108	68.7	73.5	110.3#
77	70.5	50.3	75.5



2,2'-oxybis(1-Chloropropane)
 Concen: 8.45 ng
 RT: 8.240 min Scan# 876
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	45	Resp:	12520
Ion	Ratio	Lower	Upper
45	100		
77	28.7	2.3	42.3
79	16.9	0.0	36.6









2-Methylphenol
 Concen: 7.74 ng
 RT: 8.163 min Scan# 863
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	107	Resp:	13630
Ion	Ratio	Lower	Upper
107	100		
108	110.4	89.6	134.4
77	54.4	31.8	47.6#
79	50.7	31.1	46.7#




Hexachloroethane
 Concen: 8.11 ng
 RT: 8.787 min Scan# 969
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	117	Resp:	6908
Ion	Ratio	Lower	Upper
117	100		
119	95.1	78.8	118.2
201	94.3	92.3	138.5






Abundance Scan 926 (8.534 min): BP005024.D\data.ms (-917) #19




n-Nitroso-di-n-propylamine
Concen: 7.98 ng
RT: 8.522 min Scan# 924
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion	Ratio	Lower	Upper
70	100		
42	36.1	32.8	49.2
101	9.7	9.6	14.4
130	21.7	24.3	36.5







Abundance Scan 920 (8.499 min): BP005024.D\data.ms (-909) #20



3+4-Methylphenols
Concen: 7.47 ng
RT: 8.487 min Scan# 918
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion	Ratio	Lower	Upper
107	100		
108	83.0	69.0	109.0
77	36.3	8.5	48.5
79	38.0	4.8	44.8









Naphthalene-d8
Concen: 20.00 ng
RT: 10.504 min Scan# 1261
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	136	Resp:	118038
Ion	Ratio	Lower	Upper
136	100		
137	9.8	8.6	13.0
54	7.0	3.6	5.4#
68	6.8	3.3	4.9#



Acetophenone
Concen: 7.53 ng
RT: 8.540 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	105	Resp:	25546
Ion	Ratio	Lower	Upper
105	100		
71	2.3	2.6	3.8#
51	20.9	13.5	20.3#
120	22.4	19.6	29.4






Abundance Scan 985 (8.881 min): BP005024.D\data.ms (-977) #23

Nitrobenzene-d5
Concen: 59.65 ng
RT: 8.875 min Scan# 984
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021







Tgt Ion:	82	Resp:	164110
Ion	Ratio	Lower	Upper
82	100		
128	39.2	45.6	68.4#
54	40.1	30.1	45.1




Abundance Scan 992 (8.922 min): BP005024.D\data.ms (-974) #24

Nitrobenzene
Concen: 7.53 ng
RT: 8.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	77	Resp:	21822
Ion	Ratio	Lower	Upper
77	100		
123	42.5	48.6	72.8#
65	14.3	10.0	15.0





Abundance Scan 1082 (9.452 min): BP005024.D\data.ms (-1025-)




Isophorone
Concen: 7.22 ng
RT: 9.446 min Scan# 1081
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Tgt Ion:	82	Resp:	36660
Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.4	9.6
138	14.8	18.7	28.1#



Abundance Scan 1081 (9.446 min): BP005059.D\data.ms (-1030-)





Abundance Scan 1112 (9.628 min): BP005024.D\data.ms (-1106-)




2-Nitrophenol
Concen: 7.07 ng
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	139	Resp:	7948
Ion	Ratio	Lower	Upper
139	100		
109	38.3	22.1	33.1#
65	63.3	26.5	39.7#



Abundance Scan 1111 (9.622 min): BP005059.D\data.ms (-1061-)





Abundance Scan 1123 (9.693 min): BP005024.D\data.ms (-1127-)




2,4-Dimethylphenol
Concen: 7.78 ng
RT: 9.693 min Scan# 1123
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Tgt Ion:	122	Resp:	14063
Ion	Ratio	Lower	Upper
122	100		
107	117.9	87.3	130.9
121	58.4	47.3	70.9



Abundance Scan 1123 (9.693 min): BP005059.D\data.ms (-1072-)





Abundance Scan 1164 (9.934 min): BP005024.D\data.ms (-1128-)





bis(2-Chloroethoxy)methane
Concen: 8.42 ng
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	93	Resp:	21639
Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.7	40.1
123	10.0	10.5	15.7#



Abundance Scan 1163 (9.928 min): BP005059.D\data.ms (-1112-)









2,4-Dichlorophenol
Concen: 7.01 ng
RT: 10.157 min Scan# 1202
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	162	Resp:	14187
Ion	Ratio	Lower	Upper
162	100		
164	68.4	45.8	85.8
98	43.4	10.2	50.2




1,2,4-Trichlorobenzene
Concen: 7.42 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	180	Resp:	16852
Ion	Ratio	Lower	Upper
180	100		
182	92.6	76.2	114.2
145	26.7	23.8	35.8






Abundance Scan 1271 (10.563 min): BP005024.D\data.ms (-1281) (-)




Naphthalene
Concen: 7.24 ng
RT: 10.557 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	128	Resp:	48292
Ion	Ratio	Lower	Upper
128	100		
129	11.9	8.8	13.2
127	15.4	10.6	15.8

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021





Abundance Scan 1147 (9.834 min): BP005024.D\data.ms (-1130) (-)




Benzoic acid
Concen: 3.32 ng
RT: 9.763 min Scan# 1135
Delta R.T. -0.065 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	122	Resp:	4455
Ion	Ratio	Lower	Upper
122	100		
105	167.4	94.1	134.1#
77	123.0	50.1	90.1#



Abundance Scan 1135 (9.763 min): BP005059.D\data.ms (-1095) (-)






Abundance Scan 1291 (10.681 min): BP005024.D\data.ms (-1239) (-)




4-Chloroaniline
Concen: 7.47 ng
RT: 10.675 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	127	Resp:	20143
Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.7	38.5
65	33.3	17.3	25.9#
92	22.3	12.9	19.3#







Abundance Scan 1319 (10.846 min): BP005024.D\data.ms (-1267) (-)



Hexachlorobutadiene
Concen: 7.28 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	225	Resp:	10649
Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.1	75.1
227	68.4	51.4	77.2









Caprolactam
Concen: 7.18 ng
RT: 11.451 min Scan# 1422
Delta R.T. -0.018 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	113	Resp:	4764
Ion	Ratio	Lower	Upper
113	100		
55	96.8	77.3	117.3
56	71.0	55.9	95.9



4-Chloro-3-methylphenol
Concen: 7.02 ng
RT: 11.804 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	107	Resp:	17161
Ion	Ratio	Lower	Upper
107	100		
144	21.4	22.7	34.1#
142	75.5	70.7	106.1






Abundance Scan 1546 (12.181 min): BP005024.D\data.ms (-1537) (-)

2-Methylnaphthalene
Concen: 7.38 ng
RT: 12.175 min Scan# 1545
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:	142	Resp:	35532
Ion	Ratio	Lower	Upper
142	100		
141	89.0	71.4	107.2
115	42.6	25.8	38.6#






Abundance Scan 1583 (12.398 min): BP005024.D\data.ms (-1537) (-)


1-Methylnaphthalene
Concen: 7.36 ng
RT: 12.398 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05



Tgt Ion:	142	Resp:	33203
Ion	Ratio	Lower	Upper
142	100		
141	86.5	74.4	111.6
116	5.3	2.9	4.3#






Abundance Scan 1919 (14.375 min): BP005024.D\data.ms (-1819) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.369 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	164	Resp:	77804
Ion	Ratio	Lower	Upper
164	100		
162	101.1	79.4	119.2
160	43.3	36.1	54.1






Abundance Scan 1609 (12.551 min): BP005024.D\data.ms (-1609) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 7.91 ng
RT: 12.546 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	216	Resp:	20039
Ion	Ratio	Lower	Upper
216	100		
214	79.8	63.0	94.6
179	21.9	16.9	25.3
108	20.8	12.6	19.0#






Abundance Scan 1605 (12.528 min): BP005024.D\data.ms (-1554) (-)




Hexachlorocyclopentadiene
Concen: 8.62 ng
RT: 12.528 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	237	Resp:	11886
Ion	Ratio	Lower	Upper
237	100		
235	54.4	43.2	83.2
272	15.4	0.0	33.3







Abundance Scan 2174 (15.875 min): BP005024.D\data.ms (-2123) (-)



2,4,6-Tribromophenol
Concen: 76.81 ng
RT: 15.869 min Scan# 2173
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	330	Resp:	77863
Ion	Ratio	Lower	Upper
330	100		
332	95.9	77.1	115.7
141	35.0	23.4	35.2









Scan 1651 (12.798 min): BP005024.D\data.ms (-1643) (-)
 2,4,6-Trichlorophenol
 Concen: 6.99 ng
 RT: 12.793 min Scan# 1650
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	196	Resp:	12248
Ion	Ratio	Lower	Upper
196	100		
198	103.5	77.0	115.6
200	33.3	25.1	37.7




Scan 1663 (12.869 min): BP005024.D\data.ms (-1647) (-)
 2,4,5-Trichlorophenol
 Concen: 7.06 ng
 RT: 12.863 min Scan# 1662
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	196	Resp:	13397
Ion	Ratio	Lower	Upper
196	100		
198	81.4	77.3	115.9
97	46.9	31.0	46.4#
132	22.4	20.9	31.3






Abundance Scan 1684 (12.993 min): BP005024.D\data.ms (-1675) (-)




2-Fluorobiphenyl
Concen: 61.31 ng
RT: 12.987 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	172	Resp:	352615
Ion	Ratio	Lower	Upper
172	100		
171	34.4	28.5	42.7
170	23.2	19.2	28.8






Abundance Scan 1720 (13.204 min): BP005024.D\data.ms (-1705) (-)




1,1'-Biphenyl
Concen: 7.99 ng
RT: 13.198 min Scan# 1719
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	154	Resp:	47424
Ion	Ratio	Lower	Upper
154	100		
153	42.1	20.4	60.4
76	13.8	0.0	29.3






Abundance Scan 1727 (13.245 min): BP005024.D\data.ms (-1710) (-)




2-Chloronaphthalene
Concen: 7.61 ng
RT: 13.240 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	162	Resp:	36749
Ion	Ratio	Lower	Upper
162	100		
127	38.7	29.9	44.9
164	34.8	26.1	39.1







Abundance Scan 1763 (13.457 min): BP005024.D\data.ms (-1744) (-)



2-Nitroaniline
Concen: 6.74 ng
RT: 13.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	65	Resp:	10378
Ion	Ratio	Lower	Upper
65	100		
92	65.6	70.6	106.0#
138	86.2	133.6	200.4#









Acenaphthylene
Concen: 7.27 ng
RT: 14.092 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05




Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	152	Resp:	54616
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.9	23.9
153	12.9	10.5	15.7




Dimethylphthalate
Concen: 7.69 ng
RT: 13.834 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	163	Resp:	46054
Ion	Ratio	Lower	Upper
163	100		
194	5.1	3.3	4.9#
164	9.6	8.2	12.2



Abundance Scan 1848 (13.957 min): BP005024.D\data.ms (-1850) (-)





2,6-Dinitrotoluene
Concen: 6.91 ng
RT: 13.945 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	165	Resp:	9759
Ion	Ratio	Lower	Upper
165	100		
63	65.1	27.8	41.6#
89	64.6	34.6	52.0#


Abundance Scan 1846 (13.945 min): BP005059.D\data.ms



Abundance Scan 1846 (13.945 min): BP005059.D\data.ms (-1796) (-)





Abundance Scan 1930 (14.439 min): BP005024.D\data.ms (-1950) (-)




Acenaphthene
Concen: 7.27 ng
RT: 14.434 min Scan# 1929
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	154	Resp:	33587
Ion	Ratio	Lower	Upper
154	100		
153	109.5	91.0	136.6
152	52.0	43.0	64.6


Abundance Scan 1929 (14.434 min): BP005059.D\data.ms



Abundance Scan 1929 (14.434 min): BP005059.D\data.ms (-1879) (-)






Abundance Scan 1904 (14.287 min): BP005024.D\data.ms (-1853) (-)




3-Nitroaniline
Concen: 6.90 ng
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:138 Resp: 8964
Ion Ratio Lower Upper
138 100
108 14.1 7.5 11.3#
92 125.7 75.0 112.6#






Abundance Scan 1939 (14.492 min): BP005024.D\data.ms (-1851) (-)




2,4-Dinitrophenol
Concen: 5.09 ng
RT: 14.492 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:184 Resp: 4356
Ion Ratio Lower Upper
184 100
63 64.5 32.3 48.5#
154 58.9 49.1 73.7



Abundance Scan 1987 (14.775 min): BP005024.D\data.ms (-1936) (-)





Dibenzenofuran
Concen: 7.40 ng
RT: 14.775 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	168	Resp:	56079
Ion	Ratio	Lower	Upper
168	100		
139	38.4	29.1	43.7
169	13.9	10.4	15.6

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021


Abundance Scan 1987 (14.775 min): BP005059.D\data.ms



Abundance Scan 1987 (14.775 min): BP005059.D\data.ms (-1936) (-)





Abundance Scan 1957 (14.598 min): BP005024.D\data.ms (-1936) (-)




4-Nitrophenol
Concen: 5.87 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	139	Resp:	6256
Ion	Ratio	Lower	Upper
139	100		
109	94.5	44.6	84.6#
65	103.3	41.6	81.6#

Abundance Scan 1956 (14.592 min): BP005059.D\data.ms




Abundance Scan 1956 (14.592 min): BP005059.D\data.ms (-1906) (-)






Abundance Scan 1982 (14.745 min): BP005024.D\data.ms (-1952) (-)

2,4-Dinitrotoluene
Concen: 7.07 ng
RT: 14.740 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:	165	Resp:	13183
Ion	Ratio	Lower	Upper
165	100		
63	48.0	20.6	31.0#
89	82.0	42.6	63.8#
182	2.0	2.4	3.6#






Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2058) (-)


Fluorene
Concen: 7.46 ng
RT: 15.422 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05



Tgt Ion:	166	Resp:	45644
Ion	Ratio	Lower	Upper
166	100		
165	89.4	77.8	116.8
167	13.5	11.0	16.4






Abundance Scan 2026 (15.004 min): BP005024.D\data.ms (-2015) (-)




2,3,4,6-Tetrachlorophenol
Concen: 6.95 ng
RT: 14.916 min Scan# 2011
Delta R.T. -0.088 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	232	Resp:	12157
Ion	Ratio	Lower	Upper
232	100		
131	37.8	34.1	51.1
130	2.4	1.8	2.8
166	26.4	24.6	36.8






Abundance Scan 2061 (15.210 min): BP005024.D\data.ms (-2050) (-)




Diethylphthalate
Concen: 7.63 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	149	Resp:	44740
Ion	Ratio	Lower	Upper
149	100		
177	22.4	18.0	27.0
150	12.7	10.0	15.0






Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2069) (-)




4-Chlorophenyl-phenylether
Concen: 7.68 ng
RT: 15.422 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	204	Resp:	24940
Ion	Ratio	Lower	Upper
204	100		
206	32.5	25.9	38.9
141	55.5	46.5	69.7






Abundance Scan 2103 (15.457 min): BP005024.D\data.ms (-2052) (-)




4-Nitroaniline
Concen: 6.76 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	138	Resp:	9031
Ion	Ratio	Lower	Upper
138	100		
92	63.5	20.7	60.7#
108	100.9	47.8	87.8#






Abundance Scan 2148 (15.722 min): BP005024.D\data.ms (-2361) (-)




Azobenzene
Concen: 7.08 ng
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	77	Resp:	47406
Ion	Ratio	Lower	Upper
77	100		
182	24.9	16.9	56.9
105	20.3	5.6	45.6
51	27.9	2.9	42.9






Abundance Scan 2388 (17.133 min): BP005024.D\data.ms (-2361) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.134 min Scan# 2388
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	188	Resp:	170386
Ion	Ratio	Lower	Upper
188	100		
94	8.8	6.2	9.2
80	9.1	6.2	9.2






Abundance Scan 2112 (15.510 min): BP005024.D\data.ms (-2165) (-)




4,6-Dinitro-2-methylphenol
Concen: 6.05 ng
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	198	Resp:	7526
Ion	Ratio	Lower	Upper
198	100		
51	33.5	6.4	46.4
105	42.7	18.9	58.9





Abundance Scan 2135 (15.645 min): BP005024.D\data.ms (-2165) (-)




n-Nitrosodiphenylamine
Concen: 7.10 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05


Tgt Ion:	169	Resp:	38900
Ion	Ratio	Lower	Upper
169	100		
168	65.6	53.8	80.6
167	34.2	28.9	43.3



Abundance Scan 2134 (15.639 min): BP005059.D\data.ms (-2083) (-)






Abundance Scan 2250 (16.322 min): BP005024.D\data.ms (-246) (-)




4-Bromophenyl-phenylether
Concen: 7.61 ng
RT: 16.322 min Scan# 2250
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	248	Resp:	15017
Ion	Ratio	Lower	Upper
248	100		
250	97.5	76.7	115.1
141	75.4	53.0	79.6






Abundance Scan 2269 (16.433 min): BP005024.D\data.ms (-246) (-)




Hexachlorobenzene
Concen: 7.39 ng
RT: 16.428 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	284	Resp:	16406
Ion	Ratio	Lower	Upper
284	100		
142	32.9	24.2	36.4
249	29.9	23.8	35.8






Abundance Scan 2298 (16.604 min): BP005024.D\data.ms (-2298) (-)




Atrazine
Concen: 6.62 ng
RT: 16.598 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	200	Resp:	12044
Ion	Ratio	Lower	Upper
200	100		
173	30.6	5.9	45.9
215	60.6	28.4	68.4







Abundance Scan 2328 (16.781 min): BP005024.D\data.ms (-2328) (-)



Pentachlorophenol
Concen: 6.37 ng
RT: 16.781 min Scan# 2328
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	266	Resp:	9337
Ion	Ratio	Lower	Upper
266	100		
268	69.5	51.1	76.7
264	67.1	50.6	76.0









Scan 2395 (17.175 min): BP005024.D\data.ms (-2344) (-)
 Phenanthrene
 Concen: 7.54 ng
 RT: 17.175 min Scan# 2395
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	178	Resp:	74467
Ion	Ratio	Lower	Upper
178	100		
176	17.3	15.3	22.9
179	15.8	12.5	18.7




Scan 2411 (17.269 min): BP005024.D\data.ms (-2474) (-)
 Anthracene
 Concen: 7.28 ng
 RT: 17.263 min Scan# 2410
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	178	Resp:	70290
Ion	Ratio	Lower	Upper
178	100		
176	18.1	15.0	22.4
179	15.5	12.5	18.7






Abundance Scan 2458 (17.545 min): BP005024.D\data.ms (-2407) (-)




Carbazole
Concen: 6.95 ng
RT: 17.539 min Scan# 2457
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	167	Resp:	63041
Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.6	26.4
139	11.7	10.2	15.4






Abundance Scan 2552 (18.098 min): BP005024.D\data.ms (-2502) (-)




Di-n-butylphthalate
Concen: 7.10 ng
RT: 18.098 min Scan# 2552
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	149	Resp:	70813
Ion	Ratio	Lower	Upper
149	100		
150	9.7	7.3	10.9
104	6.8	3.4	5.0#






Abundance Scan 2731 (19.151 min): BP005024.D\data.ms (-2731) (-)




Fluoranthene
Concen: 7.20 ng
RT: 19.151 min Scan# 2731
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	202	Resp:	83720
Ion	Ratio	Lower	Upper
202	100		
101	11.5	0.0	29.4
203	18.6	0.0	37.8






Abundance Scan 3084 (21.227 min): BP005024.D\data.ms (-3075) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.227 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	240	Resp:	175611
Ion	Ratio	Lower	Upper
240	100		
120	9.9	7.8	11.6
236	24.1	20.8	31.2






Abundance Scan 2763 (19.339 min): BP005024.D\data.ms (-2712) (-)




Benzidine
Concen: 7.63 ng
RT: 19.339 min Scan# 2763
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:184 Resp: 29390
Ion Ratio Lower Upper
184 100
185 13.0 11.5 17.3
183 13.7 9.6 14.4







Abundance Scan 2791 (19.504 min): BP005024.D\data.ms (-2740) (-)



Pyrene
Concen: 7.33 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:202 Resp: 88436
Ion Ratio Lower Upper
202 100
200 19.9 17.1 25.7
203 17.0 14.2 21.2









Terphenyl-d14
 Concen: 62.28 ng
 RT: 19.704 min Scan# 2825
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
 ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	244	Resp:	600828
Ion	Ratio	Lower	Upper
244	100		
212	6.7	5.8	8.6
122	9.9	7.8	11.8




Butylbenzylphthalate
 Concen: 6.71 ng
 RT: 20.386 min Scan# 2941
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	149	Resp:	30056
Ion	Ratio	Lower	Upper
149	100		
91	76.1	45.4	68.0#
206	23.1	17.1	25.7






Abundance Scan 3081 (21.210 min): BP005024.D\data.ms (-3081) (-)




Benzo(a)anthracene
Concen: 7.46 ng
RT: 21.210 min Scan# 3081
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	228	Resp:	89021
Ion	Ratio	Lower	Upper
228	100		
226	24.2	22.1	33.1
229	20.2	15.9	23.9







Abundance Scan 3071 (21.151 min): BP005024.D\data.ms (-3071) (-)



3,3'-Dichlorobenzidine
Concen: 7.75 ng
RT: 21.151 min Scan# 3071
Delta R.T. -0.000 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	252	Resp:	31085
Ion	Ratio	Lower	Upper
252	100		
254	64.1	51.0	76.4
126	10.3	7.9	11.9









Scan 3090 (21.263 min): BP005024.D\data.ms (-3088) (-)
Chrysene
 Concen: 7.31 ng
 RT: 21.263 min Scan# 3090
 Delta R.T. -0.006 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	228	Resp:	85573
Ion	Ratio	Lower	Upper
228	100		
226	28.9	24.4	36.6
229	19.2	15.8	23.6




Scan 3070 (21.145 min): BP005024.D\data.ms (-3082) (-)
Bis(2-ethylhexyl)phthalate
 Concen: 6.91 ng
 RT: 21.145 min Scan# 3070
 Delta R.T. -0.000 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05

Tgt Ion:	149	Resp:	45390
Ion	Ratio	Lower	Upper
149	100		
167	31.9	22.6	34.0
279	5.6	4.2	6.2






Abundance Scan 3220 (22.027 min): BP005024.D\data.ms (-3485) (-)




Di-n-octyl phthalate
Concen: 6.87 ng
RT: 22.027 min Scan# 3220
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	149	Resp:	76095
Ion	Ratio	Lower	Upper
149	100		
167	1.8	1.3	1.9
43	7.6	3.7	5.5#






Abundance Scan 3473 (23.516 min): BP005024.D\data.ms (-3485) (-)




Perylene-d12
Concen: 20.00 ng
RT: 23.516 min Scan# 3473
Delta R.T. -0.006 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	264	Resp:	176279
Ion	Ratio	Lower	Upper
264	100		
260	22.0	18.8	28.2
265	21.8	17.4	26.2






Abundance Scan 3872 (25.862 min): BP005024.D\data.ms (-3837) (-)




Indeno(1,2,3-cd)pyrene
Concen: 7.38 ng
RT: 25.862 min Scan# 3872
Delta R.T. -0.012 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021

Tgt Ion:	276	Resp:	97601
Ion	Ratio	Lower	Upper
276	100		
138	21.3	18.7	28.1
277	25.7	20.4	30.6







Abundance Scan 3355 (22.821 min): BP005024.D\data.ms (-3349) (-)



Benzo(b)fluoranthene
Concen: 7.36 ng
RT: 22.816 min Scan# 3354
Delta R.T. -0.012 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Tgt Ion:	252	Resp:	92109
Ion	Ratio	Lower	Upper
252	100		
253	24.5	17.6	26.4
125	9.4	7.3	10.9









Scan 3363 (22.868 min): BP005024.D\data.ms (-3363) (-)
 Benzo(k)fluoranthene
 Concen: 7.22 ng
 RT: 22.863 min Scan# 3362
 Delta R.T. -0.012 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05




Instrument :
 BNA_P
ClientSampleId :
 MDL-WATER-03-QT1-2021

Tgt Ion:	252	Resp:	86623
Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.8	26.8
125	9.5	7.4	11.2



Scan 3456 (23.415 min): BP005024.D\data.ms (-3456) (-)
 Benzo(a)pyrene
 Concen: 7.07 ng
 RT: 23.410 min Scan# 3455
 Delta R.T. -0.012 min
 Lab File: BP005059.D
 Acq: 26 Mar 2021 13:05


Tgt Ion:	252	Resp:	79385
Ion	Ratio	Lower	Upper
252	100		
253	23.4	17.4	26.0
125	10.2	8.2	12.4






Abundance Scan 3875 (25.880 min): BP005024.D\data.ms (-3875) (-)

Dibenzo(a,h)anthracene
Concen: 7.19 ng
RT: 25.868 min Scan# 3873
Delta R.T. -0.018 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05

Instrument :
BNA_P
ClientSampleId :
MDL-WATER-03-QT1-2021




Tgt Ion:	278	Resp:	82131
Ion	Ratio	Lower	Upper
278	100		
139	10.9	12.4	18.6#
279	24.5	19.0	28.6



Abundance Scan 3995 (26.586 min): BP005024.D\data.ms (-3995) (-)

Benzo(g,h,i)perylene
Concen: 7.19 ng
RT: 26.574 min Scan# 3993
Delta R.T. -0.018 min
Lab File: BP005059.D
Acq: 26 Mar 2021 13:05



Tgt Ion:	276	Resp:	79309
Ion	Ratio	Lower	Upper
276	100		
277	21.3	19.0	28.6
138	20.5	16.6	24.8

