

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032621\
 Data File : BP005060.D
 Acq On : 26 Mar 2021 13:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Mar 26 14:21:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:21:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	30983	20.00	ng	# 0.00
21) Naphthalene-d8	10.510	136	119666	20.00	ng	# 0.00
39) Acenaphthene-d10	14.375	164	82071	20.00	ng	0.00
64) Phenanthrene-d10	17.134	188	169240	20.00	ng	# 0.00
76) Chrysene-d12	21.227	240	167325	20.00	ng	0.00
86) Perylene-d12	23.516	264	167941	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.311	112	158322	78.64	ng	0.00
7) Phenol-d6	6.893	99	232145	80.50	ng	0.00
23) Nitrobenzene-d5	8.875	82	219141	78.57	ng	0.00
42) 2,4,6-Tribromophenol	15.875	330	85349	79.82	ng	0.00
45) 2-Fluorobiphenyl	12.987	172	468540	77.23	ng	0.00
79) Terphenyl-d14	19.704	244	749054	81.49	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.252	88	32393	36.95	ng	96
3) Pyridine	3.652	79	90177	41.37	ng	94
4) n-Nitrosodimethylamine	3.564	42	31088	39.89	ng	84
6) Aniline	7.052	93	130861	40.04	ng	# 88
8) 2-Chlorophenol	7.281	128	83486	39.32	ng	# 80
9) Benzaldehyde	6.864	77	53763	40.92	ng	# 78
10) Phenol	6.922	94	117400	39.72	ng	82
11) bis(2-Chloroethyl)ether	7.152	93	85808	39.66	ng	96
12) 1,3-Dichlorobenzene	7.605	146	92469	39.19	ng	# 87
13) 1,4-Dichlorobenzene	7.752	146	93056	38.85	ng	# 93
14) 1,2-Dichlorobenzene	8.069	146	89837	39.10	ng	# 94
15) Benzyl Alcohol	7.963	79	90219	40.55	ng	# 84
16) 2,2'-oxybis(1-Chloropr...	8.252	45	63287	39.87	ng	97
17) 2-Methylphenol	8.163	107	75869	40.22	ng	# 91
18) Hexachloroethane	8.787	117	35452	38.84	ng	90
19) n-Nitroso-di-n-propyla...	8.528	70	76593	40.94	ng	# 91
20) 3+4-Methylphenols	8.493	107	105431	40.94	ng	93
22) Acetophenone	8.540	105	137367	39.93	ng	# 93
24) Nitrobenzene	8.916	77	116524	39.65	ng	# 76
25) Isophorone	9.446	82	210118	40.82	ng	# 89
26) 2-Nitrophenol	9.628	139	46809	41.07	ng	# 64
27) 2,4-Dimethylphenol	9.693	122	72772	39.70	ng	89
28) bis(2-Chloroethoxy)met...	9.928	93	106990	41.07	ng	97
29) 2,4-Dichlorophenol	10.157	162	82764	40.35	ng	93
30) 1,2,4-Trichlorobenzene	10.369	180	91562	39.75	ng	98
31) Naphthalene	10.557	128	270894	40.07	ng	99
32) Benzoic acid	9.834	122	54326	39.92	ng	# 72
33) 4-Chloroaniline	10.675	127	111639	40.82	ng	# 88
34) Hexachlorobutadiene	10.840	225	58721	39.61	ng	98
35) Caprolactam	11.469	113	28725	42.68	ng	91
36) 4-Chloro-3-methylphenol	11.804	107	102611	41.41	ng	# 81
37) 2-Methylnaphthalene	12.175	142	195189	39.99	ng	# 95
38) 1-Methylnaphthalene	12.399	142	184959	40.42	ng	99
40) 1,2,4,5-Tetrachloroben...	12.551	216	102323	38.29	ng	98
41) Hexachlorocyclopentadiene	12.528	237	58600	40.31	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032621\
 Data File : BP005060.D
 Acq On : 26 Mar 2021 13:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Quant Time: Mar 26 14:21:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 26 13:21:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.793	196	73771	39.93	ng	98
44) 2,4,5-Trichlorophenol	12.863	196	78277	39.12	ng	# 94
46) 1,1'-Biphenyl	13.198	154	238280	38.06	ng	95
47) 2-Chloronaphthalene	13.240	162	197078	38.68	ng	97
48) 2-Nitroaniline	13.451	65	64469	39.71	ng	# 57
49) Acenaphthylene	14.093	152	305915	38.60	ng	99
50) Dimethylphthalate	13.840	163	244874	38.74	ng	98
51) 2,6-Dinitrotoluene	13.957	165	58510	39.27	ng	# 62
52) Acenaphthene	14.440	154	187109	38.37	ng	97
53) 3-Nitroaniline	14.287	138	54788	39.96	ng	# 65
54) 2,4-Dinitrophenol	14.493	184	34701	38.46	ng	# 78
55) Dibenzofuran	14.775	168	309260	38.69	ng	99
56) 4-Nitrophenol	14.598	139	45877	40.79	ng	# 48
57) 2,4-Dinitrotoluene	14.745	165	78539	39.92	ng	# 59
58) Fluorene	15.428	166	251607	38.99	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.004	232	70276	38.08	ng	99
60) Diethylphthalate	15.204	149	240484	38.89	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	130947	38.23	ng	100
62) 4-Nitroaniline	15.457	138	57320	40.69	ng	# 60
63) Azobenzene	15.716	77	266974	37.77	ng	85
65) 4,6-Dinitro-2-methylph...	15.510	198	53066	42.97	ng	86
66) n-Nitrosodiphenylamine	15.640	169	219236	40.30	ng	96
67) 4-Bromophenyl-phenylether	16.322	248	78544	40.06	ng	95
68) Hexachlorobenzene	16.434	284	86652	39.30	ng	96
69) Atrazine	16.604	200	75799	41.95	ng	94
70) Pentachlorophenol	16.781	266	59897	41.14	ng	98
71) Phenanthrene	17.175	178	384405	39.18	ng	99
72) Anthracene	17.269	178	377693	39.40	ng	98
73) Carbazole	17.539	167	357946	39.74	ng	99
74) Di-n-butylphthalate	18.104	149	406480	41.02	ng	# 98
75) Fluoranthene	19.151	202	456392	39.54	ng	98
77) Benzidine	19.339	184	164010	44.68	ng	100
78) Pyrene	19.504	202	468466	40.74	ng	99
80) Butylbenzylphthalate	20.386	149	182477	42.77	ng	# 76
81) Benzo(a)anthracene	21.216	228	460946	40.55	ng	98
82) 3,3'-Dichlorobenzidine	21.151	252	158831	41.57	ng	100
83) Chrysene	21.269	228	449989	40.33	ng	97
84) Bis(2-ethylhexyl)phtha...	21.145	149	266548	42.61	ng	97
85) Di-n-octyl phthalate	22.027	149	445455	42.24	ng	# 94
87) Indeno(1,2,3-cd)pyrene	25.868	276	501153	39.80	ng	97
88) Benzo(b)fluoranthene	22.827	252	466789	39.14	ng	98
89) Benzo(k)fluoranthene	22.869	252	451975	39.55	ng	99
90) Benzo(a)pyrene	23.416	252	425943	39.83	ng	100
91) Dibenzo(a,h)anthracene	25.880	278	436212	40.06	ng	96
92) Benzo(g,h,i)perylene	26.586	276	419268	39.90	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 26 14:21:47 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 26 13:21:24 2021
Response via : Initial Calibration









Scan 788 (7.722 min): BP005024.D\data.ms (-779) #1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.716 min Scan# 787
 Delta R.T. 0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleID :
 SSTDCCC040EC


Tgt Ion:	152	Resp:	30983
Ion	Ratio	Lower	Upper
152	100		
150	152.2	123.8	185.6
115	71.1	43.8	65.6#



Scan 28 (3.252 min): BP005024.D\data.ms (-23) #2
 1,4-Dioxane
 Concen: 36.95 ng
 RT: 3.252 min Scan# 28
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	88	Resp:	32393
Ion	Ratio	Lower	Upper
88	100		
58	57.5	44.2	66.4
43	21.3	15.2	22.8









Pyridine
Concen: 41.37 ng
RT: 3.652 min Scan# 96
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	79	Resp:	90177
Ion	Ratio	Lower	Upper
79	100		
52	45.8	41.3	61.9
51	25.4	21.2	31.8



n-Nitrosodimethylamine
Concen: 39.89 ng
RT: 3.564 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	42	Resp:	31088
Ion	Ratio	Lower	Upper
42	100		
74	161.7	148.6	223.0
44	6.6	5.3	7.9









2-Fluorophenol
 Concen: 78.64 ng
 RT: 5.311 min Scan# 378
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	112	Resp:	158322
Ion	Ratio	Lower	Upper
112	100		
64	56.2	32.8	49.2#
63	30.8	17.0	25.4#



Aniline
 Concen: 40.04 ng
 RT: 7.052 min Scan# 674
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	93	Resp:	130861
Ion	Ratio	Lower	Upper
93	100		
66	40.0	25.4	38.0#
65	18.4	12.5	18.7









Phenol-d6
 Concen: 80.50 ng
 RT: 6.893 min Scan# 647
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	99	Resp:	232145
Ion	Ratio	Lower	Upper
99	100		
42	14.3	8.6	13.0#
71	40.7	23.4	35.2#



2-Chlorophenol
 Concen: 39.32 ng
 RT: 7.281 min Scan# 713
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	128	Resp:	83486
Ion	Ratio	Lower	Upper
128	100		
130	32.3	13.0	53.0
64	52.9	10.4	50.4#









Benzaldehyde
 Concen: 40.92 ng
 RT: 6.864 min Scan# 642
 Delta R.T. -0.005 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	77	Resp:	53763
Ion	Ratio	Lower	Upper
77	100		
105	94.5	97.1	137.1#
106	89.2	93.7	133.7#



Phenol
 Concen: 39.72 ng
 RT: 6.922 min Scan# 652
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	94	Resp:	117400
Ion	Ratio	Lower	Upper
94	100		
65	30.3	2.3	42.3
66	41.3	10.9	50.9









Scan 691 (7.152 min): BP005024.D\data.ms (-684) #1
 bis(2-Chloroethyl)ether
 Concen: 39.66 ng
 RT: 7.152 min Scan# 691
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	93	Resp:	85808
Ion	Ratio	Lower	Upper
93	100		
63	63.2	39.0	79.0
95	31.9	12.8	52.8



Scan 769 (7.610 min): BP005024.D\data.ms (-760) #12
 1,3-Dichlorobenzene
 Concen: 39.19 ng
 RT: 7.605 min Scan# 768
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	146	Resp:	92469
Ion	Ratio	Lower	Upper
146	100		
148	61.6	52.5	78.7
75	39.6	18.1	27.1#









1,4-Dichlorobenzene
Concen: 38.85 ng
RT: 7.752 min Scan# 793
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	146	Resp:	93056
Ion	Ratio	Lower	Upper
146	100		
148	65.2	51.3	76.9
111	46.0	28.9	43.3#



1,2-Dichlorobenzene
Concen: 39.10 ng
RT: 8.069 min Scan# 847
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	146	Resp:	89837
Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.6	77.4
111	47.9	31.0	46.6#









Benzyl Alcohol
 Concen: 40.55 ng
 RT: 7.963 min Scan# 829
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	79	Resp:	90219
Ion	Ratio	Lower	Upper
79	100		
108	69.9	73.5	110.3#
77	66.4	50.3	75.5



2,2'-oxybis(1-Chloropropane)
 Concen: 39.87 ng
 RT: 8.252 min Scan# 878
 Delta R.T. 0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	45	Resp:	63287
Ion	Ratio	Lower	Upper
45	100		
77	23.4	2.3	42.3
79	17.8	0.0	36.6









2-Methylphenol
Concen: 40.22 ng
RT: 8.163 min Scan# 863
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	107	Resp:	75869
Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.6	134.4
77	49.9	31.8	47.6#
79	54.9	31.1	46.7#



Hexachloroethane
Concen: 38.84 ng
RT: 8.787 min Scan# 969
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	117	Resp:	35452
Ion	Ratio	Lower	Upper
117	100		
119	93.8	78.8	118.2
201	100.5	92.3	138.5









#19
 n-Nitroso-di-n-propylamine
 Concen: 40.94 ng
 RT: 8.528 min Scan# 925
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	70	Resp:	76593
Ion	Ratio	Lower	Upper
70	100		
42	42.0	32.8	49.2
101	11.5	9.6	14.4
130	18.5	24.3	36.5#



#20
 3+4-Methylphenols
 Concen: 40.94 ng
 RT: 8.493 min Scan# 919
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	107	Resp:	105431
Ion	Ratio	Lower	Upper
107	100		
108	90.6	69.0	109.0
77	37.9	8.5	48.5
79	32.0	4.8	44.8









Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.510 min Scan# 1262
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	136	Resp:	119666
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.6	13.0
54	5.9	3.6	5.4#
68	5.8	3.3	4.9#



Acetophenone
 Concen: 39.93 ng
 RT: 8.540 min Scan# 927
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39


Tgt Ion:	105	Resp:	137367
Ion	Ratio	Lower	Upper
105	100		
71	4.5	2.6	3.8#
51	20.9	13.5	20.3#
120	21.9	19.6	29.4





Abundance Scan 985 (8.881 min): BP005024.D\data.ms (-977) #23

Nitrobenzene-d5
Concen: 78.57 ng
RT: 8.875 min Scan# 984
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC



Tgt Ion:	82	Resp:	219141
Ion	Ratio	Lower	Upper
82	100		
128	38.5	45.6	68.4#
54	42.0	30.1	45.1





Abundance Scan 984 (8.875 min): BP005060.D\data.ms (-934) (-)




Abundance Scan 992 (8.922 min): BP005024.D\data.ms (-974) #24

Nitrobenzene
Concen: 39.65 ng
RT: 8.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	77	Resp:	116524
Ion	Ratio	Lower	Upper
77	100		
123	39.8	48.6	72.8#
65	14.2	10.0	15.0



Abundance Scan 991 (8.916 min): BP005060.D\data.ms (-941) (-)





Abundance Scan 1082 (9.452 min): BP005024.D\data.ms (-1025-)




Isophorone
Concen: 40.82 ng
RT: 9.446 min Scan# 1081
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	82	Resp:	210118
Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.4	9.6
138	16.1	18.7	28.1#



Abundance Scan 1081 (9.446 min): BP005060.D\data.ms (-1030-)





Abundance Scan 1112 (9.628 min): BP005024.D\data.ms (-1106-)




2-Nitrophenol
Concen: 41.07 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	139	Resp:	46809
Ion	Ratio	Lower	Upper
139	100		
109	38.9	22.1	33.1#
65	59.9	26.5	39.7#



Abundance Scan 1112 (9.628 min): BP005060.D\data.ms (-1061-)






Abundance Scan 1123 (9.693 min): BP005024.D\data.ms (-1127-)




2,4-Dimethylphenol
Concen: 39.70 ng
RT: 9.693 min Scan# 1123
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	122	Resp:	72772
Ion	Ratio	Lower	Upper
122	100		
107	126.7	87.3	130.9
121	58.6	47.3	70.9







Abundance Scan 1164 (9.934 min): BP005024.D\data.ms (-1128-)



bis(2-Chloroethoxy)methane
Concen: 41.07 ng
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	93	Resp:	106990
Ion	Ratio	Lower	Upper
93	100		
95	32.0	26.7	40.1
123	11.5	10.5	15.7









2,4-Dichlorophenol
Concen: 40.35 ng
RT: 10.157 min Scan# 1202
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	162	Resp:	82764
Ion	Ratio	Lower	Upper
162	100		
164	64.2	45.8	85.8
98	40.8	10.2	50.2



1,2,4-Trichlorobenzene
Concen: 39.75 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	180	Resp:	91562
Ion	Ratio	Lower	Upper
180	100		
182	92.9	76.2	114.2
145	28.7	23.8	35.8









Naphthalene
Concen: 40.07 ng
RT: 10.557 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	128	Resp:	270894
Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.8	13.2
127	13.4	10.6	15.8



Benzoic acid
Concen: 39.92 ng
RT: 9.834 min Scan# 1147
Delta R.T. 0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	122	Resp:	54326
Ion	Ratio	Lower	Upper
122	100		
105	131.7	94.1	134.1
77	107.9	50.1	90.1#









4-Chloroaniline
 Concen: 40.82 ng
 RT: 10.675 min Scan# 1290
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:	127	Resp:	111639
Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.7	38.5
65	33.6	17.3	25.9#
92	23.7	12.9	19.3#



Hexachlorobutadiene
 Concen: 39.61 ng
 RT: 10.840 min Scan# 1318
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	225	Resp:	58721
Ion	Ratio	Lower	Upper
225	100		
223	61.6	50.1	75.1
227	63.0	51.4	77.2









#35 (-)
 Caprolactam
 Concen: 42.68 ng
 RT: 11.469 min Scan# 1425
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	113	Resp:	28725
Ion	Ratio	Lower	Upper
113	100		
55	103.5	77.3	117.3
56	87.1	55.9	95.9



#36 (-)
 4-Chloro-3-methylphenol
 Concen: 41.41 ng
 RT: 11.804 min Scan# 1482
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	107	Resp:	102611
Ion	Ratio	Lower	Upper
107	100		
144	22.6	22.7	34.1#
142	68.5	70.7	106.1#



Abundance Scan 1546 (12.181 min): BP005024.D\data.ms (-137) (-)

2-Methylnaphthalene

Concen: 39.99 ng

RT: 12.175 min Scan# 1545

Delta R.T. -0.006 min

Lab File: BP005060.D


Acq: 26 Mar 2021 13:39

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040EC





Tgt Ion:142 Resp: 195189

Ion Ratio Lower Upper


142 100

141 88.4 71.4 107.2

115 40.5 25.8 38.6#



Abundance Scan 1545 (12.175 min): BP005060.D\data.ms (-1495) (-)



Abundance Scan 1583 (12.398 min): BP005024.D\data.ms (-137) (-)

1-Methylnaphthalene


Concen: 40.42 ng

RT: 12.399 min Scan# 1583

Delta R.T. -0.000 min

Lab File: BP005060.D

Acq: 26 Mar 2021 13:39





Tgt Ion:142 Resp: 184959

Ion Ratio Lower Upper


142 100

141 91.6 74.4 111.6


116 4.2 2.9 4.3



Abundance Scan 1583 (12.399 min): BP005060.D\data.ms (-1532) (-)






Abundance Scan 1919 (14.375 min): BP005024.D\data.ms (-1868) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.375 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	164	Resp:	82071
Ion Ratio	Lower	Upper	
164	100		
162	96.5	79.4	119.2
160	39.7	36.1	54.1






Abundance Scan 1609 (12.551 min): BP005024.D\data.ms (-1558) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 38.29 ng
RT: 12.551 min Scan# 1609
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	216	Resp:	102323
Ion Ratio	Lower	Upper	
216	100		
214	79.2	63.0	94.6
179	20.0	16.9	25.3
108	18.9	12.6	19.0






Abundance Scan 1605 (12.528 min): BP005024.D\data.ms (-1554) (-)




Hexachlorocyclopentadiene
Concen: 40.31 ng
RT: 12.528 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	237	Resp:	58600
Ion	Ratio	Lower	Upper
237	100		
235	59.8	43.2	83.2
272	12.1	0.0	33.3







Abundance Scan 2174 (15.875 min): BP005024.D\data.ms (-2123) (-)



2,4,6-Tribromophenol
Concen: 79.82 ng
RT: 15.875 min Scan# 2174
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	330	Resp:	85349
Ion	Ratio	Lower	Upper
330	100		
332	95.8	77.1	115.7
141	35.4	23.4	35.2#









Scan 1651 (12.798 min): BP005024.D\data.ms (-1643) (-)
 2,4,6-Trichlorophenol
 Concen: 39.93 ng
 RT: 12.793 min Scan# 1650
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	196	Resp:	73771
Ion	Ratio	Lower	Upper
196	100		
198	95.0	77.0	115.6
200	29.5	25.1	37.7




Scan 1663 (12.869 min): BP005024.D\data.ms (-1647) (-)
 2,4,5-Trichlorophenol
 Concen: 39.12 ng
 RT: 12.863 min Scan# 1662
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	196	Resp:	78277
Ion	Ratio	Lower	Upper
196	100		
198	96.5	77.3	115.9
97	53.6	31.0	46.4#
132	25.4	20.9	31.3






Abundance Scan 1684 (12.993 min): BP005024.D\data.ms (-1675) (-)




2-Fluorobiphenyl
Concen: 77.23 ng
RT: 12.987 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	172	Resp:	468540
Ion	Ratio	Lower	Upper
172	100		
171	35.0	28.5	42.7
170	22.6	19.2	28.8






Abundance Scan 1720 (13.204 min): BP005024.D\data.ms (-1705) (-)




1,1'-Biphenyl
Concen: 38.06 ng
RT: 13.198 min Scan# 1719
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	154	Resp:	238280
Ion	Ratio	Lower	Upper
154	100		
153	42.4	20.4	60.4
76	13.4	0.0	29.3





Abundance Scan 1727 (13.245 min): BP005024.D\data.ms (-1710) (-)




2-Chloronaphthalene
Concen: 38.68 ng
RT: 13.240 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	162	Resp:	197078
Ion	Ratio	Lower	Upper
162	100		
127	39.8	29.9	44.9
164	33.5	26.1	39.1



Abundance Scan 1726 (13.240 min): BP005060.D\data.ms (-1675) (-)





Abundance Scan 1763 (13.457 min): BP005024.D\data.ms (-1744) (-)





2-Nitroaniline
Concen: 39.71 ng
RT: 13.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	65	Resp:	64469
Ion	Ratio	Lower	Upper
65	100		
92	68.9	70.6	106.0#
138	92.9	133.6	200.4#



Abundance Scan 1762 (13.451 min): BP005060.D\data.ms (-1711) (-)









Acenaphthylene
Concen: 38.60 ng
RT: 14.093 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	152	Resp:	305915
Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.9	23.9
153	12.8	10.5	15.7




Dimethylphthalate
Concen: 38.74 ng
RT: 13.840 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	163	Resp:	244874
Ion	Ratio	Lower	Upper
163	100		
194	4.9	3.3	4.9
164	9.6	8.2	12.2






Abundance Scan 1848 (13.957 min): BP005024.D\data.ms (-1796) (-)




2,6-Dinitrotoluene
Concen: 39.27 ng
RT: 13.957 min Scan# 1848
Delta R.T. 0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	165	Resp:	58510
Ion	Ratio	Lower	Upper
165	100		
63	62.2	27.8	41.6#
89	62.6	34.6	52.0#






Abundance Scan 1930 (14.439 min): BP005024.D\data.ms (-1796) (-)



Acenaphthene
Concen: 38.37 ng
RT: 14.440 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	154	Resp:	187109
Ion	Ratio	Lower	Upper
154	100		
153	110.5	91.0	136.6
152	51.8	43.0	64.6






Abundance Scan 1904 (14.287 min): BP005024.D\data.ms (-1853) (-)

3-Nitroaniline
Concen: 39.96 ng
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC







Tgt Ion:	138	Resp:	54788
Ion	Ratio	Lower	Upper
138	100		
108	12.6	7.5	11.3#
92	130.0	75.0	112.6#



Abundance Scan 1939 (14.492 min): BP005024.D\data.ms (-1851) (-)

2,4-Dinitrophenol
Concen: 38.46 ng
RT: 14.493 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	184	Resp:	34701
Ion	Ratio	Lower	Upper
184	100		
63	72.6	32.3	48.5#
154	63.2	49.1	73.7





Abundance Scan 1987 (14.775 min): BP005024.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 38.69 ng
RT: 14.775 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC



Tgt Ion:	168	Resp:	309260
Ion	Ratio	Lower	Upper
168	100		
139	37.2	29.1	43.7
169	13.2	10.4	15.6





Abundance Scan 1987 (14.775 min): BP005060.D\data.ms (-1936) (-)




Abundance Scan 1957 (14.598 min): BP005024.D\data.ms (-1956) (-)

4-Nitrophenol
Concen: 40.79 ng
RT: 14.598 min Scan# 1957
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	139	Resp:	45877
Ion	Ratio	Lower	Upper
139	100		
109	96.5	44.6	84.6#
65	110.7	41.6	81.6#




Abundance Scan 1957 (14.598 min): BP005060.D\data.ms (-1906) (-)






Abundance Scan 1982 (14.745 min): BP005024.D\data.ms (-1952) (-)

2,4-Dinitrotoluene
Concen: 39.92 ng
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC







Tgt Ion:	165	Resp:	78539
Ion	Ratio	Lower	Upper
165	100		
63	49.0	20.6	31.0#
89	82.1	42.6	63.8#
182	3.6	2.4	3.6




Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2058) (-)

Fluorene
Concen: 38.99 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	166	Resp:	251607
Ion	Ratio	Lower	Upper
166	100		
165	94.9	77.8	116.8
167	13.4	11.0	16.4









Scan 2026 (15.004 min): BP005024.D\data.ms (-2015) (-)
 2,3,4,6-Tetrachlorophenol
 Concen: 38.08 ng
 RT: 15.004 min Scan# 2026
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	232	Resp:	70276
Ion	Ratio	Lower	Upper
232	100		
131	41.2	34.1	51.1
130	2.8	1.8	2.8
166	30.6	24.6	36.8




Scan 2061 (15.210 min): BP005024.D\data.ms (-2015) (-)
 Diethylphthalate
 Concen: 38.89 ng
 RT: 15.204 min Scan# 2060
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	149	Resp:	240484
Ion	Ratio	Lower	Upper
149	100		
177	22.2	18.0	27.0
150	12.4	10.0	15.0






Abundance Scan 2098 (15.428 min): BP005024.D\data.ms (-2069) (-)




4-Chlorophenyl-phenylether
Concen: 38.23 ng
RT: 15.422 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:204 Resp: 130947
Ion Ratio Lower Upper
204 100
206 32.5 25.9 38.9
141 58.5 46.5 69.7






Abundance Scan 2103 (15.457 min): BP005024.D\data.ms (-2052) (-)




4-Nitroaniline
Concen: 40.69 ng
RT: 15.457 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:138 Resp: 57320
Ion Ratio Lower Upper
138 100
92 56.2 20.7 60.7
108 106.8 47.8 87.8#





Abundance Scan 2148 (15.722 min): BP005024.D\data.ms (-2361) (-)




Azobenzene
Concen: 37.77 ng
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	77	Resp:	266974
Ion	Ratio	Lower	Upper
77	100		
182	23.3	16.9	56.9
105	19.8	5.6	45.6
51	25.8	2.9	42.9



Abundance Scan 2147 (15.716 min): BP005060.D\data.ms (-2097) (-)





Abundance Scan 2388 (17.133 min): BP005024.D\data.ms (-2359) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.134 min Scan# 2388
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	188	Resp:	169240
Ion	Ratio	Lower	Upper
188	100		
94	9.7	6.2	9.2#
80	10.2	6.2	9.2#



Abundance Scan 2388 (17.134 min): BP005060.D\data.ms (-2337) (-)






Abundance Scan 2112 (15.510 min): BP005024.D\data.ms (-2112) (-)




4,6-Dinitro-2-methylphenol
Concen: 42.97 ng
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	198	Resp:	53066
Ion	Ratio	Lower	Upper
198	100		
51	37.6	6.4	46.4
105	44.2	18.9	58.9






Abundance Scan 2135 (15.645 min): BP005024.D\data.ms (-2135) (-)




n-Nitrosodiphenylamine
Concen: 40.30 ng
RT: 15.640 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	169	Resp:	219236
Ion	Ratio	Lower	Upper
169	100		
168	64.5	53.8	80.6
167	33.6	28.9	43.3





Abundance Scan 2250 (16.322 min): BP005024.D\data.ms (-246) (-)




4-Bromophenyl-phenylether
Concen: 40.06 ng
RT: 16.322 min Scan# 2250
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	248	Resp:	78544
Ion	Ratio	Lower	Upper
248	100		
250	100.4	76.7	115.1
141	70.4	53.0	79.6



Abundance Scan 2250 (16.322 min): BP005060.D\data.ms (-2199) (-)





Abundance Scan 2269 (16.433 min): BP005024.D\data.ms (-246) (-)




Hexachlorobenzene
Concen: 39.30 ng
RT: 16.434 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:	284	Resp:	86652
Ion	Ratio	Lower	Upper
284	100		
142	31.3	24.2	36.4
249	32.9	23.8	35.8



Abundance Scan 2269 (16.434 min): BP005060.D\data.ms (-2218) (-)






Abundance Scan 2298 (16.604 min): BP005024.D\data.ms (-2247) (-)




Atrazine
Concen: 41.95 ng
RT: 16.604 min Scan# 2298
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	200	Resp:	75799
Ion	Ratio	Lower	Upper
200	100		
173	24.5	5.9	45.9
215	53.8	28.4	68.4







Abundance Scan 2328 (16.781 min): BP005024.D\data.ms (-2247) (-)



Pentachlorophenol
Concen: 41.14 ng
RT: 16.781 min Scan# 2328
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	266	Resp:	59897
Ion	Ratio	Lower	Upper
266	100		
268	63.6	51.1	76.7
264	60.9	50.6	76.0









Phenanthrene
Concen: 39.18 ng
RT: 17.175 min Scan# 2395
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	178	Resp:	384405
Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.3	22.9
179	15.1	12.5	18.7



Anthracene
Concen: 39.40 ng
RT: 17.269 min Scan# 2411
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	178	Resp:	377693
Ion	Ratio	Lower	Upper
178	100		
176	17.6	15.0	22.4
179	14.7	12.5	18.7









Scan 2458 (17.545 min): BP005024.D\data.ms (-2473)
 Carbazole
 Concen: 39.74 ng
 RT: 17.539 min Scan# 2457
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	167	Resp:	357946
Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.6	26.4
139	12.5	10.2	15.4




Scan 2552 (18.098 min): BP005024.D\data.ms (-2573)
 Di-n-butylphthalate
 Concen: 41.02 ng
 RT: 18.104 min Scan# 2553
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	149	Resp:	406480
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	5.6	3.4	5.0#






Abundance Scan 2731 (19.151 min): BP005024.D\data.ms (-275) (-)




Fluoranthene
Concen: 39.54 ng
RT: 19.151 min Scan# 2731
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	202	Resp:	456392
Ion	Ratio	Lower	Upper
202	100		
101	10.9	0.0	29.4
203	17.5	0.0	37.8






Abundance Scan 3084 (21.227 min): BP005024.D\data.ms (-3075) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.227 min Scan# 3084
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	240	Resp:	167325
Ion	Ratio	Lower	Upper
240	100		
120	9.3	7.8	11.6
236	25.3	20.8	31.2





Abundance Scan 2763 (19.339 min): BP005024.D\data.ms (-2712) (-)




Benzidine
Concen: 44.68 ng
RT: 19.339 min Scan# 2763
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:184 Resp: 164010
Ion Ratio Lower Upper
184 100
185 14.4 11.5 17.3
183 12.3 9.6 14.4



Abundance Scan 2763 (19.339 min): BP005060.D\data.ms (-2712) (-)





Abundance Scan 2791 (19.504 min): BP005024.D\data.ms (-2740) (-)




Pyrene
Concen: 40.74 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39


Tgt Ion:202 Resp: 468466
Ion Ratio Lower Upper
202 100
200 20.7 17.1 25.7
203 17.9 14.2 21.2



Abundance Scan 2791 (19.504 min): BP005060.D\data.ms (-2740) (-)






Abundance Scan 2825 (19.704 min): BP005024.D\data.ms (-2825) (-)




Terphenyl-d14
Concen: 81.49 ng
RT: 19.704 min Scan# 2825
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	244	Resp:	749054
Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.6
122	9.6	7.8	11.8






Abundance Scan 2941 (20.386 min): BP005024.D\data.ms (-2941) (-)




Butylbenzylphthalate
Concen: 42.77 ng
RT: 20.386 min Scan# 2941
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	149	Resp:	182477
Ion	Ratio	Lower	Upper
149	100		
91	79.1	45.4	68.0#
206	24.7	17.1	25.7






Abundance Scan 3081 (21.210 min): BP005024.D\data.ms (-3081) (-)




Benzo(a)anthracene
Concen: 40.55 ng
RT: 21.216 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	228	Resp:	460946
Ion	Ratio	Lower	Upper
228	100		
226	26.0	22.1	33.1
229	19.6	15.9	23.9







Abundance Scan 3071 (21.151 min): BP005024.D\data.ms (-3071) (-)



3,3'-Dichlorobenzidine
Concen: 41.57 ng
RT: 21.151 min Scan# 3071
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	252	Resp:	158831
Ion	Ratio	Lower	Upper
252	100		
254	63.4	51.0	76.4
126	9.8	7.9	11.9









Scan 3090 (21.263 min): BP005024.D\data.ms (-3088) (-)
Chrysene
 Concen: 40.33 ng
 RT: 21.269 min Scan# 3091
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	228	Resp:	449989
Ion	Ratio	Lower	Upper
228	100		
226	28.4	24.4	36.6
229	18.9	15.8	23.6




Scan 3070 (21.145 min): BP005024.D\data.ms (-3082) (-)
Bis(2-ethylhexyl)phthalate
 Concen: 42.61 ng
 RT: 21.145 min Scan# 3070
 Delta R.T. -0.000 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39

Tgt Ion:	149	Resp:	266548
Ion	Ratio	Lower	Upper
149	100		
167	30.1	22.6	34.0
279	6.1	4.2	6.2






Abundance Scan 3220 (22.027 min): BP005024.D\data.ms (-3185) (-)




Di-n-octyl phthalate
Concen: 42.24 ng
RT: 22.027 min Scan# 3220
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	149	Resp:	445455
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.4	3.7	5.5#






Abundance Scan 3473 (23.516 min): BP005024.D\data.ms (-3485) (-)




Perylene-d12
Concen: 20.00 ng
RT: 23.516 min Scan# 3473
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	264	Resp:	167941
Ion	Ratio	Lower	Upper
264	100		
260	24.2	18.8	28.2
265	21.7	17.4	26.2






Abundance Scan 3872 (25.862 min): BP005024.D\data.ms (-3872) (-)




Indeno(1,2,3-cd)pyrene
Concen: 39.80 ng
RT: 25.868 min Scan# 3873
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	276	Resp:	501153
Ion	Ratio	Lower	Upper
276	100		
138	20.8	18.7	28.1
277	25.7	20.4	30.6







Abundance Scan 3355 (22.821 min): BP005024.D\data.ms (-3355) (-)



Benzo(b)fluoranthene
Concen: 39.14 ng
RT: 22.827 min Scan# 3356
Delta R.T. -0.000 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Tgt Ion:	252	Resp:	466789
Ion	Ratio	Lower	Upper
252	100		
253	20.9	17.6	26.4
125	8.2	7.3	10.9









Scan 3363 (22.868 min): BP005024.D\data.ms (-3363) (-)
 Benzo(k)fluoranthene
 Concen: 39.55 ng
 RT: 22.869 min Scan# 3363
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	252	Resp:	451975
Ion	Ratio	Lower	Upper
252	100		
253	22.9	17.8	26.8
125	9.2	7.4	11.2



Scan 3456 (23.415 min): BP005024.D\data.ms (-3456) (-)
 Benzo(a)pyrene
 Concen: 39.83 ng
 RT: 23.416 min Scan# 3456
 Delta R.T. -0.006 min
 Lab File: BP005060.D
 Acq: 26 Mar 2021 13:39


Tgt Ion:	252	Resp:	425943
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	10.1	8.2	12.4






Abundance Scan 3875 (25.880 min): BP005024.D\data.ms (-3875) (-)

Dibenzo(a,h)anthracene
Concen: 40.06 ng
RT: 25.880 min Scan# 3875
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC




Tgt Ion:	278	Resp:	436212
Ion	Ratio	Lower	Upper
278	100		
139	12.8	12.4	18.6
279	22.8	19.0	28.6



Abundance Scan 3995 (26.586 min): BP005024.D\data.ms (-3995) (-)

Benzo(g,h,i)perylene
Concen: 39.90 ng
RT: 26.586 min Scan# 3995
Delta R.T. -0.006 min
Lab File: BP005060.D
Acq: 26 Mar 2021 13:39



Tgt Ion:	276	Resp:	419268
Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.0	28.6
138	17.4	16.6	24.8

