

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032822\
 Data File : BP009566.D
 Acq On : 28 Mar 2022 14:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 03:11:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 06:34:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	43.830	-9.6	121	0.00
3	Pyridine	40.000	43.509	-8.8	117	0.00
4	n-Nitrosodimethylamine	40.000	42.982	-7.5	117	0.00
5 S	2-Fluorophenol	80.000	83.342	-4.2	113	0.00
6	Aniline	40.000	39.716	0.7	107	0.00
7 S	Phenol-d6	80.000	79.792	0.3	109	0.00
8	2-Chlorophenol	40.000	39.399	1.5	107	0.00
9	Benzaldehyde	40.000	40.515	-1.3	113	0.00
10 C	Phenol	40.000	40.020	-0.1	109	0.00
11	bis(2-Chloroethyl)ether	40.000	40.742	-1.9	111	0.00
12	1,3-Dichlorobenzene	40.000	39.716	0.7	108	0.00
13 C	1,4-Dichlorobenzene	40.000	39.346	1.6	108	0.00
14	1,2-Dichlorobenzene	40.000	39.138	2.2	107	0.00
15	Benzyl Alcohol	40.000	37.433	6.4	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.821	-9.6	120	0.00
17	2-Methylphenol	40.000	38.357	4.1	105	0.00
18	Hexachloroethane	40.000	39.389	1.5	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.784	0.5	109	0.00
20	3+4-Methylphenols	40.000	38.690	3.3	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	40.045	-0.1	106	0.00
23 S	Nitrobenzene-d5	80.000	80.605	-0.8	106	0.00
24	Nitrobenzene	40.000	40.358	-0.9	107	0.00
25	Isophorone	40.000	40.558	-1.4	108	0.00
26 C	2-Nitrophenol	40.000	39.328	1.7	103	0.00
27	2,4-Dimethylphenol	40.000	40.201	-0.5	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.390	-3.5	109	0.00
29 C	2,4-Dichlorophenol	40.000	40.144	-0.4	105	0.00
30	1,2,4-Trichlorobenzene	40.000	39.369	1.6	104	0.00
31	Naphthalene	40.000	39.523	1.2	106	0.00
32	Benzoic acid	40.000	40.545	-1.4	104	-0.02
33	4-Chloroaniline	40.000	39.540	1.2	105	0.00
34 C	Hexachlorobutadiene	40.000	38.057	4.9	101	0.00
35	Caprolactam	40.000	38.474	3.8	106	-0.01
36 C	4-Chloro-3-methylphenol	40.000	38.020	4.9	102	0.00
37	2-Methylnaphthalene	40.000	39.342	1.6	106	0.00
38	1-Methylnaphthalene	40.000	39.373	1.6	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.769	3.1	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	31.829	20.4	84	0.00
42 S	2,4,6-Tribromophenol	80.000	70.802	11.5	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.054	2.4	103	0.00
44	2,4,5-Trichlorophenol	40.000	38.797	3.0	103	0.00
45 S	2-Fluorobiphenyl	80.000	78.707	1.6	105	0.00
46	1,1'-Biphenyl	40.000	39.678	0.8	106	0.00
47	2-Chloronaphthalene	40.000	39.573	1.1	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032822\
 Data File : BP009566.D
 Acq On : 28 Mar 2022 14:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 29 03:11:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 06:34:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.856	0.4	105	0.00
49	Acenaphthylene	40.000	39.329	1.7	106	0.00
50	Dimethylphthalate	40.000	39.298	1.8	106	0.00
51	2,6-Dinitrotoluene	40.000	39.079	2.3	104	0.00
52 C	Acenaphthene	40.000	39.477	1.3	106	0.00
53	3-Nitroaniline	40.000	39.251	1.9	105	0.00
54 P	2,4-Dinitrophenol	40.000	37.533	6.2	104	0.00
55	Dibenzofuran	40.000	39.180	2.1	106	0.00
56 P	4-Nitrophenol	40.000	39.235	1.9	102	0.00
57	2,4-Dinitrotoluene	40.000	38.973	2.6	103	0.00
58	Fluorene	40.000	39.329	1.7	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.165	4.6	102	0.00
60	Diethylphthalate	40.000	39.311	1.7	107	0.00
61	4-Chlorophenyl-phenylether	40.000	38.819	3.0	105	0.00
62	4-Nitroaniline	40.000	39.780	0.5	105	0.00
63	Azobenzene	40.000	41.932	-4.8	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.888	-2.2	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.376	1.6	107	0.00
67	4-Bromophenyl-phenylether	40.000	37.717	5.7	103	0.00
68	Hexachlorobenzene	40.000	36.516	8.7	99	0.00
69	Atrazine	40.000	39.469	1.3	103	0.00
70 C	Pentachlorophenol	40.000	39.797	0.5	101	0.00
71	Phenanthrene	40.000	39.639	0.9	109	0.00
72	Anthracene	40.000	39.931	0.2	109	0.00
73	Carbazole	40.000	40.171	-0.4	109	0.00
74	Di-n-butylphthalate	40.000	40.746	-1.9	109	0.00
75 C	Fluoranthene	40.000	40.349	-0.9	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzydine	40.000	30.917	22.7	93	0.00
78	Pyrene	40.000	38.200	4.5	110	0.00
79 S	Terphenyl-d14	80.000	74.843	6.4	107	0.00
80	Butylbenzylphthalate	40.000	39.244	1.9	113	0.00
81	Benzo(a)anthracene	40.000	39.727	0.7	117	0.00
82	3,3'-Dichlorobenzidine	40.000	37.895	5.3	112	0.00
83	Chrysene	40.000	39.745	0.6	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.301	-3.3	120	0.00
85 c	Di-n-octyl phthalate	40.000	41.821	-4.6	123	0.00
86 I	Perylene-d12	20.000	20.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.844	2.9	114	0.00
88	Benzo(b)fluoranthene	40.000	40.365	-0.9	120	0.00
89	Benzo(k)fluoranthene	40.000	40.774	-1.9	119	0.00
90 C	Benzo(a)pyrene	40.000	39.987	0.0	118	0.00
91	Dibenzo(a,h)anthracene	40.000	38.792	3.0	114	0.00
92	Benzo(g,h,i)perylene	40.000	38.481	3.8	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032822\
Data File : BP009566.D
Acq On : 28 Mar 2022 14:54
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 29 03:11:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 25 06:34:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0