

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032823\
 Data File : BP014338.D
 Acq On : 28 Mar 2023 11:49
 Operator : CG/JU
 Sample : SST01042
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010625

Quant Time: Mar 29 00:59:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 29 00:54:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	151532	20.000	ng/u1	0.00
20) Naphthalene-d8	10.846	136	610406	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.675	164	399548	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	800445	20.000	ng/u1	0.00
79) Chrysene-d12	21.528	240	589730	20.000	ng/u1	0.00
88) Perylene-d12	24.051	264	635293	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	15280	3.910	ng/uL	0.00
4) Pyridine-d5	3.834	84	85419	8.597	ng/u1	0.00
7) Phenol-d5	7.181	99	124407	8.902	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.346	67	85175	9.499	ng/u1	0.00
11) 2-Chlorophenol-d4	7.552	132	93894	9.296	ng/u1	0.00
15) 4-Methylphenol-d8	8.740	113	100486	8.914	ng/u1	0.00
21) Nitrobenzene-d5	9.199	128	46353	9.407	ng/u1	0.00
24) 2-Nitrophenol-d4	9.928	143	50866	9.015	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	97203	9.459	ng/u1	0.00
31) 4-Chloroaniline-d4	10.993	131	126690	9.360	ng/u1	0.00
46) Dimethylphthalate-d6	14.081	166	313294	9.671	ng/u1	0.00
49) Acenaphthylene-d8	14.369	160	345524	9.612	ng/u1	0.00
54) 4-Nitrophenol-d4	14.898	143	35276	7.208	ng/u1	0.01
60) Fluorene-d10	15.669	176	254273	9.412	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	48321	8.204	ng/u1	0.00
73) Anthracene-d10	17.540	188	374462	9.667	ng/u1	0.00
81) Pyrene-d10	19.769	212	392316	9.904	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	316847	9.387	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.434	88	17063	3.974	ng/uL	95
5) Pyridine	3.852	79	93042	8.722	ng/u1	96
6) Benzaldehyde	7.164	77	80149	11.682	ng/u1	98
8) Phenol	7.211	94	128406	8.857	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.440	93	109647	9.412	ng/u1	98
12) 2-Chlorophenol	7.587	128	98781	9.396	ng/u1	99
13) 2-Methylphenol	8.475	108	98154	9.079	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.546	45	148864	9.733	ng/u1	99
16) Acetophenone	8.858	105	176343	9.473	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.834	70	95429	9.490	ng/u1	97
18) 4-Methylphenol	8.805	108	108010	9.097	ng/u1	94
19) Hexachloroethane	9.111	117	43384	9.445	ng/u1	98
22) Nitrobenzene	9.240	77	136663	9.654	ng/u1	98
23) Isophorone	9.769	82	244790	9.381	ng/u1	99
25) 2-Nitrophenol	9.958	139	55718	9.355	ng/u1	94
26) 2,4-Dimethylphenol	10.016	107	127665	9.623	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.252	93	150069	9.592	ng/u1	100
29) 2,4-Dichlorophenol	10.499	162	94740	9.368	ng/u1	94
30) Naphthalene	10.899	128	333727	9.815	ng/u1	99
32) 4-Chloroaniline	11.022	127	129318	9.432	ng/u1	98
33) Hexachlorobutadiene	11.175	225	79714	9.949	ng/u1	98
34) Caprolactam	11.816	113	23424	8.020	ng/u1	99
35) 4-Chloro-3-methylphenol	12.140	107	119511	9.608	ng/u1	99
36) 2-Methylnaphthalene	12.504	142	229552	9.962	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032823\
 Data File : BP014338.D
 Acq On : 28 Mar 2023 11:49
 Operator : CG/JU
 Sample : SSTD01042
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010625


Quant Time: Mar 29 00:59:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 29 00:54:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.722	142	227066	9.936	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.869	216	136603	9.589	ng/ul	98
40) Hexachlorocyclopentadiene	12.840	237	65201	7.068	ng/ul	97
41) 2,4,6-Trichlorophenol	13.116	196	81291	9.232	ng/ul	95
42) 2,4,5-Trichlorophenol	13.193	196	88883	9.471	ng/ul	97
43) 1,1'-Biphenyl	13.510	154	316329	9.800	ng/ul	99
44) 2-Chloronaphthalene	13.551	162	249257	9.766	ng/ul	98
45) 2-Nitroaniline	13.769	65	69947	8.898	ng/ul	98
47) Dimethylphthalate	14.128	163	310650	9.591	ng/ul	99
48) 2,6-Dinitrotoluene	14.257	165	56904	8.992	ng/ul	95
50) Acenaphthylene	14.398	152	373240	9.749	ng/ul	99
51) 3-Nitroaniline	14.598	138	49038	8.649	ng/ul	99
52) Acenaphthene	14.740	153	258919	9.571	ng/ul	99
53) 2,4-Dinitrophenol	14.810	184	26770	6.532	ng/ul	94
55) 4-Nitrophenol	14.910	109	32652	6.760	ng/ul	90
56) Dibenzofuran	15.075	168	371673	9.773	ng/ul	98
57) 2,4-Dinitrotoluene	15.046	165	81018	9.025	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.304	232	77716	9.214	ng/ul	98
59) Diethylphthalate	15.493	149	305521	9.492	ng/ul	99
61) Fluorene	15.728	166	291913	9.470	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.716	204	157684	9.521	ng/ul	97
63) 4-Nitroaniline	15.763	138	33214	7.532	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.810	198	47699	8.357	ng/ul#	96
67) N-Nitrosodiphenylamine	15.934	169	247877	9.953	ng/ul	99
68) 4-Bromophenyl-phenylether	16.616	248	97348	9.773	ng/ul	99
69) Hexachlorobenzene	16.734	284	111577	9.752	ng/ul	99
70) Atrazine	16.892	200	81290	8.967	ng/ul	99
71) Pentachlorophenol	17.092	266	54000	8.147	ng/ul	97
72) Phenanthrene	17.481	178	434065	9.747	ng/ul	100
74) Anthracene	17.575	178	439666	9.796	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.475	216	142277	10.098	ng/uL	99
76) Pentachlorobenzene	14.993	250	141118	10.038	ng/uL	99
77) Carbazole	17.845	167	358251	9.538	ng/ul	99
78) Di-n-butylphthalate	18.375	149	433425	9.177	ng/ul	100
80) Fluoranthene	19.439	202	467475	9.981	ng/ul	99
82) Pyrene	19.792	202	493402	10.090	ng/ul	99
83) Butylbenzylphthalate	20.651	149	154318	9.162	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.439	252	114556	8.783	ng/ul	100
85) Benzo(a)anthracene	21.510	228	400872	9.593	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.404	149	209127	9.080	ng/ul	99
87) Chrysene	21.563	228	384485	9.744	ng/ul	98
89) Di-n-octyl phthalate	22.369	149	339199	8.228	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	399367	9.308	ng/ul	97
91) Benzo(k)fluoranthene	23.322	252	397380	9.536	ng/ul	99
93) Benzo(a)pyrene	23.933	252	347674	9.414	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.698	276	505450	9.948	ng/ul	98
95) Dibenzo(a,h)anthracene	26.715	278	417099	9.805	ng/ul	98
96) Benzo(g,h,i)perylene	27.515	276	414734	10.014	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 29 00:59:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 29 00:54:55 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.974 ng/uL
RT: 3.434 min Scan# 41
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 88	Resp: 17063
Ion Ratio	Lower Upper
88 100	
43 32.5	24.4 36.6
58 67.3	57.4 86.0



#5
Pyridine
Concen: 8.722 ng/uL
RT: 3.852 min Scan# 113
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 79	Resp: 93042
Ion Ratio	Lower Upper
79 100	
52 79.6	61.4 92.2
51 38.6	28.5 42.7









#6
Benzaldehyde
Concen: 11.682 ng/ul
RT: 7.164 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 77 Resp: 80149
Ion Ratio Lower Upper
77 100
105 90.0 70.6 105.8
106 84.2 66.5 99.7



#8
Phenol
Concen: 8.857 ng/ul
RT: 7.211 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 94 Resp: 128406
Ion Ratio Lower Upper
94 100
65 32.0 25.2 37.8
66 47.8 36.2 54.4









#10
Bis(2-Chloroethyl)ether
Concen: 9.412 ng/ul
RT: 7.440 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 93 Resp: 109647
Ion Ratio Lower Upper
93 100
63 76.8 63.2 94.8
95 31.4 26.0 39.0



#12
2-Chlorophenol
Concen: 9.396 ng/ul
RT: 7.587 min Scan# 748
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 128 Resp: 98781
Ion Ratio Lower Upper
128 100
64 51.7 41.4 62.0
130 30.8 24.1 36.1









#13
2-Methylphenol
Concen: 9.079 ng/ul
RT: 8.475 min Scan# 898
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:108 Resp: 98154
Ion Ratio Lower Upper
108 100
107 92.3 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 9.733 ng/ul
RT: 8.546 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 45 Resp: 148864
Ion Ratio Lower Upper
45 100
77 16.7 12.8 19.2
79 12.2 9.6 14.4









#16
Acetophenone
Concen: 9.473 ng/ul
RT: 8.858 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:105 Resp: 176343
Ion Ratio Lower Upper
105 100
77 84.1 67.3 100.9
51 27.6 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 9.490 ng/ul
RT: 8.834 min Scan# 960
Delta R.T. -0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 70 Resp: 95429
Ion Ratio Lower Upper
70 100
42 54.8 41.1 61.7
101 9.6 7.0 10.6
130 18.1 14.6 21.8









#18
4-Methylphenol
Concen: 9.097 ng/ul
RT: 8.805 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:108 Resp: 108010
Ion Ratio Lower Upper
108 100
107 117.8 89.0 133.6



#19
Hexachloroethane
Concen: 9.445 ng/ul
RT: 9.111 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:117 Resp: 43384
Ion Ratio Lower Upper
117 100
201 113.1 90.2 135.4
199 71.1 54.2 81.2









#22
Nitrobenzene
Concen: 9.654 ng/ul
RT: 9.240 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 77 Resp: 136663
Ion Ratio Lower Upper
77 100
123 36.6 30.3 45.5
65 14.0 11.0 16.4



#23
Isophorone
Concen: 9.381 ng/ul
RT: 9.769 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 82 Resp: 244790
Ion Ratio Lower Upper
82 100
95 8.0 6.3 9.5
138 15.6 12.2 18.2









#25
2-Nitrophenol
Concen: 9.355 ng/ul
RT: 9.958 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion	Ratio	Lower	Upper
139	100		
65	68.3	49.7	74.5
109	41.0	31.1	46.7



#26
2,4-Dimethylphenol
Concen: 9.623 ng/ul
RT: 10.016 min Scan# 1161
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion	Ratio	Lower	Upper
107	100		
121	45.2	37.0	55.6
122	76.2	59.8	89.6









#27
Bis(2-Chloroethoxy)methane
Concen: 9.592 ng/ul
RT: 10.252 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 93 Resp: 150069
Ion Ratio Lower Upper
93 100
95 32.5 25.8 38.8
123 10.3 8.5 12.7



#29
2,4-Dichlorophenol
Concen: 9.368 ng/ul
RT: 10.499 min Scan# 1243
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:162 Resp: 94740
Ion Ratio Lower Upper
162 100
164 68.2 50.1 75.1
98 40.8 30.6 45.8









#30
Naphthalene
Concen: 9.815 ng/ul
RT: 10.899 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:128 Resp: 333727
Ion Ratio Lower Upper
128 100
129 11.4 8.9 13.3
127 13.1 11.0 16.6



#32
4-Chloroaniline
Concen: 9.432 ng/ul
RT: 11.022 min Scan# 1332
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:127 Resp: 129318
Ion Ratio Lower Upper
127 100
129 33.0 25.4 38.0









#33
Hexachlorobutadiene
Concen: 9.949 ng/ul
RT: 11.175 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:225 Resp: 79714
Ion Ratio Lower Upper
225 100
223 60.1 49.3 73.9
227 60.4 49.6 74.4



#34
Caprolactam
Concen: 8.020 ng/ul
RT: 11.816 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:113 Resp: 23424
Ion Ratio Lower Upper
113 100
55 164.0 131.2 196.8
56 128.6 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 9.608 ng/ul
RT: 12.140 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:107 Resp: 119511
Ion Ratio Lower Upper
107 100
144 22.9 17.3 25.9
142 67.9 54.6 81.8



#36
2-Methylnaphthalene
Concen: 9.962 ng/ul
RT: 12.504 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:142 Resp: 229552
Ion Ratio Lower Upper
142 100
141 86.9 71.6 107.4









#37
1-Methylnaphthalene
Concen: 9.936 ng/ul
RT: 12.722 min Scan# 1621
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:142 Resp: 227066
Ion Ratio Lower Upper
142 100
141 95.7 76.6 114.8
116 4.9 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.589 ng/ul
RT: 12.869 min Scan# 1646
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:216 Resp: 136603
Ion Ratio Lower Upper
216 100
214 78.0 61.0 91.6
179 21.4 16.6 24.8
108 18.5 13.8 20.6









#40
Hexachlorocyclopentadiene
Concen: 7.068 ng/ul
RT: 12.840 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:237	Resp:	65201
Ion Ratio	Lower	Upper
237	100	
235	61.7	51.0 76.4
272	11.9	10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 9.232 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:196	Resp:	81291
Ion Ratio	Lower	Upper
196	100	
198	101.1	76.6 114.8
200	30.4	25.4 38.2





#42

2,4,5-Trichlorophenol

Concen: 9.471 ng/ul

RT: 13.193 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:196 Resp: 88883

Ion Ratio Lower Upper

196 100

198 96.2 79.8 119.8

200 32.3 25.6 38.4



#43

1,1'-Biphenyl

Concen: 9.800 ng/ul

RT: 13.510 min Scan# 1755

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:154 Resp: 316329


Ion Ratio Lower Upper

154 100

153 40.9 32.4 48.6

76 12.5 10.7 16.1









#44
2-Chloronaphthalene
Concen: 9.766 ng/ul
RT: 13.551 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:162 Resp: 249257
Ion Ratio Lower Upper
162 100
164 31.3 26.1 39.1
127 43.3 33.7 50.5



#45
2-Nitroaniline
Concen: 8.898 ng/ul
RT: 13.769 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion: 65 Resp: 69947
Ion Ratio Lower Upper
65 100
92 63.6 53.7 80.5
138 88.9 71.7 107.5





#47

Dimethylphthalate

Concen: 9.591 ng/ul

RT: 14.128 min Scan# 1860

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:163 Resp: 310650

Ion Ratio Lower Upper

163 100

194 4.0 3.1 4.7

164 10.5 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 8.992 ng/ul

RT: 14.257 min Scan# 1882

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:165 Resp: 56904


Ion Ratio Lower Upper

165 100

89 70.3 52.7 79.1

121 23.1 17.9 26.9





#50

Acenaphthylene

Concen: 9.749 ng/ul

RT: 14.398 min Scan# 1906

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:152 Resp: 373240

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.4

153 13.3 10.2 15.4



#51

3-Nitroaniline

Concen: 8.649 ng/ul

RT: 14.598 min Scan# 1940

Delta R.T. 0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:138 Resp: 49038


Ion Ratio Lower Upper

138 100

108 11.1 9.9 14.9

92 133.0 107.6 161.4









#52
Acenaphthene
Concen: 9.571 ng/ul
RT: 14.740 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:153 Resp: 258919
Ion Ratio Lower Upper
153 100
152 48.6 38.2 57.4
154 88.4 70.2 105.2



#53
2,4-Dinitrophenol
Concen: 6.532 ng/ul
RT: 14.810 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:184 Resp: 26770
Ion Ratio Lower Upper
184 100
63 67.2 57.1 85.7
154 63.4 55.3 82.9









#55
4-Nitrophenol
Concen: 6.760 ng/ul
RT: 14.910 min Scan# 1993
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:109 Resp: 32652
Ion Ratio Lower Upper
109 100
139 101.0 77.4 116.2
65 137.4 97.3 145.9



#56
Dibenzofuran
Concen: 9.773 ng/ul
RT: 15.075 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:168 Resp: 371673
Ion Ratio Lower Upper
168 100
139 41.1 32.1 48.1









#57
2,4-Dinitrotoluene
Concen: 9.025 ng/ul
RT: 15.046 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:165 Resp: 81018
Ion Ratio Lower Upper
165 100
63 55.4 41.6 62.4
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.214 ng/ul
RT: 15.304 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:232 Resp: 77716
Ion Ratio Lower Upper
232 100
131 45.4 35.0 52.4
130 2.1 1.9 2.9
166 30.3 24.1 36.1









#59
Diethylphthalate
Concen: 9.492 ng/ul
RT: 15.493 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:149 Resp: 305521
Ion Ratio Lower Upper
149 100
177 22.3 18.2 27.2
150 12.1 9.8 14.8



#61
Fluorene
Concen: 9.470 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:166 Resp: 291913
Ion Ratio Lower Upper
166 100
165 99.5 79.2 118.8
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 9.521 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:204 Resp: 157684
Ion Ratio Lower Upper
204 100
206 32.8 26.9 40.3
141 61.6 51.3 76.9



#63
4-Nitroaniline
Concen: 7.532 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:138 Resp: 33214
Ion Ratio Lower Upper
138 100
92 55.8 47.8 71.6
108 101.3 82.3 123.5









#66
4,6-Dinitro-2-methylphenol
Concen: 8.357 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:198 Resp: 47699
Ion Ratio Lower Upper
198 100
182 6.6 3.5 5.3#
77 32.2 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 9.953 ng/ul
RT: 15.934 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:169 Resp: 247877
Ion Ratio Lower Upper
169 100
168 68.2 54.2 81.2
167 36.1 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 9.773 ng/ul
RT: 16.616 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:248 Resp: 97348
Ion Ratio Lower Upper
248 100
250 97.8 78.1 117.1
141 81.6 66.0 99.0



#69
Hexachlorobenzene
Concen: 9.752 ng/ul
RT: 16.734 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:284 Resp: 111577
Ion Ratio Lower Upper
284 100
142 34.3 26.6 39.8
249 32.0 25.7 38.5





#70

Atrazine

Concen: 8.967 ng/ul

RT: 16.892 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:200 Resp: 81290

Ion Ratio Lower Upper

200 100

173 25.9 20.6 30.8

215 46.1 37.2 55.8



#71

Pentachlorophenol

Concen: 8.147 ng/ul

RT: 17.092 min Scan# 2364

Delta R.T. 0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:266 Resp: 54000


Ion Ratio Lower Upper

266 100

264 61.3 51.0 76.4

268 62.1 51.4 77.2





#72

Phenanthrene

Concen: 9.747 ng/ul

RT: 17.481 min Scan# 2430

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:178 Resp: 434065

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4

176 19.1 15.3 22.9



#74

Anthracene

Concen: 9.796 ng/ul

RT: 17.575 min Scan# 2446

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:178 Resp: 439666


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 18.8 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 10.098 ng/uL

RT: 13.475 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:216 Resp: 142277

Ion Ratio Lower Upper

216 100

214 77.0 61.0 91.6

218 48.7 38.7 58.1



#76

Pentachlorobenzene

Concen: 10.038 ng/uL

RT: 14.993 min Scan# 2007

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:250 Resp: 141118


Ion Ratio Lower Upper

250 100

248 62.6 49.0 73.6

252 64.4 50.7 76.1









#77
 Carbazole
 Concen: 9.538 ng/ul
 RT: 17.845 min Scan# 2492
 Delta R.T. 0.000 min
 Lab File: BP014338.D
 Acq: 28 Mar 2023 11:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010625


Tgt Ion:167 Resp: 358251
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.2
 139 13.6 11.3 16.9



#78
 Di-n-butylphthalate
 Concen: 9.177 ng/ul
 RT: 18.375 min Scan# 2582
 Delta R.T. 0.000 min
 Lab File: BP014338.D
 Acq: 28 Mar 2023 11:49

Tgt Ion:149 Resp: 433425
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.0
 104 6.0 4.8 7.2





#80

Fluoranthene

Concen: 9.981 ng/ul

RT: 19.439 min Scan# 2763

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:202 Resp: 467475

Ion Ratio Lower Upper

202 100

101 10.6 8.2 12.2

100 7.9 6.4 9.6



#82

Pyrene

Concen: 10.090 ng/ul

RT: 19.792 min Scan# 2823

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:202 Resp: 493402


Ion Ratio Lower Upper

202 100

101 12.3 9.7 14.5

100 10.5 8.0 12.0





#83

Butylbenzylphthalate

Concen: 9.162 ng/ul

RT: 20.651 min Scan# 2969

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:149 Resp: 154318

Ion Ratio Lower Upper

149 100

91 76.1 61.8 92.8

206 21.7 17.8 26.6



#84

3,3'-Dichlorobenzidine

Concen: 8.783 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:252 Resp: 114556


Ion Ratio Lower Upper

252 100

254 64.2 51.4 77.0

126 9.9 7.8 11.6





#85

Benzo(a)anthracene

Concen: 9.593 ng/ul

RT: 21.510 min Scan# 3114

Delta R.T. 0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:228 Resp: 400872

Ion Ratio Lower Upper

228 100

229 19.0 15.6 23.4

226 26.4 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.080 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:149 Resp: 209127


Ion Ratio Lower Upper

149 100

167 27.2 21.9 32.9

279 4.0 3.7 5.5





#87

Chrysene

Concen: 9.744 ng/ul

RT: 21.563 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:228 Resp: 384485

Ion Ratio Lower Upper

228 100

226 31.0 23.8 35.8

229 19.6 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 8.228 ng/ul




RT: 22.369 min Scan# 3261


Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Tgt Ion:149 Resp: 339199





#90

Benzo(b)fluoranthene

Concen: 9.308 ng/ul

RT: 23.274 min Scan# 3414

Delta R.T. 0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:252 Resp: 399367

Ion Ratio Lower Upper

252 100

253 20.8 18.0 27.0

125 8.2 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 9.536 ng/ul

RT: 23.322 min Scan# 3423

Delta R.T. 0.000 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:252 Resp: 397380


Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 9.1 6.9 10.3





#93

Benzo(a)pyrene

Concen: 9.414 ng/ul

RT: 23.933 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:252 Resp: 347674

Ion Ratio Lower Upper

252 100

253 22.6 16.9 25.3

125 9.6 6.8 10.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.948 ng/ul

RT: 26.698 min Scan# 3997

Delta R.T. 0.006 min

Lab File: BP014338.D

Acq: 28 Mar 2023 11:49




Tgt Ion:276 Resp: 505450


Ion Ratio Lower Upper

276 100

138 19.6 14.9 22.3

277 24.9 20.3 30.5









#95
Dibenzo(a,h)anthracene
Concen: 9.805 ng/ul
RT: 26.715 min Scan# 4000
Delta R.T. 0.006 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Instrument :
BNA_P
ClientSampleId :
SSTD010625

Tgt Ion:278 Resp: 417099
Ion Ratio Lower Upper
278 100
139 12.0 10.4 15.6
279 24.5 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 10.014 ng/ul
RT: 27.515 min Scan# 4136
Delta R.T. 0.000 min
Lab File: BP014338.D
Acq: 28 Mar 2023 11:49

Tgt Ion:276 Resp: 414734
Ion Ratio Lower Upper
276 100
138 15.2 13.0 19.4
277 23.0 18.4 27.6

