

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032824\
 Data File : BP020021.D
 Acq On : 28 Mar 2024 22:23
 Operator : MA/JU
 Sample : P1856-04 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E0LZ1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/29/2024
 Supervised By :Jagrut Upadhyay 03/29/2024

Quant Time: Mar 28 23:27:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 28 14:08:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	89980	20.000	ng/u1	0.00
20) Naphthalene-d8	10.369	136	364288	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.257	164	252722	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.063	188	553378	20.000	ng/u1	-0.01
79) Chrysene-d12	21.545	240	510609	20.000	ng/u1	0.00
88) Perylene-d12	24.827	264	537034	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	6665	3.245	ng/uL	0.00
4) Pyridine-d5	3.664	84	14693m	2.369	ng/u1	0.02
7) Phenol-d5	6.811	99	60153	7.927	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	6.969	67	227848	46.084	ng/u1	0.00
11) 2-Chlorophenol-d4	7.152	132	171027	31.381	ng/u1	0.00
15) 4-Methylphenol-d8	8.310	113	116046	19.765	ng/u1	0.00
21) Nitrobenzene-d5	8.758	128	125296	44.874	ng/u1	0.00
24) 2-Nitrophenol-d4	9.469	143	126110	40.367	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.999	165	200110	35.073	ng/u1	0.00
31) 4-Chloroaniline-d4	10.540	131	31520m	4.238	ng/u1	0.02
46) Dimethylphthalate-d6	13.657	166	849743	47.957	ng/u1	0.00
49) Acenaphthylene-d8	13.940	160	935013	45.173	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	20060m	7.201	ng/u1	0.05
60) Fluorene-d10	15.269	176	726935	48.161	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.416	200	126177	38.092	ng/u1	0.00
73) Anthracene-d10	17.175	188	1215829	49.634	ng/u1	0.00
81) Pyrene-d10	19.569	212	1445732	48.258	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.604	264	1342282	52.026	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032824\
Data File : BP020021.D
Acq On : 28 Mar 2024 22:23
Operator : MA/JU
Sample : P1856-04 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E0LZ1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/29/2024
Supervised By :Jagrut Upadhyay 03/29/2024

Quant Time: Mar 28 23:27:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 28 14:08:48 2024
Response via : Initial Calibration

