

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032922\
 Data File : BP009596.D
 Acq On : 29 Mar 2022 16:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 04:08:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 04:00:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	44.182	-10.5	115	0.00
3	Pyridine	40.000	42.316	-5.8	107	0.00
4	n-Nitrosodimethylamine	40.000	42.910	-7.3	110	0.00
5 S	2-Fluorophenol	80.000	85.345	-6.7	109	0.00
6	Aniline	40.000	40.220	-0.5	103	0.00
7 S	Phenol-d6	80.000	83.994	-5.0	108	0.00
8	2-Chlorophenol	40.000	40.627	-1.6	104	0.00
9	Benzaldehyde	40.000	41.935	-4.8	109	0.00
10 C	Phenol	40.000	42.344	-5.9	109	0.00
11	bis(2-Chloroethyl)ether	40.000	40.180	-0.4	103	0.00
12	1,3-Dichlorobenzene	40.000	39.740	0.6	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.658	0.9	102	0.00
14	1,2-Dichlorobenzene	40.000	39.835	0.4	103	0.00
15	Benzyl Alcohol	40.000	38.380	4.0	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.594	-6.5	110	0.00
17	2-Methylphenol	40.000	40.229	-0.6	103	0.00
18	Hexachloroethane	40.000	39.842	0.4	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.873	0.3	103	0.00
20	3+4-Methylphenols	40.000	40.666	-1.7	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	39.824	0.4	101	0.00
23 S	Nitrobenzene-d5	80.000	80.897	-1.1	102	0.00
24	Nitrobenzene	40.000	40.597	-1.5	103	0.00
25	Isophorone	40.000	40.553	-1.4	104	0.00
26 C	2-Nitrophenol	40.000	39.372	1.6	98	0.00
27	2,4-Dimethylphenol	40.000	40.989	-2.5	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.338	-0.8	102	0.00
29 C	2,4-Dichlorophenol	40.000	40.690	-1.7	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.920	2.7	98	0.00
31	Naphthalene	40.000	39.799	0.5	102	0.00
32	Benzoic acid	40.000	39.838	0.4	98	0.00
33	4-Chloroaniline	40.000	39.899	0.3	101	0.00
34 C	Hexachlorobutadiene	40.000	36.462	8.8	92	0.00
35	Caprolactam	40.000	41.268	-3.2	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.652	0.9	102	0.00
37	2-Methylnaphthalene	40.000	39.545	1.1	102	0.00
38	1-Methylnaphthalene	40.000	39.667	0.8	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.839	5.4	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.120	2.2	107	0.00
42 S	2,4,6-Tribromophenol	80.000	70.690	11.6	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.136	2.2	101	0.00
44	2,4,5-Trichlorophenol	40.000	38.681	3.3	100	0.00
45 S	2-Fluorobiphenyl	80.000	76.920	3.8	100	0.00
46	1,1'-Biphenyl	40.000	38.995	2.5	102	0.00
47	2-Chloronaphthalene	40.000	39.113	2.2	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032922\
 Data File : BP009596.D
 Acq On : 29 Mar 2022 16:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 30 04:08:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 04:00:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.477	-3.7	107	0.00
49	Acenaphthylene	40.000	39.799	0.5	104	0.00
50	Dimethylphthalate	40.000	38.753	3.1	102	0.00
51	2,6-Dinitrotoluene	40.000	39.586	1.0	102	0.00
52 C	Acenaphthene	40.000	39.817	0.5	105	0.00
53	3-Nitroaniline	40.000	38.626	3.4	101	0.00
54 P	2,4-Dinitrophenol	40.000	35.145	12.1	94	0.00
55	Dibenzofuran	40.000	39.520	1.2	104	0.00
56 P	4-Nitrophenol	40.000	38.505	3.7	97	0.00
57	2,4-Dinitrotoluene	40.000	40.070	-0.2	103	0.00
58	Fluorene	40.000	39.425	1.4	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.223	4.4	100	0.00
60	Diethylphthalate	40.000	38.296	4.3	101	0.00
61	4-Chlorophenyl-phenylether	40.000	38.076	4.8	101	0.00
62	4-Nitroaniline	40.000	40.612	-1.5	105	0.00
63	Azobenzene	40.000	41.792	-4.5	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.493	-1.2	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.691	0.8	105	0.00
67	4-Bromophenyl-phenylether	40.000	37.398	6.5	99	0.00
68	Hexachlorobenzene	40.000	36.286	9.3	96	0.00
69	Atrazine	40.000	39.103	2.2	99	0.00
70 C	Pentachlorophenol	40.000	37.647	5.9	93	0.00
71	Phenanthrene	40.000	39.890	0.3	106	0.00
72	Anthracene	40.000	39.995	0.0	106	0.00
73	Carbazole	40.000	40.710	-1.8	108	0.00
74	Di-n-butylphthalate	40.000	38.917	2.7	102	0.00
75 C	Fluoranthene	40.000	40.060	-0.2	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	33.848	15.4	98	0.00
78	Pyrene	40.000	38.909	2.7	108	0.00
79 S	Terphenyl-d14	80.000	73.971	7.5	101	0.00
80	Butylbenzylphthalate	40.000	38.248	4.4	105	0.00
81	Benzo(a)anthracene	40.000	39.454	1.4	111	0.00
82	3,3'-Dichlorobenzidine	40.000	37.356	6.6	106	0.00
83	Chrysene	40.000	39.710	0.7	112	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.184	4.5	106	0.00
85 c	Di-n-octyl phthalate	40.000	38.561	3.6	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.783	3.0	108	0.00
88	Benzo(b)fluoranthene	40.000	39.500	1.3	112	0.00
89	Benzo(k)fluoranthene	40.000	40.352	-0.9	113	0.00
90 C	Benzo(a)pyrene	40.000	39.771	0.6	112	0.00
91	Dibenzo(a,h)anthracene	40.000	38.384	4.0	107	0.00
92	Benzo(g,h,i)perylene	40.000	38.867	2.8	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032922\
Data File : BP009596.D
Acq On : 29 Mar 2022 16:01
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 30 04:08:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 30 04:00:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0