

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
 Data File : BP005143.D
 Acq On : 01 Apr 2021 19:07
 Operator : CG/JU
 Sample : M1885-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SG2-8034-8044

Quant Time: Apr 02 01:59:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	25393	20.00	ng	# 0.00
21) Naphthalene-d8	10.493	136	98004	20.00	ng	# 0.00
39) Acenaphthene-d10	14.357	164	64289	20.00	ng	0.00
64) Phenanthrene-d10	17.122	188	139882	20.00	ng	# 0.00
76) Chrysene-d12	21.221	240	144460	20.00	ng	# 0.00
86) Perylene-d12	23.504	264	149597	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.305	112	2390	1.45	ng	0.00
7) Phenol-d6	6.881	99	14451	6.11	ng	0.00
23) Nitrobenzene-d5	8.863	82	38086	16.67	ng	0.00
42) 2,4,6-Tribromophenol	15.869	330	162	0.19	ng	0.00
45) 2-Fluorobiphenyl	12.975	172	80257	16.89	ng	0.00
79) Terphenyl-d14	19.692	244	134170	16.91	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
Data File : BP005143.D
Acq On : 01 Apr 2021 19:07
Operator : CG/JU
Sample : M1885-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044

Quant Time: Apr 02 01:59:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:23:57 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.705 min Scan# 785
Delta R.T. -0.000 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07




Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044


Tgt Ion:152 Resp: 25393
Ion Ratio Lower Upper
152 100
150 162.4 123.8 185.6
115 68.2 43.8 65.6#



#5
2-Fluorophenol
Concen: 1.45 ng
RT: 5.305 min Scan# 377
Delta R.T. -0.000 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07

Tgt Ion:112 Resp: 2390
Ion Ratio Lower Upper
112 100
64 60.0 32.8 49.2#
63 33.2 17.0 25.4#









#7
Phenol-d6
Concen: 6.11 ng
RT: 6.881 min Scan# 645
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07




Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044


Tgt Ion:	99	Resp:	14451
Ion	Ratio	Lower	Upper
99	100		
42	18.3	8.6	13.0#
71	38.3	23.4	35.2#



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.493 min Scan# 1259
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07

Tgt Ion:	136	Resp:	98004
Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.6	13.0
54	6.4	3.6	5.4#
68	7.0	3.3	4.9#









#23
Nitrobenzene-d5
Concen: 16.67 ng
RT: 8.863 min Scan# 982
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07




Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044


Tgt Ion: 82 Resp: 38086
Ion Ratio Lower Upper
82 100
128 37.0 45.6 68.4#
54 42.3 30.1 45.1



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.357 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07

Tgt Ion:164 Resp: 64289
Ion Ratio Lower Upper
164 100
162 98.0 79.4 119.2
160 41.3 36.1 54.1









#42
2,4,6-Tribromophenol
Concen: 0.19 ng
RT: 15.869 min Scan# 2173
Delta R.T. 0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07




Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044


Tgt Ion:	330	Resp:	162
Ion Ratio	Lower	Upper	
330	100		
332	84.6	77.1	115.7
141	30.2	23.4	35.2



#45
2-Fluorobiphenyl
Concen: 16.89 ng
RT: 12.975 min Scan# 1681
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07

Tgt Ion:	172	Resp:	80257
Ion Ratio	Lower	Upper	
172	100		
171	34.5	28.5	42.7
170	22.0	19.2	28.8









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.122 min Scan# 2386
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07




Instrument :
BNA_P
ClientSampleId :
SG2-8034-8044


Tgt Ion:188 Resp: 139882
Ion Ratio Lower Upper
188 100
94 8.1 6.2 9.2
80 10.6 6.2 9.2#



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.221 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BP005143.D
Acq: 01 Apr 2021 19:07

Tgt Ion:240 Resp: 144460
Ion Ratio Lower Upper
240 100
120 7.7 7.8 11.6#
236 25.6 20.8 31.2









#79
 Terphenyl-d14
 Concen: 16.91 ng
 RT: 19.692 min Scan# 2823
 Delta R.T. -0.006 min
 Lab File: BP005143.D
 Acq: 01 Apr 2021 19:07

Instrument :
 BNA_P
 ClientSampleId :
 SG2-8034-8044

Tgt Ion:244	Resp:	134170
Ion Ratio	Lower	Upper
244	100	
212	7.4	5.8 8.6
122	10.0	7.8 11.8



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 23.504 min Scan# 3471
 Delta R.T. -0.006 min
 Lab File: BP005143.D
 Acq: 01 Apr 2021 19:07

Tgt Ion:264	Resp:	149597
Ion Ratio	Lower	Upper
264	100	
260	24.4	18.8 28.2
265	21.1	17.4 26.2

