

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
 Data File : BP005144.D
 Acq On : 01 Apr 2021 19:41
 Operator : CG/JU
 Sample : M1885-08
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SG2-8033-8043

Quant Time: Apr 02 01:59:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	25435	20.00	ng	0.00
21) Naphthalene-d8	10.493	136	97794	20.00	ng	# 0.00
39) Acenaphthene-d10	14.357	164	63923	20.00	ng	0.00
64) Phenanthrene-d10	17.122	188	136902	20.00	ng	# 0.00
76) Chrysene-d12	21.216	240	143342	20.00	ng	-0.01
86) Perylene-d12	23.498	264	147608	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.299	112	28747	17.39	ng	0.00
7) Phenol-d6	6.881	99	93586	39.53	ng	0.00
23) Nitrobenzene-d5	8.863	82	117288	51.46	ng	0.00
42) 2,4,6-Tribromophenol	15.857	330	2530	3.04	ng	0.00
45) 2-Fluorobiphenyl	12.975	172	236677	50.09	ng	0.00
79) Terphenyl-d14	19.692	244	397420	50.47	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.822	163	13293	2.70	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
Data File : BP005144.D
Acq On : 01 Apr 2021 19:41
Operator : CG/JU
Sample : M1885-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SG2-8033-8043

Quant Time: Apr 02 01:59:56 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:23:57 2021
Response via : Initial Calibration

