

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
 Data File : BP005130.D
 Acq On : 01 Apr 2021 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:24:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.566	0.564	0.4	114	0.00
3	Pyridine	1.407	1.487	-5.7	107	0.00
4	n-Nitrosodimethylamine	0.503	0.510	-1.4	107	0.00
5 S	2-Fluorophenol	1.300	1.347	-3.6	112	0.00
6	Aniline	2.110	2.234	-5.9	113	0.00
7 S	Phenol-d6	1.862	1.944	-4.4	111	0.00
8	2-Chlorophenol	1.371	1.442	-5.2	112	0.00
9	Benzaldehyde	0.948	0.917	3.3	110	0.00
10 C	Phenol	1.908	1.966	-3.0	109	0.00
11	bis(2-Chloroethyl)ether	1.397	1.437	-2.9	110	0.00
12	1,3-Dichlorobenzene	1.523	1.511	0.8	106	0.00
13 C	1,4-Dichlorobenzene	1.546	1.539	0.5	107	0.00
14	1,2-Dichlorobenzene	1.483	1.505	-1.5	110	0.00
15	Benzyl Alcohol	1.436	1.509	-5.1	109	0.00
16	2,2'-oxybis(1-Chloropropane	1.025	1.170	-14.1	121	0.00
17	2-Methylphenol	1.218	1.297	-6.5	112	0.00
18	Hexachloroethane	0.589	0.584	0.8	107	0.00
19 P	n-Nitroso-di-n-propylamine	1.208	1.266	-4.8	109	0.00
20	3+4-Methylphenols	1.663	1.775	-6.7	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
22	Acetophenone	0.575	0.557	3.1	106	0.00
23 S	Nitrobenzene-d5	0.466	0.451	3.2	105	0.00
24	Nitrobenzene	0.491	0.469	4.5	106	0.00
25	Isophorone	0.860	0.873	-1.5	110	0.00
26 C	2-Nitrophenol	0.191	0.203	-6.3	114	0.00
27	2,4-Dimethylphenol	0.306	0.300	2.0	107	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.442	-1.6	108	0.00
29 C	2,4-Dichlorophenol	0.343	0.339	1.2	106	0.00
30	1,2,4-Trichlorobenzene	0.385	0.366	4.9	104	0.00
31	Naphthalene	1.130	1.109	1.9	108	0.00
32	Benzoic acid	0.227	0.258	-13.7	118	0.00
33	4-Chloroaniline	0.457	0.455	0.4	109	0.00
34 C	Hexachlorobutadiene	0.248	0.233	6.0	102	0.00
35	Caprolactam	0.112	0.123	-9.8	117	0.00
36 C	4-Chloro-3-methylphenol	0.414	0.409	1.2	106	0.00
37	2-Methylnaphthalene	0.816	0.788	3.4	105	0.00
38	1-Methylnaphthalene	0.765	0.731	4.4	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.651	0.609	6.5	99	0.00
41 P	Hexachlorocyclopentadiene	0.354	0.348	1.7	100	0.00
42 S	2,4,6-Tribromophenol	0.261	0.253	3.1	103	0.00
43 C	2,4,6-Trichlorophenol	0.450	0.451	-0.2	104	0.00
44	2,4,5-Trichlorophenol	0.488	0.478	2.0	103	0.00
45 S	2-Fluorobiphenyl	1.478	1.403	5.1	101	0.00
46	1,1'-Biphenyl	1.525	1.469	3.7	103	0.00
47	2-Chloronaphthalene	1.242	1.190	4.2	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
 Data File : BP005130.D
 Acq On : 01 Apr 2021 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:24:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.396	0.406	-2.5	105	0.00
49	Acenaphthylene	1.931	1.870	3.2	103	0.00
50	Dimethylphthalate	1.540	1.462	5.1	103	0.00
51	2,6-Dinitrotoluene	0.363	0.361	0.6	105	0.00
52 C	Acenaphthene	1.188	1.141	4.0	103	0.00
53	3-Nitroaniline	0.334	0.347	-3.9	105	0.00
54 P	2,4-Dinitrophenol	0.220	0.237	-7.7	111	0.00
55	Dibenzofuran	1.948	1.833	5.9	101	0.00
56 P	4-Nitrophenol	0.274	0.299	-9.1	110	0.00
57	2,4-Dinitrotoluene	0.479	0.483	-0.8	104	0.00
58	Fluorene	1.573	1.498	4.8	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.450	0.433	3.8	100	0.00
60	Diethylphthalate	1.507	1.462	3.0	102	0.00
61	4-Chlorophenyl-phenylether	0.835	0.762	8.7	100	0.00
62	4-Nitroaniline	0.343	0.369	-7.6	109	0.00
63	Azobenzene	1.722	1.605	6.8	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.146	0.162	-11.0	104	0.00
66 c	n-Nitrosodiphenylamine	0.643	0.655	-1.9	102	0.00
67	4-Bromophenyl-phenylether	0.232	0.236	-1.7	100	0.00
68	Hexachlorobenzene	0.261	0.255	2.3	97	0.00
69	Atrazine	0.214	0.222	-3.7	98	0.00
70 C	Pentachlorophenol	0.172	0.180	-4.7	104	0.00
71	Phenanthrene	1.159	1.148	0.9	98	0.00
72	Anthracene	1.133	1.149	-1.4	101	0.00
73	Carbazole	1.064	1.085	-2.0	101	0.00
74	Di-n-butylphthalate	1.171	1.253	-7.0	104	0.00
75 C	Fluoranthene	1.364	1.347	1.2	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.439	0.502	-14.4	98	0.00
78	Pyrene	1.374	1.385	-0.8	99	0.00
79 S	Terphenyl-d14	1.099	1.087	1.1	97	0.00
80	Butylbenzylphthalate	0.510	0.576	-12.9	105	0.00
81	Benzo(a)anthracene	1.359	1.356	0.2	99	0.00
82	3,3'-Dichlorobenzidine	0.457	0.491	-7.4	103	0.00
83	Chrysene	1.334	1.320	1.0	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.748	0.826	-10.4	103	0.00
85 c	Di-n-octyl phthalate	1.261	1.420	-12.6	107	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	1.500	1.481	1.3	100	0.00
88	Benzo(b)fluoranthene	1.420	1.342	5.5	96	0.00
89	Benzo(k)fluoranthene	1.361	1.309	3.8	96	0.00
90 C	Benzo(a)pyrene	1.273	1.238	2.7	98	0.00
91	Dibenzo(a,h)anthracene	1.297	1.279	1.4	100	0.00
92	Benzo(g,h,i)perylene	1.251	1.248	0.2	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040121\
Data File : BP005130.D
Acq On : 01 Apr 2021 11:40
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 01 12:24:29 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:23:57 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0