

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040122\
 Data File : BP009652.D
 Acq On : 01 Apr 2022 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 22:42:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 04:36:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.01
2	1,4-Dioxane	40.000	43.118	-7.8	119	0.00
3	Pyridine	40.000	44.539	-11.3	120	0.00
4	n-Nitrosodimethylamine	40.000	44.690	-11.7	122	0.00
5 S	2-Fluorophenol	80.000	87.752	-9.7	119	0.00
6	Aniline	40.000	42.324	-5.8	115	0.00
7 S	Phenol-d6	80.000	85.667	-7.1	117	0.01
8	2-Chlorophenol	40.000	41.505	-3.8	113	0.00
9	Benzaldehyde	40.000	45.669	-14.2	125	0.00
10 C	Phenol	40.000	42.900	-7.2	117	0.01
11	bis(2-Chloroethyl)ether	40.000	42.611	-6.5	117	0.00
12	1,3-Dichlorobenzene	40.000	40.912	-2.3	112	0.01
13 C	1,4-Dichlorobenzene	40.000	40.798	-2.0	112	0.01
14	1,2-Dichlorobenzene	40.000	40.471	-1.2	111	0.01
15	Benzyl Alcohol	40.000	40.651	-1.6	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	45.331	-13.3	124	0.01
17	2-Methylphenol	40.000	41.240	-3.1	113	0.00
18	Hexachloroethane	40.000	41.265	-3.2	113	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.294	-0.7	111	0.00
20	3+4-Methylphenols	40.000	40.939	-2.3	111	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.01
22	Acetophenone	40.000	41.741	-4.4	111	0.00
23 S	Nitrobenzene-d5	80.000	86.009	-7.5	113	0.00
24	Nitrobenzene	40.000	43.088	-7.7	114	0.00
25	Isophorone	40.000	42.009	-5.0	112	0.01
26 C	2-Nitrophenol	40.000	42.177	-5.4	110	0.00
27	2,4-Dimethylphenol	40.000	42.636	-6.6	112	0.01
28	bis(2-Chloroethoxy)methane	40.000	42.685	-6.7	113	0.01
29 C	2,4-Dichlorophenol	40.000	41.746	-4.4	109	0.01
30	1,2,4-Trichlorobenzene	40.000	40.543	-1.4	107	0.01
31	Naphthalene	40.000	40.909	-2.3	109	0.01
32	Benzoic acid	40.000	43.716	-9.3	112	0.01
33	4-Chloroaniline	40.000	40.979	-2.4	108	0.01
34 C	Hexachlorobutadiene	40.000	38.403	4.0	101	0.01
35	Caprolactam	40.000	40.090	-0.2	110	0.01
36 C	4-Chloro-3-methylphenol	40.000	39.419	1.5	106	0.00
37	2-Methylnaphthalene	40.000	40.170	-0.4	108	0.01
38	1-Methylnaphthalene	40.000	40.023	-0.1	107	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	40.949	-2.4	103	0.01
41 P	Hexachlorocyclopentadiene	40.000	40.319	-0.8	107	0.02
42 S	2,4,6-Tribromophenol	80.000	70.796	11.5	90	0.01
43 C	2,4,6-Trichlorophenol	40.000	41.616	-4.0	104	0.01
44	2,4,5-Trichlorophenol	40.000	40.063	-0.2	101	0.00
45 S	2-Fluorobiphenyl	80.000	81.690	-2.1	103	0.01
46	1,1'-Biphenyl	40.000	41.809	-4.5	106	0.01
47	2-Chloronaphthalene	40.000	41.828	-4.6	106	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040122\
 Data File : BP009652.D
 Acq On : 01 Apr 2022 11:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 22:42:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 04:36:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.693	-11.7	112	0.00
49	Acenaphthylene	40.000	41.610	-4.0	106	0.01
50	Dimethylphthalate	40.000	40.091	-0.2	103	0.01
51	2,6-Dinitrotoluene	40.000	40.734	-1.8	102	0.00
52 C	Acenaphthene	40.000	41.240	-3.1	105	0.01
53	3-Nitroaniline	40.000	42.170	-5.4	106	0.00
54 P	2,4-Dinitrophenol	40.000	36.549	8.6	95	0.00
55	Dibenzofuran	40.000	40.752	-1.9	104	0.01
56 P	4-Nitrophenol	40.000	40.272	-0.7	99	0.00
57	2,4-Dinitrotoluene	40.000	40.052	-0.1	100	0.01
58	Fluorene	40.000	40.276	-0.7	104	0.02
59	2,3,4,6-Tetrachlorophenol	40.000	38.273	4.3	97	0.01
60	Diethylphthalate	40.000	39.641	0.9	102	0.01
61	4-Chlorophenyl-phenylether	40.000	39.240	1.9	101	0.01
62	4-Nitroaniline	40.000	42.200	-5.5	106	0.01
63	Azobenzene	40.000	43.004	-7.5	109	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.01
65	4,6-Dinitro-2-methylphenol	40.000	42.303	-5.8	97	0.01
66 c	n-Nitrosodiphenylamine	40.000	42.201	-5.5	103	0.01
67	4-Bromophenyl-phenylether	40.000	39.347	1.6	97	0.01
68	Hexachlorobenzene	40.000	38.025	4.9	93	0.01
69	Atrazine	40.000	40.619	-1.5	96	0.01
70 C	Pentachlorophenol	40.000	37.255	6.9	86	0.01
71	Phenanthrene	40.000	41.282	-3.2	102	0.02
72	Anthracene	40.000	41.839	-4.6	103	0.01
73	Carbazole	40.000	42.011	-5.0	103	0.01
74	Di-n-butylphthalate	40.000	41.440	-3.6	101	0.01
75 C	Fluoranthene	40.000	41.036	-2.6	102	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.02
77	Benzidine	40.000	42.839	-7.1	113	0.01
78	Pyrene	40.000	40.423	-1.1	102	0.01
79 S	Terphenyl-d14	80.000	76.636	4.2	96	0.01
80	Butylbenzylphthalate	40.000	41.166	-2.9	104	0.01
81	Benzo(a)anthracene	40.000	40.648	-1.6	105	0.01
82	3,3'-Dichlorobenzidine	40.000	40.699	-1.7	105	0.01
83	Chrysene	40.000	41.085	-2.7	106	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	41.600	-4.0	106	0.01
85 c	Di-n-octyl phthalate	40.000	42.663	-6.7	110	0.02
86 I	Perylene-d12	20.000	20.000	0.0	101	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.334	-0.8	105	0.04
88	Benzo(b)fluoranthene	40.000	40.473	-1.2	106	0.02
89	Benzo(k)fluoranthene	40.000	42.377	-5.9	110	0.02
90 C	Benzo(a)pyrene	40.000	41.287	-3.2	108	0.02
91	Dibenzo(a,h)anthracene	40.000	40.381	-1.0	105	0.04
92	Benzo(g,h,i)perylene	40.000	40.416	-1.0	105	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040122\
Data File : BP009652.D
Acq On : 01 Apr 2022 11:40
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 01 22:42:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 31 04:36:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0