


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040125\
Data File : BP024204.D
Acq On : 01 Apr 2025 09:16
Operator : RC/JU
Sample : DFTPP620
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP621

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030425.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Apr 01 12:35:32 2025



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.2	150056	PASS
68	69	0.00	2	1.7	3122	PASS
69	198	0.00	100	35.2	181231	PASS
70	69	0.00	2	0.4	752	PASS
127	198	10	80	46.7	240492	PASS
197	198	0.00	2	0.3	1456	PASS
198	198	100	100	100.0	514426	PASS
199	198	5	9	6.9	35276	PASS
275	198	10	60	27.5	141373	PASS
365	198	1	100	3.6	18623	PASS
441	198	0.01	100	13.7	70512	PASS
442	198	50	100	89.8	462120	PASS
443	442	15	24	19.6	90741	PASS

m/z	Abundance
35.90	132.0
37.00	74.0
37.90	82.0
39.00	392.0
39.90	2765.0
41.00	594.0
42.00	192.0
43.10	1008.0
44.00	4276.0
44.90	320.0
46.10	62.0
47.00	65.0
47.70	9.0
49.00	58.0
50.10	179.0
50.90	237.0
52.00	126.0
52.90	239.0
54.00	174.0
54.90	742.0
56.00	85.0
57.00	564.0
58.00	192.0
58.90	352.0
59.80	44.0
60.90	108.0
61.90	62.0
62.90	137.0
64.00	49.0
65.00	137.0
66.10	128.0
66.90	336.0
68.00	196.0
69.00	678.0
69.90	335.0
70.90	379.0
71.90	85.0
73.00	7547.0
74.00	875.0
75.00	515.0
76.00	245.0
77.10	385.0
78.00	298.0
79.00	264.0
80.20	136.0
81.10	474.0
82.00	464.0
83.00	366.0
83.90	275.0
85.00	377.0
85.80	85.0
87.00	240.0
87.90	111.0
88.70	105.0
90.10	60.0
91.00	455.0
92.00	174.0
93.20	315.0
93.80	136.0
94.90	372.0
95.90	710.0
96.90	411.0
98.00	179.0
99.00	129.0
100.30	23.0
101.10	64.0
101.80	87.0
102.80	282.0
103.90	117.0
104.90	197.0
105.90	120.0
106.80	62.0
107.20	99.0
108.00	85.0
109.00	206.0
110.00	99.0
111.10	225.0
112.00	53.0
113.00	28.0
114.20	11.0
115.00	270.0
116.00	83.0
117.00	365.0
118.00	171.0
118.90	363.0
119.80	94.0
121.10	218.0
121.80	65.0
123.10	98.0
123.90	74.0
125.10	115.0
125.80	188.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

127.10	105.0
128.00	207.0
128.90	278.0
129.90	149.0
130.90	242.0
131.90	88.0
133.00	607.0
133.90	168.0
135.10	295.0
135.90	65.0
137.00	180.0
137.80	55.0
138.90	201.0
140.00	41.0
140.90	172.0
141.80	83.0
143.00	108.0
144.00	46.0
144.80	234.0
145.90	54.0
147.00	3369.0
148.00	627.0
149.00	466.0
149.90	71.0
150.80	101.0
151.80	112.0
153.10	136.0
153.80	95.0
155.00	111.0
155.90	38.0
156.20	45.0
157.00	72.0
157.90	39.0
159.20	45.0
160.00	85.0
161.10	250.0
162.00	46.0
162.90	160.0
164.00	94.0
164.90	350.0
166.00	149.0
167.00	218.0
168.10	66.0
168.90	60.0
169.90	15.0
171.00	115.0
171.80	14.0
172.10	24.0
173.10	62.0
174.00	16.0
174.80	64.0
175.30	65.0
176.00	120.0
177.10	230.0
178.00	53.0
178.90	146.0
179.80	29.0
180.30	20.0
181.00	64.0
182.10	34.0
182.90	79.0
183.50	16.0
184.20	23.0
185.20	27.0
186.00	64.0
187.00	47.0
187.80	22.0
188.90	178.0
189.80	42.0
190.90	675.0
191.90	278.0
192.90	579.0
194.00	153.0
195.00	89.0
196.10	141.0
197.00	81.0
197.90	70.0
198.90	36.0
199.90	45.0
200.90	48.0
201.90	16.0
203.00	96.0
203.90	113.0
205.10	267.0
206.10	222.0
207.00	4515.0
208.00	1431.0
209.00	881.0
209.90	192.0
211.00	97.0
211.80	54.0
212.90	12.0
213.90	27.0
214.30	10.0
215.20	52.0
215.90	20.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

217.30	25.0
217.90	47.0
219.10	54.0
219.90	44.0
221.00	2016.0
222.00	565.0
223.00	443.0
224.00	145.0
225.10	62.0
225.80	62.0
227.00	22.0
227.90	10.0
228.90	12.0
229.40	11.0
230.00	8.0
231.40	17.0
232.00	16.0
232.50	18.0
233.30	8.0
233.70	13.0
234.40	18.0
235.10	23.0
235.80	18.0
236.80	9.0
237.80	15.0
239.10	124.0
240.00	11.0
240.70	18.0
241.10	14.0
242.00	32.0
242.70	16.0
244.30	25.0
244.80	11.0
246.00	20.0
246.90	22.0
247.60	4.0
248.20	21.0
248.90	255.0
250.00	94.0
250.80	128.0
251.90	88.0
253.10	23.0
254.10	20.0
254.80	14.0
256.10	24.0
256.90	18.0
257.50	3.0
258.10	17.0
258.50	8.0
259.40	10.0
260.10	11.0
260.70	3.0
261.20	5.0
262.50	14.0
263.00	15.0
263.70	20.0
265.00	404.0
266.00	98.0
267.00	498.0
267.90	152.0
268.90	29.0
270.10	12.0
270.60	24.0
271.10	17.0
271.90	3.0
272.80	19.0
273.40	17.0
274.30	16.0
275.00	6.0
275.80	16.0
276.70	6.0
279.00	49.0
279.50	15.0
281.00	3585.0
282.00	835.0
282.90	639.0
283.90	85.0
285.00	21.0
286.10	1.0
286.80	21.0
287.80	2.0
288.90	11.0
289.80	13.0
290.70	12.0
291.40	2.0
292.20	18.0
292.90	53.0
293.70	14.0
295.00	227.0
295.80	135.0
297.20	67.0
298.20	19.0
298.80	16.0
300.20	16.0
301.60	16.0
302.60	2.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

303.10	19.0
304.00	15.0
304.90	2.0
305.90	2.0
306.50	2.0
306.90	4.0
307.60	10.0
308.20	7.0
308.90	16.0
310.10	19.0
311.00	25.0
311.70	2.0
312.80	8.0
313.40	11.0
314.50	4.0
315.20	2.0
315.70	11.0
316.60	6.0
317.70	7.0
318.30	1.0
319.00	11.0
319.70	2.0
320.50	24.0
321.30	2.0
322.00	1.0
322.30	1.0
323.00	66.0
324.90	340.0
325.90	65.0
326.90	359.0
328.00	128.0
328.90	64.0
330.00	8.0
331.30	13.0
331.80	5.0
332.40	1.0
333.30	3.0
334.00	2.0
335.00	6.0
335.80	2.0
336.60	11.0
337.30	14.0
338.00	5.0
339.10	20.0
339.90	7.0
340.90	1115.0
341.90	349.0
342.90	273.0
343.90	74.0
344.60	7.0
345.20	18.0
345.80	22.0
346.50	1.0
347.50	2.0
348.30	21.0
349.10	3.0
350.20	4.0
350.90	14.0
351.70	2.0
352.20	4.0
352.90	13.0
353.70	3.0
355.00	1506.0
356.00	512.0
356.90	234.0
357.80	87.0
358.80	28.0
359.30	5.0
360.00	2.0
360.80	9.0
361.60	3.0
362.10	2.0
362.40	2.0
363.20	2.0
363.80	16.0
364.70	22.0
366.00	2.0
366.70	4.0
367.10	3.0
367.80	3.0
369.00	86.0
370.00	25.0
371.10	9.0
372.10	15.0
373.00	8.0
373.90	5.0
374.40	3.0
375.40	2.0
375.70	2.0
376.60	3.0
377.50	5.0
378.70	4.0
379.60	17.0
380.80	20.0
381.70	16.0
382.30	3.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

383.00	66.0
383.50	15.0
384.00	14.0
384.90	127.0
386.20	22.0
386.90	28.0
387.70	14.0
388.60	17.0
389.00	8.0
390.00	2.0
390.80	2.0
391.40	2.0
392.30	2.0
393.00	3.0
393.40	4.0
394.80	2.0
395.30	12.0
396.10	11.0
396.90	19.0
397.60	1.0
399.00	96.0
401.00	243.0
402.00	65.0
402.80	73.0
403.90	19.0
404.40	12.0
405.60	6.0
406.20	13.0
407.70	12.0
409.00	14.0
409.40	4.0
409.90	1.0
410.60	2.0
411.20	2.0
411.90	5.0
413.00	31.0
413.50	3.0
414.70	271.0
415.80	139.0
416.80	85.0
418.00	19.0
418.90	31.0
420.20	7.0
421.20	3.0
421.70	2.0
422.50	1.0
423.00	2.0
423.50	13.0
424.50	2.0
425.20	2.0
426.10	2.0
426.70	1.0
427.60	18.0
429.00	726.0
429.90	229.0
431.00	161.0
432.00	121.0
433.10	20.0
434.10	18.0
434.90	1.0
435.40	2.0
435.90	2.0
436.80	3.0
437.40	15.0
438.10	0.0
438.50	3.0
439.10	3.0
439.90	16.0
440.60	12.0
441.20	19.0
441.90	3.0
442.60	2.0
443.30	3.0
443.80	3.0
444.10	2.0
445.00	42.0
445.90	11.0
446.60	3.0
447.10	10.0
448.10	11.0
448.60	21.0
449.60	3.0
450.20	2.0
451.50	3.0
452.80	3.0
453.50	4.0
454.20	3.0
454.90	6.0
456.00	11.0
456.80	17.0
458.00	26.0
458.90	91.0
460.80	200.0
461.90	74.0
462.90	101.0
463.70	28.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

465.00	18.0
465.50	4.0
466.30	2.0
467.00	8.0
467.80	2.0
468.60	3.0
469.30	2.0
470.00	3.0
470.90	21.0
471.80	12.0
472.40	10.0
473.40	5.0
474.10	18.0
474.90	163.0
475.80	81.0
476.80	11.0
477.50	31.0
479.30	3.0
480.00	22.0
481.20	2.0
482.60	3.0
483.10	1.0
483.60	11.0
484.30	3.0
485.30	3.0
485.90	4.0
487.00	13.0
487.70	8.0
488.90	23.0
489.40	7.0
489.90	12.0
490.50	2.0
491.30	16.0
492.70	3.0
493.20	3.0
493.80	6.0
494.70	5.0
495.50	4.0
496.00	12.0
496.80	6.0
497.80	15.0
499.00	2.0
500.20	4.0
501.10	11.0
503.00	350.0
504.10	193.0
505.00	182.0
506.10	47.0
506.80	4.0
507.20	16.0
507.90	18.0
508.80	3.0
509.70	4.0
510.90	3.0
511.30	2.0
511.70	2.0
512.70	5.0
513.60	2.0
514.40	10.0
515.10	3.0
515.90	14.0
516.60	13.0
517.50	15.0
518.10	13.0
518.70	13.0
519.80	17.0
520.50	12.0
521.10	17.0
522.00	6.0
522.80	6.0
524.00	3.0
524.60	2.0
525.30	3.0
525.90	2.0
526.80	2.0
527.60	2.0
528.20	3.0
529.00	2.0
529.90	4.0
531.00	18.0
531.90	4.0
532.90	49.0
533.60	17.0
534.90	25.0
535.80	13.0
537.00	13.0
538.70	2.0
539.40	5.0
540.50	16.0
541.10	3.0
542.10	2.0
542.60	2.0
543.60	2.0
544.50	4.0
545.20	2.0
545.80	3.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

547.10	12.0
548.00	3.0
549.10	8.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

m/z	Abundance
36.00	242.0
37.00	660.0
38.00	1802.0
39.10	9566.0
39.90	2903.0
41.00	1173.0
41.80	172.0
43.00	1006.0
44.00	4161.0
45.00	505.0
45.80	120.0
46.90	99.0
48.10	85.0
48.90	824.0
50.00	34968.0
51.00	132224.0
52.00	6891.0
52.90	461.0
54.10	294.0
55.00	1611.0
56.00	4289.0
57.00	10166.0
57.90	537.0
59.00	410.0
59.90	195.0
61.00	2096.0
62.00	2166.0
63.00	6057.0
64.00	837.0
65.00	3251.0
65.90	437.0
66.90	858.0
68.00	2859.0
69.00	154304.0
70.10	1071.0
71.00	549.0
72.00	247.0
73.00	6383.0
74.00	15948.0
75.00	24632.0
76.00	8471.0
77.00	187968.0
78.00	13892.0
79.00	12158.0
80.00	9072.0
81.00	12520.0
82.00	3689.0
82.90	3486.0
84.00	829.0
84.90	2297.0
86.00	3845.0
87.00	1978.0
87.90	708.0
88.90	506.0
89.90	96.0
91.00	2611.0
92.00	3865.0
93.00	19800.0
93.90	1716.0
95.00	908.0
96.00	1885.0
97.00	885.0
98.00	14815.0
99.00	16400.0
100.00	1368.0
100.90	6799.0
101.90	553.0
102.90	2440.0
104.00	4563.0
105.00	4537.0
106.00	1150.0
107.00	52632.0
108.00	8592.0
109.00	1736.0
110.00	102032.0
111.00	15099.0
111.90	1978.0
112.90	793.0
114.10	216.0
114.90	522.0
115.90	2801.0
117.00	41232.0
118.00	3783.0
118.90	847.0
119.90	708.0
120.90	458.0
122.00	3037.0
122.90	5677.0
123.90	2667.0
125.00	2150.0
126.10	956.0
127.00	189120.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

128.00	14686.0
128.90	68608.0
130.00	6665.0
131.00	1481.0
132.00	756.0
132.80	1200.0
133.90	1762.0
134.90	5292.0
135.90	2491.0
137.00	4492.0
137.90	700.0
139.00	465.0
139.90	690.0
140.90	8762.0
142.00	2748.0
142.90	2045.0
143.90	599.0
144.90	627.0
146.00	1692.0
147.00	6535.0
147.90	10843.0
148.90	3017.0
149.90	870.0
151.10	1277.0
151.80	789.0
152.90	2718.0
153.90	2176.0
155.00	5283.0
156.00	7552.0
157.00	1696.0
157.90	1792.0
158.90	1354.0
159.90	2678.0
161.00	4120.0
161.90	1254.0
162.90	611.0
163.80	625.0
164.90	3040.0
166.00	2385.0
167.00	18120.0
167.90	16912.0
168.90	2156.0
169.90	699.0
170.90	800.0
171.90	1904.0
172.90	2033.0
173.90	3443.0
175.00	6885.0
176.00	2159.0
176.90	3132.0
178.00	1192.0
178.90	12848.0
180.00	10594.0
181.00	4678.0
182.00	618.0
182.90	463.0
184.00	1061.0
185.00	6106.0
186.00	47776.0
187.00	14369.0
188.00	1665.0
188.90	2893.0
189.90	630.0
191.00	1797.0
191.90	4301.0
193.00	5498.0
193.90	1167.0
195.00	518.0
196.00	11419.0
196.90	1823.0
197.90	373312.0
198.90	26656.0
199.90	1904.0
201.40	1604.0
203.00	2251.0
204.00	12863.0
205.00	21736.0
206.00	92536.0
207.00	17536.0
208.00	4006.0
208.90	1872.0
211.00	3415.0
212.90	243.0
213.80	207.0
214.90	1131.0
215.90	1944.0
216.90	24608.0
217.90	3041.0
218.90	360.0
221.00	17720.0
221.90	3697.0
222.90	5395.0
224.00	52032.0
225.00	14280.0
226.00	1667.0
227.00	20064.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

228.00	2919.0
229.00	4536.0
230.00	504.0
231.00	1857.0
232.00	283.0
233.00	406.0
234.00	1336.0
234.90	1803.0
236.00	1105.0
236.90	2235.0
237.90	309.0
238.90	921.0
239.90	897.0
240.90	964.0
242.00	2720.0
242.90	2666.0
244.00	42224.0
245.00	6068.0
245.90	8120.0
246.90	1973.0
248.00	285.0
248.90	1563.0
250.00	348.0
250.90	346.0
251.90	408.0
253.00	979.0
254.00	1432.0
255.00	209600.0
256.00	32272.0
256.90	2817.0
257.90	12019.0
259.00	1911.0
260.00	550.0
261.00	294.0
261.90	46.0
263.00	184.0
263.90	321.0
264.90	4649.0
265.90	831.0
266.90	694.0
267.90	225.0
269.00	123.0
269.90	332.0
271.00	902.0
271.90	454.0
273.00	5908.0
274.00	16608.0
274.90	97032.0
276.00	12573.0
276.90	8031.0
277.90	1545.0
278.90	348.0
281.00	3068.0
282.00	1050.0
283.00	1481.0
284.00	882.0
285.00	1383.0
285.90	412.0
286.90	39.0
288.00	150.0
288.90	320.0
289.90	311.0
290.90	270.0
291.80	382.0
293.00	1827.0
293.90	475.0
294.90	685.0
295.90	25416.0
297.00	3746.0
297.90	398.0
298.80	117.0
300.00	55.0
300.80	333.0
302.00	385.0
303.00	3042.0
303.90	811.0
305.00	169.0
306.10	13.0
307.10	51.0
308.00	283.0
308.90	239.0
310.00	468.0
310.90	130.0
311.80	129.0
313.00	395.0
314.00	1595.0
315.00	3092.0
315.90	1596.0
317.00	382.0
317.80	16.0
318.90	56.0
319.90	105.0
320.90	780.0
322.10	361.0
323.00	8916.0
324.00	1822.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

324.90	366.0
326.00	282.0
326.90	2088.0
327.90	895.0
328.80	238.0
329.90	53.0
331.90	579.0
333.00	1100.0
334.00	6170.0
334.90	1548.0
335.80	151.0
336.90	19.0
337.90	10.0
338.90	214.0
339.90	228.0
340.90	1887.0
341.80	483.0
342.90	304.0
343.80	88.0
345.90	1875.0
346.90	454.0
347.80	104.0
348.90	12.0
350.00	97.0
350.80	101.0
352.00	2424.0
353.00	1686.0
354.00	2701.0
355.00	1370.0
356.00	352.0
357.00	211.0
357.90	85.0
358.90	200.0
359.90	78.0
360.50	38.0
361.20	41.0
362.40	13.0
364.90	11950.0
365.90	2070.0
366.80	83.0
367.20	72.0
368.20	18.0
368.70	17.0
370.00	275.0
371.00	689.0
372.00	4765.0
373.00	1294.0
374.00	165.0
374.60	16.0
375.50	2.0
376.90	156.0
377.90	29.0
378.60	16.0
379.20	3.0
379.60	3.0
380.10	5.0
380.90	21.0
381.20	21.0
383.00	1033.0
384.00	409.0
384.90	204.0
385.70	74.0
386.20	57.0
386.90	14.0
388.00	11.0
388.90	46.0
390.00	708.0
390.90	578.0
391.90	339.0
393.10	60.0
393.90	7.0
394.70	24.0
395.60	16.0
396.20	6.0
396.90	93.0
397.90	8.0
398.90	146.0
399.80	35.0
400.80	607.0
401.90	1779.0
402.90	2833.0
404.00	1091.0
404.90	242.0
405.90	70.0
406.40	18.0
407.00	11.0
407.70	4.0
408.20	11.0
408.90	14.0
409.90	91.0
410.80	15.0
411.40	3.0
412.90	18.0
413.50	9.0
415.00	470.0
416.00	141.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

416.80	101.0
417.60	14.0
418.00	14.0
418.70	12.0
419.50	32.0
421.00	2244.0
422.00	2465.0
423.00	15934.0
424.00	5338.0
425.00	888.0
426.10	74.0
426.90	22.0
428.10	5.0
429.00	361.0
429.90	161.0
431.10	78.0
432.30	27.0
432.70	22.0
433.20	9.0
434.00	3.0
434.80	2.0
435.50	10.0
435.90	8.0
436.50	7.0
438.00	210.0
438.80	156.0
439.90	142.0
441.00	43184.0
442.00	276544.0
443.00	56840.0
444.00	5667.0
445.00	300.0
445.80	18.0
446.60	6.0
447.30	5.0
448.70	12.0
449.30	4.0
450.50	4.0
451.30	10.0
451.80	5.0
452.90	2.0
453.30	1.0
453.90	16.0
454.90	9.0
456.20	17.0
457.40	2.0
458.20	52.0
459.10	39.0
460.00	40.0
460.70	164.0
461.80	75.0
461.90	74.0
462.90	64.0
463.70	15.0
464.20	13.0
464.90	13.0
465.70	10.0
466.50	2.0
467.30	3.0
468.70	2.0
469.10	2.0
469.60	2.0
470.50	4.0
471.80	4.0
472.60	5.0
473.00	5.0
473.50	6.0
474.90	87.0
476.00	43.0
477.00	36.0
478.00	9.0
478.70	11.0
479.50	7.0
480.20	8.0
481.20	2.0
481.90	7.0
482.60	3.0
483.60	8.0
484.60	3.0
485.10	2.0
485.60	2.0
486.10	2.0
486.60	3.0
487.00	7.0
487.70	14.0
488.20	5.0
489.10	27.0
490.10	18.0
490.80	4.0
492.00	1.0
492.60	2.0
493.00	2.0
493.40	2.0
494.20	4.0
495.00	1.0
495.40	2.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

496.00	2.0
497.20	4.0
497.70	1.0
498.50	2.0
498.80	2.0
499.70	8.0
500.30	4.0
501.00	1.0
501.60	1.0
502.10	4.0
503.00	161.0
504.10	69.0
505.10	192.0
505.80	21.0
506.70	5.0
507.20	15.0
507.70	2.0
508.50	2.0
509.10	2.0
509.90	3.0
510.10	3.0
510.70	3.0
511.60	3.0
512.30	7.0
513.00	4.0
513.80	17.0
514.70	6.0
515.20	8.0
516.60	11.0
517.00	12.0
517.90	8.0
518.90	15.0
519.50	4.0
520.20	8.0
521.40	9.0
522.40	3.0
523.10	3.0
524.10	5.0
524.60	1.0
525.80	4.0
526.90	3.0
527.30	2.0
527.90	2.0
528.40	5.0
530.50	6.0
531.50	2.0
532.00	1.0
532.70	56.0
533.30	13.0
534.00	21.0
534.80	22.0
535.80	15.0
536.80	2.0
538.30	3.0
539.10	2.0
539.80	7.0
540.80	4.0
541.20	2.0
541.80	5.0
542.40	2.0
542.80	2.0
543.80	4.0
544.70	3.0
545.40	2.0
545.90	2.0
546.40	3.0
546.80	2.0
547.70	4.0
549.00	44.0
549.60	10.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

m/z	Abundance
35.10	23.0
35.90	293.0
37.00	890.0
38.00	1970.0
39.00	12413.0
39.90	3138.0
41.10	1234.0
41.90	276.0
43.00	1000.0
44.00	4204.0
45.00	679.0
45.90	96.0
46.80	120.0
48.10	139.0
49.10	1315.0
50.00	43048.0
51.00	169088.0
52.00	8849.0
53.00	531.0
54.10	149.0
55.00	1783.0
56.00	5626.0
57.00	13423.0
58.00	676.0
58.90	474.0
59.90	228.0
61.00	2950.0
62.00	2649.0
63.00	7403.0
64.00	1153.0
65.00	3652.0
66.00	443.0
66.90	932.0
68.00	3688.0
69.00	205568.0
70.00	1215.0
71.00	678.0
72.00	140.0
73.00	6371.0
74.00	20528.0
75.00	33016.0
76.00	12155.0
77.00	251392.0
78.00	18112.0
79.00	15857.0
80.00	12333.0
81.00	16880.0
82.00	4830.0
83.00	4638.0
83.90	850.0
85.00	3201.0
86.00	4821.0
87.00	2577.0
87.90	1123.0
89.00	713.0
89.90	111.0
91.00	4134.0
92.00	4042.0
93.00	25816.0
94.00	1998.0
95.00	809.0
95.90	1853.0
97.00	912.0
98.00	19360.0
99.00	19312.0
99.90	1688.0
101.00	9659.0
101.90	528.0
102.90	3387.0
104.00	5752.0
105.00	5380.0
106.00	1914.0
107.00	73712.0
108.00	11750.0
109.00	2096.0
110.00	144256.0
111.00	21800.0
111.90	2888.0
113.00	876.0
114.10	212.0
114.90	497.0
116.00	3842.0
117.00	57976.0
118.00	4979.0
118.90	1019.0
119.90	1121.0
120.90	619.0
122.00	4843.0
123.00	7712.0
124.00	3597.0
125.00	3529.0
126.00	997.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

127.00	275328.0
128.00	20664.0
129.00	103136.0
130.00	9274.0
131.00	1961.0
132.00	1148.0
132.80	857.0
134.00	3345.0
135.00	8456.0
135.90	3320.0
137.00	5428.0
137.90	1077.0
138.90	744.0
140.00	1471.0
140.90	13109.0
141.90	4379.0
142.90	3692.0
143.90	930.0
145.00	774.0
146.00	2265.0
147.00	9020.0
148.00	15366.0
149.00	4279.0
150.00	980.0
151.10	1638.0
151.80	1317.0
153.00	4110.0
153.90	3179.0
155.00	7709.0
156.00	10958.0
157.10	2144.0
157.90	2592.0
158.90	2077.0
160.00	4360.0
160.90	6855.0
161.90	2158.0
162.90	933.0
164.00	926.0
164.90	5807.0
166.00	3920.0
167.00	27800.0
168.00	20664.0
169.00	3221.0
170.00	825.0
171.00	1139.0
172.00	2568.0
173.00	3306.0
173.90	6056.0
175.00	11178.0
175.90	3312.0
176.90	5607.0
178.00	1707.0
178.90	19392.0
180.00	15406.0
181.00	6988.0
181.90	921.0
182.90	751.0
184.00	1646.0
185.00	10069.0
186.00	76824.0
187.00	22432.0
188.00	2759.0
189.00	4445.0
189.90	831.0
191.00	2745.0
192.00	6859.0
193.00	7787.0
194.00	1919.0
195.00	926.0
196.00	17240.0
197.90	595200.0
198.90	39808.0
200.00	3550.0
201.40	2750.0
203.00	3946.0
204.00	20112.0
205.00	32520.0
206.00	145920.0
207.00	25224.0
208.00	5416.0
209.00	2238.0
210.20	2390.0
210.90	5615.0
212.90	440.0
213.90	101.0
215.00	1723.0
216.00	2979.0
216.90	36776.0
217.90	4835.0
219.00	471.0
221.00	27480.0
223.00	8563.0
224.00	82600.0
225.00	20408.0
226.00	2719.0
227.00	33048.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

227.90	4473.0
229.00	6955.0
229.90	894.0
231.00	3279.0
231.90	435.0
232.90	594.0
233.90	2081.0
234.90	2258.0
235.90	1619.0
237.00	3482.0
238.10	354.0
239.00	1518.0
239.90	1297.0
240.90	1847.0
242.00	4391.0
243.00	4596.0
244.00	68160.0
245.00	9177.0
246.00	11831.0
246.90	2464.0
248.00	546.0
248.90	2715.0
250.00	414.0
251.00	645.0
251.90	779.0
252.90	1355.0
254.00	1997.0
255.00	332608.0
256.00	49976.0
257.00	4707.0
257.90	18360.0
258.90	3350.0
260.00	663.0
261.00	807.0
261.80	83.0
263.00	170.0
263.90	535.0
265.00	7608.0
265.90	1094.0
266.90	765.0
267.90	246.0
268.90	125.0
270.10	347.0
271.00	1218.0
271.90	1086.0
272.90	9911.0
274.00	26600.0
275.00	156352.0
276.00	21568.0
277.00	13139.0
277.90	2257.0
278.90	485.0
279.80	86.0
281.00	2998.0
282.00	1199.0
283.00	1823.0
283.90	997.0
285.00	2204.0
286.20	333.0
286.90	68.0
287.80	165.0
289.10	482.0
289.80	516.0
291.00	543.0
292.00	889.0
293.00	3345.0
293.90	947.0
295.90	42352.0
296.90	6544.0
297.90	390.0
299.00	95.0
299.90	60.0
301.00	582.0
301.90	564.0
302.10	572.0
303.00	4939.0
304.00	1545.0
305.00	241.0
305.70	24.0
306.70	84.0
306.90	74.0
307.90	683.0
308.90	586.0
310.00	600.0
310.90	134.0
311.80	151.0
312.90	333.0
314.00	2257.0
314.90	4301.0
316.00	2709.0
317.00	608.0
317.90	110.0
318.80	96.0
319.80	194.0
321.00	1438.0
322.10	552.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

323.00	13949.0
324.00	2942.0
324.90	395.0
325.80	456.0
326.90	3069.0
327.90	1490.0
329.00	269.0
329.90	55.0
330.70	68.0
332.00	994.0
332.90	1443.0
334.00	9272.0
335.00	2434.0
335.80	333.0
336.90	15.0
337.70	14.0
339.00	322.0
340.00	391.0
341.00	2892.0
342.00	825.0
343.00	318.0
344.00	68.0
345.90	3087.0
346.90	604.0
347.80	62.0
350.00	135.0
352.00	4621.0
353.00	3337.0
354.00	4508.0
355.00	1996.0
355.90	432.0
357.00	385.0
358.90	331.0
359.90	133.0
361.00	119.0
361.90	21.0
362.70	26.0
364.90	20864.0
365.90	3009.0
366.80	202.0
367.70	4.0
368.80	83.0
369.90	577.0
371.00	1259.0
372.00	7398.0
373.00	1715.0
374.00	222.0
374.70	15.0
375.30	7.0
376.90	231.0
377.80	33.0
378.30	23.0
378.70	20.0
380.40	6.0
381.10	19.0
381.50	14.0
383.00	2392.0
384.00	557.0
384.90	253.0
385.70	109.0
386.80	56.0
387.80	68.0
388.90	92.0
390.00	1123.0
390.90	942.0
392.00	471.0
393.10	153.0
393.90	12.0
394.50	98.0
395.00	41.0
395.70	38.0
397.00	131.0
397.70	9.0
398.80	113.0
399.90	55.0
400.90	603.0
401.90	3621.0
402.90	4742.0
404.00	1916.0
404.90	347.0
405.90	110.0
406.60	9.0
407.10	15.0
407.50	4.0
408.10	13.0
408.80	13.0
409.60	94.0
410.00	87.0
410.80	13.0
411.50	2.0
412.30	9.0
412.90	14.0
413.40	3.0
415.00	386.0
416.20	117.0
417.00	105.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

417.80	7.0
418.70	13.0
419.60	74.0
420.90	4026.0
422.00	3837.0
423.00	28176.0
424.00	8136.0
425.00	1129.0
425.90	85.0
427.60	19.0
428.00	20.0
428.90	383.0
430.10	189.0
431.20	104.0
432.10	19.0
433.10	26.0
433.80	3.0
434.40	2.0
435.50	19.0
436.10	6.0
437.00	33.0
438.00	175.0
438.90	109.0
441.00	73584.0
442.00	492864.0
443.00	97632.0
444.00	9865.0
445.00	454.0
445.90	50.0
447.50	14.0
448.30	1.0
448.90	13.0
449.60	1.0
450.10	1.0
450.70	1.0
451.20	10.0
452.10	16.0
453.00	3.0
453.40	3.0
453.80	13.0
454.60	15.0
455.40	2.0
455.70	3.0
456.30	11.0
457.10	7.0
457.80	15.0
458.20	52.0
459.00	61.0
460.00	38.0
460.70	198.0
462.00	56.0
462.70	46.0
463.50	8.0
464.20	15.0
464.90	14.0
465.50	6.0
466.20	7.0
466.70	2.0
467.10	2.0
468.10	2.0
468.70	2.0
469.10	2.0
469.70	2.0
470.20	2.0
470.80	1.0
471.50	7.0
471.80	12.0
472.70	14.0
473.70	4.0
474.80	127.0
475.70	40.0
476.80	20.0
477.20	17.0
478.70	3.0
479.10	3.0
480.00	2.0
480.80	2.0
481.40	1.0
482.10	3.0
482.80	2.0
483.60	3.0
483.90	3.0
485.30	3.0
486.80	2.0
487.80	11.0
488.60	10.0
489.20	6.0
489.70	10.0
490.20	11.0
491.60	2.0
492.10	10.0
493.30	3.0
493.80	2.0
494.20	1.0
494.80	5.0
495.80	2.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

496.50	3.0
497.10	1.0
497.70	1.0
498.70	1.0
499.30	2.0
499.90	9.0
500.80	3.0
501.80	3.0
503.10	196.0
504.10	83.0
505.10	129.0
506.00	61.0
506.50	6.0
507.10	10.0
508.20	1.0
509.00	10.0
509.90	13.0
510.90	3.0
511.90	3.0
512.40	1.0
514.00	18.0
515.00	15.0
515.90	3.0
516.60	13.0
517.90	16.0
518.90	15.0
520.00	3.0
521.10	2.0
521.70	3.0
523.20	4.0
524.40	4.0
524.90	2.0
525.40	1.0
525.90	2.0
526.40	3.0
528.00	11.0
529.00	2.0
529.70	3.0
530.50	4.0
531.50	2.0
532.00	2.0
532.70	38.0
533.60	13.0
534.20	29.0
535.10	28.0
535.80	15.0
536.20	17.0
537.30	16.0
537.90	16.0
539.30	12.0
540.00	3.0
541.40	2.0
541.90	2.0
542.40	2.0
542.90	2.0
543.60	2.0
544.10	2.0
544.70	17.0
545.50	13.0
546.50	10.0
547.10	1.0
547.70	3.0
549.00	50.0
549.70	15.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

m/z	Abundance
36.00	176.0
37.00	666.0
38.00	1794.0
39.00	11068.0
40.00	3280.0
41.00	1193.0
42.00	237.0
43.00	940.0
44.00	4300.0
44.90	555.0
45.90	103.0
46.80	80.0
48.10	97.0
49.00	1105.0
50.00	38824.0
51.00	149568.0
52.00	8251.0
53.00	378.0
53.90	149.0
55.00	1601.0
56.00	5069.0
57.00	11016.0
57.90	784.0
59.00	344.0
60.00	204.0
61.00	2295.0
62.00	2490.0
63.00	7287.0
64.00	887.0
65.00	3920.0
65.90	332.0
67.10	659.0
68.00	3406.0
69.00	185856.0
70.00	976.0
71.10	655.0
71.90	184.0
73.00	6113.0
74.00	19536.0
75.00	31280.0
76.00	10012.0
77.00	224320.0
78.00	15665.0
79.00	13839.0
80.00	10805.0
81.00	15616.0
82.00	4334.0
83.00	3963.0
84.00	804.0
84.90	3147.0
86.00	4617.0
87.00	2451.0
87.90	775.0
88.90	495.0
90.00	124.0
90.90	3810.0
92.00	4330.0
93.00	24408.0
94.00	1773.0
94.90	874.0
96.00	1926.0
96.90	817.0
98.00	18512.0
99.00	17648.0
100.00	1479.0
100.90	8865.0
102.00	680.0
103.00	3106.0
104.00	5328.0
105.00	5343.0
106.00	2068.0
107.00	68856.0
108.00	10404.0
109.00	2359.0
110.00	129888.0
111.00	18808.0
111.90	2322.0
112.90	936.0
114.10	243.0
115.00	602.0
116.00	3878.0
117.00	53872.0
118.00	4029.0
118.90	912.0
119.90	970.0
120.90	543.0
122.00	4256.0
122.90	7146.0
124.00	3383.0
125.00	3137.0
126.10	1098.0
127.00	257344.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

128.00	18976.0
129.00	94904.0
129.90	8356.0
131.00	1704.0
131.90	1037.0
133.00	988.0
133.90	2622.0
135.00	8232.0
136.00	3568.0
137.00	5382.0
137.90	1082.0
139.00	692.0
139.90	1146.0
141.00	12491.0
142.00	3985.0
142.90	2536.0
143.90	843.0
144.90	858.0
145.90	2496.0
147.00	8721.0
148.00	15383.0
148.90	3561.0
149.90	1111.0
150.90	1646.0
151.10	1627.0
151.80	1067.0
152.90	3971.0
153.90	3090.0
155.00	7371.0
156.00	10830.0
157.00	2028.0
158.00	2372.0
158.90	1766.0
160.00	4142.0
161.00	6320.0
161.90	1685.0
163.00	718.0
163.90	715.0
164.90	4635.0
166.00	4433.0
167.00	26504.0
168.00	17512.0
169.00	2756.0
169.90	793.0
170.90	1024.0
171.90	1981.0
173.00	2861.0
174.00	5644.0
175.00	10698.0
176.00	2773.0
176.90	4826.0
178.00	1594.0
178.90	19688.0
180.00	14004.0
180.90	6204.0
181.90	1175.0
183.00	607.0
184.00	1547.0
185.00	9832.0
186.00	75184.0
187.00	21952.0
188.00	2063.0
189.00	4276.0
189.90	918.0
190.90	3014.0
192.00	6974.0
193.00	7372.0
194.00	1529.0
194.90	863.0
196.00	16432.0
196.80	2787.0
197.90	574976.0
198.90	39472.0
199.90	3514.0
201.40	2920.0
202.90	3658.0
204.00	18976.0
205.00	34320.0
206.00	146176.0
207.00	24456.0
208.00	5900.0
209.00	2528.0
210.10	2337.0
211.00	5599.0
213.00	444.0
213.80	250.0
215.00	1477.0
216.00	3398.0
216.90	37560.0
218.00	4716.0
218.90	488.0
221.00	27616.0
223.00	9486.0
224.00	79840.0
225.00	20192.0
226.00	2567.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

227.00	31368.0
228.00	4573.0
229.00	6857.0
230.00	992.0
231.00	2804.0
231.80	463.0
232.90	691.0
234.00	2308.0
234.90	2338.0
235.90	1662.0
237.00	3334.0
238.00	440.0
239.00	1452.0
240.00	1143.0
241.00	1995.0
242.00	4226.0
243.00	4905.0
244.00	69616.0
245.00	9013.0
246.00	12130.0
247.00	2271.0
247.90	619.0
249.00	2203.0
250.00	677.0
251.00	809.0
252.00	581.0
252.90	1209.0
254.00	2064.0
255.00	351936.0
256.00	50608.0
257.00	4605.0
258.00	18096.0
259.00	2932.0
259.80	524.0
261.00	668.0
261.80	78.0
263.00	269.0
263.90	626.0
264.90	8510.0
266.00	1473.0
266.90	833.0
267.90	248.0
268.90	189.0
270.00	493.0
270.80	1134.0
272.00	942.0
273.00	11584.0
274.00	29200.0
275.00	170752.0
276.00	21776.0
276.90	13789.0
277.90	2419.0
279.00	655.0
279.80	155.0
281.00	3323.0
282.00	1027.0
283.00	1881.0
284.00	1247.0
284.90	2398.0
286.00	558.0
286.90	96.0
287.90	237.0
289.10	526.0
290.00	366.0
290.90	471.0
291.90	685.0
292.90	3306.0
293.90	944.0
295.90	46296.0
296.90	6407.0
297.90	463.0
299.00	79.0
300.00	55.0
301.00	540.0
301.90	727.0
303.00	5328.0
303.90	1588.0
305.00	314.0
305.70	90.0
306.60	98.0
307.90	594.0
308.90	461.0
309.90	580.0
310.80	232.0
311.90	110.0
313.00	494.0
314.00	2590.0
314.90	5488.0
316.00	3050.0
316.90	572.0
317.90	66.0
318.70	62.0
319.80	197.0
320.90	1617.0
321.90	887.0
323.00	16045.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

324.00	3110.0
324.90	505.0
326.00	417.0
326.90	3104.0
328.00	1847.0
328.80	339.0
330.00	47.0
330.90	109.0
332.00	1251.0
333.00	1541.0
334.00	10370.0
334.90	2983.0
336.00	306.0
337.00	28.0
337.70	20.0
338.90	207.0
339.80	294.0
341.00	2521.0
341.90	612.0
343.00	417.0
343.70	58.0
344.10	119.0
344.80	107.0
345.90	3835.0
346.90	759.0
347.90	82.0
348.60	25.0
349.90	132.0
350.90	238.0
352.00	5340.0
353.00	3174.0
354.00	5295.0
354.90	1902.0
355.90	422.0
356.90	205.0
357.80	175.0
358.90	454.0
360.00	67.0
360.80	150.0
362.00	17.0
362.60	19.0
363.20	8.0
364.90	23120.0
365.90	3522.0
366.90	298.0
367.70	13.0
369.80	516.0
371.00	1364.0
372.00	9488.0
373.00	2137.0
373.90	212.0
375.00	10.0
376.00	14.0
376.90	349.0
378.00	28.0
378.60	10.0
379.20	14.0
379.90	11.0
380.80	24.0
381.60	21.0
383.00	2682.0
384.00	760.0
384.90	276.0
385.70	89.0
386.80	70.0
387.80	98.0
390.00	1363.0
390.90	913.0
392.00	807.0
392.70	131.0
393.80	3.0
394.50	76.0
395.10	52.0
395.70	38.0
396.80	109.0
398.60	75.0
399.10	90.0
401.00	637.0
402.00	4382.0
403.00	5744.0
403.90	2314.0
405.00	392.0
405.80	44.0
406.30	20.0
406.90	18.0
407.80	31.0
408.70	15.0
409.00	14.0
409.80	182.0
411.00	18.0
411.90	11.0
413.10	13.0
413.60	17.0
414.80	361.0
416.10	153.0
417.00	86.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

417.70	8.0
418.60	12.0
419.20	46.0
419.80	121.0
421.00	4966.0
422.00	4684.0
423.00	35048.0
424.00	8871.0
424.90	1117.0
426.00	134.0
427.30	12.0
427.90	16.0
429.10	510.0
430.90	158.0
432.10	44.0
432.90	7.0
433.50	11.0
434.10	12.0
434.90	12.0
436.20	28.0
437.00	24.0
437.90	236.0
439.10	180.0
441.00	94824.0
442.00	616960.0
443.00	117752.0
444.00	10746.0
444.90	521.0
445.90	54.0
447.00	29.0
448.10	13.0
448.60	13.0
450.00	20.0
450.70	1.0
451.30	2.0
451.70	3.0
452.20	2.0
453.40	3.0
454.50	7.0
455.40	7.0
455.90	3.0
456.60	21.0
457.60	29.0
458.80	68.0
459.90	50.0
460.90	147.0
461.90	93.0
462.70	45.0
463.40	21.0
464.10	17.0
465.20	2.0
465.80	3.0
466.90	3.0
467.30	2.0
467.90	1.0
468.50	2.0
469.10	2.0
469.60	2.0
470.10	4.0
471.00	13.0
471.70	5.0
472.50	1.0
473.20	3.0
474.00	2.0
475.10	98.0
475.90	22.0
476.90	20.0
477.60	1.0
478.10	16.0
478.50	11.0
479.10	4.0
479.60	4.0
480.50	3.0
481.80	8.0
482.80	3.0
483.90	8.0
484.60	2.0
485.10	1.0
485.60	3.0
486.60	29.0
487.20	16.0
488.50	5.0
489.00	12.0
490.00	13.0
490.10	13.0
490.90	5.0
492.10	23.0
492.70	1.0
493.40	21.0
494.10	3.0
494.60	1.0
495.30	4.0
496.10	2.0
496.80	4.0
497.50	2.0
498.60	3.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621

499.50	5.0
500.80	15.0
501.50	3.0
503.10	359.0
504.10	88.0
505.20	60.0
506.00	67.0
506.60	4.0
507.20	1.0
507.90	3.0
509.10	11.0
509.70	5.0
510.60	3.0
511.40	4.0
512.60	3.0
513.60	10.0
514.40	3.0
515.00	12.0
516.10	24.0
517.00	21.0
517.60	21.0
519.10	17.0
520.20	3.0
520.90	23.0
521.80	15.0
522.40	2.0
523.80	4.0
524.90	3.0
525.30	3.0
526.10	3.0
526.80	2.0
527.40	14.0
528.30	3.0
529.00	1.0
530.00	5.0
530.90	16.0
531.50	2.0
532.60	19.0
533.60	14.0
535.00	33.0
536.00	18.0
536.80	11.0
537.30	1.0
538.00	16.0
538.70	2.0
539.40	3.0
540.00	1.0
541.10	4.0
541.40	5.0
542.00	3.0
542.40	1.0
542.90	4.0
543.30	2.0
543.90	2.0
544.70	10.0
545.50	3.0
546.00	3.0
546.80	1.0
548.00	14.0
549.10	16.0

Instrument :
BNA_P
ClientSampleId :
DFTPP621