

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040221\
 Data File : BP005149.D
 Acq On : 02 Apr 2021 09:00
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 02 11:06:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	40.249	-0.6	91	0.00
3	Pyridine	40.000	43.540	-8.8	87	0.00
4	n-Nitrosodimethylamine	40.000	39.210	2.0	82	0.00
5 S	2-Fluorophenol	80.000	79.392	0.8	85	0.00
6	Aniline	40.000	40.172	-0.4	84	0.00
7 S	Phenol-d6	80.000	79.798	0.3	84	0.00
8	2-Chlorophenol	40.000	39.397	1.5	83	0.00
9	Benzaldehyde	40.000	35.595	11.0	80	0.00
10 C	Phenol	40.000	40.334	-0.8	84	0.00
11	bis(2-Chloroethyl)ether	40.000	39.603	1.0	84	0.00
12	1,3-Dichlorobenzene	40.000	37.918	5.2	80	0.00
13 C	1,4-Dichlorobenzene	40.000	38.664	3.3	82	0.00
14	1,2-Dichlorobenzene	40.000	38.686	3.3	83	0.00
15	Benzyl Alcohol	40.000	39.596	1.0	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.353	-5.9	88	0.00
17	2-Methylphenol	40.000	40.790	-2.0	85	0.00
18	Hexachloroethane	40.000	38.217	4.5	82	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.059	2.4	80	0.00
20	3+4-Methylphenols	40.000	40.069	-0.2	82	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	39.238	1.9	81	0.00
23 S	Nitrobenzene-d5	80.000	78.066	2.4	79	0.00
24	Nitrobenzene	40.000	37.895	5.3	78	0.00
25	Isophorone	40.000	40.918	-2.3	83	-0.01
26 C	2-Nitrophenol	40.000	42.050	-5.1	84	0.00
27	2,4-Dimethylphenol	40.000	40.226	-0.6	83	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.306	-0.8	80	0.00
29 C	2,4-Dichlorophenol	40.000	39.379	1.6	79	0.00
30	1,2,4-Trichlorobenzene	40.000	38.244	4.4	79	0.00
31	Naphthalene	40.000	39.168	2.1	81	0.00
32	Benzoic acid	40.000	42.332	-5.8	82	-0.03
33	4-Chloroaniline	40.000	39.879	0.3	82	0.00
34 C	Hexachlorobutadiene	40.000	37.263	6.8	76	0.00
35	Caprolactam	40.000	41.090	-2.7	82	-0.01
36 C	4-Chloro-3-methylphenol	40.000	39.709	0.7	80	0.00
37	2-Methylnaphthalene	40.000	38.925	2.7	79	0.00
38	1-Methylnaphthalene	40.000	39.569	1.1	80	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.616	6.0	76	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.744	3.1	76	0.00
42 S	2,4,6-Tribromophenol	80.000	77.433	3.2	78	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.563	1.1	78	0.00
44	2,4,5-Trichlorophenol	40.000	38.723	3.2	78	0.00
45 S	2-Fluorobiphenyl	80.000	76.327	4.6	78	0.00
46	1,1'-Biphenyl	40.000	38.710	3.2	79	0.00
47	2-Chloronaphthalene	40.000	38.829	2.9	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040221\
 Data File : BP005149.D
 Acq On : 02 Apr 2021 09:00
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 02 11:06:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:23:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.636	-1.6	79	0.00
49	Acenaphthylene	40.000	38.719	3.2	78	0.00
50	Dimethylphthalate	40.000	37.433	6.4	77	0.00
51	2,6-Dinitrotoluene	40.000	40.312	-0.8	81	0.00
52 C	Acenaphthene	40.000	37.967	5.1	77	0.00
53	3-Nitroaniline	40.000	42.631	-6.6	82	0.00
54 P	2,4-Dinitrophenol	40.000	40.182	-0.5	79	0.00
55	Dibenzofuran	40.000	38.461	3.8	79	0.00
56 P	4-Nitrophenol	40.000	43.646	-9.1	84	0.00
57	2,4-Dinitrotoluene	40.000	40.676	-1.7	80	0.00
58	Fluorene	40.000	38.135	4.7	77	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.356	4.1	76	0.00
60	Diethylphthalate	40.000	38.106	4.7	76	0.00
61	4-Chlorophenyl-phenylether	40.000	37.633	5.9	78	0.00
62	4-Nitroaniline	40.000	42.491	-6.2	82	0.00
63	Azobenzene	40.000	37.931	5.2	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	80	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.668	-9.2	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.091	-0.2	79	0.00
67	4-Bromophenyl-phenylether	40.000	39.739	0.7	77	0.00
68	Hexachlorobenzene	40.000	38.562	3.6	75	0.00
69	Atrazine	40.000	41.354	-3.4	77	0.00
70 C	Pentachlorophenol	40.000	40.710	-1.8	79	0.00
71	Phenanthrene	40.000	39.379	1.6	77	0.00
72	Anthracene	40.000	39.880	0.3	78	0.00
73	Carbazole	40.000	40.965	-2.4	80	0.00
74	Di-n-butylphthalate	40.000	41.426	-3.6	79	0.00
75 C	Fluoranthene	40.000	39.507	1.2	78	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	81	0.00
77	Benzydine	40.000	49.015	-22.5	85	0.00
78	Pyrene	40.000	39.734	0.7	78	0.00
79 S	Terphenyl-d14	80.000	77.669	2.9	76	0.00
80	Butylbenzylphthalate	40.000	43.362	-8.4	81	0.00
81	Benzo(a)anthracene	40.000	40.483	-1.2	81	0.00
82	3,3'-Dichlorobenzidine	40.000	42.358	-5.9	81	0.00
83	Chrysene	40.000	40.332	-0.8	81	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.843	-9.6	82	0.00
85 c	Di-n-octyl phthalate	40.000	45.115	-12.8	86	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.678	0.8	82	-0.01
88	Benzo(b)fluoranthene	40.000	38.096	4.8	79	0.00
89	Benzo(k)fluoranthene	40.000	38.911	2.7	80	0.00
90 C	Benzo(a)pyrene	40.000	39.142	2.1	81	0.00
91	Dibenzo(a,h)anthracene	40.000	39.757	0.6	82	-0.01
92	Benzo(g,h,i)perylene	40.000	39.819	0.5	82	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040221\
Data File : BP005149.D
Acq On : 02 Apr 2021 09:00
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 02 11:06:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:23:57 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0