

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040323\
 Data File : BP014480.D
 Acq On : 03 Apr 2023 11:18
 Operator : CG/JU
 Sample : PB151840BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK840

Quant Time: Apr 04 00:51:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 31 04:41:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	118968	20.000	ng/ul	0.00
20) Naphthalene-d8	10.840	136	527313	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.669	164	361605	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.434	188	841765	20.000	ng/ul	0.00
79) Chrysene-d12	21.522	240	755030	20.000	ng/ul	0.00
88) Perylene-d12	24.045	264	736818	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	18694	6.094	ng/uL	0.00
4) Pyridine-d5	3.823	84	252278	32.089	ng/ul	0.00
7) Phenol-d5	7.187	99	328489	29.939	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.346	67	220882	31.377	ng/ul	0.00
11) 2-Chlorophenol-d4	7.552	132	247937	31.265	ng/ul	0.01
15) 4-Methylphenol-d8	8.746	113	266558	30.119	ng/ul	0.02
21) Nitrobenzene-d5	9.199	128	124352	29.212	ng/ul	0.00
24) 2-Nitrophenol-d4	9.922	143	146175	29.988	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.469	165	251906	28.376	ng/ul	0.01
31) 4-Chloroaniline-d4	10.993	131	348587	29.811	ng/ul	0.01
46) Dimethylphthalate-d6	14.075	166	908584	30.989	ng/ul	0.00
49) Acenaphthylene-d8	14.363	160	979988	30.123	ng/ul	0.00
54) 4-Nitrophenol-d4	14.916	143	95614	21.586	ng/ul	0.04
60) Fluorene-d10	15.669	176	771328	31.547	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	114049	18.413	ng/ul	0.00
73) Anthracene-d10	17.534	188	1263886	31.025	ng/ul	0.00
81) Pyrene-d10	19.763	212	1517626	29.925	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.880	264	1254237	32.039	ng/ul	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040323\
Data File : BP014480.D
Acq On : 03 Apr 2023 11:18
Operator : CG/JU
Sample : PB151840BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK840

Quant Time: Apr 04 00:51:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 31 04:41:29 2023
Response via : Initial Calibration

