

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040323\
 Data File : BP014536.D
 Acq On : 04 Apr 2023 19:44
 Operator : CG/JU
 Sample : 02158-23
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PMW-3(20230331)

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 04/05/2023
 Supervised By : Jagrut Upadhyay 04/05/2023

Quant Time: Apr 04 23:47:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 04 23:43:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	110350	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	446099	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.657	164	273594	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.416	188	622142	20.000	ng/u1	0.00
79) Chrysene-d12	21.504	240	666476	20.000	ng/u1	0.00
88) Perylene-d12	24.016	264	782584	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	13782	4.843	ng/uL	0.00
4) Pyridine-d5	3.822	84	74610	10.231	ng/u1	0.01
7) Phenol-d5	7.181	99	62791	6.170	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.328	67	182116	27.891	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	164572	22.374	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	113904	13.875	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	98446	27.336	ng/u1	0.00
24) 2-Nitrophenol-d4	9.905	143	108950	26.420	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.457	165	174074	23.179	ng/u1	0.00
31) 4-Chloroaniline-d4	10.975	131	197101	19.925	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	628411	28.328	ng/u1	0.00
49) Acenaphthylene-d8	14.351	160	702287	28.531	ng/u1	0.00
54) 4-Nitrophenol-d4	14.928	143	20210m	6.030	ng/u1	0.04
60) Fluorene-d10	15.651	176	541490	29.271	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.787	200	94632	20.672	ng/u1	0.00
73) Anthracene-d10	17.516	188	915704	30.413	ng/u1	0.00
81) Pyrene-d10	19.745	212	1163778	25.996	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.851	264	1283292	30.864	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040323\
Data File : BP014536.D
Acq On : 04 Apr 2023 19:44
Operator : CG/JU
Sample : 02158-23
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PMW-3(20230331)

Quant Time: Apr 04 23:47:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 04 23:43:56 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 04/05/2023
Supervised By :Jagrut Upadhyay 04/05/2023

