

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040422\
 Data File : BP009674.D
 Acq On : 04 Apr 2022 12:14
 Operator : CG/JU
 Sample : PB143797BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB143797BL

Quant Time: Apr 04 23:23:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 04:36:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	253637	20.000	ng	-0.01
21) Naphthalene-d8	10.522	136	945070	20.000	ng	-0.01
39) Acenaphthene-d10	14.381	164	529170	20.000	ng	-0.01
64) Phenanthrene-d10	17.145	188	1010103	20.000	ng	-0.01
76) Chrysene-d12	21.263	240	1030072	20.000	ng	-0.01
86) Perylene-d12	23.639	264	1148449	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.334	112	2204720	151.764	ng	-0.01
7) Phenol-d6	6.922	99	2649955	134.884	ng	-0.01
23) Nitrobenzene-d5	8.887	82	1698786	95.521	ng	-0.01
42) 2,4,6-Tribromophenol	15.886	330	918004	105.128	ng	0.00
45) 2-Fluorobiphenyl	12.998	172	3607525	94.508	ng	-0.01
79) Terphenyl-d14	19.727	244	5014430	87.376	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040422\
Data File : BP009674.D
Acq On : 04 Apr 2022 12:14
Operator : CG/JU
Sample : PB143797BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB143797BL

Quant Time: Apr 04 23:23:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 31 04:36:51 2022
Response via : Initial Calibration

