

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014569.D
 Acq On : 05 Apr 2023 22:13
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020662

Quant Time: Apr 06 01:00:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	114686	20.000	ng/u1	0.00
20) Naphthalene-d8	10.804	136	490127	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.640	164	331198	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	755129	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	796036	20.000	ng/u1	0.00
88) Perylene-d12	23.992	264	889908	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	27575	9.324	ng/uL	0.00
4) Pyridine-d5	3.799	84	170676	22.520	ng/u1	0.00
7) Phenol-d5	7.152	99	217045	20.521	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.311	67	153158	22.569	ng/u1	0.00
11) 2-Chlorophenol-d4	7.516	132	166186	21.739	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	176504	20.688	ng/u1	0.00
21) Nitrobenzene-d5	9.163	128	84318	21.310	ng/u1	0.00
24) 2-Nitrophenol-d4	9.887	143	91355	20.163	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	168599	20.433	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	218660	20.118	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	567698	21.140	ng/u1	0.00
49) Acenaphthylene-d8	14.334	160	626715	21.033	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	77057	18.994	ng/u1	0.00
60) Fluorene-d10	15.634	176	478329	21.359	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	104845	18.869	ng/u1	0.00
73) Anthracene-d10	17.498	188	761629	20.841	ng/u1	0.00
81) Pyrene-d10	19.728	212	930948	17.411	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.827	264	997726	21.102	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	29046	8.937	ng/uL	94
5) Pyridine	3.817	79	182289	22.984	ng/u1	95
6) Benzaldehyde	7.128	77	138292	26.305	ng/u1	99
8) Phenol	7.181	94	228335	20.811	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.405	93	190981	21.660	ng/u1	97
12) 2-Chlorophenol	7.552	128	170143	21.384	ng/u1	94
13) 2-Methylphenol	8.440	108	171391	20.946	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.516	45	269134	23.248	ng/u1	99
16) Acetophenone	8.822	105	298525	21.190	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.799	70	170799	22.441	ng/u1	95
18) 4-Methylphenol	8.769	108	181548	20.203	ng/u1	97
19) Hexachloroethane	9.063	117	78539	22.591	ng/u1	93
22) Nitrobenzene	9.205	77	249325	21.934	ng/u1	97
23) Isophorone	9.734	82	459860	21.947	ng/u1	99
25) 2-Nitrophenol	9.922	139	96666	20.214	ng/u1	97
26) 2,4-Dimethylphenol	9.981	107	227731	21.276	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.210	93	264159	21.028	ng/u1	98
29) 2,4-Dichlorophenol	10.463	162	168236	20.718	ng/u1	96
30) Naphthalene	10.857	128	575887	21.094	ng/u1	98
32) 4-Chloroaniline	10.981	127	227206	20.639	ng/u1	99
33) Hexachlorobutadiene	11.128	225	132082	20.531	ng/u1	97
34) Caprolactam	11.787	113	21167	8.823	ng/u1	85
35) 4-Chloro-3-methylphenol	12.104	107	206182	20.643	ng/u1	99
36) 2-Methylnaphthalene	12.463	142	382722	20.686	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014569.D
 Acq On : 05 Apr 2023 22:13
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020662

Quant Time: Apr 06 01:00:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.681	142	353398	19.259	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.828	216	213493	18.080	ng/ul	100
40) Hexachlorocyclopentadiene	12.798	237	125495	16.411	ng/ul	99
41) 2,4,6-Trichlorophenol	13.075	196	129267	17.690	ng/ul	95
42) 2,4,5-Trichlorophenol	13.157	196	139391	17.919	ng/ul	95
43) 1,1'-Biphenyl	13.469	154	480174	17.945	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	380982	18.007	ng/ul	99
45) 2-Nitroaniline	13.734	65	131938	20.137	ng/ul	91
47) Dimethylphthalate	14.093	163	566785	21.109	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	113420	21.621	ng/ul	94
50) Acenaphthylene	14.363	152	665185	20.961	ng/ul	99
51) 3-Nitroaniline	14.563	138	100615	21.409	ng/ul	96
52) Acenaphthene	14.704	153	471541	21.028	ng/ul	98
53) 2,4-Dinitrophenol	14.775	184	86828	25.560	ng/ul	89
55) 4-Nitrophenol	14.881	109	77552	19.369	ng/ul	86
56) Dibenzofuran	15.040	168	647474	20.539	ng/ul	98
57) 2,4-Dinitrotoluene	15.016	165	172541	23.188	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.269	232	145693	20.838	ng/ul#	96
59) Diethylphthalate	15.451	149	593360	22.238	ng/ul	98
61) Fluorene	15.687	166	550094	21.529	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.681	204	284753	20.743	ng/ul	98
63) 4-Nitroaniline	15.734	138	80082	20.972	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.781	198	100849	18.729	ng/ul	100
67) N-Nitrosodiphenylamine	15.898	169	464434	19.768	ng/ul	98
68) 4-Bromophenyl-phenylether	16.581	248	182824	19.457	ng/ul	99
69) Hexachlorobenzene	16.692	284	211216	19.569	ng/ul	94
70) Atrazine	16.857	200	180763	21.135	ng/ul	97
71) Pentachlorophenol	17.051	266	135751	21.710	ng/ul	99
72) Phenanthrene	17.445	178	881666	20.987	ng/ul	99
74) Anthracene	17.534	178	887701	20.965	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.434	216	213801	16.085	ng/uL	99
76) Pentachlorobenzene	14.957	250	247979	18.697	ng/uL	98
77) Carbazole	17.810	167	770423	21.743	ng/ul	99
78) Di-n-butylphthalate	18.339	149	1007351	22.610	ng/ul	99
80) Fluoranthene	19.404	202	1081892	17.113	ng/ul	99
82) Pyrene	19.757	202	1137682	17.237	ng/ul	100
83) Butylbenzylphthalate	20.616	149	486472	21.397	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.404	252	391004	22.210	ng/ul	99
85) Benzo(a)anthracene	21.469	228	1181437	20.944	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.369	149	705252	22.684	ng/ul	99
87) Chrysene	21.527	228	1139633	21.396	ng/ul	99
89) Di-n-octyl phthalate	22.322	149	1260147	21.821	ng/ul	100
90) Benzo(b)fluoranthene	23.221	252	1235000	20.548	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	1223505	20.960	ng/ul	99
93) Benzo(a)pyrene	23.874	252	1088274	21.036	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.615	276	1456848	20.469	ng/ul	99
95) Dibenzo(a,h)anthracene	26.627	278	1198208	20.107	ng/ul	98
96) Benzo(g,h,i)perylene	27.421	276	1194321	20.586	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014569.D
 Acq On : 05 Apr 2023 22:13
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020662

Quant Time: Apr 06 01:00:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.937 ng/uL
RT: 3.405 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:	88	Resp:	29046
Ion	Ratio	Lower	Upper
88	100		
43	32.9	24.4	36.6
58	77.5	57.4	86.0



#5
Pyridine
Concen: 22.984 ng/uL
RT: 3.817 min Scan# 107
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:	79	Resp:	182289
Ion	Ratio	Lower	Upper
79	100		
52	82.3	61.4	92.2
51	37.2	28.5	42.7









#6
Benzaldehyde
Concen: 26.305 ng/ul
RT: 7.128 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion: 77 Resp: 138292
Ion Ratio Lower Upper
77 100
105 87.1 70.6 105.8
106 82.7 66.5 99.7



#8
Phenol
Concen: 20.811 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion: 94 Resp: 228335
Ion Ratio Lower Upper
94 100
65 31.1 25.2 37.8
66 48.4 36.2 54.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.660 ng/ul
RT: 7.405 min Scan# 719
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion: 93 Resp: 190981
Ion Ratio Lower Upper
93 100
63 82.2 63.2 94.8
95 33.3 26.0 39.0



#12
2-Chlorophenol
Concen: 21.384 ng/ul
RT: 7.552 min Scan# 742
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:128 Resp: 170143
Ion Ratio Lower Upper
128 100
64 56.1 41.4 62.0
130 32.6 24.1 36.1





#13

2-Methylphenol

Concen: 20.946 ng/ul

RT: 8.440 min Scan# 896

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :





SSTD020662

Tgt Ion:108 Resp: 171391

Ion Ratio Lower Upper

108 100

107 87.1 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 23.248 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion: 45 Resp: 269134


Ion Ratio Lower Upper

45 100

77 15.5 12.8 19.2

79 12.3 9.6 14.4









#16
 Acetophenone
 Concen: 21.190 ng/ul
 RT: 8.822 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: BP014569.D
 Acq: 05 Apr 2023 22:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020662


Tgt Ion:105 Resp: 298525
 Ion Ratio Lower Upper
 105 100
 77 87.6 67.3 100.9
 51 31.3 23.0 34.4



#17
 N-Nitroso-di-n-propylamine
 Concen: 22.441 ng/ul
 RT: 8.799 min Scan# 954
 Delta R.T. 0.000 min
 Lab File: BP014569.D
 Acq: 05 Apr 2023 22:13

Tgt Ion: 70 Resp: 170799
 Ion Ratio Lower Upper
 70 100
 42 56.4 41.1 61.7
 101 8.9 7.0 10.6
 130 17.3 14.6 21.8





#18

4-Methylphenol

Concen: 20.203 ng/ul

RT: 8.769 min Scan# 949

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :





SSTD020662

Tgt Ion:108 Resp: 181548

Ion Ratio Lower Upper

108 100

107 114.6 89.0 133.6



#19

Hexachloroethane

Concen: 22.591 ng/ul

RT: 9.063 min Scan# 999

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:117 Resp: 78539


Ion Ratio Lower Upper

117 100

201 105.8 90.2 135.4

199 61.8 54.2 81.2









#22
Nitrobenzene
Concen: 21.934 ng/ul
RT: 9.205 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion: 77 Resp: 249325
Ion Ratio Lower Upper
77 100
123 36.0 30.3 45.5
65 13.4 11.0 16.4



#23
Isophorone
Concen: 21.947 ng/ul
RT: 9.734 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion: 82 Resp: 459860
Ion Ratio Lower Upper
82 100
95 7.5 6.3 9.5
138 14.6 12.2 18.2









#25
2-Nitrophenol
Concen: 20.214 ng/ul
RT: 9.922 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:139 Resp: 96666
Ion Ratio Lower Upper
139 100
65 65.8 49.7 74.5
109 39.3 31.1 46.7



#26
2,4-Dimethylphenol
Concen: 21.276 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:107 Resp: 227731
Ion Ratio Lower Upper
107 100
121 43.5 37.0 55.6
122 72.9 59.8 89.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.028 ng/ul
 RT: 10.210 min Scan# 1197
 Delta R.T. 0.000 min
 Lab File: BP014569.D
 Acq: 05 Apr 2023 22:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020662


Tgt Ion: 93 Resp: 264159
 Ion Ratio Lower Upper
 93 100
 95 32.9 25.8 38.8
 123 9.4 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 20.718 ng/ul
 RT: 10.463 min Scan# 1237
 Delta R.T. 0.000 min
 Lab File: BP014569.D
 Acq: 05 Apr 2023 22:13

Tgt Ion: 162 Resp: 168236
 Ion Ratio Lower Upper
 162 100
 164 65.0 50.1 75.1
 98 41.1 30.6 45.8









#30
Naphthalene
Concen: 21.094 ng/ul
RT: 10.857 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:128 Resp: 575887
Ion Ratio Lower Upper
128 100
129 10.4 8.9 13.3
127 13.2 11.0 16.6



#32
4-Chloroaniline
Concen: 20.639 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:127 Resp: 227206
Ion Ratio Lower Upper
127 100
129 31.3 25.4 38.0









#33
Hexachlorobutadiene
Concen: 20.531 ng/ul
RT: 11.128 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:225 Resp: 132082
Ion Ratio Lower Upper
225 100
223 58.2 49.3 73.9
227 64.0 49.6 74.4



#34
Caprolactam
Concen: 8.823 ng/ul
RT: 11.787 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:113 Resp: 21167
Ion Ratio Lower Upper
113 100
55 182.0 131.2 196.8
56 144.8 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 20.643 ng/ul
RT: 12.104 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:107 Resp: 206182
Ion Ratio Lower Upper
107 100
144 21.0 17.3 25.9
142 67.8 54.6 81.8



#36
2-Methylnaphthalene
Concen: 20.686 ng/ul
RT: 12.463 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:142 Resp: 382722
Ion Ratio Lower Upper
142 100
141 89.5 71.6 107.4





#37

1-Methylnaphthalene

Concen: 19.259 ng/ul

RT: 12.681 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:142 Resp: 353398

Ion Ratio Lower Upper

142 100

141 92.9 76.6 114.8

116 4.8 3.6 5.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.080 ng/ul

RT: 12.828 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Tgt Ion:216 Resp: 213493




Ion Ratio Lower Upper


216 100

214 76.7 61.0 91.6

179 20.6 16.6 24.8

108 17.5 13.8 20.6





#40

Hexachlorocyclopentadiene

Concen: 16.411 ng/ul

RT: 12.798 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:237 Resp: 125495

Ion Ratio Lower Upper

237 100

235 64.5 51.0 76.4

272 11.8 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 17.690 ng/ul

RT: 13.075 min Scan# 1681

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:196 Resp: 129267


Ion Ratio Lower Upper

196 100

198 100.9 76.6 114.8

200 30.3 25.4 38.2





#42

2,4,5-Trichlorophenol

Concen: 17.919 ng/ul

RT: 13.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:196 Resp: 139391

Ion Ratio Lower Upper

196 100

198 94.7 79.8 119.8

200 30.6 25.6 38.4



#43

1,1'-Biphenyl

Concen: 17.945 ng/ul

RT: 13.469 min Scan# 1748

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:154 Resp: 480174


Ion Ratio Lower Upper

154 100

153 41.1 32.4 48.6

76 14.3 10.7 16.1





#44

2-Chloronaphthalene

Concen: 18.007 ng/ul

RT: 13.516 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:162 Resp: 380982

Ion Ratio Lower Upper

162 100

164 33.6 26.1 39.1

127 41.8 33.7 50.5



#45

2-Nitroaniline

Concen: 20.137 ng/ul

RT: 13.734 min Scan# 1793

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion: 65 Resp: 131938


Ion Ratio Lower Upper

65 100

92 61.5 53.7 80.5

138 79.8 71.7 107.5









#47
Dimethylphthalate
Concen: 21.109 ng/ul
RT: 14.093 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:163 Resp: 566785
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.2 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 21.621 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:165 Resp: 113420
Ion Ratio Lower Upper
165 100
89 71.3 52.7 79.1
121 24.2 17.9 26.9









#50
Acenaphthylene
Concen: 20.961 ng/ul
RT: 14.363 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:152 Resp: 665185
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 21.409 ng/ul
RT: 14.563 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:138 Resp: 100615
Ion Ratio Lower Upper
138 100
108 12.9 9.9 14.9
92 139.9 107.6 161.4





#52

Acenaphthene

Concen: 21.028 ng/ul

RT: 14.704 min Scan# 1960

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:153 Resp: 471541

Ion Ratio Lower Upper

153 100

152 49.4 38.2 57.4

154 89.3 70.2 105.2



#53

2,4-Dinitrophenol

Concen: 25.560 ng/ul

RT: 14.775 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:184 Resp: 86828


Ion Ratio Lower Upper

184 100

63 83.0 57.1 85.7

154 75.4 55.3 82.9









#55
4-Nitrophenol
Concen: 19.369 ng/ul
RT: 14.881 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:109 Resp: 77552
Ion Ratio Lower Upper
109 100
139 106.8 77.4 116.2
65 140.9 97.3 145.9



#56
Dibenzofuran
Concen: 20.539 ng/ul
RT: 15.040 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:168 Resp: 647474
Ion Ratio Lower Upper
168 100
139 41.6 32.1 48.1









#57
2,4-Dinitrotoluene
Concen: 23.188 ng/ul
RT: 15.016 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:165 Resp: 172541
Ion Ratio Lower Upper
165 100
63 55.3 41.6 62.4
182 2.6 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.838 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:232 Resp: 145693
Ion Ratio Lower Upper
232 100
131 45.1 35.0 52.4
130 3.2 1.9 2.9#
166 33.5 24.1 36.1









#59
Diethylphthalate
Concen: 22.238 ng/ul
RT: 15.451 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:149 Resp: 593360
Ion Ratio Lower Upper
149 100
177 21.5 18.2 27.2
150 12.8 9.8 14.8



#61
Fluorene
Concen: 21.529 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:166 Resp: 550094
Ion Ratio Lower Upper
166 100
165 98.5 79.2 118.8
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 20.743 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:204 Resp: 284753
Ion Ratio Lower Upper
204 100
206 31.7 26.9 40.3
141 62.8 51.3 76.9



#63
4-Nitroaniline
Concen: 20.972 ng/ul
RT: 15.734 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:138 Resp: 80082
Ion Ratio Lower Upper
138 100
92 65.8 47.8 71.6
108 113.4 82.3 123.5









#66
4,6-Dinitro-2-methylphenol
Concen: 18.729 ng/ul
RT: 15.781 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:198 Resp: 100849
Ion Ratio Lower Upper
198 100
182 4.5 3.5 5.3
77 30.5 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 19.768 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:169 Resp: 464434
Ion Ratio Lower Upper
169 100
168 65.1 54.2 81.2
167 36.9 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.457 ng/ul
RT: 16.581 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:248 Resp: 182824
Ion Ratio Lower Upper
248 100
250 97.4 78.1 117.1
141 80.1 66.0 99.0



#69
Hexachlorobenzene
Concen: 19.569 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:284 Resp: 211216
Ion Ratio Lower Upper
284 100
142 38.2 26.6 39.8
249 33.7 25.7 38.5





#70

Atrazine

Concen: 21.135 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:200 Resp: 180763

Ion Ratio Lower Upper

200 100

173 27.7 20.6 30.8

215 48.6 37.2 55.8



#71

Pentachlorophenol

Concen: 21.710 ng/ul

RT: 17.051 min Scan# 2357

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:266 Resp: 135751


Ion Ratio Lower Upper

266 100

264 64.6 51.0 76.4

268 63.8 51.4 77.2





#72

Phenanthrene

Concen: 20.987 ng/ul

RT: 17.445 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:178 Resp: 881666

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4

176 18.8 15.3 22.9



#74

Anthracene

Concen: 20.965 ng/ul

RT: 17.534 min Scan# 2439

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:178 Resp: 887701


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 18.5 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.085 ng/uL

RT: 13.434 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:216 Resp: 213801

Ion Ratio Lower Upper

216 100

214 77.9 61.0 91.6

218 48.4 38.7 58.1



#76

Pentachlorobenzene

Concen: 18.697 ng/uL

RT: 14.957 min Scan# 2001

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:250 Resp: 247979


Ion Ratio Lower Upper

250 100

248 61.8 49.0 73.6

252 66.0 50.7 76.1





#77

Carbazole

Concen: 21.743 ng/ul

RT: 17.810 min Scan# 2486

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:167 Resp: 770423

Ion Ratio Lower Upper

167 100

166 22.0 17.4 26.2

139 13.6 11.3 16.9



#78

Di-n-butylphthalate

Concen: 22.610 ng/ul

RT: 18.339 min Scan# 2576

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:149 Resp: 1007351


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.0

104 6.2 4.8 7.2





#80

Fluoranthene

Concen: 17.113 ng/ul

RT: 19.404 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:202 Resp: 1081892

Ion Ratio Lower Upper

202 100

101 10.4 8.2 12.2

100 7.8 6.4 9.6



#82

Pyrene

Concen: 17.237 ng/ul

RT: 19.757 min Scan# 2817

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:202 Resp: 1137682


Ion Ratio Lower Upper

202 100

101 12.2 9.7 14.5

100 10.1 8.0 12.0





#83

Butylbenzylphthalate

Concen: 21.397 ng/ul

RT: 20.616 min Scan# 2963

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:149 Resp: 486472

Ion Ratio Lower Upper

149 100

91 81.8 61.8 92.8

206 22.1 17.8 26.6



#84

3,3'-Dichlorobenzidine

Concen: 22.210 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:252 Resp: 391004


Ion Ratio Lower Upper

252 100

254 63.6 51.4 77.0

126 10.5 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.944 ng/ul

RT: 21.469 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:228 Resp: 1181437

Ion Ratio Lower Upper

228 100

229 19.0 15.6 23.4

226 26.8 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.684 ng/ul

RT: 21.369 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:149 Resp: 705252


Ion Ratio Lower Upper

149 100

167 27.1 21.9 32.9

279 5.0 3.7 5.5









#87
Chrysene
Concen: 21.396 ng/ul
RT: 21.527 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:228 Resp: 1139633
Ion Ratio Lower Upper
228 100
226 29.6 23.8 35.8
229 19.1 16.1 24.1



#89
Di-n-octyl phthalate
Concen: 21.821 ng/ul
RT: 22.322 min Scan# 3253
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:149 Resp: 1260147





#90

Benzo(b)fluoranthene

Concen: 20.548 ng/ul

RT: 23.221 min Scan# 3406

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:252 Resp: 1235000

Ion Ratio Lower Upper

252 100

253 22.7 18.0 27.0

125 8.0 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.960 ng/ul

RT: 23.274 min Scan# 3415

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13




Tgt Ion:252 Resp: 1223505


Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 7.7 6.9 10.3









#93
Benzo(a)pyrene
Concen: 21.036 ng/ul
RT: 23.874 min Scan# 3517
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13




Instrument :
BNA_P
ClientSampleId :
SSTD020662


Tgt Ion:252 Resp: 1088274
Ion Ratio Lower Upper
252 100
253 21.6 16.9 25.3
125 9.0 6.8 10.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.469 ng/ul
RT: 26.615 min Scan# 3983
Delta R.T. 0.000 min
Lab File: BP014569.D
Acq: 05 Apr 2023 22:13

Tgt Ion:276 Resp: 1456848
Ion Ratio Lower Upper
276 100
138 19.4 14.9 22.3
277 25.8 20.3 30.5





#95

Dibenzo(a,h)anthracene

Concen: 20.107 ng/ul

RT: 26.627 min Scan# 3989

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Instrument :

BNA_P

ClientSampleId :

SSTD020662





Tgt Ion:278 Resp: 1198208

Ion Ratio Lower Upper

278 100

139 11.9 10.4 15.6

279 24.3 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.586 ng/ul

RT: 27.421 min Scan# 4120

Delta R.T. 0.000 min

Lab File: BP014569.D

Acq: 05 Apr 2023 22:13

Tgt Ion:276 Resp: 1194321

Ion Ratio Lower Upper

276 100

138 17.5 13.0 19.4

277 24.1 18.4 27.6

