

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014574.D
 Acq On : 06 Apr 2023 01:01
 Operator : CG/JU
 Sample : PB151910BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS910

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 04/06/2023
 Supervised By :Jagrut Upadhyay 04/06/2023

Quant Time: Apr 06 02:13:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	153190	20.000	ng/u1	0.00
20) Naphthalene-d8	10.804	136	657602	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.639	164	427351	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	866968	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	767489	20.000	ng/u1	0.00
88) Perylene-d12	23.986	264	835387	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	18518	4.688	ng/uL	0.00
4) Pyridine-d5	3.793	84	246973m	24.397	ng/u1	0.00
7) Phenol-d5	7.151	99	344798	24.405	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	234389	25.858	ng/u1	0.00
11) 2-Chlorophenol-d4	7.516	132	266169	26.066	ng/u1	0.00
15) 4-Methylphenol-d8	8.704	113	283446	24.872	ng/u1	0.00
21) Nitrobenzene-d5	9.163	128	132601	24.978	ng/u1	0.00
24) 2-Nitrophenol-d4	9.887	143	152239	25.044	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.428	165	277884	25.101	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	288443	19.780	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	851447	24.572	ng/u1	0.00
49) Acenaphthylene-d8	14.333	160	970995	25.255	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	114732	21.917	ng/u1	0.00
60) Fluorene-d10	15.633	176	713762	24.701	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	156533	24.538	ng/u1	0.00
73) Anthracene-d10	17.498	188	1061002	25.288	ng/u1	0.00
81) Pyrene-d10	19.727	212	1199527	23.268	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.827	264	1131691	25.498	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	45124	10.395	ng/uL#	85
5) Pyridine	3.816	79	267358	25.237	ng/u1	97
6) Benzaldehyde	7.128	77	149853	21.339	ng/u1	99
8) Phenol	7.181	94	383952	26.198	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.404	93	311477	26.447	ng/u1	98
12) 2-Chlorophenol	7.546	128	282357	26.567	ng/u1	91
13) 2-Methylphenol	8.440	108	289123	26.453	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.516	45	442899	28.642	ng/u1	98
16) Acetophenone	8.822	105	489730	26.024	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.798	70	285095	28.043	ng/u1	93
18) 4-Methylphenol	8.769	108	315887	26.316	ng/u1	98
19) Hexachloroethane	9.063	117	133474	28.743	ng/u1	94
22) Nitrobenzene	9.204	77	422042	27.673	ng/u1	99
23) Isophorone	9.728	82	764125	27.181	ng/u1	99
25) 2-Nitrophenol	9.922	139	166925	26.016	ng/u1	94
26) 2,4-Dimethylphenol	9.975	107	386665	26.924	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.210	93	440180	26.116	ng/u1	99
29) 2,4-Dichlorophenol	10.457	162	283682	26.038	ng/u1	96
30) Naphthalene	10.857	128	960608	26.225	ng/u1	99
32) 4-Chloroaniline	10.981	127	257150	17.410	ng/u1	99
33) Hexachlorobutadiene	11.128	225	223166	25.855	ng/u1	99
34) Caprolactam	11.781	113	85517	26.569	ng/u1	83
35) 4-Chloro-3-methylphenol	12.104	107	359927	26.859	ng/u1	97
36) 2-Methylnaphthalene	12.463	142	658665	26.534	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014574.D
 Acq On : 06 Apr 2023 01:01
 Operator : CG/JU
 Sample : PB151910BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS910

Quant Time: Apr 06 02:13:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 04/06/2023
 Supervised By :Jagrut Upadhyay 04/06/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.681	142	677197	27.507	ng/u1	98
39) 1,2,4,5-Tetrachloroben...	12.828	216	400824	26.307	ng/u1	98
40) Hexachlorocyclopentadiene	12.798	237	253954	25.737	ng/u1	98
41) 2,4,6-Trichlorophenol	13.075	196	240862	25.546	ng/u1	98
42) 2,4,5-Trichlorophenol	13.157	196	266337	26.534	ng/u1	99
43) 1,1'-Biphenyl	13.469	154	911979	26.414	ng/u1	99
44) 2-Chloronaphthalene	13.516	162	719186	26.345	ng/u1	100
45) 2-Nitroaniline	13.733	65	250596	29.642	ng/u1	95
47) Dimethylphthalate	14.092	163	922016	26.613	ng/u1	99
48) 2,6-Dinitrotoluene	14.222	165	187762	27.740	ng/u1	94
50) Acenaphthylene	14.363	152	1106647	27.026	ng/u1	99
51) 3-Nitroaniline	14.563	138	153024	25.235	ng/u1	96
52) Acenaphthene	14.704	153	748028	25.852	ng/u1	99
53) 2,4-Dinitrophenol	14.769	184	98619	22.499	ng/u1#	85
55) 4-Nitrophenol	14.875	109	129986	25.160	ng/u1	95
56) Dibenzofuran	15.039	168	1063307	26.141	ng/u1	99
57) 2,4-Dinitrotoluene	15.016	165	266768	27.784	ng/u1	93
58) 2,3,4,6-Tetrachlorophenol	15.269	232	229592	25.449	ng/u1	99
59) Diethylphthalate	15.457	149	915636	26.596	ng/u1	99
61) Fluorene	15.686	166	871919	26.446	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.680	204	465269	26.267	ng/u1	97
63) 4-Nitroaniline	15.727	138	141905	28.802	ng/u1	96
66) 4,6-Dinitro-2-methylph...	15.780	198	158405	25.623	ng/u1	99
67) N-Nitrosodiphenylamine	15.898	169	707938	26.245	ng/u1	100
68) 4-Bromophenyl-phenylether	16.580	248	275375	25.526	ng/u1	98
69) Hexachlorobenzene	16.698	284	314725	25.398	ng/u1	99
70) Atrazine	16.857	200	72073	7.340	ng/u1	99
71) Pentachlorophenol	17.051	266	157253	21.905	ng/u1	97
72) Phenanthrene	17.445	178	1313277	27.228	ng/u1	100
74) Anthracene	17.533	178	1319118	27.135	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.433	216	399394	26.171	ng/uL	97
76) Pentachlorobenzene	14.957	250	403572	26.503	ng/uL	98
77) Carbazole	17.810	167	1105032	27.163	ng/u1	99
78) Di-n-butylphthalate	18.339	149	1459334	28.529	ng/u1	100
80) Fluoranthene	19.404	202	1489950	24.444	ng/u1	99
82) Pyrene	19.757	202	1535521	24.129	ng/u1	100
83) Butylbenzylphthalate	20.615	149	641253	29.254	ng/u1	98
84) 3,3'-Dichlorobenzidine	21.398	252	397949	23.445	ng/u1	99
85) Benzo(a)anthracene	21.468	228	1459579	26.837	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.368	149	887555	29.610	ng/u1	99
87) Chrysene	21.527	228	1389063	27.049	ng/u1	99
89) Di-n-octyl phthalate	22.321	149	1514182	27.931	ng/u1	100
90) Benzo(b)fluoranthene	23.221	252	1517891	26.903	ng/u1	98
91) Benzo(k)fluoranthene	23.268	252	1467679	26.784	ng/u1	98
93) Benzo(a)pyrene	23.874	252	1317816	27.136	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.609	276	1823602	27.294	ng/u1	99
95) Dibenzo(a,h)anthracene	26.627	278	1500280	26.820	ng/u1	99
96) Benzo(g,h,i)perylene	27.421	276	1477415	27.128	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040523\
 Data File : BP014574.D
 Acq On : 06 Apr 2023 01:01
 Operator : CG/JU
 Sample : PB151910BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS910

Quant Time: Apr 06 02:13:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 06 00:55:13 2023
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By : Christian Giraldo 04/06/2023
 Supervised By : Jagrut Upadhyay 04/06/2023





#2
1,4-Dioxane
Concen: 10.395 ng/uL
RT: 3.399 min Scan# 30
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 88 Resp: 45124
Ion Ratio Lower Upper
88 100
43 36.9 24.4 36.6
58 85.8 57.4 86.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#5
Pyridine
Concen: 25.237 ng/uL
RT: 3.816 min Scan# 107
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion: 79 Resp: 267358
Ion Ratio Lower Upper
79 100
52 76.3 61.4 92.2
51 40.1 28.5 42.7





#6
Benzaldehyde
Concen: 21.339 ng/ul
RT: 7.128 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 77 Resp: 149853
Ion Ratio Lower Upper
77 100
105 88.0 70.6 105.8
106 81.5 66.5 99.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#8
Phenol
Concen: 26.198 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion: 94 Resp: 383952
Ion Ratio Lower Upper
94 100
65 33.6 25.2 37.8
66 48.5 36.2 54.4





#10
Bis(2-Chloroethyl)ether
Concen: 26.447 ng/ul
RT: 7.404 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 93 Resp: 311477
Ion Ratio Lower Upper
93 100
63 81.3 63.2 94.8
95 32.8 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#12
2-Chlorophenol
Concen: 26.567 ng/ul
RT: 7.546 min Scan# 741
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01



Tgt Ion:128 Resp: 282357
Ion Ratio Lower Upper
128 100
64 59.5 41.4 62.0
130 32.7 24.1 36.1





#13

2-Methylphenol

Concen: 26.453 ng/ul

RT: 8.440 min Scan# 893

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910

Tgt Ion:108 Resp: 289123

Ion Ratio Lower Upper

108 100





107 91.4 71.2 106.8

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.642 ng/ul

RT: 8.516 min Scan# 906

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01




Tgt Ion: 45 Resp: 442899


Ion Ratio Lower Upper

45 100

77 14.9 12.8 19.2


79 11.6 9.6 14.4





#16
Acetophenone
Concen: 26.024 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 105 Resp: 489730
Ion Ratio Lower Upper
105 100
77 86.9 67.3 100.9
51 31.3 23.0 34.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#17
N-Nitroso-di-n-propylamine
Concen: 28.043 ng/ul
RT: 8.798 min Scan# 954
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion: 70 Resp: 285095
Ion Ratio Lower Upper
70 100
42 58.1 41.1 61.7
101 9.0 7.0 10.6
130 17.2 14.6 21.8





#18
4-Methylphenol
Concen: 26.316 ng/ul
RT: 8.769 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:108 Resp: 31588
Ion Ratio Lower Upper
108 100
107 113.6 89.0 133.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#19
Hexachloroethane
Concen: 28.743 ng/ul
RT: 9.063 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:117 Resp: 133474
Ion Ratio Lower Upper
117 100
201 105.4 90.2 135.4
199 64.8 54.2 81.2





#22
Nitrobenzene
Concen: 27.673 ng/ul
RT: 9.204 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 77 Resp: 422042
Ion Ratio Lower Upper
77 100
123 38.6 30.3 45.5
65 13.9 11.0 16.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#23
Isophorone
Concen: 27.181 ng/ul
RT: 9.728 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01



Tgt Ion: 82 Resp: 764125
Ion Ratio Lower Upper
82 100
95 7.8 6.3 9.5
138 14.6 12.2 18.2









#25
2-Nitrophenol
Concen: 26.016 ng/ul
RT: 9.922 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:139 Resp: 166925
Ion Ratio Lower Upper
139 100
65 66.1 49.7 74.5
109 42.7 31.1 46.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#26
2,4-Dimethylphenol
Concen: 26.924 ng/ul
RT: 9.975 min Scan# 1154
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:107 Resp: 386665
Ion Ratio Lower Upper
107 100
121 44.7 37.0 55.6
122 74.5 59.8 89.6





#27
Bis(2-Chloroethoxy)methane
Concen: 26.116 ng/ul
RT: 10.210 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion: 93 Resp: 440180
Ion Ratio Lower Upper
93 100
95 32.8 25.8 38.8
123 9.8 8.5 12.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#29
2,4-Dichlorophenol
Concen: 26.038 ng/ul
RT: 10.457 min Scan# 1236
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion:162 Resp: 283682
Ion Ratio Lower Upper
162 100
164 64.9 50.1 75.1
98 41.6 30.6 45.8





#30
Naphthalene
Concen: 26.225 ng/ul
RT: 10.857 min Scan# 1304
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:128 Resp: 960608
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.1 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#32
4-Chloroaniline
Concen: 17.410 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:127 Resp: 257150
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.0





#33
Hexachlorobutadiene
Concen: 25.855 ng/ul
RT: 11.128 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion: 225 Resp: 223166
Ion Ratio Lower Upper
225 100
223 62.6 49.3 73.9
227 62.2 49.6 74.4




Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 04/06/2023
Supervised By : Jagrut Upadhyay 04/06/2023



#34
Caprolactam
Concen: 26.569 ng/ul
RT: 11.781 min Scan# 1461
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion: 113 Resp: 85517
Ion Ratio Lower Upper
113 100
55 187.6 131.2 196.8
56 144.8 101.1 151.7





#35
4-Chloro-3-methylphenol
Concen: 26.859 ng/ul
RT: 12.104 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:107 Resp: 359927
Ion Ratio Lower Upper
107 100
144 22.5 17.3 25.9
142 70.7 54.6 81.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#36
2-Methylnaphthalene
Concen: 26.534 ng/ul
RT: 12.463 min Scan# 1577
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:142 Resp: 658665
Ion Ratio Lower Upper
142 100
141 89.2 71.6 107.4





#37
1-Methylnaphthalene
Concen: 27.507 ng/ul
RT: 12.681 min Scan# 1614
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:142 Resp: 677197
Ion Ratio Lower Upper
142 100
141 94.2 76.6 114.8
116 5.0 3.6 5.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.307 ng/ul
RT: 12.828 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:216 Resp: 400824
Ion Ratio Lower Upper
216 100
214 77.3 61.0 91.6
179 22.2 16.6 24.8
108 17.8 13.8 20.6





#40
Hexachlorocyclopentadiene
Concen: 25.737 ng/ul
RT: 12.798 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:237 Resp: 253954
Ion Ratio Lower Upper
237 100
235 62.3 51.0 76.4
272 12.4 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#41
2,4,6-Trichlorophenol
Concen: 25.546 ng/ul
RT: 13.075 min Scan# 1681
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion:196 Resp: 240862
Ion Ratio Lower Upper
196 100
198 94.6 76.6 114.8
200 30.2 25.4 38.2





#42
2,4,5-Trichlorophenol
Concen: 26.534 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:196 Resp: 266337
Ion Ratio Lower Upper
196 100
198 100.6 79.8 119.8
200 32.4 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#43
1,1'-Biphenyl
Concen: 26.414 ng/ul
RT: 13.469 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01



Tgt Ion:154 Resp: 911979
Ion Ratio Lower Upper
154 100
153 40.4 32.4 48.6
76 14.5 10.7 16.1





#44

2-Chloronaphthalene

Concen: 26.345 ng/ul

RT: 13.516 min Scan# 1756

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910

Tgt Ion:162 Resp: 719186

Ion Ratio Lower Upper

162 100

164 32.4 26.1 39.1





127 42.0 33.7 50.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#45

2-Nitroaniline

Concen: 29.642 ng/ul

RT: 13.733 min Scan# 1793

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01




Tgt Ion: 65 Resp: 250596


Ion Ratio Lower Upper

65 100

92 64.3 53.7 80.5

138 83.1 71.7 107.5









#47
Dimethylphthalate
Concen: 26.613 ng/ul
RT: 14.092 min Scan# 1854
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:163 Resp: 922016
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.1 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#48
2,6-Dinitrotoluene
Concen: 27.740 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:165 Resp: 187762
Ion Ratio Lower Upper
165 100
89 71.2 52.7 79.1
121 23.4 17.9 26.9









#50
Acenaphthylene
Concen: 27.026 ng/ul
RT: 14.363 min Scan# 1900
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:152 Resp: 110664
Ion Ratio Lower Upper
152 100
151 19.7 16.2 24.4
153 12.7 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#51
3-Nitroaniline
Concen: 25.235 ng/ul
RT: 14.563 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:138 Resp: 153024
Ion Ratio Lower Upper
138 100
108 12.9 9.9 14.9
92 139.2 107.6 161.4





#52
Acenaphthene
Concen: 25.852 ng/ul
RT: 14.704 min Scan# 1958
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:153 Resp: 748028
Ion Ratio Lower Upper
153 100
152 47.2 38.2 57.4
154 88.5 70.2 105.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#53
2,4-Dinitrophenol
Concen: 22.499 ng/ul
RT: 14.769 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion:184 Resp: 98619
Ion Ratio Lower Upper
184 100
63 89.6 57.1 85.7#
154 74.8 55.3 82.9





#55
4-Nitrophenol
Concen: 25.160 ng/ul
RT: 14.875 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:109 Resp: 129986
Ion Ratio Lower Upper
109 100
139 98.0 77.4 116.2
65 131.2 97.3 145.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#56
Dibenzofuran
Concen: 26.141 ng/ul
RT: 15.039 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion:168 Resp: 1063307
Ion Ratio Lower Upper
168 100
139 40.4 32.1 48.1





#57
2,4-Dinitrotoluene
Concen: 27.784 ng/ul
RT: 15.016 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:165 Resp: 266768
Ion Ratio Lower Upper
165 100
63 57.4 41.6 62.4
182 2.9 2.4 3.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#58
2,3,4,6-Tetrachlorophenol
Concen: 25.449 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01



Tgt Ion:232 Resp: 229592
Ion Ratio Lower Upper
232 100
131 43.2 35.0 52.4
130 2.4 1.9 2.9
166 29.5 24.1 36.1









#59
Diethylphthalate
Concen: 26.596 ng/ul
RT: 15.457 min Scan# 2086
Delta R.T. 0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:149 Resp: 915636
Ion Ratio Lower Upper
149 100
177 22.2 18.2 27.2
150 13.0 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#61
Fluorene
Concen: 26.446 ng/ul
RT: 15.686 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:166 Resp: 871919
Ion Ratio Lower Upper
166 100
165 98.5 79.2 118.8
167 13.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 26.267 ng/ul
RT: 15.680 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:204 Resp: 465269
Ion Ratio Lower Upper
204 100
206 32.7 26.9 40.3
141 61.7 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#63
4-Nitroaniline
Concen: 28.802 ng/ul
RT: 15.727 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01


Tgt Ion:138 Resp: 141905
Ion Ratio Lower Upper
138 100
92 59.7 47.8 71.6
108 109.2 82.3 123.5





#66
4,6-Dinitro-2-methylphenol
Concen: 25.623 ng/ul
RT: 15.780 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:198 Resp: 158405
Ion Ratio Lower Upper
198 100
182 3.8 3.5 5.3
77 30.0 24.4 36.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#67
N-Nitrosodiphenylamine
Concen: 26.245 ng/ul
RT: 15.898 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:169 Resp: 707938
Ion Ratio Lower Upper
169 100
168 67.9 54.2 81.2
167 36.4 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 25.526 ng/ul
RT: 16.580 min Scan# 2277
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Instrument :
BNA_P
ClientSampleId :
SLCS910

Tgt Ion:248 Resp: 275375
Ion Ratio Lower Upper
248 100
250 98.0 78.1 117.1
141 78.5 66.0 99.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#69
Hexachlorobenzene
Concen: 25.398 ng/ul
RT: 16.698 min Scan# 2297
Delta R.T. 0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:284 Resp: 314725
Ion Ratio Lower Upper
284 100
142 33.8 26.6 39.8
249 32.3 25.7 38.5





#70

Atrazine

Concen: 7.340 ng/ul

RT: 16.857 min Scan# 2324

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :





SLCS910

Tgt Ion:	200	Resp:	72073
Ion Ratio	Lower	Upper	
200	100		
173	24.8	20.6	30.8
215	47.2	37.2	55.8

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#71

Pentachlorophenol

Concen: 21.905 ng/ul




RT: 17.051 min Scan# 2357


Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Tgt Ion:	266	Resp:	157253
Ion Ratio	Lower	Upper	
266	100		
264	61.4	51.0	76.4
268	62.5	51.4	77.2





#72

Phenanthrene

Concen: 27.228 ng/ul

RT: 17.445 min Scan# 2424

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910

Tgt Ion:178 Resp: 131327

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4





176 19.3 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#74

Anthracene

Concen: 27.135 ng/ul

RT: 17.533 min Scan# 2439

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01




Tgt Ion:178 Resp: 1319118


Ion Ratio Lower Upper

178 100

179 15.0 12.2 18.4


176 19.2 14.9 22.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 26.171 ng/uL
RT: 13.433 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:216 Resp: 399394
Ion Ratio Lower Upper
216 100
214 79.4 61.0 91.6
218 49.5 38.7 58.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#76
Pentachlorobenzene
Concen: 26.503 ng/uL
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Tgt Ion:250 Resp: 403572
Ion Ratio Lower Upper
250 100
248 63.7 49.0 73.6
252 63.3 50.7 76.1





#77
 Carbazole
 Concen: 27.163 ng/ul
 RT: 17.810 min Scan# 2486
 Delta R.T. -0.000 min
 Lab File: BP014574.D
 Acq: 06 Apr 2023 01:01




Instrument :
 BNA_P
 ClientSampleId :
 SLCS910




Tgt Ion:167 Resp: 1105032
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.4 26.2
 139 13.4 11.3 16.9

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 04/06/2023
 Supervised By :Jagrut Upadhyay 04/06/2023




#78
 Di-n-butylphthalate
 Concen: 28.529 ng/ul
 RT: 18.339 min Scan# 2576
 Delta R.T. -0.000 min
 Lab File: BP014574.D
 Acq: 06 Apr 2023 01:01




Tgt Ion:149 Resp: 1459334
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.0
 104 6.0 4.8 7.2





#80
Fluoranthene
Concen: 24.444 ng/ul
RT: 19.404 min Scan# 2757
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:202 Resp: 1489950
Ion Ratio Lower Upper
202 100
101 9.9 8.2 12.2
100 7.8 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#82
Pyrene
Concen: 24.129 ng/ul
RT: 19.757 min Scan# 2817
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:202 Resp: 1535521
Ion Ratio Lower Upper
202 100
101 12.1 9.7 14.5
100 10.2 8.0 12.0





#83

Butylbenzylphthalate

Concen: 29.254 ng/ul

RT: 20.615 min Scan# 2963

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910





Tgt Ion:	149	Resp:	641253
Ion Ratio	Lower	Upper	
149	100		
91	78.4	61.8	92.8
206	20.8	17.8	26.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#84

3,3'-Dichlorobenzidine

Concen: 23.445 ng/ul




RT: 21.398 min Scan# 3096


Delta R.T. -0.006 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01


Tgt Ion:	252	Resp:	397949
Ion Ratio	Lower	Upper	
252	100		
254	64.7	51.4	77.0
126	10.6	7.8	11.6





#85
Benzo(a)anthracene
Concen: 26.837 ng/ul
RT: 21.468 min Scan# 3108
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion:228 Resp: 1459579
Ion Ratio Lower Upper
228 100
229 19.8 15.6 23.4
226 27.4 21.8 32.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023



#86
Bis(2-ethylhexyl)phthalate
Concen: 29.610 ng/ul
RT: 21.368 min Scan# 3091
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion:149 Resp: 887555
Ion Ratio Lower Upper
149 100
167 27.9 21.9 32.9
279 4.9 3.7 5.5





#87

Chrysene

Concen: 27.049 ng/ul

RT: 21.527 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910

Tgt Ion:228 Resp: 1389063

Ion Ratio Lower Upper

228 100

226 30.4 23.8 35.8





229 19.7 16.1 24.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#89

Di-n-octyl phthalate

Concen: 27.931 ng/ul




RT: 22.321 min Scan# 3253


Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01


Tgt Ion:149 Resp: 1514182





#90
Benzo(b)fluoranthene
Concen: 26.903 ng/ul
RT: 23.221 min Scan# 3406
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910




Tgt Ion:252 Resp: 151789
Ion Ratio Lower Upper
252 100
253 21.6 18.0 27.0
125 7.9 6.8 10.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 04/06/2023
Supervised By :Jagrut Upadhyay 04/06/2023




#91
Benzo(k)fluoranthene
Concen: 26.784 ng/ul
RT: 23.268 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01



Tgt Ion:252 Resp: 1467679
Ion Ratio Lower Upper
252 100
253 22.7 17.3 25.9
125 7.9 6.9 10.3





#93

Benzo(a)pyrene

Concen: 27.136 ng/ul

RT: 23.874 min Scan# 3517

Delta R.T. -0.000 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01

Instrument :

BNA_P

ClientSampleId :

SLCS910

Tgt Ion:252 Resp: 1317816

Ion Ratio Lower Upper

252 100

253 21.6 16.9 25.3





125 9.2 6.8 10.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 04/06/2023

Supervised By :Jagrut Upadhyay 04/06/2023



#94

Indeno(1,2,3-cd)pyrene

Concen: 27.294 ng/ul

RT: 26.609 min Scan# 3982

Delta R.T. -0.006 min

Lab File: BP014574.D

Acq: 06 Apr 2023 01:01




Tgt Ion:276 Resp: 1823602


Ion Ratio Lower Upper

276 100

138 19.4 14.9 22.3


277 25.3 20.3 30.5





#95
Dibenzo(a,h)anthracene
Concen: 26.820 ng/ul
RT: 26.627 min Scan# 3985
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01




Instrument :
BNA_P
ClientSampleId :
SLCS910



Tgt Ion: 278 Resp: 1500280
Ion Ratio Lower Upper
278 100
139 12.6 10.4 15.6
279 24.2 18.9 28.3

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 04/06/2023
Supervised By : Jagrut Upadhyay 04/06/2023



#96
Benzo(g,h,i)perylene
Concen: 27.128 ng/ul
RT: 27.421 min Scan# 4120
Delta R.T. -0.000 min
Lab File: BP014574.D
Acq: 06 Apr 2023 01:01

Tgt Ion: 276 Resp: 1477415
Ion Ratio Lower Upper
276 100
138 17.0 13.0 19.4
277 23.4 18.4 27.6

