

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040721\
 Data File : BP005212.D
 Acq On : 07 Apr 2021 09:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 07 11:03:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 12:33:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.566	0.683	-20.7	85	0.00
3	Pyridine	1.407	1.719	-22.2	76	0.00
4	n-Nitrosodimethylamine	0.503	0.557	-10.7	72	0.00
5 S	2-Fluorophenol	1.300	1.386	-6.6	71	0.00
6	Aniline	2.110	2.203	-4.4	68	0.00
7 S	Phenol-d6	1.862	1.950	-4.7	68	0.00
8	2-Chlorophenol	1.371	1.431	-4.4	68	0.00
9	Benzaldehyde	0.948	0.889	6.2	65	0.00
10 C	Phenol	1.908	2.012	-5.5	69	0.00
11	bis(2-Chloroethyl)ether	1.397	1.458	-4.4	68	0.00
12	1,3-Dichlorobenzene	1.523	1.527	-0.3	66	0.00
13 C	1,4-Dichlorobenzene	1.546	1.524	1.4	65	0.00
14	1,2-Dichlorobenzene	1.483	1.469	0.9	66	0.00
15	Benzyl Alcohol	1.436	1.478	-2.9	65	0.00
16	2,2'-oxybis(1-Chloropropane	1.025	1.169	-14.0	74	0.00
17	2-Methylphenol	1.218	1.269	-4.2	68	0.00
18	Hexachloroethane	0.589	0.587	0.3	66	0.00
19 P	n-Nitroso-di-n-propylamine	1.208	1.277	-5.7	68	0.00
20	3+4-Methylphenols	1.663	1.763	-6.0	67	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
22	Acetophenone	0.575	0.582	-1.2	67	0.00
23 S	Nitrobenzene-d5	0.466	0.458	1.7	65	0.00
24	Nitrobenzene	0.491	0.470	4.3	64	0.00
25	Isophorone	0.860	0.882	-2.6	67	0.00
26 C	2-Nitrophenol	0.191	0.202	-5.8	69	0.00
27	2,4-Dimethylphenol	0.306	0.309	-1.0	67	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.448	-3.0	66	0.00
29 C	2,4-Dichlorophenol	0.343	0.336	2.0	64	0.00
30	1,2,4-Trichlorobenzene	0.385	0.370	3.9	64	0.00
31	Naphthalene	1.130	1.099	2.7	65	0.00
32	Benzoic acid	0.227	0.234	-3.1	65	0.00
33	4-Chloroaniline	0.457	0.473	-3.5	68	0.00
34 C	Hexachlorobutadiene	0.248	0.223	10.1	59	0.00
35	Caprolactam	0.112	0.126	-12.5	73	0.00
36 C	4-Chloro-3-methylphenol	0.414	0.416	-0.5	65	0.00
37	2-Methylnaphthalene	0.816	0.797	2.3	64	0.00
38	1-Methylnaphthalene	0.765	0.762	0.4	65	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
40	1,2,4,5-Tetrachlorobenzene	0.651	0.616	5.4	62	0.00
41 P	Hexachlorocyclopentadiene	0.354	0.302	14.7	54	0.00
42 S	2,4,6-Tribromophenol	0.261	0.249	4.6	63	0.00
43 C	2,4,6-Trichlorophenol	0.450	0.441	2.0	63	0.00
44	2,4,5-Trichlorophenol	0.488	0.475	2.7	64	0.00
45 S	2-Fluorobiphenyl	1.478	1.398	5.4	63	0.00
46	1,1'-Biphenyl	1.525	1.467	3.8	64	0.00
47	2-Chloronaphthalene	1.242	1.192	4.0	65	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040721\
 Data File : BP005212.D
 Acq On : 07 Apr 2021 09:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 07 11:03:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 12:33:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.396	0.418	-5.6	68	0.00
49	Acenaphthylene	1.931	1.888	2.2	64	0.00
50	Dimethylphthalate	1.540	1.478	4.0	65	0.00
51	2,6-Dinitrotoluene	0.363	0.363	0.0	65	0.00
52 C	Acenaphthene	1.188	1.163	2.1	65	0.00
53	3-Nitroaniline	0.334	0.356	-6.6	67	0.00
54 P	2,4-Dinitrophenol	0.220	0.223	-1.4	65	0.00
55	Dibenzofuran	1.948	1.878	3.6	65	0.00
56 P	4-Nitrophenol	0.274	0.304	-10.9	69	0.00
57	2,4-Dinitrotoluene	0.479	0.501	-4.6	67	0.00
58	Fluorene	1.573	1.502	4.5	63	0.00
59	2,3,4,6-Tetrachlorophenol	0.450	0.429	4.7	62	0.00
60	Diethylphthalate	1.507	1.444	4.2	62	0.00
61	4-Chlorophenyl-phenylether	0.835	0.776	7.1	63	0.00
62	4-Nitroaniline	0.343	0.389	-13.4	71	0.00
63	Azobenzene	1.722	1.649	4.2	63	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
65	4,6-Dinitro-2-methylphenol	0.146	0.163	-11.6	67	0.00
66 c	n-Nitrosodiphenylamine	0.643	0.654	-1.7	65	0.00
67	4-Bromophenyl-phenylether	0.232	0.236	-1.7	64	0.00
68	Hexachlorobenzene	0.261	0.250	4.2	61	0.00
69	Atrazine	0.214	0.222	-3.7	63	0.00
70 C	Pentachlorophenol	0.172	0.166	3.5	62	0.00
71	Phenanthrene	1.159	1.141	1.6	63	0.00
72	Anthracene	1.133	1.160	-2.4	66	0.00
73	Carbazole	1.064	1.113	-4.6	67	0.00
74	Di-n-butylphthalate	1.171	1.239	-5.8	66	0.00
75 C	Fluoranthene	1.364	1.382	-1.3	65	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
77	Benzydine	0.439	0.496	-13.0	66	0.00
78	Pyrene	1.374	1.332	3.1	64	0.00
79 S	Terphenyl-d14	1.099	1.050	4.5	63	0.00
80	Butylbenzylphthalate	0.510	0.565	-10.8	70	0.00
81	Benzo(a)anthracene	1.359	1.363	-0.3	68	0.00
82	3,3'-Dichlorobenzidine	0.457	0.468	-2.4	66	0.00
83	Chrysene	1.334	1.316	1.3	66	0.00
84	Bis(2-ethylhexyl)phthalate	0.748	0.824	-10.2	69	0.00
85 c	Di-n-octyl phthalate	1.261	1.431	-13.5	73	0.00
86 I	Perylene-d12	1.000	1.000	0.0	70	0.00
87	Indeno(1,2,3-cd)pyrene	1.500	1.492	0.5	69	0.00
88	Benzo(b)fluoranthene	1.420	1.355	4.6	66	0.00
89	Benzo(k)fluoranthene	1.361	1.302	4.3	66	0.00
90 C	Benzo(a)pyrene	1.273	1.243	2.4	67	0.00
91	Dibenzo(a,h)anthracene	1.297	1.282	1.2	68	0.00
92	Benzo(g,h,i)perylene	1.251	1.244	0.6	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040721\
Data File : BP005212.D
Acq On : 07 Apr 2021 09:49
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 07 11:03:47 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 12:33:08 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0