

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
 Data File : BP005257.D
 Acq On : 09 Apr 2021 18:10
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP040921

Quant Time: Apr 09 18:47:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 18:44:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	41.651	-4.1	112	0.01
3	Pyridine	40.000	43.832	-9.6	110	0.00
4	n-Nitrosodimethylamine	40.000	43.958	-9.9	117	0.00
5 S	2-Fluorophenol	80.000	83.245	-4.1	114	0.00
6	Aniline	40.000	40.834	-2.1	112	0.00
7 S	Phenol-d6	80.000	81.518	-1.9	112	0.00
8	2-Chlorophenol	40.000	41.207	-3.0	112	0.00
9	Benzaldehyde	40.000	40.821	-2.1	114	0.00
10 C	Phenol	40.000	40.592	-1.5	112	0.00
11	bis(2-Chloroethyl)ether	40.000	41.922	-4.8	114	0.00
12	1,3-Dichlorobenzene	40.000	39.897	0.3	111	0.00
13 C	1,4-Dichlorobenzene	40.000	40.641	-1.6	111	0.00
14	1,2-Dichlorobenzene	40.000	40.051	-0.1	110	0.00
15	Benzyl Alcohol	40.000	40.354	-0.9	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.145	2.1	110	0.01
17	2-Methylphenol	40.000	40.591	-1.5	112	0.00
18	Hexachloroethane	40.000	40.290	-0.7	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.578	-3.9	114	0.00
20	3+4-Methylphenols	40.000	40.329	-0.8	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	41.525	-3.8	113	0.00
23 S	Nitrobenzene-d5	80.000	86.016	-7.5	113	0.00
24	Nitrobenzene	40.000	42.501	-6.3	114	0.00
25	Isophorone	40.000	41.745	-4.4	112	0.00
26 C	2-Nitrophenol	40.000	41.803	-4.5	107	0.00
27	2,4-Dimethylphenol	40.000	42.341	-5.9	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.553	-3.9	111	0.00
29 C	2,4-Dichlorophenol	40.000	41.785	-4.5	112	0.00
30	1,2,4-Trichlorobenzene	40.000	41.580	-3.9	113	0.00
31	Naphthalene	40.000	41.250	-3.1	112	0.00
32	Benzoic acid	40.000	40.671	-1.7	106	0.00
33	4-Chloroaniline	40.000	41.791	-4.5	111	0.00
34 C	Hexachlorobutadiene	40.000	41.263	-3.2	113	0.00
35	Caprolactam	40.000	41.934	-4.8	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.528	-3.8	110	0.00
37	2-Methylnaphthalene	40.000	40.991	-2.5	110	0.00
38	1-Methylnaphthalene	40.000	41.145	-2.9	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.540	-6.3	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.343	-8.4	111	0.00
42 S	2,4,6-Tribromophenol	80.000	85.275	-6.6	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.498	-8.7	109	0.00
44	2,4,5-Trichlorophenol	40.000	41.945	-4.9	106	0.00
45 S	2-Fluorobiphenyl	80.000	84.121	-5.2	110	0.00
46	1,1'-Biphenyl	40.000	42.345	-5.9	110	0.00
47	2-Chloronaphthalene	40.000	42.367	-5.9	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
 Data File : BP005257.D
 Acq On : 09 Apr 2021 18:10
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP040921

Quant Time: Apr 09 18:47:29 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 18:44:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.490	-1.2	112	0.00
49	Acenaphthylene	40.000	42.391	-6.0	110	0.00
50	Dimethylphthalate	40.000	42.995	-7.5	112	0.00
51	2,6-Dinitrotoluene	40.000	44.867	-12.2	111	0.00
52 C	Acenaphthene	40.000	41.820	-4.6	109	0.00
53	3-Nitroaniline	40.000	40.926	-2.3	108	0.00
54 P	2,4-Dinitrophenol	40.000	40.134	-0.3	110	0.00
55	Dibenzofuran	40.000	42.134	-5.3	110	0.00
56 P	4-Nitrophenol	40.000	43.086	-7.7	107	0.00
57	2,4-Dinitrotoluene	40.000	43.764	-9.4	108	0.00
58	Fluorene	40.000	42.659	-6.6	111	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.883	-7.2	113	0.00
60	Diethylphthalate	40.000	43.516	-8.8	111	0.00
61	4-Chlorophenyl-phenylether	40.000	41.899	-4.7	108	0.00
62	4-Nitroaniline	40.000	42.524	-6.3	109	0.00
63	Azobenzene	40.000	44.030	-10.1	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.997	-2.5	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.961	-2.4	111	0.00
67	4-Bromophenyl-phenylether	40.000	40.584	-1.5	111	0.00
68	Hexachlorobenzene	40.000	40.518	-1.3	109	0.00
69	Atrazine	40.000	42.955	-7.4	112	0.00
70 C	Pentachlorophenol	40.000	41.939	-4.8	110	0.00
71	Phenanthrene	40.000	41.267	-3.2	112	0.00
72	Anthracene	40.000	41.858	-4.6	111	0.00
73	Carbazole	40.000	41.204	-3.0	108	0.00
74	Di-n-butylphthalate	40.000	42.569	-6.4	111	0.00
75 C	Fluoranthene	40.000	41.198	-3.0	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	42.341	-5.9	104	0.00
78	Pyrene	40.000	40.174	-0.4	109	0.00
79 S	Terphenyl-d14	80.000	80.698	-0.9	110	0.00
80	Butylbenzylphthalate	40.000	39.622	0.9	105	0.00
81	Benzo(a)anthracene	40.000	40.211	-0.5	108	0.00
82	3,3'-Dichlorobenzidine	40.000	39.996	0.0	103	0.00
83	Chrysene	40.000	39.563	1.1	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.506	-1.3	108	0.00
85 c	Di-n-octyl phthalate	40.000	39.340	1.6	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.161	-2.9	107	0.00
88	Benzo(b)fluoranthene	40.000	41.414	-3.5	106	0.00
89	Benzo(k)fluoranthene	40.000	40.865	-2.2	105	0.00
90 C	Benzo(a)pyrene	40.000	41.225	-3.1	106	0.00
91	Dibenzo(a,h)anthracene	40.000	41.202	-3.0	106	0.00
92	Benzo(g,h,i)perylene	40.000	41.346	-3.4	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
Data File : BP005257.D
Acq On : 09 Apr 2021 18:10
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP040921

Quant Time: Apr 09 18:47:29 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 18:44:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0