



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
 Data File : BP005248.D
 Acq On : 09 Apr 2021 13:04
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Apr 09 18:44:11 2021



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.3	12570	PASS
68	69	0.00	2	1.7	311	PASS
69	198	0.00	100	59.7	18616	PASS
70	69	0.00	2	0.2	33	PASS
127	198	10	80	48.9	15257	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	31174	PASS
199	198	5	9	6.7	2092	PASS
275	198	10	60	27.6	8609	PASS
365	198	1	100	3.7	1147	PASS
441	198	0.01	100	11.2	3504	PASS
442	442	100	100	100.0	21110	PASS
443	442	15	24	19.9	4194	PASS

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/9/2021	BNA_P	<u>BP005248.D</u>
Compound Name	Response	Retention Time
DDT	114793	20.51
DDD	309	20.1
DDE	299	19.59
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
608	115401	0.53

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
Data File : BP005248.D
Acq On : 09 Apr 2021 13:04
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 09 13:44:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 12:33:08 2021
Response via : Initial Calibration



TIC: BP005248.D\data.ms

(77) Benzidine

19.316min (-0.006) 494321.25 ng




response 206065

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.99
183.00	12.00	12.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP040921\
 Data File : BP005248.D
 Acq On : 09 Apr 2021 13:04
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Apr 09 13:44:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP032321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 12:33:08 2021
 Response via : Initial Calibration



TIC: BP005248.D\data.ms

(70) Pentachlorophenol (C)

16.763min (-0.006) 362288.77 ng

response 34284

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.90	60.93
264.00	63.30	61.07
0.00	0.00	0.00