

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041023\
 Data File : BP014625.D
 Acq On : 10 Apr 2023 10:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020668

Quant Time: Apr 11 00:54:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 00:52:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	131676	20.000	ng/u1	0.00
20) Naphthalene-d8	10.799	136	569515	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	347129	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	689388	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	586234	20.000	ng/u1	0.00
88) Perylene-d12	23.986	264	675156	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	30218	8.899	ng/uL	0.00
4) Pyridine-d5	3.793	84	201234	23.126	ng/u1	0.00
7) Phenol-d5	7.146	99	243031	20.013	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	176134	22.606	ng/u1	0.00
11) 2-Chlorophenol-d4	7.510	132	186986	21.304	ng/u1	0.00
15) 4-Methylphenol-d8	8.699	113	198985	20.314	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	92409	20.099	ng/u1	0.00
24) 2-Nitrophenol-d4	9.881	143	104941	19.933	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	192910	20.120	ng/u1	0.00
31) 4-Chloroaniline-d4	10.946	131	241598	19.130	ng/u1	0.00
46) Dimethylphthalate-d6	14.040	166	569581	20.237	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	653532	20.926	ng/u1	0.00
54) 4-Nitrophenol-d4	14.857	143	63861	15.019	ng/u1	0.00
60) Fluorene-d10	15.628	176	468674	19.968	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	89373	17.619	ng/u1	0.00
73) Anthracene-d10	17.492	188	692414	20.754	ng/u1	0.00
81) Pyrene-d10	19.727	212	740821	18.813	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.821	264	747638	20.843	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	34443	9.230	ng/uL	97
5) Pyridine	3.811	79	208866	22.937	ng/u1	90
6) Benzaldehyde	7.122	77	158514	26.261	ng/u1	99
8) Phenol	7.169	94	256305	20.346	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.399	93	219103	21.643	ng/u1	96
12) 2-Chlorophenol	7.540	128	191796	20.995	ng/u1	94
13) 2-Methylphenol	8.428	108	192193	20.458	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.505	45	328197	24.692	ng/u1	99
16) Acetophenone	8.816	105	335530	20.743	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.793	70	191921	21.963	ng/u1	92
18) 4-Methylphenol	8.763	108	210106	20.364	ng/u1	100
19) Hexachloroethane	9.057	117	93358	23.389	ng/u1	92
22) Nitrobenzene	9.199	77	278891	21.115	ng/u1	96
23) Isophorone	9.722	82	507720	20.854	ng/u1	98
25) 2-Nitrophenol	9.916	139	110737	19.928	ng/u1	97
26) 2,4-Dimethylphenol	9.969	107	261286	21.008	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.204	93	301676	20.667	ng/u1	98
29) 2,4-Dichlorophenol	10.451	162	190178	20.155	ng/u1	95
30) Naphthalene	10.846	128	664188	20.937	ng/u1	99
32) 4-Chloroaniline	10.975	127	246537	19.274	ng/u1	99
33) Hexachlorobutadiene	11.122	225	151331	20.244	ng/u1	95
34) Caprolactam	11.775	113	44698	16.035	ng/u1	85
35) 4-Chloro-3-methylphenol	12.098	107	218041	18.787	ng/u1	98
36) 2-Methylnaphthalene	12.457	142	432130	20.100	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041023\
 Data File : BP014625.D
 Acq On : 10 Apr 2023 10:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020668

Quant Time: Apr 11 00:54:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 00:52:14 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	435982	20.448	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.822	216	250268	20.221	ng/ul	99
40) Hexachlorocyclopentadiene	12.793	237	147239	18.371	ng/ul	96
41) 2,4,6-Trichlorophenol	13.069	196	149249	19.488	ng/ul	96
42) 2,4,5-Trichlorophenol	13.151	196	159534	19.567	ng/ul	97
43) 1,1'-Biphenyl	13.463	154	571892	20.392	ng/ul	99
44) 2-Chloronaphthalene	13.510	162	464238	20.936	ng/ul	100
45) 2-Nitroaniline	13.728	65	145193	21.143	ng/ul	92
47) Dimethylphthalate	14.087	163	564962	20.076	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	110793	20.151	ng/ul	91
50) Acenaphthylene	14.357	152	708674	21.307	ng/ul	100
51) 3-Nitroaniline	14.557	138	93000	18.881	ng/ul	97
52) Acenaphthene	14.698	153	493386	20.992	ng/ul	99
53) 2,4-Dinitrophenol	14.775	184	47644	13.381	ng/ul	89
55) 4-Nitrophenol	14.869	109	65588	15.629	ng/ul	93
56) Dibenzofuran	15.034	168	677596	20.508	ng/ul	100
57) 2,4-Dinitrotoluene	15.010	165	155512	19.940	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.263	232	139828	19.081	ng/ul	94
59) Diethylphthalate	15.451	149	562654	20.120	ng/ul	98
61) Fluorene	15.681	166	548584	20.484	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.675	204	287055	19.951	ng/ul	98
63) 4-Nitroaniline	15.728	138	64013	15.995	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.775	198	89951	18.298	ng/ul	96
67) N-Nitrosodiphenylamine	15.892	169	447774	20.876	ng/ul	98
68) 4-Bromophenyl-phenylether	16.575	248	166321	19.388	ng/ul	99
69) Hexachlorobenzene	16.692	284	192999	19.587	ng/ul	100
70) Atrazine	16.851	200	155166	19.872	ng/ul	98
71) Pentachlorophenol	17.051	266	97937	17.156	ng/ul	97
72) Phenanthrene	17.439	178	804452	20.975	ng/ul	99
74) Anthracene	17.528	178	812090	21.008	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.428	216	254621	20.982	ng/uL	98
76) Pentachlorobenzene	14.951	250	248636	20.534	ng/uL	99
77) Carbazole	17.804	167	646393	19.982	ng/ul	100
78) Di-n-butylphthalate	18.333	149	846544	20.812	ng/ul	100
80) Fluoranthene	19.398	202	880638	18.915	ng/ul	96
82) Pyrene	19.751	202	918489	18.896	ng/ul	98
83) Butylbenzylphthalate	20.610	149	348414	20.809	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.398	252	271603	20.949	ng/ul	98
85) Benzo(a)anthracene	21.463	228	858964	20.677	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.363	149	499126	21.800	ng/ul	99
87) Chrysene	21.521	228	843935	21.515	ng/ul	99
89) Di-n-octyl phthalate	22.321	149	873195	19.930	ng/ul	100
90) Benzo(b)fluoranthene	23.215	252	918691	20.147	ng/ul	99
91) Benzo(k)fluoranthene	23.263	252	902920	20.388	ng/ul	98
93) Benzo(a)pyrene	23.868	252	821533	20.931	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.604	276	1130874	20.943	ng/ul	98
95) Dibenzo(a,h)anthracene	26.615	278	916449	20.271	ng/ul	99
96) Benzo(g,h,i)perylene	27.409	276	915915	20.809	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041023\
 Data File : BP014625.D
 Acq On : 10 Apr 2023 10:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020668

Quant Time: Apr 11 00:54:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 00:52:14 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.230 ng/uL
RT: 3.399 min Scan# 30
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:	88	Resp:	34443
Ion	Ratio	Lower	Upper
88	100		
43	33.0	24.4	36.6
58	69.5	57.4	86.0



#5
Pyridine
Concen: 22.937 ng/uL
RT: 3.811 min Scan# 106
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:	79	Resp:	208866
Ion	Ratio	Lower	Upper
79	100		
52	86.7	61.4	92.2
51	39.6	28.5	42.7









#6
Benzaldehyde
Concen: 26.261 ng/ul
RT: 7.122 min Scan# 669
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion: 77 Resp: 158514
Ion Ratio Lower Upper
77 100
105 89.1 70.6 105.8
106 84.3 66.5 99.7



#8
Phenol
Concen: 20.346 ng/ul
RT: 7.169 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion: 94 Resp: 256305
Ion Ratio Lower Upper
94 100
65 32.0 25.2 37.8
66 47.9 36.2 54.4





#10

Bis(2-Chloroethyl)ether

Concen: 21.643 ng/ul

RT: 7.399 min Scan# 716

Delta R.T. -0.006 min

Lab File: BP014625.D


Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668






Tgt Ion: 93 Resp: 219103

Ion	Ratio	Lower	Upper
93	100		
63	83.3	63.2	94.8
95	32.2	26.0	39.0

93 100

63 83.3 63.2 94.8

95 32.2 26.0 39.0



#12

2-Chlorophenol

Concen: 20.995 ng/ul

RT: 7.540 min Scan# 740

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:128 Resp: 191796


Ion	Ratio	Lower	Upper
128	100		
64	57.3	41.4	62.0
130	31.3	24.1	36.1

128 100

64 57.3 41.4 62.0

130 31.3 24.1 36.1





#13

2-Methylphenol

Concen: 20.458 ng/ul

RT: 8.428 min Scan# 891

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :





SSTD020668

Tgt Ion:108 Resp: 192193

Ion Ratio Lower Upper

108 100

107 91.4 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.692 ng/ul

RT: 8.505 min Scan# 904

Delta R.T. -0.012 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion: 45 Resp: 328197


Ion Ratio Lower Upper

45 100

77 15.9 12.8 19.2

79 11.6 9.6 14.4









#16
Acetophenone
Concen: 20.743 ng/ul
RT: 8.816 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:105 Resp: 335530
Ion Ratio Lower Upper
105 100
77 87.0 67.3 100.9
51 32.7 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.963 ng/ul
RT: 8.793 min Scan# 953
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion: 70 Resp: 191921
Ion Ratio Lower Upper
70 100
42 59.2 41.1 61.7
101 8.9 7.0 10.6
130 16.6 14.6 21.8









#18
4-Methylphenol
Concen: 20.364 ng/ul
RT: 8.763 min Scan# 948
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:108 Resp: 210106
Ion Ratio Lower Upper
108 100
107 110.9 89.0 133.6



#19
Hexachloroethane
Concen: 23.389 ng/ul
RT: 9.057 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:117 Resp: 93358
Ion Ratio Lower Upper
117 100
201 103.3 90.2 135.4
199 62.4 54.2 81.2









#22
Nitrobenzene
Concen: 21.115 ng/ul
RT: 9.199 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion: 77 Resp: 278891
Ion Ratio Lower Upper
77 100
123 35.5 30.3 45.5
65 15.2 11.0 16.4



#23
Isophorone
Concen: 20.854 ng/ul
RT: 9.722 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion: 82 Resp: 507720
Ion Ratio Lower Upper
82 100
95 8.0 6.3 9.5
138 14.2 12.2 18.2









#25
2-Nitrophenol
Concen: 19.928 ng/ul
RT: 9.916 min Scan# 1144
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:139 Resp: 110737
Ion Ratio Lower Upper
139 100
65 62.4 49.7 74.5
109 43.5 31.1 46.7



#26
2,4-Dimethylphenol
Concen: 21.008 ng/ul
RT: 9.969 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:107 Resp: 261286
Ion Ratio Lower Upper
107 100
121 44.6 37.0 55.6
122 73.4 59.8 89.6





#27

Bis(2-Chloroethoxy)methane

Concen: 20.667 ng/ul

RT: 10.204 min Scan# 1193

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion: 93 Resp: 301676

Ion Ratio Lower Upper

93 100

95 31.2 25.8 38.8

123 9.6 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 20.155 ng/ul

RT: 10.451 min Scan# 1235

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:162 Resp: 190178


Ion Ratio Lower Upper

162 100

164 65.2 50.1 75.1

98 43.5 30.6 45.8









#30
Naphthalene
Concen: 20.937 ng/ul
RT: 10.846 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:128 Resp: 664188
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 13.4 11.0 16.6



#32
4-Chloroaniline
Concen: 19.274 ng/ul
RT: 10.975 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:127 Resp: 246537
Ion Ratio Lower Upper
127 100
129 31.3 25.4 38.0









#33
Hexachlorobutadiene
Concen: 20.244 ng/ul
RT: 11.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:225 Resp: 151331
Ion Ratio Lower Upper
225 100
223 64.1 49.3 73.9
227 66.5 49.6 74.4



#34
Caprolactam
Concen: 16.035 ng/ul
RT: 11.775 min Scan# 1460
Delta R.T. -0.012 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:113 Resp: 44698
Ion Ratio Lower Upper
113 100
55 183.8 131.2 196.8
56 143.4 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 18.787 ng/ul
RT: 12.098 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:107 Resp: 218041
Ion Ratio Lower Upper
107 100
144 22.5 17.3 25.9
142 69.7 54.6 81.8



#36
2-Methylnaphthalene
Concen: 20.100 ng/ul
RT: 12.457 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:142 Resp: 432130
Ion Ratio Lower Upper
142 100
141 90.9 71.6 107.4





#37

1-Methylnaphthalene

Concen: 20.448 ng/ul

RT: 12.675 min Scan# 1613

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:142 Resp: 435982

Ion Ratio Lower Upper

142 100

141 91.7 76.6 114.8

116 4.6 3.6 5.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.221 ng/ul

RT: 12.822 min Scan# 1638

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Tgt Ion:216 Resp: 250268




Ion Ratio Lower Upper


216 100

214 76.3 61.0 91.6

179 21.8 16.6 24.8

108 19.0 13.8 20.6









#40
Hexachlorocyclopentadiene
Concen: 18.371 ng/ul
RT: 12.793 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:237 Resp: 147239
Ion Ratio Lower Upper
237 100
235 60.5 51.0 76.4
272 12.3 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 19.488 ng/ul
RT: 13.069 min Scan# 1680
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:196 Resp: 149249
Ion Ratio Lower Upper
196 100
198 92.5 76.6 114.8
200 29.4 25.4 38.2





#42

2,4,5-Trichlorophenol

Concen: 19.567 ng/ul

RT: 13.151 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:196 Resp: 159534

Ion Ratio Lower Upper

196 100

198 96.4 79.8 119.8

200 32.4 25.6 38.4



#43

1,1'-Biphenyl

Concen: 20.392 ng/ul

RT: 13.463 min Scan# 1747

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:154 Resp: 571892


Ion Ratio Lower Upper

154 100

153 40.3 32.4 48.6

76 15.1 10.7 16.1









#44
2-Chloronaphthalene
Concen: 20.936 ng/ul
RT: 13.510 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:162 Resp: 464238
Ion Ratio Lower Upper
162 100
164 32.5 26.1 39.1
127 41.8 33.7 50.5



#45
2-Nitroaniline
Concen: 21.143 ng/ul
RT: 13.728 min Scan# 1792
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion: 65 Resp: 145193
Ion Ratio Lower Upper
65 100
92 62.6 53.7 80.5
138 80.8 71.7 107.5









#47
Dimethylphthalate
Concen: 20.076 ng/ul
RT: 14.087 min Scan# 1853
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:163 Resp: 564962
Ion Ratio Lower Upper
163 100
194 4.3 3.1 4.7
164 10.2 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 20.151 ng/ul
RT: 14.216 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:165 Resp: 110793
Ion Ratio Lower Upper
165 100
89 75.5 52.7 79.1
121 23.2 17.9 26.9





#50

Acenaphthylene

Concen: 21.307 ng/ul

RT: 14.357 min Scan# 1899

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:152 Resp: 708674

Ion Ratio Lower Upper

152 100

151 20.0 16.2 24.4

153 12.8 10.2 15.4



#51

3-Nitroaniline

Concen: 18.881 ng/ul

RT: 14.557 min Scan# 1933

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:138 Resp: 93000


Ion Ratio Lower Upper

138 100

108 12.5 9.9 14.9

92 138.9 107.6 161.4





#52

Acenaphthene

Concen: 20.992 ng/ul

RT: 14.698 min Scan# 1957

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:153 Resp: 493386

Ion Ratio Lower Upper

153 100

152 46.8 38.2 57.4

154 86.6 70.2 105.2



#53

2,4-Dinitrophenol

Concen: 13.381 ng/ul

RT: 14.775 min Scan# 1970

Delta R.T. 0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:184 Resp: 47644


Ion Ratio Lower Upper

184 100

63 83.2 57.1 85.7

154 75.9 55.3 82.9









#55
4-Nitrophenol
Concen: 15.629 ng/ul
RT: 14.869 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:109 Resp: 65588
Ion Ratio Lower Upper
109 100
139 100.1 77.4 116.2
65 131.8 97.3 145.9



#56
Dibenzofuran
Concen: 20.508 ng/ul
RT: 15.034 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:168 Resp: 677596
Ion Ratio Lower Upper
168 100
139 40.3 32.1 48.1









#57
2,4-Dinitrotoluene
Concen: 19.940 ng/ul
RT: 15.010 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:165 Resp: 155512
Ion Ratio Lower Upper
165 100
63 59.3 41.6 62.4
182 2.6 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.081 ng/ul
RT: 15.263 min Scan# 2053
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:232 Resp: 139828
Ion Ratio Lower Upper
232 100
131 47.7 35.0 52.4
130 2.8 1.9 2.9
166 32.9 24.1 36.1









#59
Diethylphthalate
Concen: 20.120 ng/ul
RT: 15.451 min Scan# 2085
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:149 Resp: 562654
Ion Ratio Lower Upper
149 100
177 21.2 18.2 27.2
150 12.4 9.8 14.8



#61
Fluorene
Concen: 20.484 ng/ul
RT: 15.681 min Scan# 2124
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:166 Resp: 548584
Ion Ratio Lower Upper
166 100
165 98.7 79.2 118.8
167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.951 ng/ul
RT: 15.675 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:204 Resp: 287055
Ion Ratio Lower Upper
204 100
206 32.3 26.9 40.3
141 62.8 51.3 76.9



#63
4-Nitroaniline
Concen: 15.995 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:138 Resp: 64013
Ion Ratio Lower Upper
138 100
92 63.2 47.8 71.6
108 111.0 82.3 123.5









#66
4,6-Dinitro-2-methylphenol
Concen: 18.298 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:198 Resp: 89951
Ion Ratio Lower Upper
198 100
182 4.9 3.5 5.3
77 32.7 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 20.876 ng/ul
RT: 15.892 min Scan# 2160
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:169 Resp: 447774
Ion Ratio Lower Upper
169 100
168 66.3 54.2 81.2
167 35.4 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.388 ng/ul
RT: 16.575 min Scan# 2276
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:248 Resp: 166321
Ion Ratio Lower Upper
248 100
250 98.8 78.1 117.1
141 82.8 66.0 99.0



#69
Hexachlorobenzene
Concen: 19.587 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:284 Resp: 192999
Ion Ratio Lower Upper
284 100
142 33.2 26.6 39.8
249 31.9 25.7 38.5









#70
Atrazine
Concen: 19.872 ng/ul
RT: 16.851 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:200 Resp: 155166
Ion Ratio Lower Upper
200 100
173 26.2 20.6 30.8
215 48.4 37.2 55.8



#71
Pentachlorophenol
Concen: 17.156 ng/ul
RT: 17.051 min Scan# 2357
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:266 Resp: 97937
Ion Ratio Lower Upper
266 100
264 60.1 51.0 76.4
268 62.9 51.4 77.2









#72
Phenanthrene
Concen: 20.975 ng/ul
RT: 17.439 min Scan# 2423
Delta R.T. -0.000 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24




Instrument :
BNA_P
ClientSampleId :
SSTD020668


Tgt Ion:178 Resp: 804452
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 18.8 15.3 22.9



#74
Anthracene
Concen: 21.008 ng/ul
RT: 17.528 min Scan# 2438
Delta R.T. -0.006 min
Lab File: BP014625.D
Acq: 10 Apr 2023 10:24

Tgt Ion:178 Resp: 812090
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 17.9 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.982 ng/uL

RT: 13.428 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:216 Resp: 254621

Ion Ratio Lower Upper

216 100

214 78.6 61.0 91.6

218 48.6 38.7 58.1



#76

Pentachlorobenzene

Concen: 20.534 ng/uL

RT: 14.951 min Scan# 2000

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:250 Resp: 248636


Ion Ratio Lower Upper

250 100

248 62.0 49.0 73.6

252 62.6 50.7 76.1









#77
 Carbazole
 Concen: 19.982 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. -0.006 min
 Lab File: BP014625.D
 Acq: 10 Apr 2023 10:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020668


Tgt Ion:167 Resp: 646393
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.4 26.2
 139 14.1 11.3 16.9



#78
 Di-n-butylphthalate
 Concen: 20.812 ng/ul
 RT: 18.333 min Scan# 2575
 Delta R.T. -0.000 min
 Lab File: BP014625.D
 Acq: 10 Apr 2023 10:24

Tgt Ion:149 Resp: 846544
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 6.0 4.8 7.2





#80

Fluoranthene

Concen: 18.915 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:202 Resp: 880638

Ion Ratio Lower Upper

202 100

101 11.6 8.2 12.2

100 9.3 6.4 9.6



#82

Pyrene

Concen: 18.896 ng/ul

RT: 19.751 min Scan# 2816

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:202 Resp: 918489


Ion Ratio Lower Upper

202 100

101 12.8 9.7 14.5

100 11.1 8.0 12.0





#83

Butylbenzylphthalate

Concen: 20.809 ng/ul

RT: 20.610 min Scan# 2962

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:149 Resp: 348414

Ion Ratio Lower Upper

149 100

91 84.5 61.8 92.8

206 20.7 17.8 26.6



#84

3,3'-Dichlorobenzidine

Concen: 20.949 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:252 Resp: 271603


Ion Ratio Lower Upper

252 100

254 62.7 51.4 77.0

126 10.0 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.677 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:228 Resp: 858964

Ion Ratio Lower Upper

228 100

229 19.9 15.6 23.4

226 27.5 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.800 ng/ul

RT: 21.363 min Scan# 3090

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:149 Resp: 499126


Ion Ratio Lower Upper

149 100

167 27.0 21.9 32.9

279 4.3 3.7 5.5





#87

Chrysene

Concen: 21.515 ng/ul

RT: 21.521 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:228 Resp: 843935

Ion Ratio Lower Upper

228 100

226 29.5 23.8 35.8

229 19.8 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 19.930 ng/ul




RT: 22.321 min Scan# 3253


Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Tgt Ion:149 Resp: 873195





#90

Benzo(b)fluoranthene

Concen: 20.147 ng/ul

RT: 23.215 min Scan# 3405

Delta R.T. -0.000 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:252 Resp: 918691

Ion Ratio Lower Upper

252 100

253 21.6 18.0 27.0

125 8.5 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.388 ng/ul

RT: 23.263 min Scan# 3413

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:252 Resp: 902920


Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 9.8 6.9 10.3





#93

Benzo(a)pyrene

Concen: 20.931 ng/ul

RT: 23.868 min Scan# 3516

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:252 Resp: 821533

Ion Ratio Lower Upper

252 100

253 21.7 16.9 25.3

125 9.3 6.8 10.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.943 ng/ul

RT: 26.604 min Scan# 3981

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24




Tgt Ion:276 Resp: 1130874


Ion Ratio Lower Upper

276 100

138 21.0 14.9 22.3

277 25.4 20.3 30.5





#95

Dibenzo(a,h)anthracene

Concen: 20.271 ng/ul

RT: 26.615 min Scan# 3983

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Instrument :

BNA_P

ClientSampleId :

SSTD020668





Tgt Ion:278 Resp: 916449

Ion Ratio Lower Upper

278 100

139 12.4 10.4 15.6

279 24.0 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.809 ng/ul

RT: 27.409 min Scan# 4118

Delta R.T. -0.006 min

Lab File: BP014625.D

Acq: 10 Apr 2023 10:24

Tgt Ion:276 Resp: 915915

Ion Ratio Lower Upper

276 100

138 17.6 13.0 19.4

277 24.8 18.4 27.6

