

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041023\
 Data File : BP014662.D
 Acq On : 11 Apr 2023 18:35
 Operator : CG/JU
 Sample : 02251-11
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 COMC9

Quant Time: Apr 11 23:02:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	109050	20.000	ng/ul	0.00
20) Naphthalene-d8	10.793	136	449029	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.628	164	273523	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	612783	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	633634	20.000	ng/ul	0.00
88) Perylene-d12	23.986	264	718005	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	6149	2.187	ng/uL	0.00
4) Pyridine-d5	3.817	84	34116	4.734	ng/ul	0.02
7) Phenol-d5	7.152	99	47664	4.739	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	184374	28.573	ng/ul	0.00
11) 2-Chlorophenol-d4	7.505	132	150587	20.716	ng/ul	0.00
15) 4-Methylphenol-d8	8.699	113	96979	11.954	ng/ul	0.00
21) Nitrobenzene-d5	9.152	128	93381	25.761	ng/ul	0.00
24) 2-Nitrophenol-d4	9.881	143	98362	23.697	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	166056	21.967	ng/ul	0.00
31) 4-Chloroaniline-d4	10.946	131	190654	19.147	ng/ul	0.00
46) Dimethylphthalate-d6	14.034	166	642948	28.991	ng/ul	0.00
49) Acenaphthylene-d8	14.322	160	681292	27.685	ng/ul	0.00
54) 4-Nitrophenol-d4	14.934	143	1679	0.501	ng/ul	0.08
60) Fluorene-d10	15.622	176	550055	29.741	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	72950	16.179	ng/ul	0.00
73) Anthracene-d10	17.492	188	928157	31.297	ng/ul	0.00
81) Pyrene-d10	19.722	212	1186683	27.882	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.821	264	1238254	32.460	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041023\
Data File : BP014662.D
Acq On : 11 Apr 2023 18:35
Operator : CG/JU
Sample : 02251-11
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
COMC9

Quant Time: Apr 11 23:02:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 11 01:21:43 2023
Response via : Initial Calibration

