

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009797.D
 Acq On : 12 Apr 2022 23:47
 Operator : CG/JU
 Sample : N2342-09
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0HY3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/13/2022

Quant Time: Apr 13 02:26:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 13 02:15:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	98566	20.000	ng/u1	0.00
20) Naphthalene-d8	10.775	136	440686	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.593	164	321814	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	737387	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.422	240	808195	20.000	ng/u1	#-0.01
88) Perylene-d12	23.892	264	794307	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	12961	5.869	ng/uL	0.00
4) Pyridine-d5	3.805	84	67230	11.041	ng/u1	0.00
7) Phenol-d5	7.116	99	71533	8.266	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	186725	36.245	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	178324	27.319	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	136623	18.913	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	121926	38.367	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	122994	35.301	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	221125	30.330	ng/u1	0.00
31) 4-Chloroaniline-d4	10.899	131	310511	29.903	ng/u1	0.00
46) Dimethylphthalate-d6	14.004	166	951930	37.657	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	1031083	34.429	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	38694m	8.440	ng/u1	-0.01
60) Fluorene-d10	15.587	176	819222	38.523	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.693	200	142255	32.565	ng/u1	0.00
73) Anthracene-d10	17.445	188	1430693	40.523	ng/u1	0.00
81) Pyrene-d10	19.675	212	1802137	40.828	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	1755277	42.434	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
Data File : BP009797.D
Acq On : 12 Apr 2022 23:47
Operator : CG/JU
Sample : N2342-09
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0HY3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022

Quant Time: Apr 13 02:26:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 13 02:15:22 2022
Response via : Initial Calibration

